

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County ANSON Instrument Location ANSON COUNTY  
Instrument Serial No. 008597 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]

Signature of Certifying Official

667

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008597  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Galloway, Charles L  
Permit Number: 0023-9771  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG034902  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	1:44pm
AIR BLK	.00	1:45pm
ACCY CHK	.07	1:46pm
AIR BLK	.00	1:47pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:47pm</b>
AIR BLK	.00	1:48pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:50pm</b>
AIR BLK	.00	1:51pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of ~~Chemical~~ Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008597      Test Record Number: 1809  
Test Date: 11/22/2021      Test Time: 1:53pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:53pm
FLO	Pass	1:53pm
FC	Pass	1:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:54pm
SRC	Pass	1:54pm
DET	Pass	1:54pm
BAR	Pass	1:54pm
BT	Pass	1:54pm

Blank Tests

Test	Status	Time
AIR	Pass	1:54pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:54pm

CRC Tests

Test	Status	Time
COMP	Pass	1:54pm
CAL	Pass	1:54pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County ANSON Instrument Location ANSON COUNTY  
Instrument Serial No. 008739 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]  
Signature of Certifying Official

667  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008739  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Galloway, Charles L  
Permit Number: 0023-9771  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG034902  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	1:45pm
AIR BLK	.00	1:46pm
ACCY CHK	.08	1:47pm
AIR BLK	.00	1:48pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:49pm</b>
AIR BLK	.00	1:50pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:52pm</b>
AIR BLK	.00	1:53pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008739      Test Record Number: 703  
Test Date: 11/22/2021      Test Time: 1:54pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:54pm
FLO	Pass	1:54pm
FC	Pass	1:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:55pm
SRC	Pass	1:55pm
DET	Pass	1:55pm
BAR	Pass	1:55pm
BT	Pass	1:55pm

Blank Tests

Test	Status	Time
AIR	Pass	1:55pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:55pm

CRC Tests

Test	Status	Time
COMP	Pass	1:55pm
CAL	Pass	1:55pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Ashe Instrument Location Ashe Co. Jail

Instrument Serial No. 008849 Jefferson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849  
Test Date: 11/09/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Burnette, Anthony J

Permit Number: 0018-4401

Effective:

10/01/2020-10/01/2022

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	3:09pm
AIR BLK	.00	3:09pm
ACCY CHK	.07	3:10pm
AIR BLK	.00	3:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:12pm</b>
AIR BLK	.00	3:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:14pm</b>
AIR BLK	.00	3:15pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849      Test Record Number: 1499  
Test Date: 11/09/2021      Test Time: 3:16pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:16pm
FLO	Pass	3:16pm
FC	Pass	3:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:16pm
SRC	Pass	3:16pm
DET	Pass	3:16pm
BAR	Pass	3:16pm
BT	Pass	3:16pm

Blank Tests

Test	Status	Time
AIR	Pass	3:17pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:17pm

CRC Tests

Test	Status	Time
COMP	Pass	3:17pm
CAL	Pass	3:17pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Avery Instrument Location Avery Co Jail  
Instrument Serial No. 008664 Newland, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Burnette, Anthony J  
Permit Number: 0018-4401  
Effective:  
10/01/2020-10/01/2022

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	4:35pm
AIR BLK	.00	4:36pm
ACCY CHK	.08	4:37pm
AIR BLK	.00	4:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:38pm</b>
AIR BLK	.00	4:39pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:41pm</b>
AIR BLK	.00	4:42pm

**Reported AC: .00 g/210L**

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664      Test Record Number: 1103  
Test Date: 11/10/2021      Test Time: 4:42pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:43pm
FLO	Pass	4:43pm
FC	Pass	4:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:44pm
SRC	Pass	4:44pm
DET	Pass	4:44pm
BAR	Pass	4:44pm
BT	Pass	4:44pm

Blank Tests

Test	Status	Time
AIR	Pass	4:44pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:44pm

CRC Tests

Test	Status	Time
COMP	Pass	4:44pm
CAL	Pass	4:44pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County CABARRUS Instrument Location KANNAPOLIS PO

Instrument Serial No. 008589 KANNAPOLIS, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

KANNAPOLIS PD CABARRUS COUNTY 120

Serial Number: 008589  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	11:29am
AIR BLK	.00	11:30am
ACCY CHK	.08	11:31am
AIR BLK	.00	11:32am
<b>SUB TEST</b>	<b>.00</b>	<b>11:32am</b>
AIR BLK	.00	11:33am
<b>SUB TEST</b>	<b>.00</b>	<b>11:35am</b>
AIR BLK	.00	11:36am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

*KANNAPOLIS PD CABARRUS COUNTY 120*

Serial Number: 008589      Test Record Number: 3309  
Test Date: 11/10/2021      Test Time: 11:38am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:38am

Temperature Tests

Test	Status	Time
FC1	Pass	11:38am
SRC	Pass	11:38am
DET	Pass	11:38am
BAR	Pass	11:38am
BT	Pass	11:38am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am


Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:39am
CAL	Pass	11:39am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Camden Instrument Location Camden Co. S.O.

Instrument Serial No. 008940 117 N. Carolina Hwy 343, Camden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ken P  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940

Test Date: 11/01/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG034901


Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	11:54am
AIR BLK	.00	11:54am
ACCY CHK	.08	11:55am
AIR BLK	.00	11:56am
SUB TEST	.00	11:57am
AIR BLK	.00	11:57am
SUB TEST	.00	11:59am
AIR BLK	.00	12:00pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940      Test Record Number: 1115  
Test Date: 11/01/2021      Test Time: 12:01pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FCL	Pass	12:02pm
SRC	Pass	12:02pm
DET	Pass	12:02pm
BAR	Pass	12:02pm
BT	Pass	12:02pm

Blank Tests

Test	Status	Time
AIR	Pass	12:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:02pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Caswell Instrument Location Caswell Co Detention Ctr  
Instrument Serial No. 008593 211 County Park Rd  
Yanceyville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CASWELL COUNTY DETENTION CENTER 160

Serial Number: 008593  
Test Date: 11/01/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Barnes, Simon S  
Permit Number: 0014-6221  
Effective:  
09/01/2020-09/01/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118804  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	7:00am
AIR BLK	.00	7:01am
ACCY CHK	.08	7:01am
AIR BLK	.00	7:02am
<b>SUB TEST</b>	<b>.00</b>	<b>7:03am</b>
AIR BLK	.00	7:04am
<b>SUB TEST</b>	<b>.00</b>	<b>7:05am</b>
AIR BLK	.00	7:06am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

CASWELL COUNTY DETENTION CENTER 160

Serial Number: 008593      Test Record Number: 1921  
Test Date: 11/01/2021      Test Time: 7:07am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	7:07am
FLO	Pass	7:07am
FC	Pass	7:07am

Temperature Tests

Test	Status	Time
FC1	Pass	7:07am
SRC	Pass	7:07am
DET	Pass	7:07am
BAR	Pass	7:07am
BT	Pass	7:07am

Blank Tests

Test	Status	Time
AIR	Pass	7:08am

Printer Tests

Test	Status	Time
PRNT	Pass	7:08am

CRC Tests

Test	Status	Time
COMP	Pass	7:08am
CAL	Pass	7:08am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)


County CRAVEN Instrument Location CRAVEN COUNTY

Instrument Serial No. 008732 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY DETENTION CENTER 240

Serial Number: 008732  
Test Date: 11/15/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG107401  
Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	3:08pm
AIR BLK	.00	3:09pm
ACCY CHK	.07	3:09pm
AIR BLK	.00	3:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:11pm</b>
AIR BLK	.00	3:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:13pm</b>
AIR BLK	.00	3:14pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CRAVEN COUNTY DETENTION CENTER 240

Serial Number: 008732      Test Record Number: 2683  
Test Date: 11/15/2021      Test Time: 3:14pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:15pm
FLO	Pass	3:15pm
FC	Pass	3:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:15pm
SRC	Pass	3:15pm
DET	Pass	3:15pm
BAR	Pass	3:15pm
BT	Pass	3:15pm

Blank Tests

Test	Status	Time
AIR	Pass	3:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:15pm

CRC Tests

Test	Status	Time
COMP	Pass	3:16pm
CAL	Pass	3:16pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County CRAVEN Instrument Location HAVELOCK POLICE DEPT

Instrument Serial No. 008800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Burns  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

(RAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG107401  
Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	3:26pm
AIR BLK	.00	3:27pm
ACCY CHK	.07	3:28pm
AIR BLK	.00	3:29pm
SUB TEST	.00	3:29pm
AIR BLK	.00	3:30pm
SUB TEST	.00	3:32pm
AIR BLK	.00	3:33pm

Reported AC: .00 g/210L

Alvin R. Barnes  
Signature of Chemical Analyst

Court CVR

Alvin R. Barnes  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

**CRAVEN COUNTY HAVELOCK PD 240**

Serial Number: 008800      Test Record Number: 1453  
Test Date: 11/23/2021      Test Time: 3:33pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:34pm
FLO	Pass	3:34pm
FC	Pass	3:34pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:34pm
SRC	Pass	3:34pm
DET	Pass	3:34pm
BAR	Pass	3:34pm
BT	Pass	3:34pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:34pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:34pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:35pm
CAL	Pass	3:35pm

Preventive Maintenance  
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County CRAVEN Instrument Location NEW BERN  
Instrument Serial No. 008817 POLICE DEPT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bano  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**CRAVEN COUNTY NEW BERN PD 240**

Serial Number: 008817

Test Date: 11/18/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 0014-6279

Effective:

10/09/2020-10/09/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107401

Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.07	11:32am
AIR BLK	.00	11:33am
SUB TEST	.00	11:34am
AIR BLK	.00	11:35am
SUB TEST	.00	11:36am
AIR BLK	.00	11:37am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817      Test Record Number: 1714  
Test Date: 11/18/2021      Test Time: 11:37am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38am
FLO	Pass	11:38am
FC	Pass	11:38am

Temperature Tests

Test	Status	Time
FC1	Pass	11:38am
SRC	Pass	11:38am
DET	Pass	11:38am
BAR	Pass	11:38am
BT	Pass	11:38am

Blank Tests

Test	Status	Time
AIR	Pass	11:39am

Printer Tests

Test	Status	Time
PRNT	Pass	11:39am

CRC Tests

Test	Status	Time
COMP	Pass	11:39am
CAL	Pass	11:39am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County CRAVEN Instrument Location MCA3 CHERRY POINT  
Instrument Serial No. 010819 PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG034901  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	3:56pm
AIR BLK	.00	3:57pm
ACCY CHK	.07	3:57pm
AIR BLK	.00	3:58pm
SUB TEST	.00	3:59pm
AIR BLK	.00	3:59pm
SUB TEST	.00	4:01pm
AIR BLK	.00	4:02pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**CRAVEN COUNTY MCAS CHERRY POINT 240**

Serial Number: 010819      Test Record Number: 701  
Test Date: 11/23/2021      Test Time: 4:02pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:03pm
FLO	Pass	4:03pm
FC	Pass	4:03pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:03pm
SRC	Pass	4:03pm
DET	Pass	4:03pm
BAR	Pass	4:03pm
BT	Pass	4:03pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:04pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	4:04pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:04pm
CAL	Pass	4:04pm

Preventive Maintenance  
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County Cumberland Instrument Location Cumberland County

Instrument Serial No. 008614 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	10:14am
AIR BLK	.00	10:15am
ACCY CHK	.07	10:16am
AIR BLK	.00	10:17am
<b>SUB TEST</b>	<b>.00</b>	<b>10:17am</b>
AIR BLK	.00	10:18am
<b>SUB TEST</b>	<b>.00</b>	<b>10:20am</b>
AIR BLK	.00	10:21am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614      Test Record Number: 4613  
Test Date: 11/23/2021      Test Time: 10:22am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:22am
FLO	Pass	10:22am
FC	Pass	10:22am

Temperature Tests

Test	Status	Time
FC1	Pass	10:22am
SRC	Pass	10:22am
DET	Pass	10:22am
BAR	Pass	10:22am
BT	Pass	10:22am

Blank Tests

Test	Status	Time
AIR	Pass	10:23am


Printer Tests

Test	Status	Time
PRNT	Pass	10:23am

CRC Tests

Test	Status	Time
COMP	Pass	10:23am
CAL	Pass	10:23am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Cumberland Instrument Location Cumberland County  
Instrument Serial No. 008632 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008632  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	9:09am
AIR BLK	.00	9:10am
ACCY CHK	.08	9:10am
AIR BLK	.00	9:11am
<b>SUB TEST</b>	<b>.00</b>	<b>9:12am</b>
AIR BLK	.00	9:13am
<b>SUB TEST</b>	<b>.00</b>	<b>9:14am</b>
AIR BLK	.00	9:15am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008632      Test Record Number: 4380  
Test Date: 11/23/2021      Test Time: 9:21am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:22am
FLO	Pass	9:22am
FC	Pass	9:22am

Temperature Tests

Test	Status	Time
FC1	Pass	9:22am
SRC	Pass	9:22am
DET	Pass	9:22am
BAR	Pass	9:22am
BT	Pass	9:22am

Blank Tests

Test	Status	Time
AIR	Pass	9:23am


Printer Tests

Test	Status	Time
PRNT	Pass	9:23am

CRC Tests

Test	Status	Time
COMP	Pass	9:23am
CAL	Pass	9:23am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Cumberland Instrument Location Cumberland County  
Instrument Serial No. 008633 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	9:04am
AIR BLK	.00	9:04am
ACCY CHK	.07	9:05am
AIR BLK	.00	9:06am
<b>SUB TEST</b>	<b>.00</b>	<b>9:07am</b>
AIR BLK	.00	9:08am
<b>SUB TEST</b>	<b>.00</b>	<b>9:09am</b>
AIR BLK	.00	9:10am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633      Test Record Number: 5986  
Test Date: 11/23/2021      Test Time: 9:12am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:12am
FLO	Pass	9:12am
FC	Pass	9:12am

Temperature Tests

Test	Status	Time
FC1	Pass	9:12am
SRC	Pass	9:12am
DET	Pass	9:12am
BAR	Pass	9:12am
BT	Pass	9:12am

Blank Tests

Test	Status	Time
AIR	Pass	9:13am


Printer Tests

Test	Status	Time
PRNT	Pass	9:13am

CRC Tests

Test	Status	Time
COMP	Pass	9:13am
CAL	Pass	9:13am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Cumberland Instrument Location Cumberland County  
Instrument Serial No. 008672 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

675

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Garner, Joel K

Permit Number: 0036-5156

Effective:

03/10/2021-03/10/2023

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118803


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	9:59am
AIR BLK	.00	10:00am
ACCY CHK	.07	10:00am
AIR BLK	.00	10:02am
<b>SUB TEST</b>	<b>.00</b>	<b>10:03am</b>
AIR BLK	.00	10:03am
<b>SUB TEST</b>	<b>.00</b>	<b>10:05am</b>
AIR BLK	.00	10:06am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672      Test Record Number: 7444  
Test Date: 11/23/2021      Test Time: 10:07am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:07am
FLO	Pass	10:07am
FC	Pass	10:07am

Temperature Tests

Test	Status	Time
FC1	Pass	10:07am
SRC	Pass	10:07am
DET	Pass	10:07am
BAR	Pass	10:07am
BT	Pass	10:07am

Blank Tests

Test	Status	Time
AIR	Pass	10:08am


Printer Tests

Test	Status	Time
PRNT	Pass	10:08am

CRC Tests

Test	Status	Time
COMP	Pass	10:08am
CAL	Pass	10:08am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Cumberland Instrument Location FORT BRAGG

Instrument Serial No. 013868 L.E.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 013868  
Test Date: 11/04/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG016803  
Exp Date: 06/16/2022

Test	g/210L	Time
DIAG	Pass	10:02am
AIR BLK	.00	10:03am
ACCY CHK	.07	10:03am
AIR BLK	.00	10:04am
<b>SUB TEST</b>	<b>.00</b>	<b>10:05am</b>
AIR BLK	.00	10:06am
<b>SUB TEST</b>	<b>.00</b>	<b>10:07am</b>
AIR BLK	.00	10:08am

Reported AC/ .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 013868      Test Record Number: 124  
Test Date: 11/04/2021      Test Time: 10:09am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:10am
FLO	Pass	10:10am
FC	Pass	10:10am

Temperature Tests

Test	Status	Time
FC1	Pass	10:10am
SRC	Pass	10:10am
DET	Pass	10:10am
BAR	Pass	10:10am
BT	Pass	10:10am

Blank Tests

Test	Status	Time
AIR	Pass	10:11am


Printer Tests

Test	Status	Time
PRNT	Pass	10:11am

CRC Tests

Test	Status	Time
COMP	Pass	10:11am
CAL	Pass	10:11am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Cumberland Instrument Location Fort Bragg

Instrument Serial No. 013870 L.E.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 013870  
Test Date: 11/04/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG009803  
Exp Date: 04/07/2022

Test	g/210L	Time
DIAG	Pass	10:17am
AIR BLK	.00	10:17am
ACCY CHK	.07	10:18am
AIR BLK	.00	10:19am
<b>SUB TEST</b>	<b>.00</b>	<b>10:19am</b>
AIR BLK	.00	10:20am
<b>SUB TEST</b>	<b>.00</b>	<b>10:22am</b>
AIR BLK	.00	10:23am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 013870      Test Record Number: 202  
Test Date: 11/04/2021      Test Time: 10:24am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:24am
FLO	Pass	10:24am
FC	Pass	10:25am

Temperature Tests

Test	Status	Time
FC1	Pass	10:25am
SRC	Pass	10:25am
DET	Pass	10:25am
BAR	Pass	10:25am
BT	Pass	10:25am

Blank Tests

Test	Status	Time
AIR	Pass	10:25am


Printer Tests

Test	Status	Time
PRNT	Pass	10:25am

CRC Tests

Test	Status	Time
COMP	Pass	10:26am
CAL	Pass	10:26am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Currituck Instrument Location Currituck Co. S.O.

Instrument Serial No. 008947 407-A Maple Rd, Maple, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CURRITUCK COUNTY CURRITUCK SO-MAPLE  
260

Serial Number: 008947  
Test Date: 11/01/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118805  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	11:00am
AIR BLK	.00	11:00am
ACCY CHK	.08	11:01am
AIR BLK	.00	11:02am
<b>SUB TEST</b>	<b>.00</b>	<b>11:03am</b>
AIR BLK	.00	11:04am
<b>SUB TEST</b>	<b>.00</b>	<b>11:05am</b>
AIR BLK	.00	11:06am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*CURRITUCK COUNTY CURRITUCK SO-MAPLE 260*

Serial Number: 008947      Test Record Number: 2908  
Test Date: 11/01/2021      Test Time: 11:07am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:07am
FLO	Pass	11:07am
FC	Pass	11:07am

Temperature Tests

Test	Status	Time
FC1	Pass	11:07am
SRC	Pass	11:07am
DET	Pass	11:07am
BAR	Pass	11:07am
BT	Pass	11:07am

Blank Tests

Test	Status	Time
AIR	Pass	11:08am

Printer Tests

Test	Status	Time
PRNT	Pass	11:08am

CRC Tests

Test	Status	Time
COMP	Pass	11:08am
CAL	Pass	11:08am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Currituck Instrument Location Currituck Co. S.D. - Corolla  
Instrument Serial No. 008949 1123 Ocean Trail, Corolla, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3<sup>rd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Yan  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: 008949  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: *Guard, Kelly G*  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: *Breath Test*


Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	10:21am
AIR BLK	.00	10:22am
ACCY CHK	.08	10:23am
AIR BLK	.00	10:24am
<b>SUB TEST</b>	<b>.00</b>	<b>10:25am</b>
AIR BLK	.00	10:25am
<b>SUB TEST</b>	<b>.00</b>	<b>10:27am</b>
AIR BLK	.00	10:28am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*CURRITUCK COUNTY SO-COROLLA 260*

Serial Number: 008949      Test Record Number: 630  
Test Date: 11/03/2021      Test Time: 10:29am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:30am
FLO	Pass	10:30am
FC	Pass	10:30am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:30am
SRC	Pass	10:30am
DET	Pass	10:30am
BAR	Pass	10:30am
BT	Pass	10:30am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:30am


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:30am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:31am
CAL	Pass	10:31am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Dare Instrument Location Dare Co. Detention Center  
Instrument Serial No. 008783 1044 Driftwood Dr, Manteo, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008783  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107501

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	12:46pm
AIR BLK	.00	12:46pm
ACCY CHK	.07	12:47pm
AIR BLK	.00	12:48pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:49pm</b>
AIR BLK	.00	12:49pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:51pm</b>
AIR BLK	.00	12:52pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008783      Test Record Number: 1055  
Test Date: 11/03/2021      Test Time: 12:53pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:53pm
FLO	Pass	12:53pm
FC	Pass	12:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:53pm
SRC	Pass	12:53pm
DET	Pass	12:53pm
BAR	Pass	12:53pm
BT	Pass	12:53pm

Blank Tests

Test	Status	Time
AIR	Pass	12:54pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:54pm

CRC Tests

Test	Status	Time
COMP	Pass	12:54pm
CAL	Pass	12:54pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Dare Instrument Location Dare Co. Detention Center  
Instrument Serial No. 008804 1044 Driftwood Dr, Manteo, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3<sup>rd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ken  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008804  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG018401  
Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	12:59pm
AIR BLK	.00	1:00pm
ACCY CHK	.07	1:00pm
AIR BLK	.00	1:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:02pm</b>
AIR BLK	.00	1:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:05pm</b>
AIR BLK	.00	1:05pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

DARE COUNTY DARE CO DETENTION CE 270

Serial Number: 008804      Test Record Number: 2609  
Test Date: 11/03/2021      Test Time: 1:07pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:07pm
FLO	Pass	1:07pm
FC	Pass	1:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:07pm
SRC	Pass	1:07pm
DET	Pass	1:07pm
BAR	Pass	1:07pm
BT	Pass	1:07pm

Blank Tests

Test	Status	Time
AIR	Pass	1:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:08pm

CRC Tests

Test	Status	Time
COMP	Pass	1:08pm
CAL	Pass	1:08pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County DAVIDSON Instrument Location DAVIDSON COUNTY JAIL

Instrument Serial No. 008845 LEXINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A

Permit Number: 0035-3799

Effective:

08/04/2020-08/04/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118803

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:06pm
AIR BLK	.00	12:07pm
ACCY CHK	.08	12:07pm
AIR BLK	.00	12:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:09pm</b>
AIR BLK	.00	12:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:12pm</b>
AIR BLK	.00	12:13pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845      Test Record Number: 3690  
Test Date: 11/22/2021      Test Time: 12:14pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:14pm
FLO	Pass	12:14pm
FC	Pass	12:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:14pm
SRC	Pass	12:14pm
DET	Pass	12:14pm
BAR	Pass	12:14pm
BT	Pass	12:14pm

Blank Tests

Test	Status	Time
AIR	Pass	12:15pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:15pm

CRC Tests

Test	Status	Time
COMP	Pass	12:15pm
CAL	Pass	12:15pm

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County DAVIDSON Instrument Location LEXINGTON POLICE

Instrument Serial No. 008883 DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	11:15am
AIR BLK	.00	11:16am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
<b>SUB TEST</b>	<b>.00</b>	<b>11:19am</b>
AIR BLK	.00	11:20am
<b>SUB TEST</b>	<b>.00</b>	<b>11:22am</b>
AIR BLK	.00	11:23am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883      Test Record Number: 2380  
Test Date: 11/22/2021      Test Time: 11:25am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:25am
FLO	Pass	11:25am
FC	Pass	11:25am

Temperature Tests

Test	Status	Time
FC1	Pass	11:25am
SRC	Pass	11:25am
DET	Pass	11:25am
BAR	Pass	11:25am
BT	Pass	11:25am

Blank Tests

Test	Status	Time
AIR	Pass	11:26am

Printer Tests

Test	Status	Time
PRNT	Pass	11:26am

CRC Tests

Test	Status	Time
COMP	Pass	11:26am
CAL	Pass	11:26am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County DUPLIN Instrument Location WALLACE

Instrument Serial No. 008858 POLICE DEPT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 02 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858  
Test Date: 11/02/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG107401  
Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	9:36am
AIR BLK	.00	9:36am
ACCY CHK	.08	9:37am
AIR BLK	.00	9:38am
SUB TEST	.00	9:39am
AIR BLK	.00	9:40am
SUB TEST	.00	9:41am
AIR BLK	.00	9:42am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*DUPLIN COUNTY WALLACE PD 300*

Serial Number: 008858      Test Record Number: 1118  
Test Date: 11/02/2021      Test Time: 9:43am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:43am
FLO	Pass	9:43am
FC	Pass	9:43am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:43am
SRC	Pass	9:43am
DET	Pass	9:43am
BAR	Pass	9:43am
BT	Pass	9:43am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:44am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:44am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:44am
CAL	Pass	9:44am

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County DUPLIN Instrument Location DUPLIN COUNTY  
Instrument Serial No. 008864 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 02 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Bane  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY DETENTION CENTER 300

Serial Number: 008864  
Test Date: 11/02/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	11:29am
AIR BLK	.00	11:29am
ACCY CHK	.08	11:30am
AIR BLK	.00	11:31am
<b>SUB TEST</b>	<b>.00</b>	<b>11:32am</b>
AIR BLK	.00	11:32am
<b>SUB TEST</b>	<b>.00</b>	<b>11:34am</b>
AIR BLK	.00	11:35am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

*DUPLIN COUNTY DETENTION CENTER 300*

Serial Number: 008864      Test Record Number: 4252  
Test Date: 11/02/2021      Test Time: 11:35am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35am
FLO	Pass	11:35am
FC	Pass	11:35am

Temperature Tests

Test	Status	Time
FC1	Pass	11:36am
SRC	Pass	11:36am
DET	Pass	11:36am
BAR	Pass	11:36am
BT	Pass	11:36am

Blank Tests

Test	Status	Time
AIR	Pass	11:36am

Printer Tests

Test	Status	Time
PRNT	Pass	11:36am

CRC Tests

Test	Status	Time
COMP	Pass	11:36am
CAL	Pass	11:36am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Edgecombe Instrument Location Edgecombe Co. Magistrate's  
Instrument Serial No. 008603 Office, 300 S. Anaconda Rd.  
Tarboro, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ken Signature of Certifying Official 643 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR  
320

Serial Number: 008603  
Test Date: 11/08/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	2:55pm
AIR BLK	.00	2:56pm
ACCY CHK	.08	2:56pm
AIR BLK	.00	2:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:58pm</b>
AIR BLK	.00	2:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:01pm</b>
AIR BLK	.00	3:01pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008603      Test Record Number: 1985  
Test Date: 11/08/2021      Test Time: 3:03pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:03pm
FLO	Pass	3:03pm
FC	Pass	3:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:03pm
SRC	Pass	3:03pm
DET	Pass	3:03pm
BAR	Pass	3:03pm
BT	Pass	3:03pm

Blank Tests

Test	Status	Time
AIR	Pass	3:04pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:04pm

CRC Tests

Test	Status	Time
COMP	Pass	3:04pm
CAL	Pass	3:04pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Edgecombe Instrument Location Edgecombe Co. Magistrate's  
Instrument Serial No. 008663 Office, 3005. Anaconda Rd,  
Tarboro, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kerr

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR  
320

Serial Number: 008663  
Test Date: 11/08/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG018401  
Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	3:13pm
AIR BLK	.00	3:14pm
ACCY CHK	.08	3:15pm
AIR BLK	.00	3:16pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:17pm</b>
AIR BLK	.00	3:18pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:20pm</b>
AIR BLK	.00	3:20pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008663      Test Record Number: 3433  
Test Date: 11/08/2021      Test Time: 3:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:22pm
FLO	Pass	3:22pm
FC	Pass	3:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:22pm
SRC	Pass	3:22pm
DET	Pass	3:22pm
BAR	Pass	3:22pm
BT	Pass	3:22pm

Blank Tests

Test	Status	Time
AIR	Pass	3:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:23pm

CRC Tests

Test	Status	Time
COMP	Pass	3:23pm
CAL	Pass	3:23pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County FORSYTH Instrument Location FORSYTH COUNTY DETENTION

Instrument Serial No. 008583 WINSTON-SALEM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official      669 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008583  
Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A

Permit Number: 0035-3799

Effective:

08/04/2020-08/04/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118803

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	8:48am
AIR BLK	.00	8:49am
ACCY CHK	.08	8:49am
AIR BLK	.00	8:50am
<b>SUB TEST</b>	<b>.00</b>	<b>8:51am</b>
AIR BLK	.00	8:52am
<b>SUB TEST</b>	<b>.00</b>	<b>8:54am</b>
AIR BLK	.00	8:55am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008583      Test Record Number: 8612  
Test Date: 11/02/2021      Test Time: 8:56am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:57am
FLO	Pass	8:57am
FC	Pass	8:57am

Temperature Tests

Test	Status	Time
FC1	Pass	8:57am
SRC	Pass	8:57am
DET	Pass	8:57am
BAR	Pass	8:57am
BT	Pass	8:57am

Blank Tests

Test	Status	Time
AIR	Pass	8:58am

Printer Tests

Test	Status	Time
PRNT	Pass	8:58am

CRC Tests

Test	Status	Time
COMP	Pass	8:58am
CAL	Pass	8:58am

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County FORSYTH Instrument Location KERNERSVILLE POLICE

Instrument Serial No. 008650 DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>RD</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	10:22am
AIR BLK	.00	10:23am
ACCY CHK	.07	10:24am
AIR BLK	.00	10:25am
<b>SUB TEST</b>	<b>.00</b>	<b>10:25am</b>
AIR BLK	.00	10:26am
<b>SUB TEST</b>	<b>.00</b>	<b>10:28am</b>
AIR BLK	.00	10:29am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650      Test Record Number: 1888  
Test Date: 11/23/2021      Test Time: 10:31am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:31am
FLO	Pass	10:31am
FC	Pass	10:31am

Temperature Tests

Test	Status	Time
FC1	Pass	10:32am
SRC	Pass	10:32am
DET	Pass	10:32am
BAR	Pass	10:32am
BT	Pass	10:32am

Blank Tests

Test	Status	Time
AIR	Pass	10:32am

Printer Tests

Test	Status	Time
PRNT	Pass	10:32am

CRC Tests

Test	Status	Time
COMP	Pass	10:32am
CAL	Pass	10:32am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County FORSYTH Instrument Location FORSYTH COUNTY DETENTION  
Instrument Serial No. 008659 WINSTON-SALEM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008659  
Test Date: 11/02/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	8:42am
AIR BLK	.00	8:43am
ACCY CHK	.08	8:44am
AIR BLK	.00	8:45am
<b>SUB TEST</b>	<b>.00</b>	<b>8:47am</b>
AIR BLK	.00	8:48am
<b>SUB TEST</b>	<b>.00</b>	<b>8:50am</b>
AIR BLK	.00	8:51am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008659      Test Record Number: 5813  
Test Date: 11/02/2021      Test Time: 8:52am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:52am
FLO	Pass	8:52am
FC	Pass	8:52am

Temperature Tests

Test	Status	Time
FC1	Pass	8:52am
SRC	Pass	8:52am
DET	Pass	8:52am
BAR	Pass	8:52am
BT	Pass	8:52am

Blank Tests

Test	Status	Time
AIR	Pass	8:53am


Printer Tests

Test	Status	Time
PRNT	Pass	8:53am

CRC Tests

Test	Status	Time
COMP	Pass	8:53am
CAL	Pass	8:53am

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County FORSYTH Instrument Location FORSYTH COUNTY DETENTION

Instrument Serial No. 008925 WINSTON-SALEM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>ND</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008925  
Test Date: 11/02/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	9:07am
AIR BLK	.00	9:08am
ACCY CHK	.08	9:08am
AIR BLK	.00	9:09am
<b>SUB TEST</b>	<b>.00</b>	<b>9:11am</b>
AIR BLK	.00	9:12am
<b>SUB TEST</b>	<b>.00</b>	<b>9:13am</b>
AIR BLK	.00	9:14am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008925      Test Record Number: 3979  
Test Date: 11/02/2021      Test Time: 9:15am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:16am
FLO	Pass	9:16am
FC	Pass	9:16am

Temperature Tests

Test	Status	Time
FC1	Pass	9:16am
SRC	Pass	9:16am
DET	Pass	9:16am
BAR	Pass	9:16am
BT	Pass	9:16am

Blank Tests

Test	Status	Time
AIR	Pass	9:16am

Printer Tests

Test	Status	Time
PRNT	Pass	9:17am

CRC Tests

Test	Status	Time
COMP	Pass	9:17am
CAL	Pass	9:17am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Granville Instrument Location Granville Co. IEC  
Instrument Serial No. 008635 525 New Commerce Dr  
Oxford, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. Bryant  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY GRANVILLE COUNTY LEC  
380

Serial Number: 008635  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG018401


Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	12:25pm
AIR BLK	.00	12:25pm
ACCY CHK	.07	12:26pm
AIR BLK	.00	12:27pm
SUB TEST	.00	12:27pm
AIR BLK	.00	12:28pm
SUB TEST	.00	12:30pm
AIR BLK	.00	12:31pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY GRANVILLE COUNTY LEC 380

Serial Number: 008635 Test Record Number: 1854  
Test Date: 11/03/2021 Test Time: 12:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:32pm
SRC	Pass	12:32pm
DET	Pass	12:32pm
BAR	Pass	12:32pm
BT	Pass	12:32pm

Blank Tests

Test	Status	Time
AIR	Pass	12:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:33pm

CRC Tests

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Granville Instrument Location Creedmoor PD  
Instrument Serial No. 008641 111 Masonic St  
Creedmoor, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



EAB  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641

Test Date: 11/29/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107502

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	11:23am
AIR BLK	.00	11:24am
ACCY CHK	.08	11:24am
AIR BLK	.00	11:25am
SUB TEST	.00	11:26am
AIR BLK	.00	11:27am
SUB TEST	.00	11:28am
AIR BLK	.00	11:29am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641      Test Record Number: 1415  
Test Date: 11/29/2021      Test Time: 11:30am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30am
FLO	Pass	11:30am
FC	Pass	11:30am

Temperature Tests

Test	Status	Time
FC1	Pass	11:31am
SRC	Pass	11:31am
DET	Pass	11:31am
BAR	Pass	11:31am
BT	Pass	11:31am

Blank Tests

Test	Status	Time
AIR	Pass	11:31am

Printer Tests

Test	Status	Time
PRNT	Pass	11:31am

CRC Tests

Test	Status	Time
COMP	Pass	11:31am
CAL	Pass	11:31am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Granville Instrument Location Granville Co. LEC  
Instrument Serial No. 008923 525 New Commercial DR  
Oxford, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. A. Smith  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GRANVILLE COUNTY GRANVILLE COUNTY LEC  
380

Serial Number: 008923  
Test Date: 11/03/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG018401

Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	2:42pm
AIR BLK	.00	2:42pm
ACCY CHK	.07	2:43pm
AIR BLK	.00	2:44pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:45pm</b>
AIR BLK	.00	2:46pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:47pm</b>
AIR BLK	.00	2:48pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GRANVILLE COUNTY GRANVILLE COUNTY LEC 380

Serial Number: 008923      Test Record Number: 2776  
Test Date: 11/03/2021      Test Time: 2:49pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:50pm
FLO	Pass	2:50pm
FC	Pass	2:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:50pm
SRC	Pass	2:50pm
DET	Pass	2:50pm
BAR	Pass	2:50pm
BT	Pass	2:50pm

Blank Tests

Test	Status	Time
AIR	Pass	2:51pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:51pm

CRC Tests

Test	Status	Time
COMP	Pass	2:51pm
CAL	Pass	2:51pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Greene Instrument Location Greene Co. S.O.

Instrument Serial No. 008670 301 W. Greene St., Snow Hill, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670  
Test Date: 11/02/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG018202


Exp Date: 06/30/2022

Test	g/210L	Time
DIAG	Pass	1:30pm
AIR BLK	.00	1:31pm
ACCY CHK	.08	1:32pm
AIR BLK	.00	1:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:33pm</b>
AIR BLK	.00	1:34pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:36pm</b>
AIR BLK	.00	1:37pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670      Test Record Number: 2155  
Test Date: 11/02/2021      Test Time: 1:39pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:39pm
FLO	Pass	1:39pm
FC	Pass	1:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:40pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:40pm

CRC Tests

Test	Status	Time
COMP	Pass	1:40pm
CAL	Pass	1:40pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008638 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638

Test Date: 11/08/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A

Permit Number: 0035-3799

Effective:

08/04/2020-08/04/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118803


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	3:02pm
AIR BLK	.00	3:03pm
ACCY CHK	.08	3:03pm
AIR BLK	.00	3:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:05pm</b>
AIR BLK	.00	3:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:08pm</b>
AIR BLK	.00	3:09pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638      Test Record Number: 5096  
Test Date: 11/08/2021      Test Time: 3:09pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:10pm
FLO	Pass	3:10pm
FC	Pass	3:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:10pm
SRC	Pass	3:10pm
DET	Pass	3:10pm
BAR	Pass	3:10pm
BT	Pass	3:10pm

Blank Tests

Test	Status	Time
AIR	Pass	3:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:11pm

CRC Tests

Test	Status	Time
COMP	Pass	3:11pm
CAL	Pass	3:11pm

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008790 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790  
Test Date: 11/08/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	3:52pm
AIR BLK	.00	3:53pm
ACCY CHK	.08	3:54pm
AIR BLK	.00	3:55pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:56pm</b>
AIR BLK	.00	3:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:00pm</b>
AIR BLK	.00	4:01pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790      Test Record Number: 7469  
Test Date: 11/08/2021      Test Time: 4:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:03pm
FLO	Pass	4:03pm
FC	Pass	4:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:03pm
SRC	Pass	4:03pm
DET	Pass	4:03pm
BAR	Pass	4:03pm
BT	Pass	4:03pm

Blank Tests

Test	Status	Time
AIR	Pass	4:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:03pm

CRC Tests

Test	Status	Time
COMP	Pass	4:04pm
CAL	Pass	4:04pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008794 GREENSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794  
Test Date: 11/08/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	3:19pm
AIR BLK	.00	3:20pm
ACCY CHK	.07	3:21pm
AIR BLK	.00	3:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:22pm</b>
AIR BLK	.00	3:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:25pm</b>
AIR BLK	.00	3:26pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794      Test Record Number: 7474  
Test Date: 11/08/2021      Test Time: 3:27pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:28pm
FLO	Pass	3:28pm
FC	Pass	3:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:28pm
SRC	Pass	3:28pm
DET	Pass	3:28pm
BAR	Pass	3:28pm
BT	Pass	3:28pm

Blank Tests

Test	Status	Time
AIR	Pass	3:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:28pm

CRC Tests

Test	Status	Time
COMP	Pass	3:29pm
CAL	Pass	3:29pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Halifax Instrument Location Roanoke Rapids PD  
Instrument Serial No. 008656 1040 Roanoke Ave  
Roanoke Rapids, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. B. Smith  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656

Test Date: 11/22/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG018401

Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	12:20pm
AIR BLK	.00	12:20pm
ACCY CHK	.07	12:21pm
AIR BLK	.00	12:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:23pm</b>
AIR BLK	.00	12:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:25pm</b>
AIR BLK	.00	12:26pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656      Test Record Number: 796  
Test Date: 11/22/2021      Test Time: 12:27pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:27pm
FLO	Pass	12:27pm
FC	Pass	12:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:28pm
SRC	Pass	12:28pm
DET	Pass	12:28pm
BAR	Pass	12:28pm
BT	Pass	12:28pm

Blank Tests

Test	Status	Time
AIR	Pass	12:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:28pm

CRC Tests

Test	Status	Time
COMP	Pass	12:28pm
CAL	Pass	12:28pm

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Halifax Instrument Location Halifax Co. S.O.  
Instrument Serial No. 008695 335 Fern Hill Lane  
Halifax, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



EABryant

Signature of Certifying Official

671

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HALIFAX CO. HALIFAX CO SO 410

Serial Number: 008695  
Test Date: 11/01/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:50pm
AIR BLK	.00	12:50pm
ACCY CHK	.08	12:51pm
AIR BLK	.00	12:52pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:52pm</b>
AIR BLK	.00	12:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:55pm</b>
AIR BLK	.00	12:56pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO. HALIFAX CO SO 410

Serial Number: 008695      Test Record Number: 3256  
Test Date: 11/01/2021      Test Time: 12:57pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:57pm
FLO	Pass	12:57pm
FC	Pass	12:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:57pm
SRC	Pass	12:57pm
DET	Pass	12:57pm
BAR	Pass	12:57pm
BT	Pass	12:57pm

Blank Tests

Test	Status	Time
AIR	Pass	12:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:58pm

CRC Tests

Test	Status	Time
COMP	Pass	12:58pm
CAL	Pass	12:58pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Henderson Instrument Location Henderson Co. Jail

Instrument Serial No. 008806 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

668  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

HENDERSON COUNTY HENDERSON COUNTY  
DETENTION 440

Serial Number: 008806  
Test Date: 11/10/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Loftis, Benjamin C

Permit Number: 0024-4987

Effective:

09/17/2020-09/17/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG007601

Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	12:28pm
AIR BLK	.00	12:29pm
ACCY CHK	.07	12:30pm
AIR BLK	.00	12:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:31pm</b>
AIR BLK	.00	12:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:34pm</b>
AIR BLK	.00	12:35pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

HENDERSON COUNTY HENDERSON COUNTY DETENTION 440

Serial Number: 008806      Test Record Number: 3006  
Test Date: 11/10/2021      Test Time: 12:35pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:35pm
FLO	Pass	12:35pm
FC	Pass	12:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:36pm
SRC	Pass	12:36pm
DET	Pass	12:36pm
BAR	Pass	12:36pm
BT	Pass	12:36pm

Blank Tests

Test	Status	Time
AIR	Pass	12:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:36pm

CRC Tests

Test	Status	Time
COMP	Pass	12:37pm
CAL	Pass	12:37pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Henderson Instrument Location Henderson Co. Jail

Instrument Serial No. 009877 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

668  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HENDERSON COUNTY HENDERSON COUNTY  
DETENTION 440

Serial Number: 008822  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Loftis, Benjamin C  
Permit Number: 0024-4987  
Effective:  
09/17/2020-09/17/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:25pm
AIR BLK	.00	12:25pm
ACCY CHK	.07	12:26pm
AIR BLK	.00	12:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:28pm</b>
AIR BLK	.00	12:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:31pm</b>
AIR BLK	.00	12:32pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HENDERSON COUNTY HENDERSON COUNTY DETENTION 440

Serial Number: 008822      Test Record Number: 2998  
Test Date: 11/10/2021      Test Time: 12:32pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:33pm
SRC	Pass	12:33pm
DET	Pass	12:33pm
BAR	Pass	12:33pm
BT	Pass	12:33pm

Blank Tests

Test	Status	Time
AIR	Pass	12:33pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:33pm

CRC Tests

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Hoke Instrument Location Hoke County

Instrument Serial No. 008852 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008852  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG034901  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	12:36pm
AIR BLK	.00	12:36pm
ACCY CHK	.08	12:37pm
AIR BLK	.00	12:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:38pm</b>
AIR BLK	.00	12:39pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:41pm</b>
AIR BLK	.00	12:42pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008852      Test Record Number: 1219  
Test Date: 11/10/2021      Test Time: 12:43pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:43pm
FLO	Pass	12:43pm
FC	Pass	12:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:43pm
SRC	Pass	12:43pm
DET	Pass	12:43pm
BAR	Pass	12:43pm
BT	Pass	12:43pm

Blank Tests

Test	Status	Time
AIR	Pass	12:44pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:44pm

CRC Tests

Test	Status	Time
COMP	Pass	12:44pm
CAL	Pass	12:44pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Hoke Instrument Location Hoke County  
Instrument Serial No. 008855 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008855  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Garner, Joel K

Permit Number: 0036-5156

Effective:

03/10/2021-03/10/2023

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG034901

Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	12:29pm
AIR BLK	.00	12:30pm
ACCY CHK	.08	12:30pm
AIR BLK	.00	12:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:32pm</b>
AIR BLK	.00	12:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:34pm</b>
AIR BLK	.00	12:35pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008855      Test Record Number: 1623  
Test Date: 11/10/2021      Test Time: 12:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:37pm
FLO	Pass	12:37pm
FC	Pass	12:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:37pm
SRC	Pass	12:37pm
DET	Pass	12:37pm
BAR	Pass	12:37pm
BT	Pass	12:37pm

Blank Tests

Test	Status	Time
AIR	Pass	12:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:37pm

CRC Tests

Test	Status	Time
COMP	Pass	12:38pm
CAL	Pass	12:38pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County Hyde Instrument Location Hyde Co. S.O.

Instrument Serial No. 008801 1233 Main St, Swan Quarter, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official      643 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801

Test Date: 11/04/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	10:22am
AIR BLK	.00	10:22am
ACCY CHK	.07	10:23am
AIR BLK	.00	10:24am
<b>SUB TEST</b>	<b>.00</b>	<b>10:24am</b>
AIR BLK	.00	10:25am
<b>SUB TEST</b>	<b>.00</b>	<b>10:27am</b>
AIR BLK	.00	10:28am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801      Test Record Number: 623  
Test Date: 11/04/2021      Test Time: 10:29am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:29am
FLO	Pass	10:29am
FC	Pass	10:29am

Temperature Tests

Test	Status	Time
FC1	Pass	10:29am
SRC	Pass	10:29am
DET	Pass	10:29am
BAR	Pass	10:29am
BT	Pass	10:29am

Blank Tests

Test	Status	Time
AIR	Pass	10:30am


Printer Tests

Test	Status	Time
PRNT	Pass	10:30am

CRC Tests

Test	Status	Time
COMP	Pass	10:30am
CAL	Pass	10:30am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Iredell Instrument Location Iredell County SO  
Instrument Serial No. 008809 Statesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008809  
Test Date: 11/05/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG035001  
Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	12:52pm
AIR BLK	.00	12:53pm
ACCY CHK	.08	12:54pm
AIR BLK	.00	12:55pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:55pm</b>
AIR BLK	.00	12:56pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:58pm</b>
AIR BLK	.00	12:59pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

IREDELL COUNTY IREDELL COUNTY SO 480

Serial Number: 008809      Test Record Number: 4801  
Test Date: 11/05/2021      Test Time: 12:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:00pm
FLO	Pass	1:00pm
FC	Pass	1:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:00pm
SRC	Pass	1:00pm
DET	Pass	1:00pm
BAR	Pass	1:00pm
BT	Pass	1:00pm

Blank Tests

Test	Status	Time
AIR	Pass	1:01pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:01pm

CRC Tests

Test	Status	Time
COMP	Pass	1:01pm
CAL	Pass	1:01pm

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Iredell Instrument Location Statesville PD  
Instrument Serial No. 008927 Statesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryan Helms  
Signature of Certifying Official

074  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

IREDELL COUNTY STATESVILLE PD 480

Serial Number: 008927  
Test Date: 11/05/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	1:57pm
AIR BLK	.00	1:58pm
ACCY CHK	.08	1:59pm
AIR BLK	.00	2:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:00pm</b>
AIR BLK	.00	2:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:03pm</b>
AIR BLK	.00	2:04pm

Reported AC: .00 g/210L

Bryce Helms  
Signature of Chemical Analyst

Court CVR

Bryce Helms  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*IREDELL COUNTY STATESVILLE PD 480*

Serial Number: 008927      Test Record Number: 724  
Test Date: 11/05/2021      Test Time: 2:10pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	2:10pm
FLO	Pass	2:10pm
FC	Pass	2:11pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	2:11pm
SRC	Pass	2:11pm
DET	Pass	2:11pm
BAR	Pass	2:11pm
BT	Pass	2:11pm

**Blank Tests**

Test	Status	Time
AIR	Pass	2:11pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	2:11pm

**CRC Tests**

Test	Status	Time
COMP	Pass	2:12pm
CAL	Pass	2:12pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County JONES Instrument Location JONES COUNTY

Instrument Serial No. 008705 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Bowers

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JONES COUNTY DETENTION CENTER 510

Serial Number: 008705  
Test Date: 11/15/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, ALVIN R  
Permit Number: 0014-6279  
Effective:  
10/09/2020-10/09/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118804  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:01pm
AIR BLK	.00	2:02pm
ACCY CHK	.08	2:03pm
AIR BLK	.00	2:04pm
SUB TEST	.00	2:04pm
AIR BLK	.00	2:05pm
SUB TEST	.00	2:07pm
AIR BLK	.00	2:08pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JONES COUNTY DETENTION CENTER 510

Serial Number: 008705      Test Record Number: 1558  
Test Date: 11/15/2021      Test Time: 2:08pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:08pm
FLO	Pass	2:08pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:09pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:09pm

CRC Tests

Test	Status	Time
COMP	Pass	2:09pm
CAL	Pass	2:09pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County Lenoir Instrument Location Kinston P.D.

Instrument Serial No. 008624 205 E. King St, Kinston, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kerr  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624

Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	9:58am
AIR BLK	.00	9:58am
ACCY CHK	.08	9:59am
AIR BLK	.00	10:00am
<b>SUB TEST</b>	<b>.00</b>	<b>10:01am</b>
AIR BLK	.00	10:02am
<b>SUB TEST</b>	<b>.00</b>	<b>10:03am</b>
AIR BLK	.00	10:04am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

*LENOIR COUNTY KINSTON PD 530*

Serial Number: 008624      Test Record Number: 1972  
Test Date: 11/02/2021      Test Time: 10:05am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:05am
FLO	Pass	10:05am
FC	Pass	10:06am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:06am


Printer Tests

Test	Status	Time
PRNT	Pass	10:06am

CRC Tests

Test	Status	Time
COMP	Pass	10:07am
CAL	Pass	10:07am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Lenoir Instrument Location Lenoir Co. S.O.  
Instrument Serial No. 008639 130 Queen St, Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ken P

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639

Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	10:57am
AIR BLK	.00	10:58am
ACCY CHK	.08	10:58am
AIR BLK	.00	10:59am
<b>SUB TEST</b>	<b>.00</b>	<b>11:00am</b>
AIR BLK	.00	11:01am
<b>SUB TEST</b>	<b>.00</b>	<b>11:02am</b>
AIR BLK	.00	11:03am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639      Test Record Number: 3587  
Test Date: 11/02/2021      Test Time: 11:04am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:04am
FLO	Pass	11:04am
FC	Pass	11:05am

Temperature Tests

Test	Status	Time
FC1	Pass	11:05am
SRC	Pass	11:05am
DET	Pass	11:05am
BAR	Pass	11:05am
BT	Pass	11:05am

Blank Tests

Test	Status	Time
AIR	Pass	11:05am


Printer Tests

Test	Status	Time
PRNT	Pass	11:05am

CRC Tests

Test	Status	Time
COMP	Pass	11:05am
CAL	Pass	11:05am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Lincoln Instrument Location Lincoln Co Courthouse  
Instrument Serial No. 008823 Lincolnton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Williams  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008823

Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG107502

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	11:45am
AIR BLK	.00	11:46am
ACCY CHK	.08	11:46am
AIR BLK	.00	11:47am
<b>SUB TEST</b>	<b>.00</b>	<b>11:48am</b>
AIR BLK	.00	11:49am
<b>SUB TEST</b>	<b>.00</b>	<b>11:50am</b>
AIR BLK	.00	11:51am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008823      Test Record Number: 1731  
Test Date: 11/02/2021      Test Time: 11:53am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:53am
FLO	Pass	11:53am
FC	Pass	11:54am

Temperature Tests

Test	Status	Time
FC1	Pass	11:54am
SRC	Pass	11:54am
DET	Pass	11:54am
BAR	Pass	11:54am
BT	Pass	11:54am

Blank Tests

Test	Status	Time
AIR	Pass	11:54am


Printer Tests

Test	Status	Time
PRNT	Pass	11:54am

CRC Tests

Test	Status	Time
COMP	Pass	11:54am
CAL	Pass	11:54am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Lincoln Instrument Location Lincoln County Jail  
Instrument Serial No. 008827 Lincolnton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LINCOLN COUNTY LINCOLN COUNTY SO 540

Serial Number: 008827

Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118804


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:27pm
AIR BLK	.00	12:28pm
ACCY CHK	.07	12:29pm
AIR BLK	.00	12:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:31pm</b>
AIR BLK	.00	12:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:33pm</b>
AIR BLK	.00	12:34pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

LINCOLN COUNTY LINCOLN COUNTY SO 540

Serial Number: 008827      Test Record Number: 3645  
Test Date: 11/02/2021      Test Time: 12:35pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:36pm
FLO	Pass	12:36pm
FC	Pass	12:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:36pm
SRC	Pass	12:36pm
DET	Pass	12:36pm
BAR	Pass	12:36pm
BT	Pass	12:36pm

Blank Tests

Test	Status	Time
AIR	Pass	12:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:37pm

CRC Tests

Test	Status	Time
COMP	Pass	12:37pm
CAL	Pass	12:37pm

Preventive Maintenance  
Status: *Pass*

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Madison Instrument Location Morris Hill Police Dept.  
Instrument Serial No. 008587 Morris Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

668  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

MADISON COUNTY MARS HILL PD 560

Serial Number: 008582

Test Date: 11/24/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Loftis, Benjamin C

Permit Number: 0024-4987

Effective:

09/17/2020-09/17/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG035001

Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	10:50am
AIR BLK	.00	10:51am
ACCY CHK	.07	10:51am
AIR BLK	.00	10:52am
<b>SUB TEST</b>	<b>.00</b>	<b>10:53am</b>
AIR BLK	.00	10:54am
<b>SUB TEST</b>	<b>.00</b>	<b>10:55am</b>
AIR BLK	.00	10:56am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MADISON COUNTY MARS HILL PD 560

Serial Number: 008582      Test Record Number: 1194  
Test Date: 11/24/2021      Test Time: 10:56am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:57am
FLO	Pass	10:57am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:58am


Printer Tests

Test	Status	Time
PRNT	Pass	10:58am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location CMPD LEC  
Instrument Serial No. 008594 Charlotte, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008594  
Test Date: 11/17/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG034902  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	12:38pm
AIR BLK	.00	12:39pm
ACCY CHK	.07	12:39pm
AIR BLK	.00	12:41pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:42pm</b>
AIR BLK	.00	12:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:44pm</b>
AIR BLK	.00	12:45pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008594      Test Record Number: 5027  
Test Date: 11/17/2021      Test Time: 12:47pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:47pm
FLO	Pass	12:47pm
FC	Pass	12:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:47pm
SRC	Pass	12:47pm
DET	Pass	12:47pm
BAR	Pass	12:47pm
BT	Pass	12:47pm

Blank Tests

Test	Status	Time
AIR	Pass	12:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:48pm

CRC Tests

Test	Status	Time
COMP	Pass	12:48pm
CAL	Pass	12:48pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location Mecklenburg County 80  
Instrument Serial No. 0081665 Charlotte, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY SHERIFFS OFFICE 590

Serial Number: 008665  
Test Date: 11/17/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118804


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:23pm
AIR BLK	.00	1:24pm
ACCY CHK	.08	1:25pm
AIR BLK	.00	1:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:26pm</b>
AIR BLK	.00	1:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:29pm</b>
AIR BLK	.00	1:29pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY SHERIFFS OFFICE 590

Serial Number: 008665      Test Record Number: 5045  
Test Date: 11/17/2021      Test Time: 1:30pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:30pm
FLO	Pass	1:30pm
FC	Pass	1:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:30pm
SRC	Pass	1:30pm
DET	Pass	1:30pm
BAR	Pass	1:30pm
BT	Pass	1:30pm

Blank Tests

Test	Status	Time
AIR	Pass	1:31pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:31pm

CRC Tests

Test	Status	Time
COMP	Pass	1:31pm
CAL	Pass	1:31pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location Mecklenburg County 80  
Instrument Serial No. 00820910 Charlotte, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryon Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY SHERIFFS OFFICE 590

Serial Number: 008690  
Test Date: 11/17/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG107401  
Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	1:18pm
AIR BLK	.00	1:19pm
ACCY CHK	.08	1:19pm
AIR BLK	.00	1:20pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:21pm</b>
AIR BLK	.00	1:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:23pm</b>
AIR BLK	.00	1:24pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY SHERIFFS OFFICE 590

Serial Number: 008690      Test Record Number: 6851  
Test Date: 11/17/2021      Test Time: 1:25pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:25pm
SRC	Pass	1:25pm
DET	Pass	1:25pm
BAR	Pass	1:25pm
BT	Pass	1:25pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location Matthews PD  
Instrument Serial No. 0086019 Matthews, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3<sup>rd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Heems  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: 008699  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG034902  
Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	1:04pm
AIR BLK	.00	1:05pm
ACCY CHK	.07	1:06pm
AIR BLK	.00	1:07pm
SUB TEST	.00	1:07pm
AIR BLK	.00	1:08pm
SUB TEST	.00	1:10pm
AIR BLK	.00	1:11pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: 008699      Test Record Number: 2940  
Test Date: 11/03/2021      Test Time: 1:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:11pm
FLO	Pass	1:11pm
FC	Pass	1:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:11pm
SRC	Pass	1:11pm
DET	Pass	1:11pm
BAR	Pass	1:11pm
BT	Pass	1:11pm

Blank Tests

Test	Status	Time
AIR	Pass	1:12pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:12pm

CRC Tests

Test	Status	Time
COMP	Pass	1:12pm
CAL	Pass	1:12pm

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location Pineville Police Department  
Instrument Serial No. 008703 Pineville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3<sup>rd</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms 674  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY PINEVILLE PD 590

Serial Number: 008703  
Test Date: 11/03/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Helms, Bryce A  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG035001  
Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:13pm
ACCY CHK	.08	12:13pm
AIR BLK	.00	12:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:15pm</b>
AIR BLK	.00	12:16pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:17pm</b>
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY PINEVILLE PD 590

Serial Number: 008703      Test Record Number: 6001  
Test Date: 11/03/2021      Test Time: 12:19pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:19pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:20pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:20pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Mecklenburg Instrument Location Huntersville PD  
Instrument Serial No. 008747 Huntersville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747

Test Date: 11/05/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG035001


Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	11:15am
AIR BLK	.00	11:16am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
<b>SUB TEST</b>	<b>.00</b>	<b>11:18am</b>
AIR BLK	.00	11:19am
<b>SUB TEST</b>	<b>.00</b>	<b>11:21am</b>
AIR BLK	.00	11:22am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747      Test Record Number: 2878  
Test Date: 11/05/2021      Test Time: 11:25am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:26am
FLO	Pass	11:26am
FC	Pass	11:26am

Temperature Tests

Test	Status	Time
FC1	Pass	11:26am
SRC	Pass	11:26am
DET	Pass	11:26am
BAR	Pass	11:26am
BT	Pass	11:26am

Blank Tests

Test	Status	Time
AIR	Pass	11:26am


Printer Tests

Test	Status	Time
PRNT	Pass	11:27am

CRC Tests

Test	Status	Time
COMP	Pass	11:27am
CAL	Pass	11:27am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Moore Instrument Location Moore County  
Instrument Serial No. 008735 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MOORE COUNTY DETENTION CENTER 620

Serial Number: 008735  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG107401  
Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	3:01pm
AIR BLK	.00	3:02pm
ACCY CHK	.08	3:03pm
AIR BLK	.00	3:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:05pm</b>
AIR BLK	.00	3:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:07pm</b>
AIR BLK	.00	3:08pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

MOORE COUNTY DETENTION CENTER 620

Serial Number: 008735      Test Record Number: 2709  
Test Date: 11/23/2021      Test Time: 3:09pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:09pm
FLO	Pass	3:09pm
FC	Pass	3:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:10pm

CRC Tests

Test	Status	Time
COMP	Pass	3:10pm
CAL	Pass	3:10pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County New Hanover Instrument Location New Hanover County  
Instrument Serial No. 008917 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]

Signature of Certifying Official

670

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY DETENTION CENTER  
640

Serial Number: 008917  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Ryan, Robert F  
Permit Number: 0084-5023  
Effective:  
07/31/2020-07/31/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG107501  
Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:47am
ACCY CHK	.07	11:47am
AIR BLK	.00	11:48am
<b>SUB TEST</b>	<b>.00</b>	<b>11:49am</b>
AIR BLK	.00	11:50am
<b>SUB TEST</b>	<b>.00</b>	<b>11:51am</b>
AIR BLK	.00	11:53am

Reported AC<sub>0</sub> .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

NEW HANOVER COUNTY DETENTION CENTER 640

Serial Number: 008917 Test Record Number: 926  
Test Date: 11/12/2021 Test Time: 12:01pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:02pm
FLO	Pass	12:02pm
FC	Pass	12:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:02pm
SRC	Pass	12:02pm
DET	Pass	12:02pm
BAR	Pass	12:02pm
BT	Pass	12:02pm

Blank Tests

Test	Status	Time
AIR	Pass	12:02pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:02pm

CRC Tests

Test	Status	Time
COMP	Pass	12:03pm
CAL	Pass	12:03pm

Preventive Maintenance  
Status: Pass



---

Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Northampton Instrument Location Northampton Co. S.O.  
Instrument Serial No. 008607 105 West Jefferson St  
Jackson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E.A. Baylath

Signature of Certifying Official

671

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT  
650

Serial Number: 008607  
Test Date: 11/04/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118804

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:08pm
AIR BLK	.00	1:08pm
ACCY CHK	.07	1:09pm
AIR BLK	.00	1:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:11pm</b>
AIR BLK	.00	1:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:13pm</b>
AIR BLK	.00	1:14pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT 650

Serial Number: 008607      Test Record Number: 1129  
Test Date: 11/04/2021      Test Time: 1:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:16pm
FLO	Pass	1:16pm
FC	Pass	1:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:17pm
SRC	Pass	1:17pm
DET	Pass	1:17pm
BAR	Pass	1:17pm
BT	Pass	1:17pm

Blank Tests

Test	Status	Time
AIR	Pass	1:17pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:17pm

CRC Tests

Test	Status	Time
COMP	Pass	1:17pm
CAL	Pass	1:17pm

Preventive Maintenance  
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County NORTHAMPTON Instrument Location NORTHAMPTON CO S.O.  
Instrument Serial No. 008688 105 WEST JEFFERSON ST  
JACKSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of NOVEMBER, 20     the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**NORTHAMPTON COUNTY SHERIFFS DEPARTMENT**  
650

Serial Number: 008688  
Test Date: 11/04/2021

Citation Number: M0000000-0

Subject's Name:

*PREVENTIVE, MAINTENANCE*

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: *Barnes, Simon S*

Permit Number: 0014-6221

Effective:

09/01/2020-09/01/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: *Breath Test*

Lot Number: AG118804

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:12pm
AIR BLK	.00	1:12pm
ACCY CHK	.07	1:13pm
AIR BLK	.00	1:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:16pm</b>
AIR BLK	.00	1:16pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:18pm</b>
AIR BLK	.00	1:19pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**  
**Forensic Tests for Alcohol Branch**  
**Department of Health and Human Services**  
**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**NORTHAMPTON COUNTY SHERIFFS DEPARTMENT 650**

Serial Number: 008688      Test Record Number: 961  
Test Date: 11/04/2021      Test Time: 1:20pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:20pm
FLO	Pass	1:20pm
FC	Pass	1:20pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:20pm
SRC	Pass	1:20pm
DET	Pass	1:20pm
BAR	Pass	1:20pm
BT	Pass	1:20pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:21pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:21pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:21pm
CAL	Pass	1:21pm

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Orange Instrument Location Hillsborough PD  
Instrument Serial No. 008799 127 Churton St  
Hillsborough, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. B. Smith

Signature of Certifying Official

671

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799

Test Date: 11/29/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118804

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:06pm
AIR BLK	.00	2:07pm
ACCY CHK	.08	2:07pm
AIR BLK	.00	2:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:09pm</b>
AIR BLK	.00	2:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:12pm</b>
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799      Test Record Number: 3607  
Test Date: 11/29/2021      Test Time: 2:13pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:14pm
FLO	Pass	2:14pm
FC	Pass	2:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:14pm
SRC	Pass	2:14pm
DET	Pass	2:14pm
BAR	Pass	2:14pm
BT	Pass	2:14pm

Blank Tests

Test	Status	Time
AIR	Pass	2:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:14pm

CRC Tests

Test	Status	Time
COMP	Pass	2:15pm
CAL	Pass	2:15pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County ORANGE Instrument Location CARRBORO PD

Instrument Serial No. 008945 100 N. Greensboro St  
CARRBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CARRBORO PD 670

Serial Number: 008945

Test Date: 11/01/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Barnes, Simon S

Permit Number: 0014-6221

Effective:

09/01/2020-09/01/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG007601


Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	10:47am
AIR BLK	.00	10:49am
ACCY CHK	.07	10:49am
AIR BLK	.00	10:50am
SUB TEST	.00	10:51am
AIR BLK	.00	10:52am
SUB TEST	.00	10:54am
AIR BLK	.00	10:55am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ORANGE COUNTY CARRBORO PD 670

Serial Number: 008945      Test Record Number: 497  
Test Date: 11/01/2021      Test Time: 10:56am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:56am
FLO	Pass	10:56am
FC	Pass	10:56am

Temperature Tests

Test	Status	Time
FC1	Pass	10:56am
SRC	Pass	10:56am
DET	Pass	10:56am
BAR	Pass	10:56am
BT	Pass	10:56am

Blank Tests

Test	Status	Time
AIR	Pass	10:57am

Printer Tests

Test	Status	Time
PRNT	Pass	10:57am

CRC Tests

Test	Status	Time
COMP	Pass	10:57am
CAL	Pass	10:57am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County PAMLICO Instrument Location PAMLICO COUNTY  
Instrument Serial No. 008640 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Bann  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PAMLICO COUNTY DETENTION CENTER 680

Serial Number: 008640

Test Date: 11/23/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 0014-6279

Effective:

10/09/2020-10/09/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG034901

Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	12:16pm
AIR BLK	.00	12:17pm
ACCY CHK	.08	12:17pm
AIR BLK	.00	12:18pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:19pm</b>
AIR BLK	.00	12:20pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:21pm</b>
AIR BLK	.00	12:22pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*PAMLICO COUNTY DETENTION CENTER 680*

Serial Number: 008640      Test Record Number: 1511  
Test Date: 11/23/2021      Test Time: 12:22pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:23pm
FLO	Pass	12:23pm
FC	Pass	12:23pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:23pm
SRC	Pass	12:23pm
DET	Pass	12:23pm
BAR	Pass	12:23pm
BT	Pass	12:23pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:23pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:23pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:24pm
CAL	Pass	12:24pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Pasquotank Instrument Location Elizabeth City, P.D.  
Instrument Serial No. 008941 315 Main St., Elizabeth City,  
NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941

Test Date: 11/24/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG034901

Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.07	11:33am
AIR BLK	.00	11:34am
<b>SUB TEST</b>	<b>.00</b>	<b>11:36am</b>
AIR BLK	.00	11:36am
<b>SUB TEST</b>	<b>.00</b>	<b>11:38am</b>
AIR BLK	.00	11:39am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941      Test Record Number: 1517  
Test Date: 11/24/2021      Test Time: 11:40am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40am
FLO	Pass	11:40am
FC	Pass	11:40am

Temperature Tests

Test	Status	Time
FC1	Pass	11:40am
SRC	Pass	11:40am
DET	Pass	11:40am
BAR	Pass	11:40am
BT	Pass	11:40am

Blank Tests

Test	Status	Time
AIR	Pass	11:41am


Printer Tests

Test	Status	Time
PRNT	Pass	11:41am

CRC Tests

Test	Status	Time
COMP	Pass	11:41am
CAL	Pass	11:41am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Pasquotank Instrument Location Pasquotank Co. Public Safety

Instrument Serial No. 008950 Bldg. 200 E. Colonial St.,  
Elizabeth City, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PASQUOTANK COUNTY PUBLIC SAFETY BLDG  
690

Serial Number: 008950  
Test Date: 11/24/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG035001  
Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	10:38am
AIR BLK	.00	10:38am
ACCY CHK	.08	10:39am
AIR BLK	.00	10:40am
<b>SUB TEST</b>	<b>.00</b>	<b>10:42am</b>
AIR BLK	.00	10:42am
<b>SUB TEST</b>	<b>.00</b>	<b>10:44am</b>
AIR BLK	.00	10:45am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008950      Test Record Number: 1890  
Test Date: 11/24/2021      Test Time: 10:46am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:46am
FLO	Pass	10:46am
FC	Pass	10:47am

Temperature Tests

Test	Status	Time
FC1	Pass	10:47am
SRC	Pass	10:47am
DET	Pass	10:47am
BAR	Pass	10:47am
BT	Pass	10:47am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am


Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:47am
CAL	Pass	10:47am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Pitt Instrument Location Ayden P.D.  
Instrument Serial No. 0086666 4144 West Ave., Ayden, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kun Signature of Certifying Official 643 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

PITT COUNTY AYDEN PD 730

Serial Number: 008666

Test Date: 11/08/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107501


Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	10:52am
AIR BLK	.00	10:53am
ACCY CHK	.08	10:54am
AIR BLK	.00	10:55am
<b>SUB TEST</b>	<b>.00</b>	<b>10:56am</b>
AIR BLK	.00	10:57am
<b>SUB TEST</b>	<b>.00</b>	<b>10:58am</b>
AIR BLK	.00	10:59am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

PITT COUNTY AYDEN PD 730

Serial Number: 008666      Test Record Number: 1363  
Test Date: 11/08/2021      Test Time: 11:00am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:00am
FLO	Pass	11:00am
FC	Pass	11:01am

Temperature Tests

Test	Status	Time
FC1	Pass	11:01am
SRC	Pass	11:01am
DET	Pass	11:01am
BAR	Pass	11:01am
BT	Pass	11:01am

Blank Tests

Test	Status	Time
AIR	Pass	11:01am


Printer Tests

Test	Status	Time
PRNT	Pass	11:01am

CRC Tests

Test	Status	Time
COMP	Pass	11:01am
CAL	Pass	11:01am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Randolph Instrument Location Randolph County  
Instrument Serial No. 008860 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008860  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118805  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:05pm
AIR BLK	.00	12:06pm
ACCY CHK	.08	12:07pm
AIR BLK	.00	12:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:08pm</b>
AIR BLK	.00	12:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:11pm</b>
AIR BLK	.00	12:12pm

Reported/AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008860      Test Record Number: 3046  
Test Date: 11/12/2021      Test Time: 12:13pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:14pm
FLO	Pass	12:14pm
FC	Pass	12:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:14pm
SRC	Pass	12:14pm
DET	Pass	12:14pm
BAR	Pass	12:14pm
BT	Pass	12:14pm

Blank Tests

Test	Status	Time
AIR	Pass	12:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:14pm

CRC Tests

Test	Status	Time
COMP	Pass	12:15pm
CAL	Pass	12:15pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County Randolph Instrument Location Randleman Police Department

Instrument Serial No. 008737

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY RANDLEMAN PD 750

Serial Number: 008737  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	10:48am
AIR BLK	.00	10:49am
ACCY CHK	.08	10:50am
AIR BLK	.00	10:51am
<b>SUB TEST</b>	<b>.00</b>	<b>10:51am</b>
AIR BLK	.00	10:52am
<b>SUB TEST</b>	<b>.00</b>	<b>10:53am</b>
AIR BLK	.00	10:54am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY RANDLEMAN PD 750

Serial Number: 008737      Test Record Number: 1310  
Test Date: 11/12/2021      Test Time: 10:55am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:55am
FLO	Pass	10:55am
FC	Pass	10:56am

Temperature Tests

Test	Status	Time
FC1	Pass	10:56am
SRC	Pass	10:56am
DET	Pass	10:56am
BAR	Pass	10:56am
BT	Pass	10:56am

Blank Tests

Test	Status	Time
AIR	Pass	10:56am


Printer Tests

Test	Status	Time
PRNT	Pass	10:56am

CRC Tests

Test	Status	Time
COMP	Pass	10:56am
CAL	Pass	10:56am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Randolph Instrument Location Liberty Police Department  
Instrument Serial No. 008830 Liberty, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY LIBERTY PD 750

Serial Number: 008830  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Garner, Joel K

Permit Number: 0036-5156

Effective:

03/10/2021-03/10/2023

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107401


Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	9:14am
AIR BLK	.00	9:15am
ACCY CHK	.08	9:15am
AIR BLK	.00	9:16am
<b>SUB TEST</b>	<b>.00</b>	<b>9:17am</b>
AIR BLK	.00	9:18am
<b>SUB TEST</b>	<b>.00</b>	<b>9:19am</b>
AIR BLK	.00	9:20am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY LIBERTY PD 750

Serial Number: 008830      Test Record Number: 733  
Test Date: 11/12/2021      Test Time: 9:26am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:26am
FLO	Pass	9:26am
FC	Pass	9:26am

Temperature Tests

Test	Status	Time
FC1	Pass	9:26am
SRC	Pass	9:26am
DET	Pass	9:26am
BAR	Pass	9:26am
BT	Pass	9:26am

Blank Tests

Test	Status	Time
AIR	Pass	9:27am


Printer Tests

Test	Status	Time
PRNT	Pass	9:27am

CRC Tests

Test	Status	Time
COMP	Pass	9:27am
CAL	Pass	9:27am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Randolph Instrument Location Randolph County  
Instrument Serial No. 008899 Detention Center

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008899  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	12:22pm
AIR BLK	.00	12:23pm
ACCY CHK	.08	12:23pm
AIR BLK	.00	12:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:25pm</b>
AIR BLK	.00	12:26pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:27pm</b>
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008899      Test Record Number: 3742  
Test Date: 11/12/2021      Test Time: 12:29pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:29pm
FLO	Pass	12:29pm
FC	Pass	12:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:30pm
SRC	Pass	12:30pm
DET	Pass	12:30pm
BAR	Pass	12:30pm
BT	Pass	12:30pm

Blank Tests

Test	Status	Time
AIR	Pass	12:30pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:30pm

CRC Tests

Test	Status	Time
COMP	Pass	12:30pm
CAL	Pass	12:30pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County RANDOLPH Instrument Location ARCHDALE POLICE

Instrument Serial No. 008791 DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>RD</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: *Oligmueller Jr., Leo A*  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:19pm
AIR BLK	.00	1:20pm
ACCY CHK	.08	1:20pm
AIR BLK	.00	1:21pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:22pm</b>
AIR BLK	.00	1:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:24pm</b>
AIR BLK	.00	1:25pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst  
Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791      Test Record Number: 1474  
Test Date: 11/23/2021      Test Time: 1:27pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:27pm
FLO	Pass	1:27pm
FC	Pass	1:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:27pm
SRC	Pass	1:27pm
DET	Pass	1:27pm
BAR	Pass	1:27pm
BT	Pass	1:27pm

Blank Tests

Test	Status	Time
AIR	Pass	1:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:28pm

CRC Tests

Test	Status	Time
COMP	Pass	1:28pm
CAL	Pass	1:28pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**


County Robeson Instrument Location Pembroke Police

Instrument Serial No. 008837 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROBESON COUNTY PEMBROKE POLICE DEPT  
770

Serial Number: 008837  
Test Date: 11/08/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	9:52am
AIR BLK	.00	9:53am
ACCY CHK	.07	9:53am
AIR BLK	.00	9:54am
<b>SUB TEST</b>	<b>.00</b>	<b>9:55am</b>
AIR BLK	.00	9:56am
<b>SUB TEST</b>	<b>.00</b>	<b>9:57am</b>
AIR BLK	.00	9:58am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY PEMBROKE POLICE DEPT 770

Serial Number: 008837      Test Record Number: 1132  
Test Date: 11/08/2021      Test Time: 9:59am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:00am
FLO	Pass	10:00am
FC	Pass	10:00am

Temperature Tests

Test	Status	Time
FC1	Pass	10:00am
SRC	Pass	10:00am
DET	Pass	10:00am
BAR	Pass	10:00am
BT	Pass	10:00am

Blank Tests

Test	Status	Time
AIR	Pass	10:01am

Printer Tests

Test	Status	Time
PRNT	Pass	10:01am

CRC Tests

Test	Status	Time
COMP	Pass	10:01am
CAL	Pass	10:01am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County ROWAN Instrument Location SALISBURY POLICE  
Instrument Serial No. 008835 DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:36pm
AIR BLK	.00	2:37pm
ACCY CHK	.08	2:38pm
AIR BLK	.00	2:39pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:39pm</b>
AIR BLK	.00	2:40pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:42pm</b>
AIR BLK	.00	2:43pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835      Test Record Number: 2802  
Test Date: 11/10/2021      Test Time: 2:46pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:46pm
FLO	Pass	2:46pm
FC	Pass	2:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:46pm
SRC	Pass	2:46pm
DET	Pass	2:46pm
BAR	Pass	2:46pm
BT	Pass	2:46pm

Blank Tests

Test	Status	Time
AIR	Pass	2:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:47pm

CRC Tests

Test	Status	Time
COMP	Pass	2:47pm
CAL	Pass	2:47pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County ROWAN Instrument Location SALISBURY POLICE

Instrument Serial No. 008868 DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10<sup>TH</sup> day of NOVEMBER, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

669  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008868  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Oligmueller Jr., Leo A  
Permit Number: 0035-3799  
Effective:  
08/04/2020-08/04/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118803  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:44pm
AIR BLK	.00	2:45pm
ACCY CHK	.08	2:46pm
AIR BLK	.00	2:47pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:48pm</b>
AIR BLK	.00	2:49pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:50pm</b>
AIR BLK	.00	2:51pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008868      Test Record Number: 3314  
Test Date: 11/10/2021      Test Time: 2:53pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:53pm
FLO	Pass	2:53pm
FC	Pass	2:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:53pm
SRC	Pass	2:53pm
DET	Pass	2:53pm
BAR	Pass	2:53pm
BT	Pass	2:53pm

Blank Tests

Test	Status	Time
AIR	Pass	2:54pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:54pm

CRC Tests

Test	Status	Time
COMP	Pass	2:54pm
CAL	Pass	2:54pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Scotland Instrument Location Laurinburg Police  
Instrument Serial No. 008834 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834

Test Date: 11/10/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Garner, Joel K

Permit Number: 0036-5156

Effective:

03/10/2021-03/10/2023

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	11:09am
AIR BLK	.00	11:10am
ACCY CHK	.08	11:10am
AIR BLK	.00	11:12am
<b>SUB TEST</b>	<b>.00</b>	<b>11:12am</b>
AIR BLK	.00	11:13am
<b>SUB TEST</b>	<b>.00</b>	<b>11:15am</b>
AIR BLK	.00	11:16am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834      Test Record Number: 1035  
Test Date: 11/10/2021      Test Time: 11:16am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:17am
FLO	Pass	11:17am
FC	Pass	11:17am

Temperature Tests

Test	Status	Time
FC1	Pass	11:17am
SRC	Pass	11:17am
DET	Pass	11:17am
BAR	Pass	11:17am
BT	Pass	11:17am

Blank Tests

Test	Status	Time
AIR	Pass	11:18am


Printer Tests

Test	Status	Time
PRNT	Pass	11:18am

CRC Tests

Test	Status	Time
COMP	Pass	11:18am
CAL	Pass	11:18am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Scotland Instrument Location Scotland Co.

Instrument Serial No. 008861 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

675  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Garner, Joel K  
Permit Number: 0036-5156  
Effective:  
03/10/2021-03/10/2023

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	9:50am
AIR BLK	.00	9:50am
ACCY CHK	.08	9:51am
AIR BLK	.00	9:52am
<b>SUB TEST</b>	<b>.00</b>	<b>9:53am</b>
AIR BLK	.00	9:54am
<b>SUB TEST</b>	<b>.00</b>	<b>9:55am</b>
AIR BLK	.00	9:56am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861      Test Record Number: 1775  
Test Date: 11/10/2021      Test Time: 9:57am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:58am
FLO	Pass	9:58am
FC	Pass	9:58am

Temperature Tests

Test	Status	Time
FC1	Pass	9:58am
SRC	Pass	9:58am
DET	Pass	9:58am
BAR	Pass	9:58am
BT	Pass	9:58am

Blank Tests

Test	Status	Time
AIR	Pass	9:58am


Printer Tests

Test	Status	Time
PRNT	Pass	9:58am

CRC Tests

Test	Status	Time
COMP	Pass	9:59am
CAL	Pass	9:59am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Stanly Instrument Location Locust PD  
Instrument Serial No. 008704 Locust, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

STANLY LOCUST PD 830

Serial Number: 008706

Test Date: 11/04/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118804

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	11:55am
AIR BLK	.00	11:56am
ACCY CHK	.08	11:56am
AIR BLK	.00	11:58am
<b>SUB TEST</b>	<b>.00</b>	<b>11:58am</b>
AIR BLK	.00	11:59am
<b>SUB TEST</b>	<b>.00</b>	<b>12:01pm</b>
AIR BLK	.00	12:02pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

STANLY LOCUST PD 830

Serial Number: 008706      Test Record Number: 3608  
Test Date: 11/04/2021      Test Time: 12:04pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:04pm
FLO	Pass	12:04pm
FC	Pass	12:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:04pm
SRC	Pass	12:04pm
DET	Pass	12:04pm
BAR	Pass	12:04pm
BT	Pass	12:04pm

Blank Tests

Test	Status	Time
AIR	Pass	12:05pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:05pm

CRC Tests

Test	Status	Time
COMP	Pass	12:05pm
CAL	Pass	12:05pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Stanly Instrument Location Stanly Co SO  
Instrument Serial No. 008824 Albemarle, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Boye Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008824  
Test Date: 11/04/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Helms, Bryce A

Permit Number: 0084-9845

Effective:

04/20/2021-04/20/2023

Officer's Name: Helms, Bryce A

Type of Agency: CP

Agency: Appstate

Test Type: Breath Test

Lot Number: AG107501

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:52am
ACCY CHK	.08	10:52am
AIR BLK	.00	10:54am
<b>SUB TEST</b>	<b>.00</b>	<b>10:55am</b>
AIR BLK	.00	10:56am
<b>SUB TEST</b>	<b>.00</b>	<b>10:57am</b>
AIR BLK	.00	10:58am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008824      Test Record Number: 1788  
Test Date: 11/04/2021      Test Time: 10:59am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:59am
FLO	Pass	10:59am
FC	Pass	10:59am

Temperature Tests

Test	Status	Time
FC1	Pass	10:59am
SRC	Pass	10:59am
DET	Pass	10:59am
BAR	Pass	10:59am
BT	Pass	10:59am

Blank Tests

Test	Status	Time
AIR	Pass	11:00am


Printer Tests

Test	Status	Time
PRNT	Pass	11:00am

CRC Tests

Test	Status	Time
COMP	Pass	11:00am
CAL	Pass	11:00am

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Stanly Instrument Location Stanly County SD  
Instrument Serial No. 008842 Albemarle, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bryce Helms  
Signature of Certifying Official

674  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008842  
Test Date: 11/04/2021

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: *Helms, Bryce A*  
Permit Number: 0084-9845  
Effective:  
04/20/2021-04/20/2023

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: *Breath Test*

Lot Number: AG107501  
Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	10:48am
AIR BLK	.00	10:48am
ACCY CHK	.08	10:49am
AIR BLK	.00	10:50am
<b>SUB TEST</b>	<b>.00</b>	<b>10:51am</b>
AIR BLK	.00	10:52am
<b>SUB TEST</b>	<b>.00</b>	<b>10:53am</b>
AIR BLK	.00	10:54am

Reported AC: .00 g/210L

*Bryce Helms*  
Signature of Chemical Analyst

Court CVR

*Bryce Helms*  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008842      Test Record Number: 2465  
Test Date: 11/04/2021      Test Time: 10:57am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:57am
FLO	Pass	10:57am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:58am


Printer Tests

Test	Status	Time
PRNT	Pass	10:58am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Surry Instrument Location Surry Co Jail  
Dobson, Nc  
Instrument Serial No. 008660

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Timothy Hanko  
Signature of Certifying Official

672  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SURRY COUNTY SURRY CO JAIL 850

Serial Number: 008660  
Test Date: 11/23/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Hanks, Timothy S  
Permit Number: 0063-3175  
Effective:  
10/14/2020-10/14/2022

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG007601  
Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	7:08pm
AIR BLK	.00	7:09pm
ACCY CHK	.07	7:09pm
AIR BLK	.00	7:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:11pm</b>
AIR BLK	.00	7:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:13pm</b>
AIR BLK	.00	7:14pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*SURRY COUNTY SURRY CO JAIL 850*

Serial Number: 008660      Test Record Number: 4447  
Test Date: 11/23/2021      Test Time: 7:18pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:18pm
FLO	Pass	7:18pm
FC	Pass	7:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:18pm
SRC	Pass	7:18pm
DET	Pass	7:18pm
BAR	Pass	7:18pm
BT	Pass	7:18pm

Blank Tests

Test	Status	Time
AIR	Pass	7:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	7:19pm

CRC Tests

Test	Status	Time
COMP	Pass	7:19pm
CAL	Pass	7:19pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Swain Instrument Location Swain Co. Jail  
Instrument Serial No. 008808 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuthbert

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

SWAIN COUNTY JAIL 860

Serial Number: 008808  
Test Date: 11/18/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Cutler, Daniel R

Permit Number: 0084-3310

Effective:

09/17/2020-09/17/2022

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG018401

Exp Date: 07/02/2022

Test	g/210L	Time
DIAG	Pass	9:54am
AIR BLK	.00	9:55am
ACCY CHK	.07	9:56am
AIR BLK	.00	9:57am
SUB TEST	.00	9:58am
AIR BLK	.00	9:59am
SUB TEST	.00	10:01am
AIR BLK	.00	10:02am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

SWAIN COUNTY JAIL 860

Serial Number: 008808      Test Record Number: 1491  
Test Date: 11/18/2021      Test Time: 10:02am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02am
FLO	Pass	10:02am
FC	Pass	10:03am

Temperature Tests

Test	Status	Time
FC1	Pass	10:03am
SRC	Pass	10:03am
DET	Pass	10:03am
BAR	Pass	10:03am
BT	Pass	10:03am

Blank Tests

Test	Status	Time
AIR	Pass	10:03am

Printer Tests

Test	Status	Time
PRNT	Pass	10:03am

CRC Tests

Test	Status	Time
COMP	Pass	10:04am
CAL	Pass	10:04am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Transylvania Instrument Location Transylvania Co. Jail  
Instrument Serial No. 008609 Bevard, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

668  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

TRANSYLVANIA COUNTY TRANSYLVANIA CO  
JAIL 870

Serial Number: 008609

Test Date: 11/10/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Loftis, Benjamin C

Permit Number: 0024-4987

Effective:

09/17/2020-09/17/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG107401

Exp Date: 03/15/2023

Test	g/210L	Time
DIAG	Pass	10:57am
AIR BLK	.00	10:58am
ACCY CHK	.08	10:58am
AIR BLK	.00	10:59am
<b>SUB TEST</b>	<b>.00</b>	<b>11:00am</b>
AIR BLK	.00	11:01am
<b>SUB TEST</b>	<b>.00</b>	<b>11:02am</b>
AIR BLK	.00	11:03am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

TRANSYLVANIA COUNTY TRANSYLVANIA CO JAIL 870

Serial Number: 008609      Test Record Number: 1000  
Test Date: 11/10/2021      Test Time: 11:03am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:04am
FLO	Pass	11:04am
FC	Pass	11:04am

Temperature Tests

Test	Status	Time
FC1	Pass	11:04am
SRC	Pass	11:04am
DET	Pass	11:04am
BAR	Pass	11:04am
BT	Pass	11:04am

Blank Tests

Test	Status	Time
AIR	Pass	11:05am

Printer Tests

Test	Status	Time
PRNT	Pass	11:05am

CRC Tests

Test	Status	Time
COMP	Pass	11:05am
CAL	Pass	11:05am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Transylvania Instrument Location Transylvania Co. Jail

Instrument Serial No. 008870 Bevard, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

6608  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

TRANSYLVANIA COUNTY TRANSYLVANIA CO  
JAIL 870

Serial Number: 008820  
Test Date: 11/10/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: Loftis, Benjamin C  
Permit Number: 0024-4987  
Effective:  
09/17/2020-09/17/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG035001  
Exp Date: 12/15/2022

Test	g/210L	Time
DIAG	Pass	10:56am
AIR BLK	.00	10:57am
ACCY CHK	.07	10:58am
AIR BLK	.00	10:59am
<b>SUB TEST</b>	<b>.00</b>	<b>10:59am</b>
AIR BLK	.00	11:00am
<b>SUB TEST</b>	<b>.00</b>	<b>11:01am</b>
AIR BLK	.00	11:02am

Reported AG: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

TRANSYLVANIA COUNTY TRANSYLVANIA CO JAIL 870

Serial Number: 008820      Test Record Number: 1412  
Test Date: 11/10/2021      Test Time: 11:03am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:03am
FLO	Pass	11:03am
FC	Pass	11:03am

Temperature Tests

Test	Status	Time
FC1	Pass	11:03am
SRC	Pass	11:03am
DET	Pass	11:03am
BAR	Pass	11:03am
BT	Pass	11:03am

Blank Tests

Test	Status	Time
AIR	Pass	11:04am

Printer Tests

Test	Status	Time
PRNT	Pass	11:04am

CRC Tests

Test	Status	Time
COMP	Pass	11:04am
CAL	Pass	11:04am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Vance Instrument Location Vance Co. S.O.  
Instrument Serial No. 008870 156 Church St  
Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E.A. Bryant

Signature of Certifying Official

671

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870  
Test Date: 11/03/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107502


Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	4:01pm
AIR BLK	.00	4:01pm
ACCY CHK	.08	4:02pm
AIR BLK	.00	4:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:04pm</b>
AIR BLK	.00	4:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:06pm</b>
AIR BLK	.00	4:07pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870      Test Record Number: 3312  
Test Date: 11/03/2021      Test Time: 4:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:08pm
FLO	Pass	4:08pm
FC	Pass	4:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:08pm
SRC	Pass	4:08pm
DET	Pass	4:08pm
BAR	Pass	4:08pm
BT	Pass	4:08pm

Blank Tests

Test	Status	Time
AIR	Pass	4:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:09pm

CRC Tests

Test	Status	Time
COMP	Pass	4:09pm
CAL	Pass	4:09pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Vance Instrument Location Vance Co. S.O.  
Instrument Serial No. 008937 156 Church St.  
Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. Bryant  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937

Test Date: 11/03/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107502

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	3:49pm
AIR BLK	.00	3:49pm
ACCY CHK	.08	3:50pm
AIR BLK	.00	3:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:51pm</b>
AIR BLK	.00	3:52pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:54pm</b>
AIR BLK	.00	3:55pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937      Test Record Number: 3010  
Test Date: 11/03/2021      Test Time: 3:57pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:57pm
FLO	Pass	3:57pm
FC	Pass	3:58pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:58pm
SRC	Pass	3:58pm
DET	Pass	3:58pm
BAR	Pass	3:58pm
BT	Pass	3:58pm

Blank Tests

Test	Status	Time
AIR	Pass	3:58pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:58pm

CRC Tests

Test	Status	Time
COMP	Pass	3:59pm
CAL	Pass	3:59pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County WAKE Instrument Location Wake Co Detention Ctr

Instrument Serial No. 008577 3301 Hammond Rd

Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Barnes, Simon S  
Permit Number: 0014-6221  
Effective:  
09/01/2020-09/01/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG118805  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:28pm
AIR BLK	.00	2:28pm
ACCY CHK	.07	2:29pm
AIR BLK	.00	2:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:31pm</b>
AIR BLK	.00	2:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:33pm</b>
AIR BLK	.00	2:34pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577      Test Record Number: 5950  
Test Date: 11/22/2021      Test Time: 2:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:37pm
FLO	Pass	2:37pm
FC	Pass	2:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:37pm
SRC	Pass	2:37pm
DET	Pass	2:37pm
BAR	Pass	2:37pm
BT	Pass	2:37pm

Blank Tests

Test	Status	Time
AIR	Pass	2:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:37pm

CRC Tests

Test	Status	Time
COMP	Pass	2:38pm
CAL	Pass	2:38pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wake Instrument Location Wake Co Detention Ctr  
Instrument Serial No. 008612 3301 HAMMOND RD  
Raleigh NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



EAB  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612

Test Date: 11/02/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG107502

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	11:23am
AIR BLK	.00	11:23am
ACCY CHK	.07	11:24am
AIR BLK	.00	11:25am
<b>SUB TEST</b>	<b>.00</b>	<b>11:25am</b>
AIR BLK	.00	11:26am
<b>SUB TEST</b>	<b>.00</b>	<b>11:28am</b>
AIR BLK	.00	11:29am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612      Test Record Number: 4867  
Test Date: 11/02/2021      Test Time: 11:29am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30am
FLO	Pass	11:30am
FC	Pass	11:30am

Temperature Tests

Test	Status	Time
FC1	Pass	11:30am
SRC	Pass	11:30am
DET	Pass	11:30am
BAR	Pass	11:30am
BT	Pass	11:30am

Blank Tests

Test	Status	Time
AIR	Pass	11:30am


Printer Tests

Test	Status	Time
PRNT	Pass	11:30am

CRC Tests

Test	Status	Time
COMP	Pass	11:31am
CAL	Pass	11:31am

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wake Instrument Location Wake Co. Durham, GA  
Instrument Serial No. 008612 3301 Hammond Rd  
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612  
Test Date: 11/22/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Barnes, Simon S

Permit Number: 0014-6221

Effective:

09/01/2020-09/01/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:29pm
AIR BLK	.00	2:29pm
ACCY CHK	.07	2:30pm
AIR BLK	.00	2:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:31pm</b>
AIR BLK	.00	2:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:34pm</b>
AIR BLK	.00	2:35pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612      Test Record Number: 4899  
Test Date: 11/22/2021      Test Time: 2:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:37pm
FLO	Pass	2:37pm
FC	Pass	2:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:37pm
SRC	Pass	2:37pm
DET	Pass	2:37pm
BAR	Pass	2:37pm
BT	Pass	2:37pm

Blank Tests

Test	Status	Time
AIR	Pass	2:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:38pm

CRC Tests

Test	Status	Time
COMP	Pass	2:38pm
CAL	Pass	2:38pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wake Instrument Location Raleigh PD Northeast Dist  
Instrument Serial No. 008623 5228 GREENS DAIRY Rd  
RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E.A. Smith  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY NORTH EAST DISTRICT 910

Serial Number: 008623

Test Date: 11/30/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG034901

Exp Date: 12/14/2022

Test	g/210L	Time
DIAG	Pass	5:55pm
AIR BLK	.00	5:56pm
ACCY CHK	.07	5:56pm
AIR BLK	.00	5:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>5:58pm</b>
AIR BLK	.00	5:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>6:00pm</b>
AIR BLK	.00	6:01pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY NORTH EAST DISTRICT 910

Serial Number: 008623      Test Record Number: 4327  
Test Date: 11/30/2021      Test Time: 6:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:02pm
FLO	Pass	6:02pm
FC	Pass	6:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:03pm
SRC	Pass	6:03pm
DET	Pass	6:03pm
BAR	Pass	6:03pm
BT	Pass	6:03pm

Blank Tests

Test	Status	Time
AIR	Pass	6:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:03pm

CRC Tests

Test	Status	Time
COMP	Pass	6:04pm
CAL	Pass	6:04pm

Preventive Maintenance  
Status: Pass

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wake Instrument Location Wake FOREST PD  
Instrument Serial No. 008700 225 S. Taylor St.  
Wake Forest, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



E. Brumby  
Signature of Certifying Official      671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700

Test Date: 11/30/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118804

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	4:19pm
AIR BLK	.00	4:20pm
ACCY CHK	.08	4:20pm
AIR BLK	.00	4:21pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:22pm</b>
AIR BLK	.00	4:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:24pm</b>
AIR BLK	.00	4:25pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700      Test Record Number: 1934  
Test Date: 11/30/2021      Test Time: 4:26pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:26pm
FLO	Pass	4:26pm
FC	Pass	4:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:27pm
SRC	Pass	4:27pm
DET	Pass	4:27pm
BAR	Pass	4:27pm
BT	Pass	4:27pm

Blank Tests

Test	Status	Time
AIR	Pass	4:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:27pm

CRC Tests

Test	Status	Time
COMP	Pass	4:27pm
CAL	Pass	4:27pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County WAKE Instrument Location Wake Co Detention Ctr  
Instrument Serial No. 005760 3301 Hammond Rd  
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Simon Stokes Perry  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760

Test Date: 11/22/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Barnes, Simon S

Permit Number: 0014-6221

Effective:

09/01/2020-09/01/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:26pm
AIR BLK	.00	2:27pm
ACCY CHK	.08	2:28pm
AIR BLK	.00	2:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:29pm</b>
AIR BLK	.00	2:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:32pm</b>
AIR BLK	.00	2:33pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760      Test Record Number: 4590  
Test Date: 11/22/2021      Test Time: 2:33pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:33pm
FLO	Pass	2:33pm
FC	Pass	2:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:34pm
SRC	Pass	2:34pm
DET	Pass	2:34pm
BAR	Pass	2:34pm
BT	Pass	2:34pm

Blank Tests

Test	Status	Time
AIR	Pass	2:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:34pm

CRC Tests

Test	Status	Time
COMP	Pass	2:34pm
CAL	Pass	2:34pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wake Instrument Location Wake Co Detention Center

Instrument Serial No. 008778 3301 HAMMOND RD  
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

467  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Barnes, Simon S  
Permit Number: 0014-6221  
Effective:  
09/01/2020-09/01/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG118805  
Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	2:27pm
AIR BLK	.00	2:28pm
ACCY CHK	.07	2:28pm
AIR BLK	.00	2:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:30pm</b>
AIR BLK	.00	2:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:32pm</b>
AIR BLK	.00	2:33pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778      Test Record Number: 6399  
Test Date: 11/22/2021      Test Time: 2:43pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:44pm
FLO	Pass	2:44pm
FC	Pass	2:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:44pm
SRC	Pass	2:44pm
DET	Pass	2:44pm
BAR	Pass	2:44pm
BT	Pass	2:44pm

Blank Tests

Test	Status	Time
AIR	Pass	2:45pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:45pm

CRC Tests

Test	Status	Time
COMP	Pass	2:45pm
CAL	Pass	2:45pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County WARREN Instrument Location WARREN Co. LEC  
Instrument Serial No. 008793 128 Rafter's Lane  
WARRENTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

671  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793

Test Date: 11/04/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Bryant Jr., Earl A

Permit Number: 0017-9707

Effective:

09/15/2020-09/15/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG107501

Exp Date: 03/16/2023

Test	g/210L	Time
DIAG	Pass	3:48pm
AIR BLK	.00	3:49pm
ACCY CHK	.07	3:50pm
AIR BLK	.00	3:50pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:51pm</b>
AIR BLK	.00	3:52pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:53pm</b>
AIR BLK	.00	3:54pm

Reported ~~AC~~: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793      Test Record Number: 1766  
Test Date: 11/04/2021      Test Time: 3:55pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:55pm
FLO	Pass	3:55pm
FC	Pass	3:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:55pm
SRC	Pass	3:55pm
DET	Pass	3:55pm
BAR	Pass	3:55pm
BT	Pass	3:55pm

Blank Tests

Test	Status	Time
AIR	Pass	3:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:56pm

CRC Tests

Test	Status	Time
COMP	Pass	3:56pm
CAL	Pass	3:56pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Watauga Instrument Location Boone PD  
Instrument Serial No. 008716 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716  
Test Date: 11/24/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Burnette, Anthony J

Permit Number: 0018-4401

Effective:

10/01/2020-10/01/2022

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	11:50am
AIR BLK	.00	11:51am
ACCY CHK	.08	11:52am
AIR BLK	.00	11:53am
<b>SUB TEST</b>	<b>.00</b>	<b>11:54am</b>
AIR BLK	.00	11:55am
<b>SUB TEST</b>	<b>.00</b>	<b>11:56am</b>
AIR BLK	.00	11:57am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716      Test Record Number: 2748  
Test Date: 11/24/2021      Test Time: 11:58am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:59am
FLO	Pass	11:59am
FC	Pass	11:59am

Temperature Tests

Test	Status	Time
FC1	Pass	11:59am
SRC	Pass	11:59am
DET	Pass	11:59am
BAR	Pass	11:59am
BT	Pass	11:59am

Blank Tests

Test	Status	Time
AIR	Pass	12:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:00pm

CRC Tests

Test	Status	Time
COMP	Pass	12:00pm
CAL	Pass	12:00pm

Preventive Maintenance  
Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II and**  
**MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wayne Instrument Location Wayne Co. Detention Center  
Instrument Serial No. 008649 207E Chestnut St., Goldsboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Xiu  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: *Guard, Kelly G*  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: *Breath Test*

Lot Number: AG107501  
Exp Date: 03/16/2023


Test	g/210L	Time
DIAG	Pass	10:52am
AIR BLK	.00	10:53am
ACCY CHK	.08	10:53am
AIR BLK	.00	10:54am
<b>SUB TEST</b>	<b>.00</b>	<b>10:55am</b>
AIR BLK	.00	10:56am
<b>SUB TEST</b>	<b>.00</b>	<b>10:58am</b>
AIR BLK	.00	10:59am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649      Test Record Number: 4888  
Test Date: 11/12/2021      Test Time: 11:00am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:00am
FLO	Pass	11:00am
FC	Pass	11:00am

Temperature Tests

Test	Status	Time
FC1	Pass	11:00am
SRC	Pass	11:00am
DET	Pass	11:00am
BAR	Pass	11:00am
BT	Pass	11:00am

Blank Tests

Test	Status	Time
AIR	Pass	11:01am


Printer Tests

Test	Status	Time
PRNT	Pass	11:01am

CRC Tests

Test	Status	Time
COMP	Pass	11:01am
CAL	Pass	11:01am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wayne Instrument Location Seymour Johnson AFB  
Instrument Serial No. 008786 1010 Vermont Garrison St.,  
Goldstboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official 643 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786  
Test Date: 11/12/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Guard, Kelly G

Permit Number: 0037-7722

Effective:

09/28/2020-09/28/2022

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG007601

Exp Date: 03/16/2022

Test	g/210L	Time
DIAG	Pass	9:47am
AIR BLK	.00	9:48am
ACCY CHK	.08	9:48am
AIR BLK	.00	9:50am
SUB TEST	.00	9:50am
AIR BLK	.00	9:51am
SUB TEST	.00	9:53am
AIR BLK	.00	9:53am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786      Test Record Number: 379  
Test Date: 11/12/2021      Test Time: 9:54am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55am
FLO	Pass	9:55am
FC	Pass	9:55am

Temperature Tests

Test	Status	Time
FC1	Pass	9:55am
SRC	Pass	9:55am
DET	Pass	9:55am
BAR	Pass	9:55am
BT	Pass	9:55am

Blank Tests

Test	Status	Time
AIR	Pass	9:56am


Printer Tests

Test	Status	Time
PRNT	Pass	9:56am

CRC Tests

Test	Status	Time
COMP	Pass	9:56am
CAL	Pass	9:56am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wayne Instrument Location Wayne Co. Detention Center  
Instrument Serial No. 008879 207 E. Chestnut St., Goldsboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official      643 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879  
Test Date: 11/12/2021

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: Guard, Kelly G  
Permit Number: 0037-7722  
Effective:  
09/28/2020-09/28/2022

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG018202  
Exp Date: 06/30/2022

Test	g/210L	Time
DIAG	Pass	10:38am
AIR BLK	.00	10:39am
ACCY CHK	.07	10:40am
AIR BLK	.00	10:41am
<b>SUB TEST</b>	<b>.00</b>	<b>10:42am</b>
AIR BLK	.00	10:43am
<b>SUB TEST</b>	<b>.00</b>	<b>10:44am</b>
AIR BLK	.00	10:45am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879      Test Record Number: 1704  
Test Date: 11/12/2021      Test Time: 10:46am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:47am
FLO	Pass	10:47am
FC	Pass	10:47am

Temperature Tests

Test	Status	Time
FC1	Pass	10:47am
SRC	Pass	10:47am
DET	Pass	10:47am
BAR	Pass	10:47am
BT	Pass	10:47am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am


Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:48am
CAL	Pass	10:48am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wilkes Instrument Location Wilkes Co. Detention

Instrument Serial No. 008843 Wilkesboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843

Test Date: 11/22/2021

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Burnette, Anthony J

Permit Number: 0018-4401

Effective:

10/01/2020-10/01/2022

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118805


Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:31pm
AIR BLK	.00	1:32pm
ACCY CHK	.08	1:33pm
AIR BLK	.00	1:34pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:35pm</b>
AIR BLK	.00	1:36pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:37pm</b>
AIR BLK	.00	1:38pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843      Test Record Number: 2600  
Test Date: 11/22/2021      Test Time: 1:39pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:40pm
FLO	Pass	1:40pm
FC	Pass	1:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:40pm
SRC	Pass	1:40pm
DET	Pass	1:40pm
BAR	Pass	1:40pm
BT	Pass	1:40pm

Blank Tests

Test	Status	Time
AIR	Pass	1:41pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:41pm

CRC Tests

Test	Status	Time
COMP	Pass	1:41pm
CAL	Pass	1:41pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II and  
MODEL INTOX EC/IR II (Enhanced with serial number 10,000 or higher)**

County Wilkes Instrument Location Wilkes Co. Detention  
Instrument Serial No. 008865 Wilkesboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II and Model Intox EC/IR II (Enhanced with serial number 10,000 or higher) to be followed at least once every four months are:

- (1) Verify the ethanol gas canister displays at least 51 pounds per square inch (psi) of pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
- (2) Verify instrument displays time and date;
- (3) Initiate breath test sequence;
- (4) Enter information as prompted;
- (5) Verify instrument accuracy;
- (6) When "PLEASE BLOW" appears, collect breath sample;
- (7) When "PLEASE BLOW" appears, collect breath sample;
- (8) Print test record;
- (9) Run diagnostic program and confirm preventive maintenance status of "Pass"; and
- (10) Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2021 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865  
Test Date: 11/22/2021

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Burnette, Anthony J

Permit Number: 0018-4401

Effective:

10/01/2020-10/01/2022

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG118805

Exp Date: 07/07/2023

Test	g/210L	Time
DIAG	Pass	1:30pm
AIR BLK	.00	1:31pm
ACCY CHK	.08	1:32pm
AIR BLK	.00	1:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:34pm</b>
AIR BLK	.00	1:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:36pm</b>
AIR BLK	.00	1:37pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007



Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865      Test Record Number: 846  
Test Date: 11/22/2021      Test Time: 1:39pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:40pm
FLO	Pass	1:40pm
FC	Pass	1:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:40pm
SRC	Pass	1:40pm
DET	Pass	1:40pm
BAR	Pass	1:40pm
BT	Pass	1:40pm

Blank Tests

Test	Status	Time
AIR	Pass	1:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:41pm

CRC Tests

Test	Status	Time
COMP	Pass	1:41pm
CAL	Pass	1:41pm

Preventive Maintenance  
Status: Pass



**Analyst**