

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alamance Instrument Location CAT Mobile Unit  
Instrument Serial No. 008898 Graham PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of October, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

660  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY BAT MOBILE UNIT 1 000

Serial Number: 008898  
Test Date: 10/26/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JASON R

Permit Number: 19145E

Effective:

03/01/2018-03/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	10:20pm
AIR BLK	.00	10:21pm
ACCY CHK	.07	10:21pm
AIR BLK	.00	10:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:23pm</b>
AIR BLK	.00	10:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:25pm</b>
AIR BLK	.00	10:26pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ALAMANCE COUNTY BAT MOBILE UNIT 1 000

Serial Number: 008898      Test Record Number: 1018  
Test Date: 10/26/2019      Test Time: 10:31pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:31pm
FLO	Pass	10:31pm
FC	Pass	10:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:31pm
SRC	Pass	10:31pm
DET	Pass	10:31pm
BAR	Pass	10:31pm
BT	Pass	10:31pm

Blank Tests

Test	Status	Time
AIR	Pass	10:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:32pm

CRC Tests

Test	Status	Time
COMP	Pass	10:32pm
CAL	Pass	10:32pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Anson Co. Instrument Location Anson Co. Sheriff Office

Instrument Serial No. 0074139 Wadesboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008739  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902202

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	3:01pm
AIR BLK	.00	3:02pm
ACCY CHK	.08	3:03pm
AIR BLK	.00	3:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:04pm</b>
AIR BLK	.00	3:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:07pm</b>
AIR BLK	.00	3:08pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008739      Test Record Number: 452  
Test Date: 10/01/2019      Test Time: 3:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:08pm
FLO	Pass	3:08pm
FC	Pass	3:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:09pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:09pm

CRC Tests

Test	Status	Time
COMP	Pass	3:09pm
CAL	Pass	3:09pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County ANSON Co. Instrument Location ANSON Co. Sheriff Office

Instrument Serial No. 008591 Wadesboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of October, 2010 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ANSON COUNTY SHERIFF'S OFFICE 030

Serial Number: 008597  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902202

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	3:00pm
AIR BLK	.00	3:00pm
ACCY CHK	.08	3:01pm
AIR BLK	.00	3:02pm
SUB TEST	.00	3:03pm
AIR BLK	.00	3:03pm
SUB TEST	.00	3:05pm
AIR BLK	.00	3:06pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*ANSON COUNTY SHERIFF'S OFFICE 030*

Serial Number: 008597      Test Record Number: 1733  
Test Date: 10/01/2019      Test Time: 3:06pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:07pm
FLO	Pass	3:07pm
FC	Pass	3:07pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:07pm
SRC	Pass	3:07pm
DET	Pass	3:07pm
BAR	Pass	3:07pm
BT	Pass	3:07pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:08pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:08pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:08pm
CAL	Pass	3:08pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Avery Instrument Location Avery Co. Jail  
Instrument Serial No. 008664 Newland, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664  
Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801


Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	11:20pm
AIR BLK	.00	11:21pm
ACCY CHK	.08	11:21pm
AIR BLK	.00	11:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:23pm</b>
AIR BLK	.00	11:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:25pm</b>
AIR BLK	.00	11:26pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*AVERY COUNTY AVERY COUNTY JAIL 050*

Serial Number: 008664      Test Record Number: 993  
Test Date: 10/04/2019      Test Time: 11:26pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:27pm
FLO	Pass	11:27pm
FC	Pass	11:27pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:27pm
SRC	Pass	11:27pm
DET	Pass	11:27pm
BAR	Pass	11:27pm
BT	Pass	11:27pm

**Blank Tests**

Test	Status	Time
AIR	Pass	11:27pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	11:28pm

**CRC Tests**

Test	Status	Time
COMP	Pass	11:28pm
CAL	Pass	11:28pm

Preventive Maintenance  
Status: Pass



**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Avery Instrument Location Banner Elk PD  
Instrument Serial No. 008724 Banner Elk, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

1649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

AVERY COUNTY BANNER ELK PD 050

Serial Number: 008724  
Test Date: 10/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902201


Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	4:43pm
AIR BLK	.00	4:44pm
ACCY CHK	.08	4:45pm
AIR BLK	.00	4:46pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:46pm</b>
AIR BLK	.00	4:47pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:49pm</b>
AIR BLK	.00	4:50pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*AVERY COUNTY BANNER ELK PD 050*

Serial Number: 008724      Test Record Number: 597  
Test Date: 10/18/2019      Test Time: 4:51pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:51pm
FLO	Pass	4:51pm
FC	Pass	4:51pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:51pm
SRC	Pass	4:51pm
DET	Pass	4:51pm
BAR	Pass	4:51pm
BT	Pass	4:51pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:52pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:52pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:52pm
CAL	Pass	4:52pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Beaufort Instrument Location Beaufort Co. Courthouse  
Instrument Serial No. 008586 102 E. 2nd St., Washington, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008586  
Test Date: 10/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	10:50am
AIR BLK	.00	10:51am
ACCY CHK	.08	10:51am
AIR BLK	.00	10:52am
<b>SUB TEST</b>	<b>.00</b>	<b>10:53am</b>
AIR BLK	.00	10:54am
<b>SUB TEST</b>	<b>.00</b>	<b>10:55am</b>
AIR BLK	.00	10:56am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008586      Test Record Number: 1526  
Test Date: 10/22/2019      Test Time: 10:57am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:58am
FLO	Pass	10:58am
FC	Pass	10:58am

Temperature Tests

Test	Status	Time
FC1	Pass	10:58am
SRC	Pass	10:58am
DET	Pass	10:58am
BAR	Pass	10:58am
BT	Pass	10:58am

Blank Tests

Test	Status	Time
AIR	Pass	10:59am

Printer Tests

Test	Status	Time
PRNT	Pass	10:59am

CRC Tests

Test	Status	Time
COMP	Pass	10:59am
CAL	Pass	10:59am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. 008909 102 E. 2<sup>nd</sup> St., Washington, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22<sup>nd</sup> day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008909

Test Date: 10/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	10:36am
AIR BLK	.00	10:37am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
SUB TEST	.00	10:39am
AIR BLK	.00	10:40am
SUB TEST	.00	10:42am
AIR BLK	.00	10:42am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008909      Test Record Number: 3405  
Test Date: 10/22/2019      Test Time: 10:43am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:44am
FLO	Pass	10:44am
FC	Pass	10:44am

Temperature Tests

Test	Status	Time
FC1	Pass	10:44am
SRC	Pass	10:44am
DET	Pass	10:44am
BAR	Pass	10:44am
BT	Pass	10:44am

Blank Tests

Test	Status	Time
AIR	Pass	10:45am

Printer Tests

Test	Status	Time
PRNT	Pass	10:45am

CRC Tests

Test	Status	Time
COMP	Pass	10:45am
CAL	Pass	10:45am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Buncombe


Instrument Location Bus Mobile Unit 12

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 27 day of October, 2011, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chad. Dow

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BUNCOMBE BAT MOBILE UNIT 02 100

Serial Number: 008973  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

05/01/2019-05/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	9:26pm
AIR BLK	.00	9:27pm
ACCY CHK	.08	9:27pm
AIR BLK	.00	9:28pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:29pm</b>
AIR BLK	.00	9:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:31pm</b>
AIR BLK	.00	9:32pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

BUNCOMBE BAT MOBILE UNIT 02 100

Serial Number: 008973      Test Record Number: 715  
Test Date: 10/24/2019      Test Time: 9:33pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:33pm
FLO	Pass	9:33pm
FC	Pass	9:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:34pm
SRC	Pass	9:34pm
DET	Pass	9:34pm
BAR	Pass	9:34pm
BT	Pass	9:34pm

Blank Tests

Test	Status	Time
AIR	Pass	9:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:34pm

CRC Tests

Test	Status	Time
COMP	Pass	9:34pm
CAL	Pass	9:34pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Burke

Instrument Location Burke-Catawba Jail


Instrument Serial No. 008831

Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008831  
Test Date: 10/23/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919902


Exp Date: 07/28/2021

Test	g/210L	Time
DIAG	Pass	3:09pm
AIR BLK	.00	3:10pm
ACCY CHK	.08	3:11pm
AIR BLK	.00	3:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:13pm</b>
AIR BLK	.00	3:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:15pm</b>
AIR BLK	.00	3:16pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

*BURKE COUNTY BURKE-CATAWBA JAIL 110*

Serial Number: 008831      Test Record Number: 2178  
Test Date: 10/23/2019      Test Time: 3:17pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:17pm
FLO	Pass	3:17pm
FC	Pass	3:17pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:17pm
SRC	Pass	3:17pm
DET	Pass	3:17pm
BAR	Pass	3:17pm
BT	Pass	3:17pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:18pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:18pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:18pm
CAL	Pass	3:18pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Burke Instrument Location Burke Catawba Jail  
Instrument Serial No. 008904 Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008904

Test Date: 10/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	3:08pm
AIR BLK	.00	3:09pm
ACCY CHK	.08	3:10pm
AIR BLK	.00	3:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:12pm</b>
AIR BLK	.00	3:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:14pm</b>
AIR BLK	.00	3:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*BURKE COUNTY BURKE-CATAWBA JAIL 110*

Serial Number: 008904      Test Record Number: 2417  
Test Date: 10/23/2019      Test Time: 3:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:17pm
FLO	Pass	3:17pm
FC	Pass	3:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:17pm
SRC	Pass	3:17pm
DET	Pass	3:17pm
BAR	Pass	3:17pm
BT	Pass	3:17pm

Blank Tests

Test	Status	Time
AIR	Pass	3:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:18pm

CRC Tests

Test	Status	Time
COMP	Pass	3:18pm
CAL	Pass	3:18pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cabarrus Instrument Location Cabarrus County SO  
Instrument Serial No. 008625 30 Corban Ave., Concord

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008625

Test Date: 10/09/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902106

Exp Date: 01/21/2021

Test	g/210L	Time
DIAG	Pass	1:09pm
AIR BLK	.00	1:10pm
ACCY CHK	.08	1:11pm
AIR BLK	.00	1:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:12pm</b>
AIR BLK	.00	1:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:16pm</b>
AIR BLK	.00	1:17pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CABARRUS COUNTY CABARRUS COUNTY SO 120

Serial Number: 008625      Test Record Number: 5368  
Test Date: 10/09/2019      Test Time: 1:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:18pm
FLO	Pass	1:18pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Caldwell Instrument Location Caldwell Co. Jail

Instrument Serial No. 008803 Lenoir, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CALDWELL COUNTY CALDWELL COUNTY JAIL

130

Serial Number: 008803

Test Date: 10/14/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	2:10pm
AIR BLK	.00	2:11pm
ACCY CHK	.07	2:12pm
AIR BLK	.00	2:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:13pm</b>
AIR BLK	.00	2:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:16pm</b>
AIR BLK	.00	2:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**CALDWELL COUNTY CALDWELL COUNTY JAIL 130**

Serial Number: 008803      Test Record Number: 632  
Test Date: 10/14/2019      Test Time: 2:20pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	2:21pm
FLO	Pass	2:21pm
FC	Pass	2:21pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	2:21pm
SRC	Pass	2:21pm
DET	Pass	2:21pm
BAR	Pass	2:21pm
BT	Pass	2:21pm

**Blank Tests**

Test	Status	Time
AIR	Pass	2:22pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	2:22pm

**CRC Tests**

Test	Status	Time
COMP	Pass	2:22pm
CAL	Pass	2:22pm

Preventive Maintenance  
Status: Pass



**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Caldwell Instrument Location Caldwell Co Jail  
Instrument Serial No. 008719 Lenoir, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CALDWELL COUNTY CALDWELL COUNTY JAIL  
130

Serial Number: 008719  
Test Date: 10/14/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	2:10pm
AIR BLK	.00	2:11pm
ACCY CHK	.07	2:12pm
AIR BLK	.00	2:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:14pm</b>
AIR BLK	.00	2:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:16pm</b>
AIR BLK	.00	2:17pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CALDWELL COUNTY CALDWELL COUNTY JAIL 130

Serial Number: 008719      Test Record Number: 2551  
Test Date: 10/14/2019      Test Time: 2:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:19pm
FLO	Pass	2:19pm
FC	Pass	2:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:20pm
SRC	Pass	2:20pm
DET	Pass	2:20pm
BAR	Pass	2:20pm
BT	Pass	2:20pm

Blank Tests

Test	Status	Time
AIR	Pass	2:20pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:20pm

CRC Tests

Test	Status	Time
COMP	Pass	2:20pm
CAL	Pass	2:20pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CARTERET Instrument Location CARTERET COUNTY  
Instrument Serial No. 008605 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alma R. Bane  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CARTERET COUNTY DETENTION CENTER 150

Serial Number: 008605

Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	12:55pm
AIR BLK	.00	12:56pm
ACCY CHK	.08	12:56pm
AIR BLK	.00	12:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:58pm</b>
AIR BLK	.00	12:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:00pm</b>
AIR BLK	.00	1:01pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*CARTERET COUNTY DETENTION CENTER 150*

Serial Number: 008605      Test Record Number: 3900  
Test Date: 10/02/2019      Test Time: 1:03pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:04pm
FLO	Pass	1:04pm
FC	Pass	1:04pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:04pm
SRC	Pass	1:04pm
DET	Pass	1:04pm
BAR	Pass	1:04pm
BT	Pass	1:04pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:04pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:04pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:05pm
CAL	Pass	1:05pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CARTERET Instrument Location CARTERET COUNTY  
Instrument Serial No. 008882 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CARTERET COUNTY DETENTION CENTER 150

Serial Number: 008882

Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901


Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	1:35pm
AIR BLK	.00	1:36pm
ACCY CHK	.08	1:37pm
AIR BLK	.00	1:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:38pm</b>
AIR BLK	.00	1:39pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:41pm</b>
AIR BLK	.00	1:42pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR



\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

*CARTERET COUNTY DETENTION CENTER 150*

Serial Number: 008882      Test Record Number: 1885  
Test Date: 10/02/2019      Test Time: 1:42pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:42pm
FLO	Pass	1:42pm
FC	Pass	1:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:43pm
SRC	Pass	1:43pm
DET	Pass	1:43pm
BAR	Pass	1:43pm
BT	Pass	1:43pm

Blank Tests

Test	Status	Time
AIR	Pass	1:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:43pm

CRC Tests

Test	Status	Time
COMP	Pass	1:43pm
CAL	Pass	1:43pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CARTERET Instrument Location MOOREHEAD CITY

Instrument Serial No. 008731 POLICE DEPT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Bowers  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	2:54pm
AIR BLK	.00	2:55pm
ACCY CHK	.07	2:55pm
AIR BLK	.00	2:56pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:57pm</b>
AIR BLK	.00	2:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:59pm</b>
AIR BLK	.00	3:00pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731      Test Record Number: 2187  
Test Date: 10/02/2019      Test Time: 3:01pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:02pm
FLO	Pass	3:02pm
FC	Pass	3:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:02pm
SRC	Pass	3:02pm
DET	Pass	3:02pm
BAR	Pass	3:02pm
BT	Pass	3:02pm

Blank Tests

Test	Status	Time
AIR	Pass	3:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:03pm

CRC Tests

Test	Status	Time
COMP	Pass	3:03pm
CAL	Pass	3:03pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chatham Co. Instrument Location Chatham Co. Detention Center  
Instrument Serial No. 008591 Pittsboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CHATHAM COUNTY DETENTION CENTER 180

Serial Number: 008591  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021

Test g/210L Time

DIAG Pass 2:39pm

AIR BLK .00 2:40pm

ACCY CHK .08 2:40pm

AIR BLK .00 2:41pm


**SUB TEST .00 2:42pm**

AIR BLK .00 2:43pm

**SUB TEST .00 2:44pm**

AIR BLK .00 2:45pm

Reported AC1 .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CHATHAM COUNTY DETENTION CENTER 180

Serial Number: 008591      Test Record Number: 2206  
Test Date: 10/02/2019      Test Time: 2:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:46pm
FLO	Pass	2:46pm
FC	Pass	2:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:46pm
SRC	Pass	2:46pm
DET	Pass	2:46pm
BAR	Pass	2:46pm
BT	Pass	2:46pm

Blank Tests

Test	Status	Time
AIR	Pass	2:47pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:47pm

CRC Tests

Test	Status	Time
COMP	Pass	2:47pm
CAL	Pass	2:47pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chatham Co. Instrument Location Siler City, P.D.  
Instrument Serial No. 008811 Siler City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CHATHAM COUNTY SILER CITY PD 180

Serial Number: 008811  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	1:55pm
AIR BLK	.00	1:56pm
ACCY CHK	.08	1:56pm
AIR BLK	.00	1:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:58pm</b>
AIR BLK	.00	1:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:00pm</b>
AIR BLK	.00	2:02pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CHATHAM COUNTY SILER CITY PD 180

Serial Number: 008811      Test Record Number: 1375  
Test Date: 10/02/2019      Test Time: 2:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:03pm
FLO	Pass	2:03pm
FC	Pass	2:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:03pm
SRC	Pass	2:03pm
DET	Pass	2:03pm
BAR	Pass	2:03pm
BT	Pass	2:03pm

Blank Tests

Test	Status	Time
AIR	Pass	2:04pm

Printer Tests


Test	Status	Time
PRNT	Pass	2:04pm

CRC Tests

Test	Status	Time
COMP	Pass	2:04pm
CAL	Pass	2:04pm

Preventive Maintenance

Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County COLUMBUS Instrument Location COLUMBUS COUNTY  
Instrument Serial No. 008875 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

COLUMBUS COUNTY DETENTION CENTER 230

Serial Number: 008875  
Test Date: 10/01/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	10:15am
AIR BLK	.00	10:15am
ACCY CHK	.08	10:16am
AIR BLK	.00	10:17am
<b>SUB TEST</b>	<b>.00</b>	<b>10:18am</b>
AIR BLK	.00	10:19am
<b>SUB TEST</b>	<b>.00</b>	<b>10:20am</b>
AIR BLK	.00	10:21am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

COLUMBUS COUNTY DETENTION CENTER 230

Serial Number: 008875      Test Record Number: 2211  
Test Date: 10/01/2019      Test Time: 10:26am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:26am
FLO	Pass	10:26am
FC	Pass	10:27am

Temperature Tests

Test	Status	Time
FC1	Pass	10:27am
SRC	Pass	10:27am
DET	Pass	10:27am
BAR	Pass	10:27am
BT	Pass	10:27am

Blank Tests

Test	Status	Time
AIR	Pass	10:27am

Printer Tests

Test	Status	Time
PRNT	Pass	10:27am

CRC Tests

Test	Status	Time
COMP	Pass	10:27am
CAL	Pass	10:27am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County COLUMBUS Instrument Location COLUMBUS COUNTY  
Instrument Serial No. 008886 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bann  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

COLUMBUS COUNTY DETENTION CENTER 230

Serial Number: 008886  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501


Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	9:46am
AIR BLK	.00	9:47am
ACCY CHK	.08	9:47am
AIR BLK	.00	9:48am
<b>SUB TEST</b>	<b>.00</b>	<b>9:49am</b>
AIR BLK	.00	9:50am
<b>SUB TEST</b>	<b>.00</b>	<b>9:52am</b>
AIR BLK	.00	9:52am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY DETENTION CENTER 230

Serial Number: 008886      Test Record Number: 1486  
Test Date: 10/01/2019      Test Time: 9:53am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:54am
FLO	Pass	9:54am
FC	Pass	9:54am

Temperature Tests

Test	Status	Time
FC1	Pass	9:54am
SRC	Pass	9:54am
DET	Pass	9:54am
BAR	Pass	9:54am
BT	Pass	9:54am

Blank Tests

Test	Status	Time
AIR	Pass	9:54am


Printer Tests

Test	Status	Time
PRNT	Pass	9:54am

CRC Tests

Test	Status	Time
COMP	Pass	9:55am
CAL	Pass	9:55am

Preventive Maintenance  
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CRAVEN Instrument Location HAVELOCK PD

Instrument Serial No. 008800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800

Test Date: 10/16/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501


Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	3:04pm
AIR BLK	.00	3:05pm
ACCY CHK	.08	3:06pm
AIR BLK	.00	3:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:07pm</b>
AIR BLK	.00	3:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:10pm</b>
AIR BLK	.00	3:11pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800      Test Record Number: 1237  
Test Date: 10/16/2019      Test Time: 3:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:13pm
FLO	Pass	3:13pm
FC	Pass	3:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:13pm
SRC	Pass	3:13pm
DET	Pass	3:13pm
BAR	Pass	3:13pm
BT	Pass	3:13pm

Blank Tests

Test	Status	Time
AIR	Pass	3:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:14pm

CRC Tests

Test	Status	Time
COMP	Pass	3:14pm
CAL	Pass	3:14pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CRAVEN Instrument Location MCAS CHERRY POINT

Instrument Serial No. 010819

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819  
Test Date: 10/16/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG831801

Exp Date: 11/14/2020

Test	g/210L	Time
DIAG	Pass	1:50pm
AIR BLK	.00	1:51pm
ACCY CHK	.08	1:51pm
AIR BLK	.00	1:52pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:53pm</b>
AIR BLK	.00	1:54pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:57pm</b>
AIR BLK	.00	1:57pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR



\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819      Test Record Number: 596  
Test Date: 10/16/2019      Test Time: 2:03pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:04pm
FLO	Pass	2:04pm
FC	Pass	2:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:04pm
SRC	Pass	2:04pm
DET	Pass	2:04pm
BAR	Pass	2:04pm
BT	Pass	2:04pm

Blank Tests

Test	Status	Time
AIR	Pass	2:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:04pm

CRC Tests

Test	Status	Time
COMP	Pass	2:05pm
CAL	Pass	2:05pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location NEW BERN PD

Instrument Serial No. 008817

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817  
Test Date: 10/16/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	10:20am
AIR BLK	.00	10:20am
ACCY CHK	.08	10:21am
AIR BLK	.00	10:22am
<b>SUB TEST</b>	<b>.00</b>	<b>10:23am</b>
AIR BLK	.00	10:24am
<b>SUB TEST</b>	<b>.00</b>	<b>10:25am</b>
AIR BLK	.00	10:26am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst



**Intox EC/IR-II: Preventive Maintenance**

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817      Test Record Number: 1502  
Test Date: 10/16/2019      Test Time: 10:27am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:28am
FLO	Pass	10:28am
FC	Pass	10:28am

Temperature Tests

Test	Status	Time
FC1	Pass	10:28am
SRC	Pass	10:28am
DET	Pass	10:28am
BAR	Pass	10:28am
BT	Pass	10:28am

Blank Tests

Test	Status	Time
AIR	Pass	10:28am


Printer Tests

Test	Status	Time
PRNT	Pass	10:28am

CRC Tests

Test	Status	Time
COMP	Pass	10:29am
CAL	Pass	10:29am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CRAVEN Instrument Location CRAVEN COUNTY

Instrument Serial No. 008732 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin Ray Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY DETENTION CENTER 240

Serial Number: 008732  
Test Date: 10/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	3:21pm
AIR BLK	.00	3:22pm
ACCY CHK	.08	3:22pm
AIR BLK	.00	3:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:24pm</b>
AIR BLK	.00	3:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:27pm</b>
AIR BLK	.00	3:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

**CRAVEN COUNTY DETENTION CENTER 240**

Serial Number: 008732      Test Record Number: 2320  
Test Date: 10/15/2019      Test Time: 3:29pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:29pm
FLO	Pass	3:29pm
FC	Pass	3:29pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:29pm
SRC	Pass	3:29pm
DET	Pass	3:29pm
BAR	Pass	3:29pm
BT	Pass	3:29pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:30pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	3:30pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:30pm
CAL	Pass	3:30pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cumberland Co. Instrument Location FT. BRAGG LEC  
Instrument Serial No. 008903 PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>th</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 008903

Test Date: 10/08/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	10:35am
AIR BLK	.00	10:36am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
<b>SUB TEST</b>	<b>.00</b>	<b>10:38am</b>
AIR BLK	.00	10:39am
<b>SUB TEST</b>	<b>.00</b>	<b>10:41am</b>
AIR BLK	.00	10:42am

Reported AC: 00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FORT BRAGG LEC 250

Serial Number: 008903      Test Record Number: 2523  
Test Date: 10/08/2019      Test Time: 10:42am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:43am
FLO	Pass	10:43am
FC	Pass	10:43am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:43am
SRC	Pass	10:43am
DET	Pass	10:43am
BAR	Pass	10:43am
BT	Pass	10:43am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:44am


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:44am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:44am
CAL	Pass	10:44am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Cumberland Co. Instrument Location FT. BRAGG LEC

Instrument Serial No. 008908 PM10

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY FT BRAGG LEC 250

Serial Number: 008908

Test Date: 10/08/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	10:36am
AIR BLK	.00	10:36am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
<b>SUB TEST</b>	<b>.00</b>	<b>10:39am</b>
AIR BLK	.00	10:40am
<b>SUB TEST</b>	<b>.00</b>	<b>10:41am</b>
AIR BLK	.00	10:42am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY FT BRAGG LEC 250

Serial Number: 008908      Test Record Number: 1814  
Test Date: 10/08/2019      Test Time: 10:43am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:43am
FLO	Pass	10:43am
FC	Pass	10:43am

Temperature Tests

Test	Status	Time
FC1	Pass	10:43am
SRC	Pass	10:43am
DET	Pass	10:43am
BAR	Pass	10:43am
BT	Pass	10:43am

Blank Tests

Test	Status	Time
AIR	Pass	10:44am

Printer Tests


Test	Status	Time
PRNT	Pass	10:44am

CRC Tests

Test	Status	Time
COMP	Pass	10:44am
CAL	Pass	10:44am

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davidson Instrument Location Lexington  
Instrument Serial No. 008883 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24<sup>th</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	11:38am
AIR BLK	.00	11:39am
ACCY CHK	.08	11:39am
AIR BLK	.00	11:40am
<b>SUB TEST</b>	<b>.00</b>	<b>11:41am</b>
AIR BLK	.00	11:42am
<b>SUB TEST</b>	<b>.00</b>	<b>11:44am</b>
AIR BLK	.00	11:45am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008883      Test Record Number: 2134  
Test Date: 10/24/2019      Test Time: 11:47am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:48am
FLO	Pass	11:48am
FC	Pass	11:48am

Temperature Tests

Test	Status	Time
FC1	Pass	11:48am
SRC	Pass	11:48am
DET	Pass	11:48am
BAR	Pass	11:48am
BT	Pass	11:48am

Blank Tests

Test	Status	Time
AIR	Pass	11:49am


Printer Tests

Test	Status	Time
PRNT	Pass	11:49am

CRC Tests

Test	Status	Time
COMP	Pass	11:49am
CAL	Pass	11:49am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davidson Instrument Location Davidson  
Instrument Serial No. 008845 County Jail  
Lexington, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24<sup>th</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Gale Farber 655  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845  
Test Date: 10/24/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	10:09am
AIR BLK	.00	10:10am
ACCY CHK	.08	10:10am
AIR BLK	.00	10:12am
SUB TEST	.00	10:12am
AIR BLK	.00	10:13am
SUB TEST	.00	10:16am
AIR BLK	.00	10:17am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845      Test Record Number: 3266  
Test Date: 10/24/2019      Test Time: 10:18am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:18am
FLO	Pass	10:18am
FC	Pass	10:18am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:18am
SRC	Pass	10:18am
DET	Pass	10:18am
BAR	Pass	10:18am
BT	Pass	10:18am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:19am

**Printer Tests**

Test	Status	Time
PRNT	Pass	10:19am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:19am
CAL	Pass	10:19am

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davidson Instrument Location Thomasville  
Instrument Serial No. 008872 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>rd</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872  
Test Date: 10/23/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: FARLEY, CYNTHIA D  
Permit Number: 24123E  
Effective:  
11/01/2018-11/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG919901  
Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	3:11pm
AIR BLK	.00	3:11pm
ACCY CHK	.07	3:12pm
AIR BLK	.00	3:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:14pm</b>
AIR BLK	.00	3:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:17pm</b>
AIR BLK	.00	3:17pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872      Test Record Number: 1447  
Test Date: 10/23/2019      Test Time: 3:27pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:28pm
FLO	Pass	3:28pm
FC	Pass	3:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:28pm
SRC	Pass	3:28pm
DET	Pass	3:28pm
BAR	Pass	3:28pm
BT	Pass	3:28pm

Blank Tests

Test	Status	Time
AIR	Pass	3:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:29pm

CRC Tests

Test	Status	Time
COMP	Pass	3:29pm
CAL	Pass	3:29pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davie Instrument Location Davie  
Instrument Serial No. 008905 County Jail

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>rd</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905  
Test Date: 10/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	1:06pm
AIR BLK	.00	1:07pm
ACCY CHK	.08	1:08pm
AIR BLK	.00	1:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:10pm</b>
AIR BLK	.00	1:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:13pm</b>
AIR BLK	.00	1:14pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905      Test Record Number: 2425  
Test Date: 10/23/2019      Test Time: 1:17pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:18pm
FLO	Pass	1:18pm
FC	Pass	1:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:18pm
SRC	Pass	1:18pm
DET	Pass	1:18pm
BAR	Pass	1:18pm
BT	Pass	1:18pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:19pm
CAL	Pass	1:19pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County DUPLIN Instrument Location WALLACE POLICE DEPT

Instrument Serial No. 008858

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Bean  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858  
Test Date: 10/04/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 15671E  
Effective:  
06/01/2019-06/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG919901  
Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	11:42am
AIR BLK	.00	11:43am
ACCY CHK	.07	11:43am
AIR BLK	.00	11:45am
<b>SUB TEST</b>	<b>.00</b>	<b>11:45am</b>
AIR BLK	.00	11:46am
<b>SUB TEST</b>	<b>.00</b>	<b>11:48am</b>
AIR BLK	.00	11:49am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858      Test Record Number: 970  
Test Date: 10/04/2019      Test Time: 11:50am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:50am
FLO	Pass	11:50am
FC	Pass	11:50am

Temperature Tests

Test	Status	Time
FC1	Pass	11:51am
SRC	Pass	11:51am
DET	Pass	11:51am
BAR	Pass	11:51am
BT	Pass	11:51am

Blank Tests

Test	Status	Time
AIR	Pass	11:51am

Printer Tests

Test	Status	Time
PRNT	Pass	11:51am

CRC Tests

Test	Status	Time
COMP	Pass	11:51am
CAL	Pass	11:51am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County DUPLIN Instrument Location DUPLIN COUNTY  
Instrument Serial No. 008864 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY DETENTION CENTER 300

Serial Number: 008864

Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG904301


Exp Date: 02/12/2021

Test	g/210L	Time
DIAG	Pass	9:02am
AIR BLK	.00	9:02am
ACCY CHK	.07	9:03am
AIR BLK	.00	9:04am
<b>SUB TEST</b>	<b>.00</b>	<b>9:04am</b>
AIR BLK	.00	9:05am
<b>SUB TEST</b>	<b>.00</b>	<b>9:07am</b>
AIR BLK	.00	9:08am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*DUPLIN COUNTY DETENTION CENTER 300*

Serial Number: 008864      Test Record Number: 3712  
Test Date: 10/04/2019      Test Time: 9:08am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:09am
FLO	Pass	9:09am
FC	Pass	9:09am

Temperature Tests

Test	Status	Time
FC1	Pass	9:09am
SRC	Pass	9:09am
DET	Pass	9:09am
BAR	Pass	9:09am
BT	Pass	9:09am

Blank Tests

Test	Status	Time
AIR	Pass	9:09am


Printer Tests

Test	Status	Time
PRNT	Pass	9:10am

CRC Tests

Test	Status	Time
COMP	Pass	9:10am
CAL	Pass	9:10am

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County ENDICOTT Instrument Location ASSEMBLY UNIT 6

Instrument Serial No. 008779 TANALU

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2011, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008779  
Test Date: 10/12/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG911506

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	3:50pm
AIR BLK	.00	3:51pm
ACCY CHK	.07	3:52pm
AIR BLK	.00	3:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:53pm</b>
AIR BLK	.00	3:54pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:55pm</b>
AIR BLK	.00	3:56pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

EDGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008779      Test Record Number: 3592  
Test Date: 10/12/2019      Test Time: 3:57pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:57pm
FLO	Pass	3:57pm
FC	Pass	3:57pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:57pm
SRC	Pass	3:57pm
DET	Pass	3:57pm
BAR	Pass	3:57pm
BT	Pass	3:57pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:58pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:58pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:58pm
CAL	Pass	3:58pm

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County ENDICOTT Instrument Location 1155 MARBLE UNIT 6

Instrument Serial No. 00V5V4 TARANCO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008584  
Test Date: 10/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	3:54pm
AIR BLK	.00	3:54pm
ACCY CHK	.07	3:55pm
AIR BLK	.00	3:56pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:56pm</b>
AIR BLK	.00	3:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:59pm</b>
AIR BLK	.00	4:00pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008584      Test Record Number: 2297  
Test Date: 10/12/2019      Test Time: 4:00pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:01pm
FLO	Pass	4:01pm
FC	Pass	4:01pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:01pm
SRC	Pass	4:01pm
DET	Pass	4:01pm
BAR	Pass	4:01pm
BT	Pass	4:01pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:02pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:02pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:02pm
CAL	Pass	4:02pm

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County EDLER COUNTY Instrument Location SAF MURDER UNIT 6

Instrument Serial No. 008570 7A1100

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WINGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008580  
Test Date: 10/12/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE N

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	3:54pm
AIR BLK	.00	3:55pm
ACCY CHK	.07	3:55pm
AIR BLK	.00	3:56pm
SUB TEST	.00	3:57pm
AIR BLK	.00	3:58pm
SUB TEST	.00	3:59pm
AIR BLK	.00	4:00pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008580      Test Record Number: 2529  
Test Date: 10/12/2019      Test Time: 4:01pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:01pm
FLO	Pass	4:01pm
FC	Pass	4:01pm

**Temperature Tests**

Test	Status	Time
FCI	Pass	4:01pm
SRC	Pass	4:01pm
DET	Pass	4:01pm
BAR	Pass	4:01pm
BT	Pass	4:01pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:02pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:02pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:02pm
CAL	Pass	4:02pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County EDGEWATER Instrument Location BAT MIDDLE UNIT 6

Instrument Serial No. 008776 THANON

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008776  
Test Date: 10/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	4:54pm
AIR BLK	.00	4:55pm
ACCY CHK	.07	4:55pm
AIR BLK	.00	4:56pm
SUB TEST	.00	4:57pm
AIR BLK	.00	4:58pm
SUB TEST	.00	4:59pm
AIR BLK	.00	5:00pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008776      Test Record Number: 3547  
Test Date: 10/12/2019      Test Time: 5:01pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	5:01pm
FLO	Pass	5:01pm
FC	Pass	5:01pm

**Temperature Tests**

Test	Status	Time
FCL	Pass	5:02pm
SRC	Pass	5:02pm
DET	Pass	5:02pm
BAR	Pass	5:02pm
BT	Pass	5:02pm

**Blank Tests**

Test	Status	Time
AIR	Pass	5:02pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	5:02pm

**CRC Tests**

Test	Status	Time
COMP	Pass	5:02pm
CAL	Pass	5:02pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County COLEMAN Instrument Location STATE MODEL UNIT C

Instrument Serial No. 009637 TANNO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

463  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008637  
Test Date: 10/12/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS'


Test Type: Breath Test

Lot Number: AG716201

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	4:52pm
AIR BLK	.00	4:53pm
ACCY CHK	.07	4:54pm
AIR BLK	.00	4:54pm
SUB TEST	.00	4:55pm
AIR BLK	.00	4:56pm
SUB TEST	.00	4:57pm
AIR BLK	.00	4:58pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008637      Test Record Number: 3033  
Test Date: 10/12/2019      Test Time: 5:01pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	5:01pm
FLO	Pass	5:01pm
FC	Pass	5:01pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	5:01pm
SRC	Pass	5:01pm
DET	Pass	5:01pm
BAR	Pass	5:01pm
BT	Pass	5:01pm

**Blank Tests**

Test	Status	Time
AIR	Pass	5:02pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	5:02pm

**CRC Tests**

Test	Status	Time
COMP	Pass	5:02pm
CAL	Pass	5:02pm

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County EDWING Instrument Location NAF MONDE UND 6

Instrument Serial No. 008676 7AAN10

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY BAT MOBILE UNIT 6 320

Serial Number: 008686  
Test Date: 10/12/2019

Citation Number: M000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test


Lot Number: AG807101

Exp Date: 03/12/2020

Test g/210L Time

DIAG	Pass	4:51pm
AIR BLK	.00	4:52pm
ACCY CHK	.07	4:53pm
AIR BLK	.00	4:53pm
SUB TEST	.00	4:54pm
AIR BLK	.00	4:55pm
SUB TEST	.00	4:57pm
AIR BLK	.00	4:58pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**EDGECOMBE COUNTY BAT MOBILE UNIT 6 320**

Serial Number: 008686      Test Record Number: 6658  
Test Date: 10/12/2019      Test Time: 5:03pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	5:04pm
FLO	Pass	5:04pm
FC	Pass	5:04pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	5:04pm
SRC	Pass	5:04pm
DET	Pass	5:04pm
BAR	Pass	5:04pm
BT	Pass	5:04pm

**Blank Tests**

Test	Status	Time
AIR	Pass	5:04pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	5:05pm

**CRC Tests**

Test	Status	Time
COMP	Pass	5:05pm
CAL	Pass	5:05pm

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gaston Instrument Location Belmont

Instrument Serial No. 008733 209 Chronicle St., Belmont

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mandy Henry  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GASTON COUNTY BELMONT PD 350

Serial Number: 008733

Test Date: 10/16/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG904301

Exp Date: 02/12/2021

Test	g/210L	Time
DIAG	Pass	10:29am
AIR BLK	.00	10:30am
ACCY CHK	.08	10:30am
AIR BLK	.00	10:31am
SUB TEST	.00	10:32am
AIR BLK	.00	10:33am
SUB TEST	.00	10:34am
AIR BLK	.00	10:35am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

GASTON COUNTY BELMONT PD 350

Serial Number: 008733      Test Record Number: 1205  
Test Date: 10/16/2019      Test Time: 10:37am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:37am
FLO	Pass	10:37am
FC	Pass	10:37am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:38am

**Printer Tests**


Test	Status	Time
PRNT	Pass	10:38am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:38am
CAL	Pass	10:38am

**Preventive Maintenance**

Status: Pass



---

**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Greensboro  
Instrument Serial No. 008725 Police Department  
100 Federal Plaza

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:12pm
AIR BLK	.00	1:13pm
ACCY CHK	.08	1:14pm
AIR BLK	.00	1:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:15pm</b>
AIR BLK	.00	1:16pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:18pm</b>
AIR BLK	.00	1:19pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

*GUILFORD COUNTY GREENSBORO PD 400*

Serial Number: 008725      Test Record Number: 4467  
Test Date: 10/01/2019      Test Time: 1:20pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:21pm
FLO	Pass	1:21pm
FC	Pass	1:21pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:21pm
SRC	Pass	1:21pm
DET	Pass	1:21pm
BAR	Pass	1:21pm
BT	Pass	1:21pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:21pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:22pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:22pm
CAL	Pass	1:22pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford

Instrument Location Greensboro Jail

Instrument Serial No. 008794

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794  
Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	1:27pm
AIR BLK	.00	1:28pm
ACCY CHK	.08	1:29pm
AIR BLK	.00	1:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:31pm</b>
AIR BLK	.00	1:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:33pm</b>
AIR BLK	.00	1:34pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794      Test Record Number: 6701  
Test Date: 10/03/2019      Test Time: 1:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:36pm
FLO	Pass	1:36pm
FC	Pass	1:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:36pm
SRC	Pass	1:36pm
DET	Pass	1:36pm
BAR	Pass	1:36pm
BT	Pass	1:36pm

Blank Tests

Test	Status	Time
AIR	Pass	1:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:37pm

CRC Tests

Test	Status	Time
COMP	Pass	1:37pm
CAL	Pass	1:37pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford


Instrument Location Greensboro Jail

Instrument Serial No. 008790

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790  
Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG911506

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	1:01pm
AIR BLK	.00	1:02pm
ACCY CHK	.08	1:03pm
AIR BLK	.00	1:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:04pm</b>
AIR BLK	.00	1:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:07pm</b>
AIR BLK	.00	1:08pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790      Test Record Number: 6577  
Test Date: 10/03/2019      Test Time: 1:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:10pm
FLO	Pass	1:10pm
FC	Pass	1:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:10pm
SRC	Pass	1:10pm
DET	Pass	1:10pm
BAR	Pass	1:10pm
BT	Pass	1:10pm

Blank Tests

Test	Status	Time
AIR	Pass	1:11pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:11pm

CRC Tests

Test	Status	Time
COMP	Pass	1:11pm
CAL	Pass	1:11pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Greensboro Jail  
Instrument Serial No. 008638

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3<sup>rd</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638  
Test Date: 10/03/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: FARLEY, CYNTHIA D  
Permit Number: 24123E  
Effective:  
11/01/2018-11/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG902106  
Exp Date: 01/21/2021

Test	g/210L	Time
DIAG	Pass	12:36pm
AIR BLK	.00	12:37pm
ACCY CHK	.08	12:37pm
AIR BLK	.00	12:39pm
SUB TEST	.00	12:40pm
AIR BLK	.00	12:41pm
SUB TEST	.00	12:42pm
AIR BLK	.00	12:43pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*GUILFORD COUNTY GREENSBORO JAIL 400*

Serial Number: 008638      Test Record Number: 4309  
Test Date: 10/03/2019      Test Time: 12:45pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:46pm
FLO	Pass	12:46pm
FC	Pass	12:46pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:46pm
SRC	Pass	12:46pm
DET	Pass	12:46pm
BAR	Pass	12:46pm
BT	Pass	12:46pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:46pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:46pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:47pm
CAL	Pass	12:47pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location BAT Mobile Unit 1  
Instrument Serial No. 008898 Greensboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

660  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

GUILFORD COUNTY BAT MOBILE UNIT 1 400

Serial Number: 008898  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JASON R

Permit Number: 19145E

Effective:

03/01/2018-03/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	10:27pm
AIR BLK	.00	10:28pm
ACCY CHK	.08	10:29pm
AIR BLK	.00	10:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:31pm</b>
AIR BLK	.00	10:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:33pm</b>
AIR BLK	.00	10:34pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*GUILFORD COUNTY BAT MOBILE UNIT 1 400*

Serial Number: 008898      Test Record Number: 1013  
Test Date: 10/24/2019      Test Time: 10:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:37pm
FLO	Pass	10:37pm
FC	Pass	10:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:38pm
SRC	Pass	10:38pm
DET	Pass	10:38pm
BAR	Pass	10:38pm
BT	Pass	10:38pm

Blank Tests

Test	Status	Time
AIR	Pass	10:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:38pm

CRC Tests

Test	Status	Time
COMP	Pass	10:38pm
CAL	Pass	10:38pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location BAT Mobile Unit 1

Instrument Serial No. 008939 Greensboro PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

660  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY BAT MOBILE UNIT 1 400

Serial Number: 008939  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JASON R

Permit Number: 19145E

Effective:

03/01/2018-03/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	10:28pm
AIR BLK	.00	10:29pm
ACCY CHK	.08	10:30pm
AIR BLK	.00	10:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:31pm</b>
AIR BLK	.00	10:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:34pm</b>
AIR BLK	.00	10:35pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*GUILFORD COUNTY BAT MOBILE UNIT 1 400*

Serial Number: 008939      Test Record Number: 972  
Test Date: 10/24/2019      Test Time: 10:36pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:36pm
FLO	Pass	10:36pm
FC	Pass	10:36pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:37pm
SRC	Pass	10:37pm
DET	Pass	10:37pm
BAR	Pass	10:37pm
BT	Pass	10:37pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:37pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:37pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:37pm
CAL	Pass	10:37pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson

Instrument Location Henderson Co Detention


Instrument Serial No. 008806

Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806  
Test Date: 10/02/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	8:47pm
AIR BLK	.00	8:48pm
ACCY CHK	.07	8:49pm
AIR BLK	.00	8:50pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:50pm</b>
AIR BLK	.00	8:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:53pm</b>
AIR BLK	.00	8:54pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806      Test Record Number: 2834  
Test Date: 10/02/2019      Test Time: 8:55pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	8:55pm
FLO	Pass	8:55pm
FC	Pass	8:55pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	8:56pm
SRC	Pass	8:56pm
DET	Pass	8:56pm
BAR	Pass	8:56pm
BT	Pass	8:56pm

**Blank Tests**

Test	Status	Time
AIR	Pass	8:56pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	8:56pm

**CRC Tests**

Test	Status	Time
COMP	Pass	8:56pm
CAL	Pass	8:56pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson

Instrument Location Henderson Co Detention


Instrument Serial No. 009822

Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HENDERSON COUNTY DETENTION 440

Serial Number: 008822  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	8:48pm
AIR BLK	.00	8:49pm
ACCY CHK	.08	8:50pm
AIR BLK	.00	8:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:52pm</b>
AIR BLK	.00	8:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:54pm</b>
AIR BLK	.00	8:55pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

*HENDERSON COUNTY DETENTION 440*

Serial Number: 008822      Test Record Number: 2450  
Test Date: 10/02/2019      Test Time: 8:59pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:00pm
FLO	Pass	9:00pm
FC	Pass	9:00pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:00pm
SRC	Pass	9:00pm
DET	Pass	9:00pm
BAR	Pass	9:00pm
BT	Pass	9:00pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:00pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:00pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:01pm
CAL	Pass	9:01pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Hertford Instrument Location Ahoskie P.D.

Instrument Serial No. 008848 705 W. Main St., Ahoskie, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Keyl  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: 008848

Test Date: 10/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	11:57am
AIR BLK	.00	11:57am
ACCY CHK	.07	11:58am
AIR BLK	.00	11:59am
SUB TEST	.00	12:00pm
AIR BLK	.00	12:01pm
SUB TEST	.00	12:02pm
AIR BLK	.00	12:03pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*HERTFORD COUNTY AHOSKIE PD 450*

Serial Number: 008848      Test Record Number: 1436  
Test Date: 10/15/2019      Test Time: 12:05pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:05pm
FLO	Pass	12:05pm
FC	Pass	12:05pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:05pm
SRC	Pass	12:05pm
DET	Pass	12:05pm
BAR	Pass	12:05pm
BT	Pass	12:05pm

Blank Tests

Test	Status	Time
AIR	Pass	12:06pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:06pm

CRC Tests

Test	Status	Time
COMP	Pass	12:06pm
CAL	Pass	12:06pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Statesville PD  
Instrument Serial No. 008619 300 S. Tradd St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. [Signature]  
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY STATESVILLE PD 480

Serial Number: 008619  
Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	9:12am
AIR BLK	.00	9:13am
ACCY CHK	.07	9:14am
AIR BLK	.00	9:15am
<b>SUB TEST</b>	<b>.00</b>	<b>9:15am</b>
AIR BLK	.00	9:16am
<b>SUB TEST</b>	<b>.00</b>	<b>9:18am</b>
AIR BLK	.00	9:19am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*IREDELL COUNTY STATESVILLE PD 480*

Serial Number: 008619      Test Record Number: 1635  
Test Date: 10/03/2019      Test Time: 9:21am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:22am
FLO	Pass	9:22am
FC	Pass	9:22am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:22am
SRC	Pass	9:22am
DET	Pass	9:22am
BAR	Pass	9:22am
BT	Pass	9:22am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:23am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:23am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:23am
CAL	Pass	9:23am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County JONES Instrument Location JONES COUNTY

Instrument Serial No. 008705 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

JONES COUNTY DETENTION CENTER 510

Serial Number: 008705

Test Date: 10/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	1:30pm
AIR BLK	.00	1:31pm
ACCY CHK	.07	1:32pm
AIR BLK	.00	1:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:33pm</b>
AIR BLK	.00	1:34pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:36pm</b>
AIR BLK	.00	1:37pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

JONES COUNTY DETENTION CENTER 510

Serial Number: 008705      Test Record Number: 1369  
Test Date: 10/15/2019      Test Time: 1:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:38pm
FLO	Pass	1:38pm
FC	Pass	1:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:38pm
SRC	Pass	1:38pm
DET	Pass	1:38pm
BAR	Pass	1:38pm
BT	Pass	1:38pm

Blank Tests

Test	Status	Time
AIR	Pass	1:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:38pm

CRC Tests

Test	Status	Time
COMP	Pass	1:39pm
CAL	Pass	1:39pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lenoir Instrument Location Kinston P.P.

Instrument Serial No. 008624 205 E. KING ST., KINSTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Lincoln Kane  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624

Test Date: 10/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 0045-5468

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	1:48pm
AIR BLK	.00	1:48pm
ACCY CHK	.08	1:49pm
AIR BLK	.00	1:50pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:50pm</b>
AIR BLK	.00	1:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:53pm</b>
AIR BLK	.00	1:54pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*LENOIR COUNTY KINSTON PD 530*

Serial Number: 008624      Test Record Number: 1795  
Test Date: 10/18/2019      Test Time: 1:54pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:55pm
FLO	Pass	1:55pm
FC	Pass	1:55pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:55pm
SRC	Pass	1:55pm
DET	Pass	1:55pm
BAR	Pass	1:55pm
BT	Pass	1:55pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:56pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:56pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:56pm
CAL	Pass	1:56pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County LENOIR Instrument Location LENOIR Co. S.O.  
Instrument Serial No. 008639 130 QUEEN ST., KINSTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18<sup>th</sup> day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kees  
Signature of Certifying Official

6457  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639  
Test Date: 10/18/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 0045-5468

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	1:22pm
AIR BLK	.00	1:22pm
ACCY CHK	.07	1:23pm
AIR BLK	.00	1:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:25pm</b>
AIR BLK	.00	1:26pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:27pm</b>
AIR BLK	.00	1:28pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*LENOIR COUNTY LENOIR CO SO 530*

Serial Number: 008639      Test Record Number: 3351  
Test Date: 10/18/2019      Test Time: 1:28pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:29pm
FLO	Pass	1:29pm
FC	Pass	1:29pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:29pm
SRC	Pass	1:29pm
DET	Pass	1:29pm
BAR	Pass	1:29pm
BT	Pass	1:29pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:30pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:30pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:30pm
CAL	Pass	1:30pm

Preventive Maintenance  
Status: Pass

  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County McDowell Instrument Location McDowell Co Jail  
Instrument Serial No. 008892 Merion, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MCDOWELL COUNTY JAIL 580

Serial Number: 008892

Test Date: 10/15/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	3:33pm
AIR BLK	.00	3:34pm
ACCY CHK	.07	3:35pm
AIR BLK	.00	3:36pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:36pm</b>
AIR BLK	.00	3:37pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:39pm</b>
AIR BLK	.00	3:40pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MCDOWELL COUNTY JAIL 580

Serial Number: 008892 Test Record Number: 799

Test Date: 10/15/2019 Test Time: 3:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:41pm
FLO	Pass	3:41pm
FC	Pass	3:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:41pm
SRC	Pass	3:41pm
DET	Pass	3:41pm
BAR	Pass	3:41pm
BT	Pass	3:41pm

Blank Tests

Test	Status	Time
AIR	Pass	3:42pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:42pm

CRC Tests

Test	Status	Time
COMP	Pass	3:42pm
CAL	Pass	3:42pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Mecklenburg County Sheriff's Office  
Instrument Serial No. 008690 804 E. 4th St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY SHERIFFS OFFICE 590

Serial Number: 008690  
Test Date: 10/09/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020


Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG902106  
Exp Date: 01/21/2021

Test	g/210L	Time
DIAG	Pass	10:21am
AIR BLK	.00	10:22am
ACCY CHK	.08	10:23am
AIR BLK	.00	10:24am
<b>SUB TEST</b>	<b>.00</b>	<b>10:25am</b>
AIR BLK	.00	10:26am
<b>SUB TEST</b>	<b>.00</b>	<b>10:27am</b>
AIR BLK	.00	10:28am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst  
Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**MECKLENBURG COUNTY SHERIFFS OFFICE 590**

Serial Number: 008690      Test Record Number: 6353  
Test Date: 10/09/2019      Test Time: 10:29am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:29am
FLO	Pass	10:29am
FC	Pass	10:30am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:30am
SRC	Pass	10:30am
DET	Pass	10:30am
BAR	Pass	10:30am
BT	Pass	10:30am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:30am


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:30am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:31am
CAL	Pass	10:31am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location CPD-LEC  
Instrument Serial No. 008091 601 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008691

Test Date: 10/10/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.08	11:33am
AIR BLK	.00	11:34am
<b>SUB TEST</b>	<b>.00</b>	<b>11:35am</b>
AIR BLK	.00	11:36am
<b>SUB TEST</b>	<b>.00</b>	<b>11:38am</b>
AIR BLK	.00	11:39am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008691      Test Record Number: 7819  
Test Date: 10/10/2019      Test Time: 11:40am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40am
FLO	Pass	11:40am
FC	Pass	11:40am

Temperature Tests

Test	Status	Time
FC1	Pass	11:41am
SRC	Pass	11:41am
DET	Pass	11:41am
BAR	Pass	11:41am
BT	Pass	11:41am

Blank Tests

Test	Status	Time
AIR	Pass	11:41am


Printer Tests

Test	Status	Time
PRNT	Pass	11:41am

CRC Tests

Test	Status	Time
COMP	Pass	11:41am
CAL	Pass	11:41am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mitchell Instrument Location Spruce Pine PD  
Instrument Serial No. 008720 Spruce Pine, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG807101


Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	12:44pm
AIR BLK	.00	12:45pm
ACCY CHK	.08	12:46pm
AIR BLK	.00	12:47pm
SUB TEST	.00	12:47pm
AIR BLK	.00	12:48pm
SUB TEST	.00	12:50pm
AIR BLK	.00	12:51pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726      Test Record Number: 1020  
Test Date: 10/24/2019      Test Time: 12:51pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:52pm
FLO	Pass	12:52pm
FC	Pass	12:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:52pm
SRC	Pass	12:52pm
DET	Pass	12:52pm
BAR	Pass	12:52pm
BT	Pass	12:52pm

Blank Tests

Test	Status	Time
AIR	Pass	12:52pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:53pm

CRC Tests

Test	Status	Time
COMP	Pass	12:53pm
CAL	Pass	12:53pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Montgomery Co.

Instrument Location Montgomery Co. Detention Center


Instrument Serial No. 0087109

Troy NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MONTGOMERY COUNTY DETENTION CENTER 610

Serial Number: 008709  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG919902

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	1:36pm
AIR BLK	.00	1:37pm
ACCY CHK	.07	1:37pm
AIR BLK	.00	1:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:39pm</b>
AIR BLK	.00	1:40pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:42pm</b>
AIR BLK	.00	1:43pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*MONTGOMERY COUNTY DETENTION CENTER 610*

Serial Number: 008709      Test Record Number: 1192  
Test Date: 10/01/2019      Test Time: 1:43pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:44pm
FLO	Pass	1:44pm
FC	Pass	1:44pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:44pm
SRC	Pass	1:44pm
DET	Pass	1:44pm
BAR	Pass	1:44pm
BT	Pass	1:44pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:45pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:45pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:45pm
CAL	Pass	1:45pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Montgomery Co. Instrument Location Montgomery Co. Detention Center  
Instrument Serial No. 078651 Troy, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

MONTGOMERY COUNTY DETENTION CENTER 610

Serial Number: 008657  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911506

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	1:31pm
AIR BLK	.00	1:32pm
ACCY CHK	.08	1:32pm
AIR BLK	.00	1:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:34pm</b>
AIR BLK	.00	1:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:36pm</b>
AIR BLK	.00	1:37pm

Reported EC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*MONTGOMERY COUNTY DETENTION CENTER 610*

Serial Number: 008657      Test Record Number: 1629  
Test Date: 10/01/2019      Test Time: 1:38pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:38pm
FLO	Pass	1:38pm
FC	Pass	1:38pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:38pm
SRC	Pass	1:38pm
DET	Pass	1:38pm
BAR	Pass	1:38pm
BT	Pass	1:38pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:39pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:39pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:39pm
CAL	Pass	1:39pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Moore Instrument Location BAT Mobile unit #5  
Instrument Serial No. 00 8824 Pinehurst

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

664  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Preventive Maintenance**

**MOORE COUNTY BAT MOBILE UNIT 5 620**

Serial Number: 008826      Test Record Number: 8152  
Test Date: 10/31/2019      Test Time: 10:16pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:17pm
FLO	Pass	10:17pm
FC	Pass	10:17pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:17pm
SRC	Pass	10:17pm
DET	Pass	10:17pm
BAR	Pass	10:17pm
BT	Pass	10:17pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:18pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:18pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:18pm
CAL	Pass	10:18pm

Preventive Maintenance  
Status: Pass



**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Subject Test**

MOORE COUNTY BAT MOBILE UNIT 5 620

Serial Number: 008826  
Test Date: 10/31/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: TODD, SHANE C  
Permit Number: 11391E  
Effective:  
07/27/2018-07/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG807101  
Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	10:09pm
AIR BLK	.00	10:10pm
ACCY CHK	.08	10:10pm
AIR BLK	.00	10:11pm
SUB TEST	.00	10:12pm
AIR BLK	.00	10:13pm
SUB TEST	.00	10:14pm
AIR BLK	.00	10:15pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Moore Instrument Location BAT Mobile units  
Instrument Serial No. 008575 Pinehurst

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

664  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MOORE COUNTY BAT MOBILE UNIT 5 620

Serial Number: 008575  
Test Date: 10/31/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TODD, SHANE C

Permit Number: 11391E

Effective:

07/27/2018-07/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	10:12pm
AIR BLK	.00	10:13pm
ACCY CHK	.08	10:13pm
AIR BLK	.00	10:14pm
SUB TEST	.00	10:15pm
AIR BLK	.00	10:16pm
SUB TEST	.00	10:17pm
AIR BLK	.00	10:18pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MOORE COUNTY BAT MOBILE UNIT 5 620

Serial Number: 008575      Test Record Number: 1192  
Test Date: 10/31/2019      Test Time: 10:19pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:19pm
FLO	Pass	10:19pm
FC	Pass	10:20pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:20pm
SRC	Pass	10:20pm
DET	Pass	10:20pm
BAR	Pass	10:20pm
BT	Pass	10:20pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:20pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:20pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:21pm
CAL	Pass	10:21pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County BNSLOW

Instrument Location CAMP LEJEUNE

Instrument Serial No. 008920

PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bann

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ONslow COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920  
Test Date: 10/31/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901


Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	2:33pm
AIR BLK	.00	2:34pm
ACCY CHK	.08	2:34pm
AIR BLK	.00	2:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:36pm</b>
AIR BLK	.00	2:37pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:38pm</b>
AIR BLK	.00	2:39pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR



\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ONslow COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920      Test Record Number: 1718  
Test Date: 10/31/2019      Test Time: 2:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:40pm
FLO	Pass	2:40pm
FC	Pass	2:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:40pm
SRC	Pass	2:40pm
DET	Pass	2:40pm
BAR	Pass	2:40pm
BT	Pass	2:40pm

Blank Tests

Test	Status	Time
AIR	Pass	2:41pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:41pm

CRC Tests

Test	Status	Time
COMP	Pass	2:41pm
CAL	Pass	2:41pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ONslow Instrument Location MCA3 NEW RIVER  
Instrument Serial No. 008917 PMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bann  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY MCAS NEW RIVER 660

Serial Number: 008917

Test Date: 10/31/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG814902


Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	1:00pm
AIR BLK	.00	1:01pm
ACCY CHK	.07	1:02pm
AIR BLK	.00	1:03pm
SUB TEST	.00	1:03pm
AIR BLK	.00	1:04pm
SUB TEST	.00	1:06pm
AIR BLK	.00	1:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY MCAS NEW RIVER 660

Serial Number: 008917 Test Record Number: 776  
Test Date: 10/31/2019 Test Time: 1:07pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:07pm
FLO	Pass	1:07pm
FC	Pass	1:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:07pm
SRC	Pass	1:07pm
DET	Pass	1:07pm
BAR	Pass	1:07pm
BT	Pass	1:07pm

Blank Tests

Test	Status	Time
AIR	Pass	1:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:08pm

CRC Tests

Test	Status	Time
COMP	Pass	1:08pm
CAL	Pass	1:08pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County PAMLICO Instrument Location PAMLICO COUNTY  
Instrument Serial No. 008640 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Rg Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PAMLICO COUNTY DETENTION CENTER 680

Serial Number: 008640  
Test Date: 10/16/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 15671E  
Effective:  
06/01/2019-06/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG814902  
Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:52am
ACCY CHK	.07	11:52am
AIR BLK	.00	11:53am
<b>SUB TEST</b>	<b>.00</b>	<b>11:54am</b>
AIR BLK	.00	11:55am
<b>SUB TEST</b>	<b>.00</b>	<b>11:56am</b>
AIR BLK	.00	11:57am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

**PAMLICO COUNTY DETENTION CENTER 680**

Serial Number: 008640      Test Record Number: 1385  
Test Date: 10/16/2019      Test Time: 11:57am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:59am

**Printer Tests**

Test	Status	Time
PRNT	Pass	11:59am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location Pasquotank Co. Public Safety  
Instrument Serial No. 008950 Bldg., 200 E. Colonial St., Elizabeth  
City, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda K. Kees  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PASQUOTANK COUNTY PUBLIC SAFETY BLDG  
690

Serial Number: 008950  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 0045-5468

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	11:18am
AIR BLK	.00	11:18am
ACCY CHK	.08	11:19am
AIR BLK	.00	11:20am
SUB TEST	.00	11:21am
AIR BLK	.00	11:21am
SUB TEST	.00	11:23am
AIR BLK	.00	11:24am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008950      Test Record Number: 1724  
Test Date: 10/02/2019      Test Time: 11:25am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:25am
FLO	Pass	11:25am
FC	Pass	11:25am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:25am
SRC	Pass	11:25am
DET	Pass	11:25am
BAR	Pass	11:25am
BT	Pass	11:25am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:26am

**Printer Tests**

Test	Status	Time
PRNT	Pass	11:26am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:26am
CAL	Pass	11:26am

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location Elizabeth City P.D.  
Instrument Serial No. 008941 315 E. MAIN ST., Elizabeth City, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of OCTOBER, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James A. Keene  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941  
Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 0045-5468

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	11:52am
AIR BLK	.00	11:53am
ACCY CHK	.08	11:54am
AIR BLK	.00	11:55am
<b>SUB TEST</b>	<b>.00</b>	<b>11:57am</b>
AIR BLK	.00	11:58am
<b>SUB TEST</b>	<b>.00</b>	<b>11:59am</b>
AIR BLK	.00	12:00pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures**

**Forensic Tests for Alcohol Branch**

**Department of Health and Human Services**

**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

PASQUOTANK COUNTY ELIZABETH CITY 690

Serial Number: 008941      Test Record Number: 1390  
Test Date: 10/02/2019      Test Time: 12:01pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:01pm
SRC	Pass	12:01pm
DET	Pass	12:01pm
BAR	Pass	12:01pm
BT	Pass	12:01pm

Blank Tests

Test	Status	Time
AIR	Pass	12:02pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:02pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance  
Status: Pass

---

**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County PENDER Instrument Location PENDER COUNTY  
Instrument Serial No. 008946 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bane  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

PENDER COUNTY DETENTION CENTER 700

Serial Number: 008946  
Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	2:21pm
AIR BLK	.00	2:22pm
ACCY CHK	.08	2:23pm
AIR BLK	.00	2:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:24pm</b>
AIR BLK	.00	2:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:27pm</b>
AIR BLK	.00	2:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

*PENDER COUNTY DETENTION CENTER 700*

Serial Number: 008946      Test Record Number: 1032  
Test Date: 10/04/2019      Test Time: 2:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:29pm
FLO	Pass	2:29pm
FC	Pass	2:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:29pm
SRC	Pass	2:29pm
DET	Pass	2:29pm
BAR	Pass	2:29pm
BT	Pass	2:29pm

Blank Tests

Test	Status	Time
AIR	Pass	2:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:30pm

CRC Tests

Test	Status	Time
COMP	Pass	2:30pm
CAL	Pass	2:30pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County PENDER Instrument Location PENDER COUNTY

Instrument Serial No. 008935 DETENTION CENTER

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 41 day of OCTOBER 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PENDER COUNTY DETENTION CENTER 700

Serial Number: 008935  
Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	2:21pm
AIR BLK	.00	2:22pm
ACCY CHK	.07	2:22pm
AIR BLK	.00	2:23pm
SUB TEST	.00	2:24pm
AIR BLK	.00	2:25pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY DETENTION CENTER 700

Serial Number: 008935      Test Record Number: 2518  
Test Date: 10/04/2019      Test Time: 2:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:28pm
FLO	Pass	2:28pm
FC	Pass	2:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:29pm
SRC	Pass	2:29pm
DET	Pass	2:29pm
BAR	Pass	2:29pm
BT	Pass	2:29pm

Blank Tests

Test	Status	Time
AIR	Pass	2:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:29pm

CRC Tests

Test	Status	Time
COMP	Pass	2:30pm
CAL	Pass	2:30pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County PENDER Instrument Location PENDER COUNTY  
Instrument Serial No. 008948 GOVERNMENT ANNEX

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PENDER COUNTY GOVERNMENT ANNEX 700

Serial Number: 008948  
Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG904301


Exp Date: 02/12/2021

Test	g/210L	Time
DIAG	Pass	4:04pm
AIR BLK	.00	4:05pm
ACCY CHK	.08	4:05pm
AIR BLK	.00	4:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:07pm</b>
AIR BLK	.00	4:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:09pm</b>
AIR BLK	.00	4:10pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY GOVERNMENT ANNEX 700

Serial Number: 008948      Test Record Number: 903  
Test Date: 10/04/2019      Test Time: 4:11pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	4:12pm
FLO	Pass	4:12pm
FC	Pass	4:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:12pm
SRC	Pass	4:12pm
DET	Pass	4:12pm
BAR	Pass	4:12pm
BT	Pass	4:12pm

Blank Tests

Test	Status	Time
AIR	Pass	4:12pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:13pm

CRC Tests

Test	Status	Time
COMP	Pass	4:13pm
CAL	Pass	4:13pm

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Person Instrument Location Person Co LEC

Instrument Serial No. 008924 120 COURT ST Roxboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barnes  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008924  
Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG821801

Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	4:38pm
AIR BLK	.00	4:39pm
ACCY CHK	.08	4:40pm
AIR BLK	.00	4:41pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:42pm</b>
AIR BLK	.00	4:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>4:44pm</b>
AIR BLK	.00	4:45pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008924      Test Record Number: 1468  
Test Date: 10/03/2019      Test Time: 4:45pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:46pm
FLO	Pass	4:46pm
FC	Pass	4:46pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:46pm
SRC	Pass	4:46pm
DET	Pass	4:46pm
BAR	Pass	4:46pm
BT	Pass	4:46pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:47pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	4:47pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:47pm
CAL	Pass	4:47pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Person Instrument Location Person Co. LEC

Instrument Serial No. 008799 120 COURT ST ROXBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Aakes Baeres

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008799  
Test Date: 10/03/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, STOKES  
Permit Number: 11434E  
Effective:  
04/01/2019-04/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG821801  
Exp Date: 08/06/2020

Test	g/210L	Time
DIAG	Pass	4:38pm
AIR BLK	.00	4:38pm
ACCY CHK	.08	4:39pm
AIR BLK	.00	4:40pm
SUB TEST	.00	4:41pm
AIR BLK	.00	4:41pm
SUB TEST	.00	4:43pm
AIR BLK	.00	4:44pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008799      Test Record Number: 3041  
Test Date: 10/03/2019      Test Time: 4:44pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:45pm
FLO	Pass	4:45pm
FC	Pass	4:45pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:45pm
SRC	Pass	4:45pm
DET	Pass	4:45pm
BAR	Pass	4:45pm
BT	Pass	4:45pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:46pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:46pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:46pm
CAL	Pass	4:46pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Ayden P.D.  
Instrument Serial No. 0086666 4144 West Ave., Ayden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PITT AYDEN PD 730

Serial Number: 008666

Test Date: 10/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	9:55am
AIR BLK	.00	9:55am
ACCY CHK	.08	9:56am
AIR BLK	.00	9:57am
<b>SUB TEST</b>	<b>.00</b>	<b>9:58am</b>
AIR BLK	.00	9:59am
<b>SUB TEST</b>	<b>.00</b>	<b>10:00am</b>
AIR BLK	.00	10:01am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

PITT AYDEN PD 730

Serial Number: 008666      Test Record Number: 1126  
Test Date: 10/28/2019      Test Time: 10:02am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02am
FLO	Pass	10:02am
FC	Pass	10:02am

Temperature Tests

Test	Status	Time
FC1	Pass	10:03am
SRC	Pass	10:03am
DET	Pass	10:03am
BAR	Pass	10:03am
BT	Pass	10:03am

Blank Tests

Test	Status	Time
AIR	Pass	10:03am


Printer Tests

Test	Status	Time
PRNT	Pass	10:03am

CRC Tests

Test	Status	Time
COMP	Pass	10:03am
CAL	Pass	10:03am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County PIST Instrument Location NAT MORDRE UNIT 6

Instrument Serial No. 008686 UNIVERSAL

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008686

Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	8:25pm
AIR BLK	.00	8:26pm
ACCY CHK	.07	8:26pm
AIR BLK	.00	8:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:28pm</b>
AIR BLK	.00	8:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:30pm</b>
AIR BLK	.00	8:31pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

PITT COUNTY BAT MOBILE UNIT 6 730

Serial Number: 008686      Test Record Number: 6647  
Test Date: 10/04/2019      Test Time: 8:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:36pm
FLO	Pass	8:36pm
FC	Pass	8:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:37pm
SRC	Pass	8:37pm
DET	Pass	8:37pm
BAR	Pass	8:37pm
BT	Pass	8:37pm

Blank Tests

Test	Status	Time
AIR	Pass	8:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:37pm

CRC Tests

Test	Status	Time
COMP	Pass	8:38pm
CAL	Pass	8:38pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008662 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>rd</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008662  
Test Date: 10/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	10:05am
AIR BLK	.00	10:06am
ACCY CHK	.08	10:07am
AIR BLK	.00	10:08am
<b>SUB TEST</b>	<b>.00</b>	<b>10:08am</b>
AIR BLK	.00	10:09am
<b>SUB TEST</b>	<b>.00</b>	<b>10:11am</b>
AIR BLK	.00	10:12am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*PITT COUNTY PITT CO DETENTION 730*

Serial Number: 008662      Test Record Number: 1130  
Test Date: 10/23/2019      Test Time: 10:13am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:13am
SRC	Pass	10:13am
DET	Pass	10:13am
BAR	Pass	10:13am
BT	Pass	10:13am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:14am

**Printer Tests**

Test	Status	Time
PRNT	Pass	10:14am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center  
Instrument Serial No. 008646 124 Detention Dr., Greenville, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646  
Test Date: 10/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	9:43am
AIR BLK	.00	9:44am
ACCY CHK	.08	9:44am
AIR BLK	.00	9:45am
<b>SUB TEST</b>	<b>.00</b>	<b>9:46am</b>
AIR BLK	.00	9:47am
<b>SUB TEST</b>	<b>.00</b>	<b>9:49am</b>
AIR BLK	.00	9:50am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646 Test Record Number: 3933  
Test Date: 10/23/2019 Test Time: 9:53am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:53am
FLO	Pass	9:53am
FC	Pass	9:53am

Temperature Tests

Test	Status	Time
FC1	Pass	9:54am
SRC	Pass	9:54am
DET	Pass	9:54am
BAR	Pass	9:54am
BT	Pass	9:54am

Blank Tests

Test	Status	Time
AIR	Pass	9:54am

Printer Tests

Test	Status	Time
PRNT	Pass	9:54am

CRC Tests

Test	Status	Time
COMP	Pass	9:54am
CAL	Pass	9:54am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008668 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23<sup>rd</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668  
Test Date: 10/23/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902106

Exp Date: 01/21/2021

Test	g/210L	Time
DIAG	Pass	9:22am
AIR BLK	.00	9:23am
ACCY CHK	.07	9:24am
AIR BLK	.00	9:25am
<b>SUB TEST</b>	<b>.00</b>	<b>9:25am</b>
AIR BLK	.00	9:26am
<b>SUB TEST</b>	<b>.00</b>	<b>9:28am</b>
AIR BLK	.00	9:29am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*PITT COUNTY PITT CO DETENTION 730*

Serial Number: 008668      Test Record Number: 3405  
Test Date: 10/23/2019      Test Time: 9:34am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:34am
FLO	Pass	9:34am
FC	Pass	9:34am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:34am
SRC	Pass	9:34am
DET	Pass	9:34am
BAR	Pass	9:34am
BT	Pass	9:34am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:35am

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:35am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:35am
CAL	Pass	9:35am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Polk

Instrument Location Polk County LEC


Instrument Serial No. 008832

164 Government Complex Dr., Columbus

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Maria D. [Signature]  
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008832

Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	12:08pm
AIR BLK	.00	12:09pm
ACCY CHK	.07	12:09pm
AIR BLK	.00	12:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:11pm</b>
AIR BLK	.00	12:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:13pm</b>
AIR BLK	.00	12:14pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008832      Test Record Number: 1546  
Test Date: 10/04/2019      Test Time: 12:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:37pm
FLO	Pass	12:37pm
FC	Pass	12:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:38pm
SRC	Pass	12:38pm
DET	Pass	12:38pm
BAR	Pass	12:38pm
BT	Pass	12:38pm

Blank Tests

Test	Status	Time
AIR	Pass	12:38pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:38pm

CRC Tests

Test	Status	Time
COMP	Pass	12:39pm
CAL	Pass	12:39pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Polk Instrument Location Polk County LEC

Instrument Serial No. 008881 167 Government Complex Dr., Columbus

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. D. [Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008881

Test Date: 10/04/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401


Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	11:50am
AIR BLK	.00	11:50am
ACCY CHK	.08	11:51am
AIR BLK	.00	11:52am
<b>SUB TEST</b>	<b>.00</b>	<b>11:53am</b>
AIR BLK	.00	11:54am
<b>SUB TEST</b>	<b>.00</b>	<b>11:55am</b>
AIR BLK	.00	11:56am

Reported ACx .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

POLK COUNTY POLK COUNTY LEC 740

Serial Number: 008881      Test Record Number: 837  
Test Date: 10/04/2019      Test Time: 11:57am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:57am
FLO	Pass	11:57am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:58am


Printer Tests

Test	Status	Time
PRNT	Pass	11:58am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD**  
**INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Co. Instrument Location Randolph Co. Detention Center  
Instrument Serial No. 008899 Asheboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008899

Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902202

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	12:22pm
AIR BLK	.00	12:23pm
ACCY CHK	.08	12:24pm
AIR BLK	.00	12:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:25pm</b>
AIR BLK	.00	12:26pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:28pm</b>
AIR BLK	.00	12:29pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**RANDOLPH COUNTY DETENTION CENTER 750**

Serial Number: 008899      Test Record Number: 3253  
Test Date: 10/02/2019      Test Time: 12:33pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:34pm
FLO	Pass	12:34pm
FC	Pass	12:34pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:34pm
SRC	Pass	12:34pm
DET	Pass	12:34pm
BAR	Pass	12:34pm
BT	Pass	12:34pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:35pm

**Printer Tests**


Test	Status	Time
PRNT	Pass	12:35pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:35pm
CAL	Pass	12:35pm

Preventive Maintenance

Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Co. Instrument Location Randolph Co. Detention Center  
Instrument Serial No. 008860 Asheboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008860

Test Date: 10/02/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG902202

Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	12:21pm
AIR BLK	.00	12:22pm
ACCY CHK	.07	12:23pm
AIR BLK	.00	12:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:24pm</b>
AIR BLK	.00	12:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:27pm</b>
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY DETENTION CENTER 750

Serial Number: 008860      Test Record Number: 2745  
Test Date: 10/02/2019      Test Time: 12:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:28pm
FLO	Pass	12:28pm
FC	Pass	12:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:29pm
SRC	Pass	12:29pm
DET	Pass	12:29pm
BAR	Pass	12:29pm
BT	Pass	12:29pm

Blank Tests

Test	Status	Time
AIR	Pass	12:29pm

Printer Tests


Test	Status	Time
PRNT	Pass	12:29pm

CRC Tests

Test	Status	Time
COMP	Pass	12:29pm
CAL	Pass	12:29pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rowan Instrument Location China Grove  
Instrument Serial No. 008862 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25<sup>th</sup> day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROWAN COUNTY CHINA GROVE PD 790

Serial Number: 008862  
Test Date: 10/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG821401

Exp Date: 08/02/2020

Test	g/210L	Time
DIAG	Pass	11:05am
AIR BLK	.00	11:05am
ACCY CHK	.07	11:06am
AIR BLK	.00	11:07am
<b>SUB TEST</b>	<b>.00</b>	<b>11:08am</b>
AIR BLK	.00	11:09am
<b>SUB TEST</b>	<b>.00</b>	<b>11:11am</b>
AIR BLK	.00	11:12am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROWAN COUNTY CHINA GROVE PD 790

Serial Number: 008862      Test Record Number: 881  
Test Date: 10/25/2019      Test Time: 11:16am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:16am
FLO	Pass	11:16am
FC	Pass	11:16am

Temperature Tests

Test	Status	Time
FC1	Pass	11:16am
SRC	Pass	11:16am
DET	Pass	11:16am
BAR	Pass	11:16am
BT	Pass	11:16am

Blank Tests

Test	Status	Time
AIR	Pass	11:17am


Printer Tests

Test	Status	Time
PRNT	Pass	11:17am

CRC Tests

Test	Status	Time
COMP	Pass	11:17am
CAL	Pass	11:17am

Preventive Maintenance  
Status: Pass

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rowan Instrument Location Salisbury  
Instrument Serial No. 008835 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25<sup>th</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

655  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835

Test Date: 10/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	1:29pm
AIR BLK	.00	1:30pm
ACCY CHK	.08	1:30pm
AIR BLK	.00	1:31pm
SUB TEST	.00	1:32pm
AIR BLK	.00	1:33pm
SUB TEST	.00	1:34pm
AIR BLK	.00	1:35pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835      Test Record Number: 2395  
Test Date: 10/25/2019      Test Time: 1:38pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:39pm
FLO	Pass	1:39pm
FC	Pass	1:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:39pm

CRC Tests

Test	Status	Time
COMP	Pass	1:40pm
CAL	Pass	1:40pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rowan Instrument Location Salisbury  
Instrument Serial No. 008868 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25<sup>th</sup> day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Dale Farley 655  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008868  
Test Date: 10/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: FARLEY, CYNTHIA D

Permit Number: 24123E

Effective:

11/01/2018-11/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	1:06pm
AIR BLK	.00	1:07pm
ACCY CHK	.07	1:08pm
AIR BLK	.00	1:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:10pm</b>
AIR BLK	.00	1:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:13pm</b>
AIR BLK	.00	1:14pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008868      Test Record Number: 3083  
Test Date: 10/25/2019      Test Time: 1:18pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:19pm
FLO	Pass	1:19pm
FC	Pass	1:19pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:19pm
SRC	Pass	1:19pm
DET	Pass	1:19pm
BAR	Pass	1:19pm
BT	Pass	1:19pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:19pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:19pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:20pm
CAL	Pass	1:20pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Scotland Co. Instrument Location Scotland Co. Sheriff's Office

Instrument Serial No. 008861 LAURINBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY SHERIFF'S OFFICE 820

Serial Number: 008861  
Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911506

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	12:53pm
AIR BLK	.00	12:53pm
ACCY CHK	.08	12:54pm
AIR BLK	.00	12:55pm
SUB TEST	.00	12:55pm
AIR BLK	.00	12:56pm
SUB TEST	.00	12:58pm
AIR BLK	.00	12:59pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*SCOTLAND COUNTY SHERIFF'S OFFICE 820*

Serial Number: 008861      Test Record Number: 1581  
Test Date: 10/03/2019      Test Time: 1:00pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:00pm
FLO	Pass	1:00pm
FC	Pass	1:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:00pm
SRC	Pass	1:00pm
DET	Pass	1:00pm
BAR	Pass	1:00pm
BT	Pass	1:00pm

Blank Tests

Test	Status	Time
AIR	Pass	1:01pm

Printer Tests


Test	Status	Time
PRNT	Pass	1:01pm

CRC Tests

Test	Status	Time
COMP	Pass	1:01pm
CAL	Pass	1:01pm

Preventive Maintenance

Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Scotland Co. Instrument Location Laurinburg P.D.  
Instrument Serial No. 008834 Laurinburg, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834

Test Date: 10/03/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 07682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG911506

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	12:18pm
AIR BLK	.00	12:19pm
ACCY CHK	.07	12:20pm
AIR BLK	.00	12:21pm
SUB TEST	.00	12:21pm
AIR BLK	.00	12:22pm
SUB TEST	.00	12:24pm
AIR BLK	.00	12:25pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

SCOTLAND COUNTY LAURINBURG PD 820

Serial Number: 008834      Test Record Number: 909  
Test Date: 10/03/2019      Test Time: 12:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:26pm
FLO	Pass	12:26pm
FC	Pass	12:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:26pm
SRC	Pass	12:26pm
DET	Pass	12:26pm
BAR	Pass	12:26pm
BT	Pass	12:26pm

Blank Tests

Test	Status	Time
AIR	Pass	12:27pm

Printer Tests


Test	Status	Time
PRNT	Pass	12:27pm

CRC Tests

Test	Status	Time
COMP	Pass	12:27pm
CAL	Pass	12:27pm

Preventive Maintenance

Status: Pass



Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Stanly Instrument Location Stanly County SO  
Instrument Serial No. 008824 126 S. 3<sup>rd</sup> St. Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008824  
Test Date: 10/07/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814902

Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	10:42am
AIR BLK	.00	10:42am
ACCY CHK	.08	10:43am
AIR BLK	.00	10:44am
<b>SUB TEST</b>	<b>.00</b>	<b>10:45am</b>
AIR BLK	.00	10:46am
<b>SUB TEST</b>	<b>.00</b>	<b>10:48am</b>
AIR BLK	.00	10:48am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008824 Test Record Number: 1463  
Test Date: 10/07/2019 Test Time: 10:51am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:52am

Temperature Tests

Test	Status	Time
FC1	Pass	10:52am
SRC	Pass	10:52am
DET	Pass	10:52am
BAR	Pass	10:52am
BT	Pass	10:52am

Blank Tests

Test	Status	Time
AIR	Pass	10:52am


Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:53am
CAL	Pass	10:53am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Stanly County SO  
Instrument Serial No. 008842 126 S. 3<sup>rd</sup> St, Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of October, 20 19, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008842

Test Date: 10/07/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test


Lot Number: AG919901

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	10:32am
AIR BLK	.00	10:33am
ACCY CHK	.08	10:34am
AIR BLK	.00	10:34am
<b>SUB TEST</b>	<b>.00</b>	<b>10:35am</b>
AIR BLK	.00	10:36am
<b>SUB TEST</b>	<b>.00</b>	<b>10:38am</b>
AIR BLK	.00	10:38am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst  
Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

STANLY COUNTY STANLY COUNTY SO 830

Serial Number: 008842      Test Record Number: 2264  
Test Date: 10/07/2019      Test Time: 10:40am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:40am
FLO	Pass	10:40am
FC	Pass	10:40am

Temperature Tests

Test	Status	Time
FC1	Pass	10:40am
SRC	Pass	10:40am
DET	Pass	10:40am
BAR	Pass	10:40am
BT	Pass	10:40am

Blank Tests

Test	Status	Time
AIR	Pass	10:41am


Printer Tests

Test	Status	Time
PRNT	Pass	10:41am

CRC Tests

Test	Status	Time
COMP	Pass	10:41am
CAL	Pass	10:41am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County VANCE Instrument Location Vance Co Sheriff's Office

Instrument Serial No. 008870 156 Church St. Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stacey Barnes  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870

Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021


Test	g/210L	Time
DIAG	Pass	3:33pm
AIR BLK	.00	3:33pm
ACCY CHK	.08	3:34pm
AIR BLK	.00	3:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:36pm</b>
AIR BLK	.00	3:37pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:39pm</b>
AIR BLK	.00	3:40pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870      Test Record Number: 2844  
Test Date: 10/01/2019      Test Time: 3:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:41pm
FLO	Pass	3:41pm
FC	Pass	3:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:41pm
SRC	Pass	3:41pm
DET	Pass	3:41pm
BAR	Pass	3:41pm
BT	Pass	3:41pm

Blank Tests

Test	Status	Time
AIR	Pass	3:42pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:42pm

CRC Tests

Test	Status	Time
COMP	Pass	3:42pm
CAL	Pass	3:42pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Vance Instrument Location Vance Co Sheriff's Office

Instrument Serial No. 008937 156 Church St. Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

002

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937  
Test Date: 10/01/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807102

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	3:18pm
AIR BLK	.00	3:19pm
ACCY CHK	.08	3:19pm
AIR BLK	.00	3:20pm
SUB TEST	.00	3:21pm
AIR BLK	.00	3:21pm
SUB TEST	.00	3:23pm
AIR BLK	.00	3:24pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937      Test Record Number: 2777  
Test Date: 10/01/2019      Test Time: 3:25pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:26pm
FLO	Pass	3:26pm
FC	Pass	3:26pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:26pm
SRC	Pass	3:26pm
DET	Pass	3:26pm
BAR	Pass	3:26pm
BT	Pass	3:26pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:27pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	3:27pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:27pm
CAL	Pass	3:27pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Vance Instrument Location Vance Co Sheriff's Office

Instrument Serial No. 008937 156 Church St Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937  
Test Date: 10/24/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG807102


Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	9:17am
AIR BLK	.00	9:18am
ACCY CHK	.08	9:18am
AIR BLK	.00	9:19am
<b>SUB TEST</b>	<b>.00</b>	<b>9:20am</b>
AIR BLK	.00	9:21am
<b>SUB TEST</b>	<b>.00</b>	<b>9:22am</b>
AIR BLK	.00	9:23am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937      Test Record Number: 2788  
Test Date: 10/24/2019      Test Time: 9:24am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:24am
FLO	Pass	9:24am
FC	Pass	9:24am

Temperature Tests

Test	Status	Time
FC1	Pass	9:25am
SRC	Pass	9:25am
DET	Pass	9:25am
BAR	Pass	9:25am
BT	Pass	9:25am

Blank Tests

Test	Status	Time
AIR	Pass	9:25am


Printer Tests

Test	Status	Time
PRNT	Pass	9:25am

CRC Tests

Test	Status	Time
COMP	Pass	9:25am
CAL	Pass	9:25am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Vance Instrument Location Vance Co Sheriff's Office

Instrument Serial No. 008870 156 Church St Henderson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870  
Test Date: 10/24/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, STOKES  
Permit Number: 11434E  
Effective:  
04/01/2019-04/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG911501  
Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	9:16am
AIR BLK	.00	9:17am
ACCY CHK	.08	9:18am
AIR BLK	.00	9:19am
<b>SUB TEST</b>	<b>.00</b>	<b>9:20am</b>
AIR BLK	.00	9:21am
<b>SUB TEST</b>	<b>.00</b>	<b>9:22am</b>
AIR BLK	.00	9:23am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870      Test Record Number: 2870  
Test Date: 10/24/2019      Test Time: 9:24am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:24am
FLO	Pass	9:24am
FC	Pass	9:24am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:24am
SRC	Pass	9:24am
DET	Pass	9:24am
BAR	Pass	9:24am
BT	Pass	9:24am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:25am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:25am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:25am
CAL	Pass	9:25am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake

Instrument Location Wake Co Detention Ctr


Instrument Serial No. 008740

3301 HAMMOND RD RALEIGH NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760

Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG919902

Exp Date: 07/18/2021


Test	g/210L	Time
DIAG	Pass	8:39am
AIR BLK	.00	8:40am
ACCY CHK	.08	8:40am
AIR BLK	.00	8:42am
<b>SUB TEST</b>	<b>.00</b>	<b>8:42am</b>
AIR BLK	.00	8:43am
<b>SUB TEST</b>	<b>.00</b>	<b>8:45am</b>
AIR BLK	.00	8:46am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760 Test Record Number: 3573  
Test Date: 10/01/2019 Test Time: 8:46am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:47am
FLO	Pass	8:47am
FC	Pass	8:47am

Temperature Tests

Test	Status	Time
FC1	Pass	8:47am
SRC	Pass	8:47am
DET	Pass	8:47am
BAR	Pass	8:47am
BT	Pass	8:47am

Blank Tests

Test	Status	Time
AIR	Pass	8:47am

Printer Tests

Test	Status	Time
PRNT	Pass	8:47am

CRC Tests

Test	Status	Time
COMP	Pass	8:48am
CAL	Pass	8:48am

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location Wake Co Detention Ctr

Instrument Serial No. 008778 3301 Hammond Rd Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 20 17 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Boeris

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778  
Test Date: 10/01/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, STOKES  
Permit Number: 11434E  
Effective:  
04/01/2019-04/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG919902  
Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	8:25am
AIR BLK	.00	8:26am
ACCY CHK	.08	8:27am
AIR BLK	.00	8:28am
<b>SUB TEST</b>	<b>.00</b>	<b>8:29am</b>
AIR BLK	.00	8:30am
<b>SUB TEST</b>	<b>.00</b>	<b>8:31am</b>
AIR BLK	.00	8:32am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778      Test Record Number: 4406  
Test Date: 10/01/2019      Test Time: 8:34am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	8:35am
FLO	Pass	8:35am
FC	Pass	8:35am

**Temperature Tests**

Test	Status	Time
FC1	Pass	8:35am
SRC	Pass	8:35am
DET	Pass	8:35am
BAR	Pass	8:35am
BT	Pass	8:35am

**Blank Tests**

Test	Status	Time
AIR	Pass	8:36am

**Printer Tests**

Test	Status	Time
PRNT	Pass	8:36am

**CRC Tests**

Test	Status	Time
COMP	Pass	8:36am
CAL	Pass	8:36am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location Wake Co Detention Ctr  
Instrument Serial No. 008577 3301 HAMMOND RD RALPH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of OCTOBER, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577  
Test Date: 10/01/2019

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, STOKES  
Permit Number: 11434E  
Effective:  
04/01/2019-04/01/2021

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG919902  
Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	8:25am
AIR BLK	.00	8:25am
ACCY CHK	.07	8:26am
AIR BLK	.00	8:27am
<b>SUB TEST</b>	<b>.00</b>	<b>8:28am</b>
AIR BLK	.00	8:28am
<b>SUB TEST</b>	<b>.00</b>	<b>8:30am</b>
AIR BLK	.00	8:31am

Reported AC: 00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577      Test Record Number: 4641  
Test Date: 10/01/2019      Test Time: 8:34am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:34am
FLO	Pass	8:34am
FC	Pass	8:34am

Temperature Tests

Test	Status	Time
FC1	Pass	8:34am
SRC	Pass	8:34am
DET	Pass	8:34am
BAR	Pass	8:34am
BT	Pass	8:34am

Blank Tests

Test	Status	Time
AIR	Pass	8:35am

Printer Tests

Test	Status	Time
PRNT	Pass	8:35am

CRC Tests

Test	Status	Time
COMP	Pass	8:35am
CAL	Pass	8:35am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Wake County Detention Ctr  
Instrument Serial No. 008612 3301 Hammond Rd Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG919902

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	8:24am
AIR BLK	.00	8:25am
ACCY CHK	.07	8:25am
AIR BLK	.00	8:26am
SUB TEST	.00	8:27am
AIR BLK	.00	8:28am
SUB TEST	.00	8:30am
AIR BLK	.00	8:31am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612      Test Record Number: 4322  
Test Date: 10/01/2019      Test Time: 8:33am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:33am
FLO	Pass	8:33am
FC	Pass	8:33am

Temperature Tests

Test	Status	Time
FC1	Pass	8:33am
SRC	Pass	8:33am
DET	Pass	8:33am
BAR	Pass	8:33am
BT	Pass	8:33am

Blank Tests

Test	Status	Time
AIR	Pass	8:34am


Printer Tests

Test	Status	Time
PRNT	Pass	8:34am

CRC Tests

Test	Status	Time
COMP	Pass	8:34am
CAL	Pass	8:34am

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE

Instrument Location CARY PD


Instrument Serial No. 008587

120 Wilkerson Ave. Cary, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY CARY PD 910

Serial Number: 008587

Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911501

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	7:44am
AIR BLK	.00	7:45am
ACCY CHK	.08	7:46am
AIR BLK	.00	7:47am
<b>SUB TEST</b>	<b>.00</b>	<b>7:47am</b>
AIR BLK	.00	7:48am
<b>SUB TEST</b>	<b>.00</b>	<b>7:49am</b>
AIR BLK	.00	7:50am

Reported AC: 00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CARY PD 910

Serial Number: 008587      Test Record Number: 4115  
Test Date: 10/01/2019      Test Time: 7:51am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:51am
FLO	Pass	7:51am
FC	Pass	7:51am

Temperature Tests

Test	Status	Time
FC1	Pass	7:51am
SRC	Pass	7:51am
DET	Pass	7:51am
BAR	Pass	7:51am
BT	Pass	7:51am

Blank Tests

Test	Status	Time
AIR	Pass	7:52am


Printer Tests

Test	Status	Time
PRNT	Pass	7:52am

CRC Tests

Test	Status	Time
COMP	Pass	7:52am
CAL	Pass	7:52am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake

Instrument Location Apex PD STATION 4


Instrument Serial No. 008621

1615 E. Williams St. Apex, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY APEX PD

Serial Number: 008621

Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	7:02am
AIR BLK	.00	7:02am
ACCY CHK	.08	7:03am
AIR BLK	.00	7:04am
<b>SUB TEST</b>	<b>.00</b>	<b>7:04am</b>
AIR BLK	.00	7:05am
<b>SUB TEST</b>	<b>.00</b>	<b>7:07am</b>
AIR BLK	.00	7:08am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY APEX PD

Serial Number: 008621      Test Record Number: 2699  
Test Date: 10/01/2019      Test Time: 7:08am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:09am
FLO	Pass	7:09am
FC	Pass	7:09am

Temperature Tests

Test	Status	Time
FC1	Pass	7:09am
SRC	Pass	7:09am
DET	Pass	7:09am
BAR	Pass	7:09am
BT	Pass	7:09am

Blank Tests

Test	Status	Time
AIR	Pass	7:10am


Printer Tests

Test	Status	Time
PRNT	Pass	7:10am

CRC Tests

Test	Status	Time
COMP	Pass	7:10am
CAL	Pass	7:10am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County WAKE Instrument Location SAF MOORE UNIT 6

Instrument Serial No. 008776 LAWREN

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

463

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008776      Test Record Number: 3540  
Test Date: 10/11/2019      Test Time: 10:12pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:13pm
FLO	Pass	10:13pm
FC	Pass	10:13pm

Temperature Tests

Test	Status	Time
FCI	Pass	10:13pm
SRC	Pass	10:13pm
DET	Pass	10:13pm
BAR	Pass	10:13pm
BT	Pass	10:13pm

Blank Tests

Test	Status	Time
AIR	Pass	10:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:13pm

CRC Tests

Test	Status	Time
COMP	Pass	10:14pm
CAL	Pass	10:14pm

Preventive Maintenance  
Status: Pass



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008776  
Test Date: 10/11/2019

Citation Number: M0000000-0

Subject's Name:  
PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG902201


Exp Date: 01/22/2021

Test	g/210L	Time
DIAG	Pass	10:03pm
AIR BLK	.00	10:04pm
ACCY CHK	.07	10:05pm
AIR BLK	.00	10:06pm
SUB TEST	.00	10:06pm
AIR BLK	.00	10:07pm
SUB TEST	.00	10:09pm
AIR BLK	.00	10:10pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location JAT MODE UNIT C  
Instrument Serial No. 609637 CALWA

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008637

Test Date: 10/11/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201

Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	10:02pm
AIR BLK	.00	10:03pm
ACCY CHK	.07	10:04pm
AIR BLK	.00	10:05pm
SUB TEST	.00	10:05pm
AIR BLK	.00	10:06pm
SUB TEST	.00	10:07pm
AIR BLK	.00	10:08pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008637      Test Record Number: 3025  
Test Date: 10/11/2019      Test Time: 10:09pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:10pm
FLO	Pass	10:10pm
FC	Pass	10:10pm

**Temperature Tests**

Test	Status	Time
FCL	Pass	10:10pm
SRC	Pass	10:10pm
DET	Pass	10:10pm
BAR	Pass	10:10pm
BT	Pass	10:10pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:11pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:11pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:11pm
CAL	Pass	10:11pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County WAKE Instrument Location NAT MORGUE UNIT 6  
Instrument Serial No. 00 9686 CANON

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008686

Test Date: 10/11/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Varnell, Bryon L

Permit Number: 0036-1210

Effective:

08/14/2019-08/14/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG807101

Exp Date: 03/12/2020

Test	g/210L	Time
DIAG	Pass	10:01pm
AIR BLK	.00	10:02pm
ACCY CHK	.07	10:02pm
AIR BLK	.00	10:03pm
SUB TEST	.00	10:04pm
AIR BLK	.00	10:05pm
SUB TEST	.00	10:06pm
AIR BLK	.00	10:07pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008686 Test Record Number: 6653  
Test Date: 10/11/2019 Test Time: 10:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:14pm
FLO	Pass	10:14pm
FC	Pass	10:14pm

Temperature Tests

Test	Status	Time
FCL	Pass	10:14pm
SRC	Pass	10:14pm
DET	Pass	10:14pm
BAR	Pass	10:14pm
BT	Pass	10:14pm

Blank Tests

Test	Status	Time
AIR	Pass	10:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:15pm

CRC Tests

Test	Status	Time
COMP	Pass	10:15pm
CAL	Pass	10:15pm

Preventive Maintenance  
Status: *Pass*



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WARREN Instrument Location WARREN Co LEC

Instrument Serial No. 008793 128 RAFTERS Lane, WARRINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of October, 2019 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

602  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793  
Test Date: 10/01/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, STOKES

Permit Number: 11434E

Effective:

04/01/2019-04/01/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:26pm
AIR BLK	.00	1:27pm
ACCY CHK	.07	1:27pm
AIR BLK	.00	1:28pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:29pm</b>
AIR BLK	.00	1:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:31pm</b>
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793 Test Record Number: 1527  
Test Date: 10/01/2019 Test Time: 1:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:33pm
FLO	Pass	1:33pm
FC	Pass	1:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:34pm

CRC Tests

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance  
Status: Pass

  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Watauga Instrument Location Watauga Co. Jail  
Instrument Serial No. 008715 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715

Test Date: 10/22/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG911506


Exp Date: 04/25/2021

Test	g/210L	Time
DIAG	Pass	3:00pm
AIR BLK	.00	3:01pm
ACCY CHK	.08	3:01pm
AIR BLK	.00	3:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:03pm</b>
AIR BLK	.00	3:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:06pm</b>
AIR BLK	.00	3:07pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715      Test Record Number: 2314  
Test Date: 10/22/2019      Test Time: 3:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:08pm
FLO	Pass	3:08pm
FC	Pass	3:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:08pm
SRC	Pass	3:08pm
DET	Pass	3:08pm
BAR	Pass	3:08pm
BT	Pass	3:08pm

Blank Tests

Test	Status	Time
AIR	Pass	3:09pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:09pm

CRC Tests

Test	Status	Time
COMP	Pass	3:09pm
CAL	Pass	3:09pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Watauga Instrument Location Boone PD  
Instrument Serial No. 008716 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of October, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716  
Test Date: 10/25/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

06/01/2019-06/01/2021

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG919902

Exp Date: 07/18/2021

Test	g/210L	Time
DIAG	Pass	1:50pm
AIR BLK	.00	1:51pm
ACCY CHK	.08	1:52pm
AIR BLK	.00	1:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:53pm</b>
AIR BLK	.00	1:54pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:55pm</b>
AIR BLK	.00	1:56pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WATAUGA COUNTY BOONE P D 940

Serial Number: 008716      Test Record Number: 2491  
Test Date: 10/25/2019      Test Time: 1:57pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:57pm
FLO	Pass	1:57pm
FC	Pass	1:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:58pm
SRC	Pass	1:58pm
DET	Pass	1:58pm
BAR	Pass	1:58pm
BT	Pass	1:58pm

Blank Tests

Test	Status	Time
AIR	Pass	1:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:58pm

CRC Tests

Test	Status	Time
COMP	Pass	1:58pm
CAL	Pass	1:58pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wayne Instrument Location Seymour Johnson A.F.B.

Instrument Serial No. 008786 1010 Vermont Garrison Rd., Goldsboro  
N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of October, 20 19 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

697  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786  
Test Date: 10/15/2019

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 11646E

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	11:21am
AIR BLK	.00	11:21am
ACCY CHK	.08	11:22am
AIR BLK	.00	11:23am
<b>SUB TEST</b>	<b>.00</b>	<b>11:24am</b>
AIR BLK	.00	11:25am
<b>SUB TEST</b>	<b>.00</b>	<b>11:26am</b>
AIR BLK	.00	11:27am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY SEYMOUR JOHNSON AFB 950

Serial Number: 008786      Test Record Number: 336  
Test Date: 10/15/2019      Test Time: 11:29am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:29am
FLO	Pass	11:29am
FC	Pass	11:29am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:29am
SRC	Pass	11:29am
DET	Pass	11:29am
BAR	Pass	11:29am
BT	Pass	11:29am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:30am


**Printer Tests**

Test	Status	Time
PRNT	Pass	11:30am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:30am
CAL	Pass	11:30am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilson

Instrument Location Wilson Co. Detention CTR.


Instrument Serial No. 008588

100 E. Greene St., Wilson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28<sup>th</sup> day of OCTOBER, 2019, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008588

Test Date: 10/28/2019

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: Keesler, Linda A

Permit Number: 0045-5468

Effective:

07/08/2019-07/08/2021

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG814901


Exp Date: 05/29/2020

Test	g/210L	Time
DIAG	Pass	2:42pm
AIR BLK	.00	2:43pm
ACCY CHK	.08	2:43pm
AIR BLK	.00	2:44pm
SUB TEST	.00	2:45pm
AIR BLK	.00	2:46pm
SUB TEST	.00	2:47pm
AIR BLK	.00	2:48pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

WILSON COUNTY DETENTION CENTER 970

Serial Number: 008588      Test Record Number: 1026  
Test Date: 10/28/2019      Test Time: 2:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:31pm
FLO	Pass	2:31pm
FC	Pass	2:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:31pm
SRC	Pass	2:31pm
DET	Pass	2:31pm
BAR	Pass	2:31pm
BT	Pass	2:31pm

Blank Tests

Test	Status	Time
AIR	Pass	2:32pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:32pm

CRC Tests

Test	Status	Time
COMP	Pass	2:32pm
CAL	Pass	2:32pm

Preventive Maintenance  
Status: Pass

  
Analyst