

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ALAMANCE Instrument Location ALAMANCE CO JAIL  
Instrument Serial No. 008913 109 S. MAPLE ST  
GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913  
Test Date: 03/01/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	1:09pm
AIR BLK	.00	1:10pm
ACCY CHK	.08	1:11pm
AIR BLK	.00	1:12pm
SUB TEST	.00	1:12pm
AIR BLK	.00	1:13pm
SUB TEST	.00	1:15pm
AIR BLK	.00	1:17pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913      Test Record Number: 3095  
Test Date: 03/01/2018      Test Time: 1:18pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:19pm
FLO	Pass	1:19pm
FC	Pass	1:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:19pm
SRC	Pass	1:19pm
DET	Pass	1:19pm
BAR	Pass	1:19pm
BT	Pass	1:19pm

Blank Tests

Test	Status	Time
AIR	Pass	1:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:19pm

CRC Tests

Test	Status	Time
COMP	Pass	1:20pm
CAL	Pass	1:20pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alamance Instrument Location Bat mobile Unit 8  
Instrument Serial No. 008615 Haw River PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner  
Signature of Certifying Official

644  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

ALAMANCE COUNTY BAT MOBILE UNIT 8 000

Serial Number: 008615  
Test Date: 03/03/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	11:41pm
AIR BLK	.00	11:42pm
ACCY CHK	.07	11:43pm
AIR BLK	.00	11:44pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:44pm</b>
AIR BLK	.00	11:45pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:47pm</b>
AIR BLK	.00	11:47pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



**Analyst**

**Intox EC/IR-II: Preventive Maintenance**

ALAMANCE COUNTY BAT MOBILE UNIT 8 000

Serial Number: 008615      Test Record Number: 5524  
Test Date: 03/03/2018      Test Time: 11:49pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:50pm
FLO	Pass	11:50pm
FC	Pass	11:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:50pm
SRC	Pass	11:50pm
DET	Pass	11:50pm
BAR	Pass	11:50pm
BT	Pass	11:50pm

Blank Tests

Test	Status	Time
AIR	Pass	11:51pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:51pm

CRC Tests

Test	Status	Time
COMP	Pass	11:51pm
CAL	Pass	11:51pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Alamance Instrument Location Bat Mobile Unit 8

Instrument Serial No. 008816 Haw River PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of March, 20 18, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY BAT MOBILE UNIT 8 000

Serial Number: 008816  
Test Date: 03/04/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	12:06am
AIR BLK	.00	12:07am
ACCY CHK	.07	12:08am
AIR BLK	.00	12:09am
SUB TEST	.00	12:10am
AIR BLK	.00	12:11am
SUB TEST	.00	12:12am
AIR BLK	.00	12:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY BAT MOBILE UNIT 8 000

Serial Number: 008816      Test Record Number: 7385  
Test Date: 03/04/2018      Test Time: 12:15am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:15am
FLO	Pass	12:15am
FC	Pass	12:15am

Temperature Tests

Test	Status	Time
FC1	Pass	12:15am
SRC	Pass	12:15am
DET	Pass	12:15am
BAR	Pass	12:15am
BT	Pass	12:15am

Blank Tests

Test	Status	Time
AIR	Pass	12:16am


Printer Tests

Test	Status	Time
PRNT	Pass	12:16am

CRC Tests

Test	Status	Time
COMP	Pass	12:16am
CAL	Pass	12:16am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Alexander Instrument Location Alexander County SD  
Instrument Serial No. 008813 91 Commercial Park Ave., Taylorsville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALEXANDER COUNTY ALEXANDER COUNTY SD  
010

Serial Number: 008813  
Test Date: 03/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:27am
AIR BLK	.00	8:28am
ACCY CHK	.08	8:29am
AIR BLK	.00	8:30am
SUB TEST	.00	8:31am
AIR BLK	.00	8:32am
SUB TEST	.00	8:33am
AIR BLK	.00	8:34am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

ALEXANDER COUNTY ALEXANDER COUNTY SD 010

Serial Number: 008813 Test Record Number: 1736  
Test Date: 03/01/2018 Test Time: 8:36am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:36am
FLO	Pass	8:36am
FC	Pass	8:36am

Temperature Tests

Test	Status	Time
FC1	Pass	8:36am
SRC	Pass	8:36am
DET	Pass	8:36am
BAR	Pass	8:36am
BT	Pass	8:36am

Blank Tests

Test	Status	Time
AIR	Pass	8:37am


Printer Tests

Test	Status	Time
PRNT	Pass	8:37am

CRC Tests

Test	Status	Time
COMP	Pass	8:37am
CAL	Pass	8:37am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Avery Instrument Location Avery Co. Jail  
Instrument Serial No. 008664 Newland, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664

Test Date: 03/23/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	2:36pm
AIR BLK	.00	2:37pm
ACCY CHK	.08	2:37pm
AIR BLK	.00	2:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:39pm</b>
AIR BLK	.00	2:40pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:41pm</b>
AIR BLK	.00	2:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664      Test Record Number: 903  
Test Date: 03/23/2018      Test Time: 2:44pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:44pm
FLO	Pass	2:44pm
FC	Pass	2:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:44pm
SRC	Pass	2:44pm
DET	Pass	2:44pm
BAR	Pass	2:44pm
BT	Pass	2:44pm

Blank Tests

Test	Status	Time
AIR	Pass	2:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:45pm

CRC Tests

Test	Status	Time
COMP	Pass	2:45pm
CAL	Pass	2:45pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Bladen Instrument Location Bladen County  
Instrument Serial No. 008818 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008818  
Test Date: 03/23/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401


Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	2:20pm
AIR BLK	.00	2:21pm
ACCY CHK	.07	2:21pm
AIR BLK	.00	2:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:23pm</b>
AIR BLK	.00	2:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:25pm</b>
AIR BLK	.00	2:27pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008818 Test Record Number: 1348  
Test Date: 03/23/2018 Test Time: 2:27pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:28pm
FLO	Pass	2:28pm
FC	Pass	2:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:28pm
SRC	Pass	2:28pm
DET	Pass	2:28pm
BAR	Pass	2:28pm
BT	Pass	2:28pm

Blank Tests

Test	Status	Time
AIR	Pass	2:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:28pm

CRC Tests

Test	Status	Time
COMP	Pass	2:29pm
CAL	Pass	2:29pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Bladen Instrument Location Bladen County  
Instrument Serial No. 008894 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008894

Test Date: 03/23/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	2:33pm
AIR BLK	.00	2:34pm
ACCY CHK	.08	2:34pm
AIR BLK	.00	2:36pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:37pm</b>
AIR BLK	.00	2:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:39pm</b>
AIR BLK	.00	2:40pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008894 Test Record Number: 1072  
Test Date: 03/23/2018 Test Time: 2:41pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:41pm
FLO	Pass	2:41pm
FC	Pass	2:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:42pm
SRC	Pass	2:42pm
DET	Pass	2:42pm
BAR	Pass	2:42pm
BT	Pass	2:42pm

Blank Tests

Test	Status	Time
AIR	Pass	2:43pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:43pm

CRC Tests

Test	Status	Time
COMP	Pass	2:43pm
CAL	Pass	2:43pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location Oak Island  
Instrument Serial No. 008648 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: 008648  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: RHODES, K  
Permit Number:  
Effective:  
00/00/0000-00/00/0000

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG721401  
Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	11:20am
AIR BLK	.00	11:21am
ACCY CHK	.08	11:22am
AIR BLK	.00	11:23am
<b>SUB TEST</b>	<b>.00</b>	<b>11:25am</b>
AIR BLK	.00	11:26am
<b>SUB TEST</b>	<b>.00</b>	<b>11:27am</b>
AIR BLK	.00	11:28am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: 008648      Test Record Number: 1558  
Test Date: 03/15/2018      Test Time: 11:29am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:30am
FLO	Pass	11:30am
FC	Pass	11:30am

Temperature Tests

Test	Status	Time
FC1	Pass	11:30am
SRC	Pass	11:30am
DET	Pass	11:30am
BAR	Pass	11:30am
BT	Pass	11:30am

Blank Tests

Test	Status	Time
AIR	Pass	11:30am


Printer Tests

Test	Status	Time
PRNT	Pass	11:31am

CRC Tests

Test	Status	Time
COMP	Pass	11:31am
CAL	Pass	11:31am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location Brunswick County  
Instrument Serial No. 008585 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008585

Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	1:11pm
AIR BLK	.00	1:11pm
ACCY CHK	.07	1:12pm
AIR BLK	.00	1:13pm
SUB TEST	.00	1:14pm
AIR BLK	.00	1:15pm
SUB TEST	.00	1:16pm
AIR BLK	.00	1:17pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008585      Test Record Number: 4038  
Test Date: 03/15/2018      Test Time: 1:23pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:23pm
FLO	Pass	1:23pm
FC	Pass	1:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:23pm
SRC	Pass	1:23pm
DET	Pass	1:23pm
BAR	Pass	1:23pm
BT	Pass	1:23pm

Blank Tests

Test	Status	Time
AIR	Pass	1:24pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:24pm

CRC Tests

Test	Status	Time
COMP	Pass	1:24pm
CAL	Pass	1:24pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location Brunswick County  
Instrument Serial No. 008602 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



A. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008602  
Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	1:08pm
AIR BLK	.00	1:09pm
ACCY CHK	.08	1:10pm
AIR BLK	.00	1:11pm
SUB TEST	.00	1:12pm
AIR BLK	.00	1:13pm
SUB TEST	.00	1:14pm
AIR BLK	.00	1:15pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008602      Test Record Number: 3990  
Test Date: 03/15/2018      Test Time: 1:22pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:23pm
FLO	Pass	1:23pm
FC	Pass	1:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:23pm
SRC	Pass	1:23pm
DET	Pass	1:23pm
BAR	Pass	1:23pm
BT	Pass	1:23pm

Blank Tests

Test	Status	Time
AIR	Pass	1:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:24pm

CRC Tests

Test	Status	Time
COMP	Pass	1:24pm
CAL	Pass	1:24pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location Sunset Beach  
Instrument Serial No. 008874 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY SUNSET BEACH PD 090

Serial Number: 008874

Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020


Test	g/210L	Time
DIAG	Pass	9:26am
AIR BLK	.00	9:26am
ACCY CHK	.08	9:27am
AIR BLK	.00	9:28am
<b>SUB TEST</b>	<b>.00</b>	<b>9:28am</b>
AIR BLK	.00	9:29am
<b>SUB TEST</b>	<b>.00</b>	<b>9:31am</b>
AIR BLK	.00	9:32am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**BRUNSWICK COUNTY SUNSET BEACH PD 090**

Serial Number: 008874      Test Record Number: 626  
Test Date: 03/15/2018      Test Time: 9:34am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:35am
FLO	Pass	9:35am
FC	Pass	9:35am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:35am
SRC	Pass	9:35am
DET	Pass	9:35am
BAR	Pass	9:35am
BT	Pass	9:35am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:36am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:36am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:36am
CAL	Pass	9:36am

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Buncombe Instrument Location Bad Mobile Unit 11

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008973  
Test Date: 03/18/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V  
Permit Number: 26632E  
Effective:  
06/01/2017-06/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG621403  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:00pm
AIR BLK	.00	10:01pm
ACCY CHK	.08	10:02pm
AIR BLK	.00	10:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:03pm</b>
AIR BLK	.00	10:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:06pm</b>
AIR BLK	.00	10:07pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008973      Test Record Number: 467  
Test Date: 03/18/2018      Test Time: 10:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:08pm
FLO	Pass	10:08pm
FC	Pass	10:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:08pm
SRC	Pass	10:08pm
DET	Pass	10:08pm
BAR	Pass	10:08pm
BT	Pass	10:08pm

Blank Tests

Test	Status	Time
AIR	Pass	10:09pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:09pm

CRC Tests

Test	Status	Time
COMP	Pass	10:09pm
CAL	Pass	10:09pm

Preventive Maintenance  
Status: Pass



Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Burke Instrument Location Burke - Catawba Jail

Instrument Serial No. 008831 Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008831

Test Date: 03/19/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734102

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:25pm
AIR BLK	.00	1:26pm
ACCY CHK	.08	1:27pm
AIR BLK	.00	1:28pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:29pm</b>
AIR BLK	.00	1:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:31pm</b>
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**BURKE COUNTY BURKE-CATAWBA JAIL 110**

Serial Number: 008831      Test Record Number: 1958  
Test Date: 03/19/2018      Test Time: 1:33pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:34pm
FLO	Pass	1:34pm
FC	Pass	1:34pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:34pm
SRC	Pass	1:34pm
DET	Pass	1:34pm
BAR	Pass	1:34pm
BT	Pass	1:34pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:34pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:34pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:35pm
CAL	Pass	1:35pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Burke Instrument Location Burke - Catawba Jail

Instrument Serial No. 008904 Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008904

Test Date: 03/19/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	1:24pm
AIR BLK	.00	1:25pm
ACCY CHK	.08	1:26pm
AIR BLK	.00	1:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:28pm</b>
AIR BLK	.00	1:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:31pm</b>
AIR BLK	.00	1:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**BURKE COUNTY BURKE-CATAWBA JAIL 110**

Serial Number: 008904      Test Record Number: 2184  
Test Date: 03/19/2018      Test Time: 1:33pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:33pm
FLO	Pass	1:33pm
FC	Pass	1:33pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:34pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:34pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance  
Status: *Pass*



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cabarrus Instrument Location Cabarrus County SD  
Instrument Serial No. 008590 30 Gordon Ave., Concord

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805802  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	9:42am
AIR BLK	.00	9:43am
ACCY CHK	.08	9:44am
AIR BLK	.00	9:45am
<b>SUB TEST</b>	<b>.00</b>	<b>9:46am</b>
AIR BLK	.00	9:47am
<b>SUB TEST</b>	<b>.00</b>	<b>9:49am</b>
AIR BLK	.00	9:50am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007



**Intox EC/IR-II: Preventive Maintenance**

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590      Test Record Number: 2913  
Test Date: 03/15/2018      Test Time: 9:51am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:51am
FLO	Pass	9:51am
FC	Pass	9:52am

Temperature Tests

Test	Status	Time
FC1	Pass	9:52am
SRC	Pass	9:52am
DET	Pass	9:52am
BAR	Pass	9:52am
BT	Pass	9:52am

Blank Tests

Test	Status	Time
AIR	Pass	9:52am


Printer Tests

Test	Status	Time
PRNT	Pass	9:52am

CRC Tests

Test	Status	Time
COMP	Pass	9:53am
CAL	Pass	9:53am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret Instrument Location Carteret County  
Instrument Serial No. 008882 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008882  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	11:25am
AIR BLK	.00	11:25am
ACCY CHK	.08	11:26am
AIR BLK	.00	11:28am
<b>SUB TEST</b>	<b>.00</b>	<b>11:28am</b>
AIR BLK	.00	11:29am
<b>SUB TEST</b>	<b>.00</b>	<b>11:30am</b>
AIR BLK	.00	11:31am

Reported AC: .00 g/210L

*R E Hall*

Signature of Chemical Analyst

Court CVR

*Randy E Hall*  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008882      Test Record Number: 1691  
Test Date: 03/13/2018      Test Time: 11:32am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32am
FLO	Pass	11:32am
FC	Pass	11:32am

Temperature Tests

Test	Status	Time
FC1	Pass	11:32am
SRC	Pass	11:32am
DET	Pass	11:32am
BAR	Pass	11:32am
BT	Pass	11:32am

Blank Tests

Test	Status	Time
AIR	Pass	11:33am

Printer Tests

Test	Status	Time
PRNT	Pass	11:33am

CRC Tests

Test	Status	Time
COMP	Pass	11:33am
CAL	Pass	11:33am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CARTERET Instrument Location CARTERET COUNTY

Instrument Serial No. 008605 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Raney E Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008605  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	11:27am
AIR BLK	.00	11:28am
ACCY CHK	.07	11:28am
AIR BLK	.00	11:30am
<b>SUB TEST</b>	<b>.00</b>	<b>11:30am</b>
AIR BLK	.00	11:31am
<b>SUB TEST</b>	<b>.00</b>	<b>11:33am</b>
AIR BLK	.00	11:34am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008605      Test Record Number: 3704  
Test Date: 03/13/2018      Test Time: 11:34am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35am
FLO	Pass	11:35am
FC	Pass	11:35am

Temperature Tests

Test	Status	Time
FC1	Pass	11:35am
SRC	Pass	11:35am
DET	Pass	11:35am
BAR	Pass	11:35am
BT	Pass	11:35am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am


Printer Tests

Test	Status	Time
PRNT	Pass	11:36am

CRC Tests

Test	Status	Time
COMP	Pass	11:36am
CAL	Pass	11:36am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret


Instrument Location Morehead City PD

Instrument Serial No. 008731

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG722408

Exp Date: 08/12/2019


Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:13pm
ACCY CHK	.08	12:13pm
AIR BLK	.00	12:14pm
SUB TEST	.00	12:15pm
AIR BLK	.00	12:16pm
SUB TEST	.00	12:17pm
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731      Test Record Number: 2018  
Test Date: 03/13/2018      Test Time: 12:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

Blank Tests

Test	Status	Time
AIR	Pass	12:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:20pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret


Instrument Location ATLANTIC BEACH PD

Instrument Serial No. 008785

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: 008785  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	12:49pm
AIR BLK	.00	12:50pm
ACCY CHK	.07	12:51pm
AIR BLK	.00	12:52pm
SUB TEST	.00	12:52pm
AIR BLK	.00	12:53pm
SUB TEST	.00	12:54pm
AIR BLK	.00	12:55pm

Reported AC: .00 g/210L

*R E Hall*

Signature of Chemical Analyst

Court CVR

*Randy E Hall*  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: 008785      Test Record Number: 1002  
Test Date: 03/13/2018      Test Time: 12:56pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:56pm
FLO	Pass	12:56pm
FC	Pass	12:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:56pm
SRC	Pass	12:56pm
DET	Pass	12:56pm
BAR	Pass	12:56pm
BT	Pass	12:56pm

Blank Tests

Test	Status	Time
AIR	Pass	12:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:57pm

CRC Tests

Test	Status	Time
COMP	Pass	12:57pm
CAL	Pass	12:57pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret

Instrument Location EMERALD ISLE PD

Instrument Serial No. 008620

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY EMERALD ISLE PD 150

Serial Number: 008620  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019


Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:41pm
ACCY CHK	.08	1:42pm
AIR BLK	.00	1:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:44pm</b>
AIR BLK	.00	1:45pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:46pm</b>
AIR BLK	.00	1:47pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**CARTERET COUNTY EMERALD ISLE PD 150**

Serial Number: 008620      Test Record Number: 1929  
Test Date: 03/13/2018      Test Time: 1:49pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:49pm
FLO	Pass	1:49pm
FC	Pass	1:49pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:50pm
SRC	Pass	1:50pm
DET	Pass	1:50pm
BAR	Pass	1:50pm
BT	Pass	1:50pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:50pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:50pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:50pm
CAL	Pass	1:50pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CATAWBA Instrument Location BAT MOBILE 7

Instrument Serial No. 008968

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008968  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 07281E

Effective:

02/01/2018-02/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:58pm
AIR BLK	.00	10:59pm
ACCY CHK	.07	11:00pm
AIR BLK	.00	11:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:01pm</b>
AIR BLK	.00	11:02pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:04pm</b>
AIR BLK	.00	11:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

Intox EC/IR-II: Preventive Maintenance

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008968      Test Record Number: 233  
Test Date: 03/16/2018      Test Time: 11:06pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:06pm
FLO	Pass	11:06pm
FC	Pass	11:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:06pm
SRC	Pass	11:06pm
DET	Pass	11:06pm
BAR	Pass	11:06pm
BT	Pass	11:06pm

Blank Tests

Test	Status	Time
AIR	Pass	11:07pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:07pm

CRC Tests

Test	Status	Time
COMP	Pass	11:07pm
CAL	Pass	11:07pm

Preventive Maintenance  
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CATAWBA Instrument Location BAF MOBILE 7

Instrument Serial No. 008971

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008971

Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 07281E

Effective:

02/01/2018-02/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	9:54pm
AIR BLK	.00	9:55pm
ACCY CHK	.07	9:56pm
AIR BLK	.00	9:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:57pm</b>
AIR BLK	.00	9:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:00pm</b>
AIR BLK	.00	10:01pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008971      Test Record Number: 206  
Test Date: 03/16/2018      Test Time: 10:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02pm
FLO	Pass	10:02pm
FC	Pass	10:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:02pm
SRC	Pass	10:02pm
DET	Pass	10:02pm
BAR	Pass	10:02pm
BT	Pass	10:02pm

Blank Tests

Test	Status	Time
AIR	Pass	10:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:03pm

CRC Tests

Test	Status	Time
COMP	Pass	10:03pm
CAL	Pass	10:03pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chatham Co. Instrument Location Chatham Co. Jail

Instrument Serial No. 008591 Pittsboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHATHAM COUNTY JAIL 180

Serial Number: 008591

Test Date: 03/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	12:21pm
AIR BLK	.00	12:22pm
ACCY CHK	.07	12:23pm
AIR BLK	.00	12:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:24pm</b>
AIR BLK	.00	12:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:27pm</b>
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

CHATHAM COUNTY JAIL 180

Serial Number: 008591      Test Record Number: 1967  
Test Date: 03/09/2018      Test Time: 12:28pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:29pm
FLO	Pass	12:29pm
FC	Pass	12:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:29pm
SRC	Pass	12:29pm
DET	Pass	12:29pm
BAR	Pass	12:29pm
BT	Pass	12:29pm

Blank Tests

Test	Status	Time
AIR	Pass	12:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:29pm

CRC Tests

Test	Status	Time
COMP	Pass	12:30pm
CAL	Pass	12:30pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CHowan Instrument Location Chowan Co. Public  
Instrument Serial No. 008588 Safety CTR. Edenboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10<sup>th</sup> day of April, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008588  
Test Date: 04/10/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA  
Permit Number: 11646E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	11:24am
AIR BLK	.00	11:25am
ACCY CHK	.08	11:26am
AIR BLK	.00	11:27am
<b>SUB TEST</b>	<b>.00</b>	<b>11:27am</b>
AIR BLK	.00	11:28am
<b>SUB TEST</b>	<b>.00</b>	<b>11:30am</b>
AIR BLK	.00	11:31am

Reported AC: 00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**CHOWAN COUNTY PUBLIC SAFETY CENTER 200**

Serial Number: 008588      Test Record Number: 954  
Test Date: 04/10/2018      Test Time: 11:32am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:32am
FLO	Pass	11:32am
FC	Pass	11:33am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:33am
SRC	Pass	11:33am
DET	Pass	11:33am
BAR	Pass	11:33am
BT	Pass	11:33am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:33am


**Printer Tests**

Test	Status	Time
PRNT	Pass	11:33am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:34am
CAL	Pass	11:34am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cleveland Instrument Location Cleveland County SD Annex  
Instrument Serial No. 008867 407 McBrayer St., Shelby

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CLEVELAND COUNTY CLEVELAND SD-ANNEX  
220

Serial Number: 008887  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805801  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	3:39pm
AIR BLK	.00	3:40pm
ACCY CHK	.08	3:41pm
AIR BLK	.00	3:42pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:43pm</b>
AIR BLK	.00	3:44pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:45pm</b>
AIR BLK	.00	3:46pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: 008887      Test Record Number: 2595  
Test Date: 03/09/2018      Test Time: 3:47pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:48pm
FLO	Pass	3:48pm
FC	Pass	3:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:48pm
SRC	Pass	3:48pm
DET	Pass	3:48pm
BAR	Pass	3:48pm
BT	Pass	3:48pm

Blank Tests

Test	Status	Time
AIR	Pass	3:49pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:49pm

CRC Tests

Test	Status	Time
COMP	Pass	3:49pm
CAL	Pass	3:49pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Columbus Instrument Location Columbus County  
Instrument Serial No. 008886 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Phoke  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



Intox EC/IR-II: Subject Test

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008886  
Test Date: 03/14/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	4:41pm
AIR BLK	.00	4:42pm
ACCY CHK	.08	4:42pm
AIR BLK	.00	4:44pm
SUB TEST	.00	4:44pm
AIR BLK	.00	4:45pm
SUB TEST	.00	4:46pm
AIR BLK	.00	4:47pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008886 Test Record Number: 1354  
Test Date: 03/14/2018 Test Time: 4:54pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	4:54pm
FLO	Pass	4:54pm
FC	Pass	4:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:54pm
SRC	Pass	4:54pm
DET	Pass	4:54pm
BAR	Pass	4:54pm
BT	Pass	4:54pm

Blank Tests

Test	Status	Time
AIR	Pass	4:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:55pm

CRC Tests

Test	Status	Time
COMP	Pass	4:55pm
CAL	Pass	4:55pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Columbus Instrument Location Columbus

Instrument Serial No. 008875 County Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



H.C. Proctor  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008875

Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	4:38pm
AIR BLK	.00	4:38pm
ACCY CHK	.08	4:39pm
AIR BLK	.00	4:40pm
SUB TEST	.00	4:41pm
AIR BLK	.00	4:42pm
SUB TEST	.00	4:44pm
AIR BLK	.00	4:45pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008875 Test Record Number: 2019  
Test Date: 03/14/2018 Test Time: 4:47pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	4:47pm
FLO	Pass	4:47pm
FC	Pass	4:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:47pm
SRC	Pass	4:47pm
DET	Pass	4:47pm
BAR	Pass	4:47pm
BT	Pass	4:47pm

Blank Tests

Test	Status	Time
AIR	Pass	4:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:48pm

CRC Tests

Test	Status	Time
COMP	Pass	4:48pm
CAL	Pass	4:48pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County CRAVEN Instrument Location HAVELOCK AS

Instrument Serial No. 008800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HALL, RANDY E  
Permit Number: 3462E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG722408  
Exp Date: 08/12/2019

Test	g/210L	Time
DIAG	Pass	9:12am
AIR BLK	.00	9:13am
ACCY CHK	.08	9:13am
AIR BLK	.00	9:14am
SUB TEST	.00	9:15am
AIR BLK	.00	9:16am
SUB TEST	.00	9:17am
AIR BLK	.00	9:18am

Reported AC: .00 g/210L

*RE Hall*

Signature of Chemical Analyst

Court CVR

*Randy E Hall*  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800      Test Record Number: 1135  
Test Date: 03/15/2018      Test Time: 9:18am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:19am
FLO	Pass	9:19am
FC	Pass	9:19am

Temperature Tests

Test	Status	Time
FC1	Pass	9:19am
SRC	Pass	9:19am
DET	Pass	9:19am
BAR	Pass	9:19am
BT	Pass	9:19am

Blank Tests

Test	Status	Time
AIR	Pass	9:20am


Printer Tests

Test	Status	Time
PRNT	Pass	9:20am

CRC Tests

Test	Status	Time
COMP	Pass	9:20am
CAL	Pass	9:20am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN

Instrument Location MCA5 CHERRY POINT


Instrument Serial No. 010819

AMC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HALL, RANDY E  
Permit Number: 3462E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG702302  
Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	9:44am
AIR BLK	.00	9:44am
ACCY CHK	.08	9:45am
AIR BLK	.00	9:46am
<b>SUB TEST</b>	<b>.00</b>	<b>9:46am</b>
AIR BLK	.00	9:47am
<b>SUB TEST</b>	<b>.00</b>	<b>9:49am</b>
AIR BLK	.00	9:50am

Reported AC: .00 g/210L

*R E Hall*

Signature of Chemical Analyst

Court CVR

*Randy E Hall*  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819      Test Record Number: 528  
Test Date: 03/15/2018      Test Time: 9:50am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:50am
FLO	Pass	9:50am
FC	Pass	9:51am

Temperature Tests

Test	Status	Time
FC1	Pass	9:51am
SRC	Pass	9:51am
DET	Pass	9:51am
BAR	Pass	9:51am
BT	Pass	9:51am

Blank Tests

Test	Status	Time
AIR	Pass	9:51am

Printer Tests

Test	Status	Time
PRNT	Pass	9:51am

CRC Tests

Test	Status	Time
COMP	Pass	9:52am
CAL	Pass	9:52am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location NEW BERN PD

Instrument Serial No. 008817

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817  
Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019


Test	g/210L	Time
DIAG	Pass	10:43am
AIR BLK	.00	10:44am
ACCY CHK	.07	10:45am
AIR BLK	.00	10:46am
<b>SUB TEST</b>	<b>.00</b>	<b>10:46am</b>
AIR BLK	.00	10:47am
<b>SUB TEST</b>	<b>.00</b>	<b>10:49am</b>
AIR BLK	.00	10:49am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817      Test Record Number: 1347  
Test Date: 03/15/2018      Test Time: 10:52am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:52am
FLO	Pass	10:52am
FC	Pass	10:52am

Temperature Tests

Test	Status	Time
FC1	Pass	10:52am
SRC	Pass	10:52am
DET	Pass	10:52am
BAR	Pass	10:52am
BT	Pass	10:52am

Blank Tests

Test	Status	Time
AIR	Pass	10:53am


Printer Tests

Test	Status	Time
PRNT	Pass	10:53am

CRC Tests

Test	Status	Time
COMP	Pass	10:53am
CAL	Pass	10:53am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location CRAVEN County

Instrument Serial No. 008732 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: 008732  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020


Test	g/210L	Time
DIAG	Pass	11:29am
AIR BLK	.00	11:30am
ACCY CHK	.08	11:31am
AIR BLK	.00	11:32am
<b>SUB TEST</b>	<b>.00</b>	<b>11:33am</b>
AIR BLK	.00	11:34am
<b>SUB TEST</b>	<b>.00</b>	<b>11:35am</b>
AIR BLK	.00	11:36am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: 008732      Test Record Number: 2031  
Test Date: 03/15/2018      Test Time: 11:37am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:37am
FLO	Pass	11:37am
FC	Pass	11:37am

Temperature Tests

Test	Status	Time
FC1	Pass	11:37am
SRC	Pass	11:37am
DET	Pass	11:37am
BAR	Pass	11:37am
BT	Pass	11:37am

Blank Tests

Test	Status	Time
AIR	Pass	11:38am


Printer Tests

Test	Status	Time
PRNT	Pass	11:38am

CRC Tests

Test	Status	Time
COMP	Pass	11:38am
CAL	Pass	11:38am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cumberland Co. Instrument Location Cumberland Co. Det. CTR.  
Instrument Serial No. 008632 Fayetteville NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008632  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805802  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:01pm
ACCY CHK	.08	12:01pm
AIR BLK	.00	12:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:04pm</b>
AIR BLK	.00	12:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:06pm</b>
AIR BLK	.00	12:07pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

*CUMBERLAND COUNTY DETENTION CENTER 250*

Serial Number: 008632      Test Record Number: 4035  
Test Date: 03/07/2018      Test Time: 12:08pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:08pm
FLO	Pass	12:08pm
FC	Pass	12:08pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:09pm
SRC	Pass	12:09pm
DET	Pass	12:09pm
BAR	Pass	12:09pm
BT	Pass	12:09pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:09pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:09pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cumberland Co. Instrument Location Cumberland Co. Det. Center

Instrument Serial No. 005672 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG805802  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	12:01pm
AIR BLK	.00	12:02pm
ACCY CHK	.08	12:03pm
AIR BLK	.00	12:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:05pm</b>
AIR BLK	.00	12:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:07pm</b>
AIR BLK	.00	12:08pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672      Test Record Number: 5914  
Test Date: 03/07/2018      Test Time: 12:12pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:13pm
FLO	Pass	12:13pm
FC	Pass	12:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:13pm
SRC	Pass	12:13pm
DET	Pass	12:13pm
BAR	Pass	12:13pm
BT	Pass	12:13pm

Blank Tests

Test	Status	Time
AIR	Pass	12:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:13pm

CRC Tests

Test	Status	Time
COMP	Pass	12:14pm
CAL	Pass	12:14pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Cumberland Co. Instrument Location Cumberland Co. Det. CTR

Instrument Serial No. 008633 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805802  
Exp Date: 02/27/2020


Test	g/210L	Time
DIAG	Pass	12:02pm
AIR BLK	.00	12:03pm
ACCY CHK	.07	12:04pm
AIR BLK	.00	12:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:06pm</b>
AIR BLK	.00	12:07pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:09pm</b>
AIR BLK	.00	12:10pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633      Test Record Number: 4686  
Test Date: 03/07/2018      Test Time: 12:11pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm

Printer Tests


Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Cumberland Co. Instrument Location Cumberland Co. Det. Center

Instrument Serial No. 0096614 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614  
Test Date: 03/07/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	12:03pm
AIR BLK	.00	12:04pm
ACCY CHK	.08	12:05pm
AIR BLK	.00	12:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:07pm</b>
AIR BLK	.00	12:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:09pm</b>
AIR BLK	.00	12:10pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614      Test Record Number: 3989  
Test Date: 03/07/2018      Test Time: 12:11pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:12pm
FLO	Pass	12:12pm
FC	Pass	12:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:12pm
SRC	Pass	12:12pm
DET	Pass	12:12pm
BAR	Pass	12:12pm
BT	Pass	12:12pm

Blank Tests

Test	Status	Time
AIR	Pass	12:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:13pm

CRC Tests

Test	Status	Time
COMP	Pass	12:13pm
CAL	Pass	12:13pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Currituck Instrument Location Currituck Co. S.O. - Corolla  
Instrument Serial No. DD 8949 1123 Ocean Trail, Corolla, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 6<sup>th</sup> day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

643  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: 008949

Test Date: 03/06/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:36am
AIR BLK	.00	11:36am
ACCY CHK	.08	11:37am
AIR BLK	.00	11:38am
<b>SUB TEST</b>	<b>.00</b>	<b>11:39am</b>
AIR BLK	.00	11:40am
<b>SUB TEST</b>	<b>.00</b>	<b>11:41am</b>
AIR BLK	.00	11:42am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**CURRITUCK COUNTY SO-COROLLA 260**

Serial Number: 008949      Test Record Number: 450  
Test Date: 03/06/2018      Test Time: 11:44am EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:45am
FLO	Pass	11:45am
FC	Pass	11:45am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:45am
SRC	Pass	11:45am
DET	Pass	11:45am
BAR	Pass	11:45am
BT	Pass	11:45am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:45am


**Printer Tests**

Test	Status	Time
PRNT	Pass	11:45am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:46am
CAL	Pass	11:46am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Wallace Police  
Instrument Serial No. 008858 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858

Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	12:41pm
AIR BLK	.00	12:42pm
ACCY CHK	.08	12:42pm
AIR BLK	.00	12:44pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:44pm</b>
AIR BLK	.00	12:45pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:47pm</b>
AIR BLK	.00	12:48pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858      Test Record Number: 838  
Test Date: 03/14/2018      Test Time: 12:50pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:50pm
FLO	Pass	12:50pm
FC	Pass	12:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:50pm
SRC	Pass	12:50pm
DET	Pass	12:50pm
BAR	Pass	12:50pm
BT	Pass	12:50pm

Blank Tests

Test	Status	Time
AIR	Pass	12:51pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:51pm

CRC Tests

Test	Status	Time
COMP	Pass	12:51pm
CAL	Pass	12:51pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Wallace Police  
Instrument Serial No. 008858 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of January, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes 601  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858

Test Date: 01/31/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	11:59am
AIR BLK	.00	12:00pm
ACCY CHK	.08	12:00pm
AIR BLK	.00	12:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:02pm</b>
AIR BLK	.00	12:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:04pm</b>
AIR BLK	.00	12:05pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858      Test Record Number: 832  
Test Date: 01/31/2018      Test Time: 12:06pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:06pm
FLO	Pass	12:06pm
FC	Pass	12:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:07pm
SRC	Pass	12:07pm
DET	Pass	12:07pm
BAR	Pass	12:07pm
BT	Pass	12:07pm

Blank Tests

Test	Status	Time
AIR	Pass	12:07pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:07pm

CRC Tests

Test	Status	Time
COMP	Pass	12:07pm
CAL	Pass	12:07pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Duplin County  
Instrument Serial No. 008864 Specifc Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



F. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: 008864  
Test Date: 03/23/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716202

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:56am
AIR BLK	.00	9:57am
ACCY CHK	.08	9:58am
AIR BLK	.00	9:59am
<b>SUB TEST</b>	<b>.00</b>	<b>10:00am</b>
AIR BLK	.00	10:01am
<b>SUB TEST</b>	<b>.00</b>	<b>10:02am</b>
AIR BLK	.00	10:04am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007



**Intox EC/IR-II: Preventive Maintenance**

DUPLIN COUNTY DUPLIN CO. SD 300

Serial Number: 008864      Test Record Number: 3330  
Test Date: 03/23/2018      Test Time: 10:05am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:05am
FLO	Pass	10:05am
FC	Pass	10:05am

Temperature Tests

Test	Status	Time
FC1	Pass	10:05am
SRC	Pass	10:05am
DET	Pass	10:05am
BAR	Pass	10:05am
BT	Pass	10:05am

Blank Tests

Test	Status	Time
AIR	Pass	10:06am


Printer Tests

Test	Status	Time
PRNT	Pass	10:06am

CRC Tests

Test	Status	Time
COMP	Pass	10:06am
CAL	Pass	10:06am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County EDGEMOND Instrument Location NAT MODEL UNIT 10

Instrument Serial No. 008580

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

EDGECOMBE COUNTY BAT MOBILE UNIT 10 320

Serial Number: 008580      Test Record Number: 2417  
Test Date: 03/10/2018      Test Time: 8:22pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:23pm
FLO	Pass	8:23pm
FC	Pass	8:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:23pm
SRC	Pass	8:23pm
DET	Pass	8:23pm
BAR	Pass	8:23pm
BT	Pass	8:23pm

Blank Tests

Test	Status	Time
AIR	Pass	8:24pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:24pm

CRC Tests

Test	Status	Time
COMP	Pass	8:24pm
CAL	Pass	8:24pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Subject Test**

EDGECOMBE COUNTY BAT MOBILE UNIT 10  
320

Serial Number: 008580  
Test Date: 03/10/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: VARNELL, BRYON L  
Permit Number: 16896E  
Effective:  
09/22/2017-09/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:10pm
AIR BLK	.00	8:11pm
ACCY CHK	.07	8:12pm
AIR BLK	.00	8:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:13pm</b>
AIR BLK	.00	8:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:15pm</b>
AIR BLK	.00	8:16pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Forsyth Instrument Location Bad mobile Unit 11

Instrument Serial No. 008470

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris Jones

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH BAT MOBILE UNIT 11 330

Serial Number: 008970

Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

06/01/2017-06/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018


Test	g/210L	Time
DIAG	Pass	8:29pm
AIR BLK	.00	8:30pm
ACCY CHK	.08	8:31pm
AIR BLK	.00	8:31pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:32pm</b>
AIR BLK	.00	8:33pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:34pm</b>
AIR BLK	.00	8:35pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

FORSYTH BAT MOBILE UNIT 11 330

Serial Number: 008970      Test Record Number: 440  
Test Date: 03/15/2018      Test Time: 8:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:37pm
FLO	Pass	8:37pm
FC	Pass	8:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:37pm
SRC	Pass	8:37pm
DET	Pass	8:37pm
BAR	Pass	8:37pm
BT	Pass	8:37pm

Blank Tests

Test	Status	Time
AIR	Pass	8:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	8:37pm

CRC Tests

Test	Status	Time
COMP	Pass	8:38pm
CAL	Pass	8:38pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Forsyth

Instrument Location Forsyth County Detention


Instrument Serial No. 008683

Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kenneth R. Byrd  
Signature of Certifying Official

657  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008683  
Test Date: 03/24/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R  
Permit Number: 22067E  
Effective:  
07/01/2016-07/01/2018

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805801  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	9:55am
AIR BLK	.00	9:55am
ACCY CHK	.08	9:56am
AIR BLK	.00	9:57am
<b>SUB TEST</b>	<b>.00</b>	<b>9:58am</b>
AIR BLK	.00	9:59am
<b>SUB TEST</b>	<b>.00</b>	<b>10:00am</b>
AIR BLK	.00	10:01am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008683      Test Record Number: 613  
Test Date: 03/24/2018      Test Time: 10:03am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:03am
FLO	Pass	10:03am
FC	Pass	10:03am

Temperature Tests

Test	Status	Time
FC1	Pass	10:04am
SRC	Pass	10:04am
DET	Pass	10:04am
BAR	Pass	10:04am
BT	Pass	10:04am

Blank Tests

Test	Status	Time
AIR	Pass	10:04am


Printer Tests

Test	Status	Time
PRNT	Pass	10:04am

CRC Tests

Test	Status	Time
COMP	Pass	10:04am
CAL	Pass	10:04am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention  
Instrument Serial No. 008660 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

657  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

FORSYTH COUNTY FORSYTH CO DETENTION  
330

Serial Number: 008660  
Test Date: 03/24/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BENFIELD II, KENNETH R  
Permit Number: 22067E  
Effective:  
07/01/2016-07/01/2018

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805801  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	10:00am
AIR BLK	.00	10:01am
ACCY CHK	.08	10:01am
AIR BLK	.00	10:02am
<b>SUB TEST</b>	<b>.00</b>	<b>10:03am</b>
AIR BLK	.00	10:04am
<b>SUB TEST</b>	<b>.00</b>	<b>10:07am</b>
AIR BLK	.00	10:07am

Reported AC: 00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008660      Test Record Number: 3945  
Test Date: 03/24/2018      Test Time: 10:08am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:09am
FLO	Pass	10:09am
FC	Pass	10:09am

Temperature Tests

Test	Status	Time
FC1	Pass	10:09am
SRC	Pass	10:09am
DET	Pass	10:09am
BAR	Pass	10:09am
BT	Pass	10:09am

Blank Tests

Test	Status	Time
AIR	Pass	10:09am

Printer Tests

Test	Status	Time
PRNT	Pass	10:09am

CRC Tests

Test	Status	Time
COMP	Pass	10:10am
CAL	Pass	10:10am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County FRANKLIN Instrument Location FRANKLIN CO LEC  
Instrument Serial No. 008933 285 T. Kemp RD  
Louisburg, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

062  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933  
Test Date: 03/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:42pm
ACCY CHK	.08	1:42pm
AIR BLK	.00	1:44pm
SUB TEST	.00	1:44pm
AIR BLK	.00	1:45pm
SUB TEST	.00	1:46pm
AIR BLK	.00	1:47pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933 Test Record Number: 960  
Test Date: 03/08/2018 Test Time: 1:48pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:49pm
FLO	Pass	1:49pm
FC	Pass	1:49pm

Temperature Tests

Test	Status	Time
FCL	Pass	1:49pm
SRC	Pass	1:49pm
DET	Pass	1:49pm
BAR	Pass	1:49pm
BT	Pass	1:49pm

Blank Tests

Test	Status	Time
AIR	Pass	1:50pm

Printer Tests


Test	Status	Time
PRNT	Pass	1:50pm

CRC Tests

Test	Status	Time
COMP	Pass	1:50pm
CAL	Pass	1:50pm

Preventive Maintenance

Status: Pass



Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County FRANKLIN Instrument Location Franklin Co LEC  
Instrument Serial No. 0089412 285 T KEMP RD  
LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942

Test Date: 03/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test


Lot Number: AG702302

Exp Date: 01/23/2019

Test g/210L Time

DIAG	Pass	11:40am
AIR BLK	.00	11:41am
ACCY CHK	.07	11:42am
AIR BLK	.00	11:43am
SUB TEST	.00	11:43am
AIR BLK	.00	11:44am
SUB TEST	.00	11:45am
AIR BLK	.00	11:45am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942      Test Record Number: 1396  
Test Date: 03/08/2018      Test Time: 1:32pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:33pm
FLO	Pass	1:33pm
FC	Pass	1:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:34pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:34pm

CRC Tests

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance  
Status: Pass



---

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Graham Instrument Location Graham Co. S.O.

Instrument Serial No. 008915 Robbinsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Orvil R. Cuth  
Signature of Certifying Official

635  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GRAHAM COUNTY GRAHAM COUNTY SD 370

Serial Number: 008915  
Test Date: 03/22/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R  
Permit Number: 08457  
Effective:  
09/01/2017-09/01/2019

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG734102  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	12:24pm
AIR BLK	.00	12:25pm
ACCY CHK	.08	12:25pm
AIR BLK	.00	12:27pm
SUB TEST	.00	12:27pm
AIR BLK	.00	12:28pm
SUB TEST	.00	12:29pm
AIR BLK	.00	12:30pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

GRAHAM COUNTY GRAHAM COUNTY SD 370

Serial Number: 008915 Test Record Number: 734  
Test Date: 03/22/2018 Test Time: 12:32pm

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:32pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:32pm
SRC	Pass	12:32pm
DET	Pass	12:32pm
BAR	Pass	12:32pm
BT	Pass	12:32pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:33pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:33pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008638

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



A. K. Quinn  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: DEAN, L K  
Permit Number: 11598E  
Effective:  
04/01/2017-04/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG721401  
Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	2:32pm
AIR BLK	.00	2:33pm
ACCY CHK	.08	2:34pm
AIR BLK	.00	2:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:36pm</b>
AIR BLK	.00	2:37pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:38pm</b>
AIR BLK	.00	2:39pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638      Test Record Number: 3393  
Test Date: 03/13/2018      Test Time: 2:40pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:40pm
FLO	Pass	2:40pm
FC	Pass	2:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:40pm
SRC	Pass	2:40pm
DET	Pass	2:40pm
BAR	Pass	2:40pm
BT	Pass	2:40pm

Blank Tests

Test	Status	Time
AIR	Pass	2:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:41pm

CRC Tests

Test	Status	Time
COMP	Pass	2:41pm
CAL	Pass	2:41pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008718

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. K. [Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008718  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: DEAN, L K  
Permit Number: 11598E  
Effective:  
04/01/2017-04/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG734102  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	2:05pm
AIR BLK	.00	2:05pm
ACCY CHK	.08	2:06pm
AIR BLK	.00	2:07pm
SUB TEST	.00	2:08pm
AIR BLK	.00	2:09pm
SUB TEST	.00	2:10pm
AIR BLK	.00	2:11pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008718      Test Record Number: 1725  
Test Date: 03/13/2018      Test Time: 2:12pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:12pm
FLO	Pass	2:12pm
FC	Pass	2:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:12pm
SRC	Pass	2:12pm
DET	Pass	2:12pm
BAR	Pass	2:12pm
BT	Pass	2:12pm

Blank Tests

Test	Status	Time
AIR	Pass	2:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:13pm

CRC Tests

Test	Status	Time
COMP	Pass	2:13pm
CAL	Pass	2:13pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County GUILFORD Instrument Location GREENSBORO JAIL

Instrument Serial No. 008794

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Keen Dease  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794  
Test Date: 03/13/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: DEAN, L K  
Permit Number: 11598E  
Effective:  
04/01/2017-04/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG734101  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	12:57pm
AIR BLK	.00	12:57pm
ACCY CHK	.07	12:58pm
AIR BLK	.00	12:59pm
SUB TEST	.00	1:00pm
AIR BLK	.00	1:01pm
SUB TEST	.00	1:02pm
AIR BLK	.00	1:03pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

**GUILFORD COUNTY GREENSBORO JAIL 400**

Serial Number: 008794      Test Record Number: 5936  
Test Date: 03/13/2018      Test Time: 1:04pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:04pm
FLO	Pass	1:04pm
FC	Pass	1:04pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:04pm
SRC	Pass	1:05pm
DET	Pass	1:05pm
BAR	Pass	1:05pm
BT	Pass	1:05pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:05pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:05pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:05pm
CAL	Pass	1:05pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford

Instrument Location Bat Mobile Unit 8

Instrument Serial No. 008615

Guilford Co 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B. Skinner  
Signature of Certifying Official

644  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008615  
Test Date: 03/30/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG716201  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	10:53pm
AIR BLK	.00	10:54pm
ACCY CHK	.07	10:54pm
AIR BLK	.00	10:55pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:57pm</b>
AIR BLK	.00	10:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:00pm</b>
AIR BLK	.00	11:00pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**GREENSBORO BAT MOBILE UNIT 8 400**

Serial Number: 008615      Test Record Number: 5534  
Test Date: 03/30/2018      Test Time: 11:02pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:02pm
FLO	Pass	11:02pm
FC	Pass	11:02pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:02pm
SRC	Pass	11:02pm
DET	Pass	11:02pm
BAR	Pass	11:02pm
BT	Pass	11:02pm

**Blank Tests**

Test	Status	Time
AIR	Pass	11:03pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	11:03pm

**CRC Tests**

Test	Status	Time
COMP	Pass	11:03pm
CAL	Pass	11:03pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County HALIFAX Instrument Location Roanoke Rapids PD  
Instrument Serial No. 008656 1040 Roanoke Ave  
Roanoke Rapids, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	2:41pm
AIR BLK	.00	2:42pm
ACCY CHK	.07	2:42pm
AIR BLK	.00	2:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:44pm</b>
AIR BLK	.00	2:45pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:46pm</b>
AIR BLK	.00	2:47pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656      Test Record Number: 602  
Test Date: 03/09/2018      Test Time: 2:48pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:48pm
FLO	Pass	2:48pm
FC	Pass	2:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:48pm
SRC	Pass	2:48pm
DET	Pass	2:48pm
BAR	Pass	2:48pm
BT	Pass	2:48pm

Blank Tests

Test	Status	Time
AIR	Pass	2:49pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:49pm

CRC Tests

Test	Status	Time
COMP	Pass	2:49pm
CAL	Pass	2:49pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County HALIFAX Instrument Location Roanoke Rapids PD

Instrument Serial No. 008635 1040 Roanoke Ave  
Roanoke Rapids, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HALIFAX CO ROANOKE RAPIDS PD 410

Serial Number: 008635  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTATIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	2:40pm
AIR BLK	.00	2:40pm
ACCY CHK	.08	2:41pm
AIR BLK	.00	2:42pm
SUB TEST	.00	2:42pm
AIR BLK	.00	2:43pm
SUB TEST	.00	2:45pm
AIR BLK	.00	2:46pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO ROANOKE RAPIDS PD 410

Serial Number: 008635 Test Record Number: 1635  
Test Date: 03/09/2018 Test Time: 2:46pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:47pm
FLO	Pass	2:47pm
FC	Pass	2:47pm

Temperature Tests

Test	Status	Time
ECI	Pass	2:47pm
SRC	Pass	2:47pm
DET	Pass	2:47pm
BAR	Pass	2:47pm
BT	Pass	2:47pm

Blank Tests

Test	Status	Time
AIR	Pass	2:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:48pm

CRC Tests

Test	Status	Time
COMP	Pass	2:48pm
CAL	Pass	2:48pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County HALIFAX Instrument Location HALIFAX CO. SHERIFF'S OFFICE

Instrument Serial No. 008695 355 Fernell Ln  
HALIFAX, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HALIFAX CO. HALIFAX CO. SD 410

Serial Number: 008695  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTATIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, SIMON S  
Permit Number: 11434E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG734101  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:29pm
AIR BLK	.00	1:30pm
ACCY CHK	.08	1:30pm
AIR BLK	.00	1:31pm
SUB TEST	.00	1:32pm
AIR BLK	.00	1:33pm
SUB TEST	.00	1:34pm
AIR BLK	.00	1:35pm

Reported/AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HALIFAX CO. HALIFAX CO. SD 410

Serial Number: 008695      Test Record Number: 2451  
Test Date: 03/09/2018      Test Time: 1:38pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:39pm
FLO	Pass	1:39pm
FC	Pass	1:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:40pm

CRC Tests

Test	Status	Time
COMP	Pass	1:40pm
CAL	Pass	1:40pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Harnett Instrument Location BAT mobile unit 212  
Instrument Serial No 008788 Harnett Co SHP

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

6600  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HARNETT COUNTY BAT MOBILE UNIT 12 420

Serial Number: 008788  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: SMITH, JASON R  
Permit Number: 19145E  
Effective:  
03/01/2018-03/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG716201  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	10:28pm
AIR BLK	.00	10:29pm
ACCY CHK	.07	10:30pm
AIR BLK	.00	10:30pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:31pm</b>
AIR BLK	.00	10:32pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:33pm</b>
AIR BLK	.00	10:34pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**HARNETT COUNTY BAT MOBILE UNIT 12 420**

Serial Number: 008788      Test Record Number: 1333  
Test Date: 03/16/2018      Test Time: 10:36pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:36pm
FLO	Pass	10:36pm
FC	Pass	10:36pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:36pm
SRC	Pass	10:36pm
DET	Pass	10:36pm
BAR	Pass	10:36pm
BT	Pass	10:36pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:37pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:37pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:37pm
CAL	Pass	10:37pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County Haywood Instrument Location Haywood Co. Jail

Instrument Serial No. 008714 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Orin R. Cutler  
Signature of Certifying Official

635  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714  
Test Date: 03/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621501


Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	11:00am
AIR BLK	.00	11:01am
ACCY CHK	.07	11:02am
AIR BLK	.00	11:03am
<b>SUB TEST</b>	<b>.00</b>	<b>11:03am</b>
AIR BLK	.00	11:04am
<b>SUB TEST</b>	<b>.00</b>	<b>11:06am</b>
AIR BLK	.00	11:06am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714      Test Record Number: 1468  
Test Date: 03/20/2018      Test Time: 11:07am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:08am
FLO	Pass	11:08am
FC	Pass	11:08am

Temperature Tests

Test	Status	Time
FC1	Pass	11:08am
SRC	Pass	11:08am
DET	Pass	11:08am
BAR	Pass	11:08am
BT	Pass	11:08am

Blank Tests

Test	Status	Time
AIR	Pass	11:09am

Printer Tests

Test	Status	Time
PRNT	Pass	11:09am

CRC Tests

Test	Status	Time
COMP	Pass	11:09am
CAL	Pass	11:09am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Haywood Instrument Location Haywood Co. Jail

Instrument Serial No. 008712 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris R. Cutler  
Signature of Certifying Official

635  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008712  
Test Date: 03/20/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 08457

Effective:

09/01/2017-09/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702401

Exp Date: 02/24/2019

Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	11:00am
ACCY CHK	.07	11:01am
AIR BLK	.00	11:02am
<b>SUB TEST</b>	<b>.00</b>	<b>11:02am</b>
AIR BLK	.00	11:03am
<b>SUB TEST</b>	<b>.00</b>	<b>11:05am</b>
AIR BLK	.00	11:06am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008712 Test Record Number: 2017  
Test Date: 03/20/2018 Test Time: 11:07am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:07am
FLO	Pass	11:07am
FC	Pass	11:08am

Temperature Tests

Test	Status	Time
FC1	Pass	11:08am
SRC	Pass	11:08am
DET	Pass	11:08am
BAR	Pass	11:08am
BT	Pass	11:08am

Blank Tests

Test	Status	Time
AIR	Pass	11:08am


Printer Tests

Test	Status	Time
PRNT	Pass	11:08am

CRC Tests

Test	Status	Time
COMP	Pass	11:08am
CAL	Pass	11:08am

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Henderson Instrument Location BA Mobile Unit 11

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris [Signature]  
Signature of Certifying Official

658  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**HENDERSON BAT MOBILE UNIT 11 440**

Serial Number: 008973  
Test Date: 03/31/2018

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V  
Permit Number: 26632E  
Effective:  
06/01/2017-06/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG621403  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	7:57pm
AIR BLK	.00	7:58pm
ACCY CHK	.08	7:59pm
AIR BLK	.00	8:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:00pm</b>
AIR BLK	.00	8:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:03pm</b>
AIR BLK	.00	8:04pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

HENDERSON BAT MOBILE UNIT 11 440

Serial Number: 008973      Test Record Number: 476  
Test Date: 03/31/2018      Test Time: 8:04pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	8:05pm
FLO	Pass	8:05pm
FC	Pass	8:05pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	8:05pm
SRC	Pass	8:05pm
DET	Pass	8:05pm
BAR	Pass	8:05pm
BT	Pass	8:05pm

**Blank Tests**

Test	Status	Time
AIR	Pass	8:06pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	8:06pm

**CRC Tests**

Test	Status	Time
COMP	Pass	8:06pm
CAL	Pass	8:06pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County HELFORD Instrument Location AHOSKIE P.D.  
Instrument Serial No. 008848 705 W. MAIN ST., AHOSKIE, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21<sup>ST</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keel  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: 008848

Test Date: 03/21/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	12:33pm
AIR BLK	.00	12:34pm
ACCY CHK	.08	12:34pm
AIR BLK	.00	12:35pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:36pm</b>
AIR BLK	.00	12:37pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:39pm</b>
AIR BLK	.00	12:40pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: 008848 Test Record Number: 1317  
Test Date: 03/21/2018 Test Time: 12:41pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:41pm
FLO	Pass	12:41pm
FC	Pass	12:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:41pm
SRC	Pass	12:41pm
DET	Pass	12:41pm
BAR	Pass	12:41pm
BT	Pass	12:41pm

Blank Tests

Test	Status	Time
AIR	Pass	12:42pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:42pm

CRC Tests

Test	Status	Time
COMP	Pass	12:42pm
CAL	Pass	12:42pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Statesville PD  
Instrument Serial No. 008619 300 S. Tradd St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY STATESVILLE PD 480

Serial Number: 008619  
Test Date: 03/01/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male

Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG621403  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	9:28am
AIR BLK	.00	9:29am
ACCY CHK	.07	9:30am
AIR BLK	.00	9:31am
SUB TEST	.00	9:31am
AIR BLK	.00	9:32am
SUB TEST	.00	9:34am
AIR BLK	.00	9:35am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**IREDELL COUNTY STATESVILLE PD 480**

Serial Number: 008619      Test Record Number: 1394  
Test Date: 03/01/2018      Test Time: 9:36am EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:36am
FLO	Pass	9:36am
FC	Pass	9:36am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:36am
SRC	Pass	9:36am
DET	Pass	9:36am
BAR	Pass	9:36am
BT	Pass	9:36am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:37am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:37am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:37am
CAL	Pass	9:37am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Iredell County SD  
Instrument Serial No. 008809 201 E. Water St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: 008809  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	12:58pm
AIR BLK	.00	12:59pm
ACCY CHK	.08	1:00pm
AIR BLK	.00	1:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:01pm</b>
AIR BLK	.00	1:02pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:03pm</b>
AIR BLK	.00	1:04pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*IREDELL COUNTY IREDELL COUNTY SD 480*

Serial Number: 008809      Test Record Number: 3840  
Test Date: 03/09/2018      Test Time: 1:05pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:06pm
FLO	Pass	1:06pm
FC	Pass	1:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:06pm
SRC	Pass	1:06pm
DET	Pass	1:06pm
BAR	Pass	1:06pm
BT	Pass	1:06pm

Blank Tests

Test	Status	Time
AIR	Pass	1:07pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:07pm

CRC Tests

Test	Status	Time
COMP	Pass	1:07pm
CAL	Pass	1:07pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County JONES Instrument Location Jones County

Instrument Serial No. 008705 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JONES COUNTY JONES COUNTY SD 510

Serial Number: 008705  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: HALL, RANDY E  
Permit Number: 3462E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG734101  
Exp Date: 12/07/2019


Test	g/210L	Time
DIAG	Pass	12:22pm
AIR BLK	.00	12:22pm
ACCY CHK	.07	12:23pm
AIR BLK	.00	12:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:25pm</b>
AIR BLK	.00	12:26pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:27pm</b>
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

JONES COUNTY JONES COUNTY SD 510

Serial Number: 008705      Test Record Number: 1204  
Test Date: 03/15/2018      Test Time: 12:29pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:29pm
FLO	Pass	12:29pm
FC	Pass	12:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:29pm
SRC	Pass	12:29pm
DET	Pass	12:29pm
BAR	Pass	12:29pm
BT	Pass	12:29pm

Blank Tests

Test	Status	Time
AIR	Pass	12:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:30pm

CRC Tests

Test	Status	Time
COMP	Pass	12:30pm
CAL	Pass	12:30pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County LENOIR Instrument Location LENOIR CO. S.D.  
Instrument Serial No. 008639 130 QUEEN ST., KINSTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA  
Permit Number: 11646E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG734101  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	10:16am
AIR BLK	.00	10:17am
ACCY CHK	.07	10:17am
AIR BLK	.00	10:19am
SUB TEST	.00	10:19am
AIR BLK	.00	10:20am
SUB TEST	.00	10:22am
AIR BLK	.00	10:23am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639      Test Record Number: 3126  
Test Date: 03/07/2018      Test Time: 10:23am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:24am
FLO	Pass	10:24am
FC	Pass	10:24am

Temperature Tests

Test	Status	Time
FC1	Pass	10:24am
SRC	Pass	10:24am
DET	Pass	10:24am
BAR	Pass	10:24am
BT	Pass	10:24am

Blank Tests

Test	Status	Time
AIR	Pass	10:25am

Printer Tests

Test	Status	Time
PRNT	Pass	10:25am

CRC Tests

Test	Status	Time
COMP	Pass	10:25am
CAL	Pass	10:25am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County LENOIR Instrument Location KINSTON P.P.  
Instrument Serial No. 008624 205 E. KING ST., KINSTON, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Jim N. Reese  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:49am
AIR BLK	.00	10:50am
ACCY CHK	.07	10:51am
AIR BLK	.00	10:52am
<b>SUB TEST</b>	<b>.00</b>	<b>10:52am</b>
AIR BLK	.00	10:53am
<b>SUB TEST</b>	<b>.00</b>	<b>10:55am</b>
AIR BLK	.00	10:56am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
**Analyst**

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



**Intox EC/IR-II: Preventive Maintenance**

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624      Test Record Number: 1670  
Test Date: 03/07/2018      Test Time: 10:56am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:57am
FLO	Pass	10:57am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:57am


Printer Tests

Test	Status	Time
PRNT	Pass	10:58am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Cornelius PD  
Instrument Serial No. 008692 21410 Catawba Ave. Cornelius

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692

Test Date: 03/01/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	1:18pm
AIR BLK	.00	1:19pm
ACCY CHK	.08	1:20pm
AIR BLK	.00	1:21pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:22pm</b>
AIR BLK	.00	1:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:24pm</b>
AIR BLK	.00	1:25pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692      Test Record Number: 2768  
Test Date: 03/01/2018      Test Time: 1:26pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:26pm
FLO	Pass	1:26pm
FC	Pass	1:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:27pm
SRC	Pass	1:27pm
DET	Pass	1:27pm
BAR	Pass	1:27pm
BT	Pass	1:27pm

Blank Tests

Test	Status	Time
AIR	Pass	1:27pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:27pm

CRC Tests

Test	Status	Time
COMP	Pass	1:28pm
CAL	Pass	1:28pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location CM PD-LEC  
Instrument Serial No. 008691 600 E. Trade St. Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008691  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	2:40pm
AIR BLK	.00	2:40pm
ACCY CHK	.08	2:41pm
AIR BLK	.00	2:42pm
SUB TEST	.00	2:43pm
AIR BLK	.00	2:44pm
SUB TEST	.00	2:45pm
AIR BLK	.00	2:46pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008691      Test Record Number: 7198  
Test Date: 03/16/2018      Test Time: 2:47pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:48pm
FLO	Pass	2:48pm
FC	Pass	2:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:48pm
SRC	Pass	2:48pm
DET	Pass	2:48pm
BAR	Pass	2:48pm
BT	Pass	2:48pm

Blank Tests

Test	Status	Time
AIR	Pass	2:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:48pm

CRC Tests

Test	Status	Time
COMP	Pass	2:49pm
CAL	Pass	2:49pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location BAT MOBILE 7  
Instrument Serial No. 008969

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008969  
Test Date: 03/17/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W  
Permit Number: 07281E  
Effective:  
02/01/2018-02/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG621404  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:06pm
AIR BLK	.00	10:07pm
ACCY CHK	.08	10:08pm
AIR BLK	.00	10:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:11pm</b>
AIR BLK	.00	10:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:13pm</b>
AIR BLK	.00	10:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008969      Test Record Number: 217  
Test Date: 03/17/2018      Test Time: 10:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:16pm
FLO	Pass	10:16pm
FC	Pass	10:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:16pm
SRC	Pass	10:16pm
DET	Pass	10:16pm
BAR	Pass	10:16pm
BT	Pass	10:16pm

Blank Tests

Test	Status	Time
AIR	Pass	10:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:17pm

CRC Tests

Test	Status	Time
COMP	Pass	10:17pm
CAL	Pass	10:17pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Matthews PD  
Instrument Serial No. 008699 1201 Crews Rd, Matthews

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hitt  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG COUNTY MATTHEWS PD 590

Serial Number: 008699  
Test Date: 03/22/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: HUTCHINSON, JOSEPH  
Permit Number: 19951E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	4:21pm
AIR BLK	.00	4:22pm
ACCY CHK	.07	4:23pm
AIR BLK	.00	4:24pm
SUB TEST	.00	4:25pm
AIR BLK	.00	4:26pm
SUB TEST	.00	4:27pm
AIR BLK	.00	4:28pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

**MECKLENBURG COUNTY MATTHEWS PD 590**

Serial Number: 008699      Test Record Number: 2628  
Test Date: 03/22/2018      Test Time: 4:29pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:29pm
FLO	Pass	4:29pm
FC	Pass	4:30pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:30pm
SRC	Pass	4:30pm
DET	Pass	4:30pm
BAR	Pass	4:30pm
BT	Pass	4:30pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:30pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:30pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:30pm
CAL	Pass	4:30pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mecklenburg Instrument Location BAO Mobile ?

Instrument Serial No. 008972

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008972  
Test Date: 03/29/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W  
Permit Number: 07281E  
Effective:  
02/01/2018-02/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG621404  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:18pm
AIR BLK	.00	11:19pm
ACCY CHK	.07	11:20pm
AIR BLK	.00	11:21pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:21pm</b>
AIR BLK	.00	11:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:24pm</b>
AIR BLK	.00	11:24pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**MECKLENBURG BAT MOBILE UNIT 7 590**

Serial Number: 008972      Test Record Number: 288  
Test Date: 03/29/2018      Test Time: 11:26pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:27pm
FLO	Pass	11:27pm
FC	Pass	11:27pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:27pm
SRC	Pass	11:27pm
DET	Pass	11:27pm
BAR	Pass	11:27pm
BT	Pass	11:27pm

**Blank Tests**

Test	Status	Time
AIR	Pass	11:28pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	11:28pm

**CRC Tests**

Test	Status	Time
COMP	Pass	11:28pm
CAL	Pass	11:28pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mecklenburg Instrument Location BAF MOBILE 7

Instrument Serial No. 008968

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29<sup>th</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008968  
Test Date: 03/29/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 07281E

Effective:

02/01/2018-02/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:16pm
AIR BLK	.00	10:17pm
ACCY CHK	.07	10:18pm
AIR BLK	.00	10:19pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:19pm</b>
AIR BLK	.00	10:20pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:21pm</b>
AIR BLK	.00	10:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008968      Test Record Number: 237  
Test Date: 03/29/2018      Test Time: 10:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:26pm
FLO	Pass	10:26pm
FC	Pass	10:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:26pm
SRC	Pass	10:26pm
DET	Pass	10:26pm
BAR	Pass	10:26pm
BT	Pass	10:26pm

Blank Tests

Test	Status	Time
AIR	Pass	10:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:27pm

CRC Tests

Test	Status	Time
COMP	Pass	10:27pm
CAL	Pass	10:27pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County MECKLENBURG Instrument Location BAT MOBILE 7

Instrument Serial No. 008971

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008971  
Test Date: 03/29/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 07281E

Effective:

02/01/2018-02/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:06pm
AIR BLK	.00	10:07pm
ACCY CHK	.07	10:08pm
AIR BLK	.00	10:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:09pm</b>
AIR BLK	.00	10:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:12pm</b>
AIR BLK	.00	10:13pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008971      Test Record Number: 211  
Test Date: 03/29/2018      Test Time: 10:14pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:14pm
FLO	Pass	10:14pm
FC	Pass	10:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:14pm
SRC	Pass	10:14pm
DET	Pass	10:14pm
BAR	Pass	10:14pm
BT	Pass	10:14pm

Blank Tests

Test	Status	Time
AIR	Pass	10:15pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:15pm

CRC Tests

Test	Status	Time
COMP	Pass	10:15pm
CAL	Pass	10:15pm

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location BAF MOBILE 7

Instrument Serial No. 008969

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

659  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008969  
Test Date: 03/30/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: INGLE, LARRY W  
Permit Number: 07281E  
Effective:  
02/01/2018-02/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG621404  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	12:32am
AIR BLK	.00	12:33am
ACCY CHK	.08	12:33am
AIR BLK	.00	12:34am
<b>SUB TEST</b>	<b>.00</b>	<b>12:35am</b>
AIR BLK	.00	12:36am
<b>SUB TEST</b>	<b>.00</b>	<b>12:37am</b>
AIR BLK	.00	12:38am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008969      Test Record Number: 220  
Test Date: 03/30/2018      Test Time: 12:42am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:42am
FLO	Pass	12:42am
FC	Pass	12:43am

Temperature Tests

Test	Status	Time
FC1	Pass	12:43am
SRC	Pass	12:43am
DET	Pass	12:43am
BAR	Pass	12:43am
BT	Pass	12:43am

Blank Tests

Test	Status	Time
AIR	Pass	12:43am

Printer Tests

Test	Status	Time
PRNT	Pass	12:43am

CRC Tests

Test	Status	Time
COMP	Pass	12:43am
CAL	Pass	12:43am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mitchell

Instrument Location Spruce Pine PD


Instrument Serial No. 008726

Spruce Pine, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726

Test Date: 03/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	1:35pm
AIR BLK	.00	1:36pm
ACCY CHK	.08	1:37pm
AIR BLK	.00	1:38pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:38pm</b>
AIR BLK	.00	1:39pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:41pm</b>
AIR BLK	.00	1:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726      Test Record Number: 912  
Test Date: 03/08/2018      Test Time: 1:43pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:43pm
FLO	Pass	1:43pm
FC	Pass	1:43pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:43pm
SRC	Pass	1:43pm
DET	Pass	1:43pm
BAR	Pass	1:43pm
BT	Pass	1:43pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:44pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:44pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:44pm
CAL	Pass	1:44pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Moore Co. Instrument Location Pinehurst Police Dept.  
Instrument Serial No. 008710 Pinehurst, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710  
Test Date: 03/08/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	12:49pm
AIR BLK	.00	12:50pm
ACCY CHK	.07	12:51pm
AIR BLK	.00	12:52pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:52pm</b>
AIR BLK	.00	12:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:55pm</b>
AIR BLK	.00	12:56pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710      Test Record Number: 1471  
Test Date: 03/08/2018      Test Time: 12:56pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:57pm
FLO	Pass	12:57pm
FC	Pass	12:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:57pm
SRC	Pass	12:57pm
DET	Pass	12:57pm
BAR	Pass	12:57pm
BT	Pass	12:57pm

Blank Tests

Test	Status	Time
AIR	Pass	12:58pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:58pm

CRC Tests

Test	Status	Time
COMP	Pass	12:58pm
CAL	Pass	12:58pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County NASH

Instrument Location NAT MODEL 10

Instrument Serial No. 009580

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9<sup>th</sup> day of MARCH, 2011, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Preventive Maintenance**

NASH COUNTY BAT MOBILE UNIT 10 630

Serial Number: 008580      Test Record Number: 2414  
Test Date: 03/09/2018      Test Time: 8:47pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:47pm
FLO	Pass	8:47pm
FC	Pass	8:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:47pm
SRC	Pass	8:47pm
DET	Pass	8:47pm
BAR	Pass	8:47pm
BT	Pass	8:47pm

Blank Tests

Test	Status	Time
AIR	Pass	8:48pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:48pm

CRC Tests

Test	Status	Time
COMP	Pass	8:48pm
CAL	Pass	8:48pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

**Intox EC/IR-II: Subject Test**

NASH COUNTY BAT MOBILE UNIT 10 630

Serial Number: 008580  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
*PREVENTIVE, MAINTENANCE*  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: VARNELL, BRYON L  
Permit Number: 16896E  
Effective:  
09/22/2017-09/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:38pm
AIR BLK	.00	8:39pm
ACCY CHK	.07	8:40pm
AIR BLK	.00	8:41pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:41pm</b>
AIR BLK	.00	8:42pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:44pm</b>
AIR BLK	.00	8:45pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER Instrument Location BAT MOBILE UNIT 9  
Instrument Serial No. 008647 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY BAT MOBILE UNIT 9  
640

Serial Number: 008647  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 15671E  
Effective:  
07/01/2017-07/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:02pm
AIR BLK	.00	8:03pm
ACCY CHK	.07	8:04pm
AIR BLK	.00	8:05pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:06pm</b>
AIR BLK	.00	8:07pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:08pm</b>
AIR BLK	.00	8:09pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008647      Test Record Number: 2394  
Test Date: 03/09/2018      Test Time: 8:10pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:10pm
FLO	Pass	8:10pm
FC	Pass	8:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:10pm
SRC	Pass	8:10pm
DET	Pass	8:10pm
BAR	Pass	8:10pm
BT	Pass	8:10pm

Blank Tests

Test	Status	Time
AIR	Pass	8:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:11pm

CRC Tests

Test	Status	Time
COMP	Pass	8:11pm
CAL	Pass	8:11pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County NEW HANOVER Instrument Location BAT MOBILE UNIT 9

Instrument Serial No. 008704 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY BAT MOBILE UNIT 9  
640

Serial Number: 008704  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 15671E  
Effective:  
07/01/2017-07/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG607601  
Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	7:41pm
AIR BLK	.00	7:42pm
ACCY CHK	.08	7:43pm
AIR BLK	.00	7:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:44pm</b>
AIR BLK	.00	7:45pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:46pm</b>
AIR BLK	.00	7:47pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008704      Test Record Number: 503  
Test Date: 03/09/2018      Test Time: 7:48pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:49pm
FLO	Pass	7:49pm
FC	Pass	7:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:49pm
SRC	Pass	7:49pm
DET	Pass	7:49pm
BAR	Pass	7:49pm
BT	Pass	7:49pm

Blank Tests

Test	Status	Time
AIR	Pass	7:49pm


Printer Tests

Test	Status	Time
PRNT	Pass	7:50pm

CRC Tests

Test	Status	Time
COMP	Pass	7:50pm
CAL	Pass	7:50pm

Preventive Maintenance  
Status: Pass



Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER

Instrument Location BAT MOBILE UNIT 9


Instrument Serial No. 008575

WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Banks

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY BAT MOBILE UNIT 9  
640

Serial Number: 008575  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R  
Permit Number: 15671E  
Effective:  
07/01/2017-07/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	7:39pm
AIR BLK	.00	7:39pm
ACCY CHK	.07	7:40pm
AIR BLK	.00	7:41pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:42pm</b>
AIR BLK	.00	7:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>7:44pm</b>
AIR BLK	.00	7:45pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008575      Test Record Number: 1066  
Test Date: 03/09/2018      Test Time: 7:46pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	7:46pm
FLO	Pass	7:46pm
FC	Pass	7:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:46pm
SRC	Pass	7:46pm
DET	Pass	7:46pm
BAR	Pass	7:46pm
BT	Pass	7:46pm

Blank Tests

Test	Status	Time
AIR	Pass	7:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:47pm

CRC Tests

Test	Status	Time
COMP	Pass	7:47pm
CAL	Pass	7:47pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location New Hanover  
Instrument Serial No. 008626 County Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY. NEW HANOVER CO SD  
640

Serial Number: 008626  
Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534902

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	3:00pm
AIR BLK	.00	3:01pm
ACCY CHK	.07	3:01pm
AIR BLK	.00	3:02pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:03pm</b>
AIR BLK	.00	3:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:05pm</b>
AIR BLK	.00	3:06pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: 008626 Test Record Number: 7040  
Test Date: 03/15/2018 Test Time: 3:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:09pm
FLO	Pass	3:09pm
FC	Pass	3:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:10pm

CRC Tests

Test	Status	Time
COMP	Pass	3:10pm
CAL	Pass	3:10pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location New Hanover  
Instrument Serial No. 008617 County Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY NEW HANOVER CO SD  
640

Serial Number: 008617  
Test Date: 03/15/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C  
Permit Number: 5329E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG721401  
Exp Date: 09/02/2019

Test	g/210L	Time
DIAG	Pass	2:55pm
AIR BLK	.00	2:56pm
ACCY CHK	.08	2:56pm
AIR BLK	.00	2:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:00pm</b>
AIR BLK	.00	3:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:02pm</b>
AIR BLK	.00	3:03pm

Reported AC: 00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: 008617      Test Record Number: 2888  
Test Date: 03/15/2018      Test Time: 3:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:09pm
FLO	Pass	3:09pm
FC	Pass	3:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:10pm

CRC Tests

Test	Status	Time
COMP	Pass	3:10pm
CAL	Pass	3:10pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Wilmington  
Instrument Serial No. 008628 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: 008628  
Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	4:17pm
AIR BLK	.00	4:18pm
ACCY CHK	.07	4:19pm
AIR BLK	.00	4:20pm
SUB TEST	.00	4:20pm
AIR BLK	.00	4:21pm
SUB TEST	.00	4:23pm
AIR BLK	.00	4:24pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: 008628      Test Record Number: 4484  
Test Date: 03/15/2018      Test Time: 4:26pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	4:27pm
FLO	Pass	4:27pm
FC	Pass	4:27pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	4:27pm
SRC	Pass	4:27pm
DET	Pass	4:27pm
BAR	Pass	4:27pm
BT	Pass	4:27pm

**Blank Tests**

Test	Status	Time
AIR	Pass	4:28pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	4:28pm

**CRC Tests**

Test	Status	Time
COMP	Pass	4:28pm
CAL	Pass	4:28pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Carolina Beach  
Instrument Serial No. 008661 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhoads  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY CAROLINA BEACH PD  
640

Serial Number: 008661

Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	9:35am
AIR BLK	.00	9:35am
ACCY CHK	.07	9:36am
AIR BLK	.00	9:37am
<b>SUB TEST</b>	<b>.00</b>	<b>9:38am</b>
AIR BLK	.00	9:39am
<b>SUB TEST</b>	<b>.00</b>	<b>9:40am</b>
AIR BLK	.00	9:42am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY CAROLINA BEACH PD 640

Serial Number: 008661 Test Record Number: 2418  
Test Date: 03/16/2018 Test Time: 9:43am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:43am
FLO	Pass	9:43am
FC	Pass	9:43am

Temperature Tests

Test	Status	Time
FC1	Pass	9:43am
SRC	Pass	9:43am
DET	Pass	9:43am
BAR	Pass	9:43am
BT	Pass	9:43am

Blank Tests

Test	Status	Time
AIR	Pass	9:44am

Printer Tests

Test	Status	Time
PRNT	Pass	9:44am

CRC Tests

Test	Status	Time
COMP	Pass	9:44am
CAL	Pass	9:44am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Wrightsville  
Instrument Serial No. 008667 Beach Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD  
640

Serial Number: 008667  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:51am
ACCY CHK	.07	10:52am
AIR BLK	.00	10:53am
<b>SUB TEST</b>	<b>.00</b>	<b>10:53am</b>
AIR BLK	.00	10:54am
<b>SUB TEST</b>	<b>.00</b>	<b>10:56am</b>
AIR BLK	.00	10:57am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

**NEW HANOVER COUNTY WRIGHTSVILLE BCH PD 640**

Serial Number: 008667      Test Record Number: 1714  
Test Date: 03/16/2018      Test Time: 10:57am EDT

System Check: Passed

**Baseline Tests**

Test	Status	Time
IR	Pass	10:58am
FLO	Pass	10:58am
FC	Pass	10:58am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:58am
SRC	Pass	10:58am
DET	Pass	10:58am
BAR	Pass	10:58am
BT	Pass	10:58am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:59am


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:59am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:59am
CAL	Pass	10:59am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Northampton Instrument Location Northampton Co Sheriff's office  
Instrument Serial No. 008688 105 W. Jefferson St  
Jackson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

642

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

**NORTHAMPTON COUNTY SHERIFFS DEPARTMENT**  
650

Serial Number: 008688  
Test Date: 03/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702302

Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	11:14am
AIR BLK	.00	11:15am
ACCY CHK	.07	11:16am
AIR BLK	.00	11:17am
SUB TEST	.00	11:17am
AIR BLK	.00	11:18am
SUB TEST	.00	11:20am
AIR BLK	.00	11:21am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures**  
**Forensic Tests for Alcohol Branch**  
**Department of Health and Human Services**  
**Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**NORTHAMPTON COUNTY SHERIFFS DEPARTMENT 650**

Serial Number: 008688      Test Record Number: 769  
Test Date: 03/09/2018      Test Time: 11:22am EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	11:22am
FLO	Pass	11:22am
FC	Pass	11:23am

**Temperature Tests**

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

**Blank Tests**

Test	Status	Time
AIR	Pass	11:23am

**Printer Tests**


Test	Status	Time
PRNT	Pass	11:23am

**CRC Tests**

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance

Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County NORTHAMPTON Instrument Location NORTHAMPTON Co. Sheriff's Office

Instrument Serial No. 008607 105 W. Jefferson ST  
JACKSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

662  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT  
650

Serial Number: 008607  
Test Date: 03/09/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, SIMON S  
Permit Number: 11434E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG702302  
Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	11:13am
AIR BLK	.00	11:14am
ACCY CHK	.07	11:14am
AIR BLK	.00	11:15am
SUB TEST	.00	11:16am
AIR BLK	.00	11:17am
SUB TEST	.00	11:19am
AIR BLK	.00	11:20am

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT 650

Serial Number: 008607      Test Record Number: 934  
Test Date: 03/09/2018      Test Time: 11:21am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:22am
FLO	Pass	11:22am
FC	Pass	11:22am

Temperature Tests

Test	Status	Time
FC1	Pass	11:22am
SRC	Pass	11:22am
DET	Pass	11:22am
BAR	Pass	11:22am
BT	Pass	11:22am

Blank Tests

Test	Status	Time
AIR	Pass	11:22am


Printer Tests

Test	Status	Time
PRNT	Pass	11:22am

CRC Tests

Test	Status	Time
COMP	Pass	11:23am
CAL	Pass	11:23am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ONslow

Instrument Location BAT MOBILE UNIT 9


Instrument Serial No. 008707

HOLLY RIDGE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 02 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ONslow COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008707  
Test Date: 03/02/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201


Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:49pm
AIR BLK	.00	9:50pm
ACCY CHK	.08	9:50pm
AIR BLK	.00	9:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:52pm</b>
AIR BLK	.00	9:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:54pm</b>
AIR BLK	.00	9:55pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

ONslow COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008707      Test Record Number: 2454  
Test Date: 03/02/2018      Test Time: 9:57pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:57pm
FLO	Pass	9:57pm
FC	Pass	9:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:57pm
SRC	Pass	9:57pm
DET	Pass	9:57pm
BAR	Pass	9:57pm
BT	Pass	9:57pm

Blank Tests

Test	Status	Time
AIR	Pass	9:58pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:58pm

CRC Tests

Test	Status	Time
COMP	Pass	9:58pm
CAL	Pass	9:58pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County ONSLOW


Instrument Location CAMP LEJEUNE PND

Instrument Serial No. 0D8920

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ONslow COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920  
Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	11:40am
AIR BLK	.00	11:41am
ACCY CHK	.08	11:42am
AIR BLK	.00	11:43am
SUB TEST	.00	11:44am
AIR BLK	.00	11:45am
SUB TEST	.00	11:46am
AIR BLK	.00	11:47am

Reported AC: .00 g/210L

*R E Hall*

Signature of Chemical Analyst

Court CVR

*Randy E Hall*  
Analyst

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920      Test Record Number: 1468  
Test Date: 03/14/2018      Test Time: 11:48am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:48am
FLO	Pass	11:48am
FC	Pass	11:48am

Temperature Tests

Test	Status	Time
FC1	Pass	11:48am
SRC	Pass	11:48am
DET	Pass	11:48am
BAR	Pass	11:48am
BT	Pass	11:48am

Blank Tests

Test	Status	Time
AIR	Pass	11:49am


Printer Tests

Test	Status	Time
PRNT	Pass	11:49am

CRC Tests

Test	Status	Time
COMP	Pass	11:49am
CAL	Pass	11:49am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ONSTOW


Instrument Location JACKSONVILLE AD

Instrument Serial No. 008930

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY JACKSONVILLE RD 660

Serial Number: 008930  
Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XY

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	12:25pm
AIR BLK	.00	12:25pm
ACCY CHK	.08	12:26pm
AIR BLK	.00	12:27pm
SUB TEST	.00	12:28pm
AIR BLK	.00	12:29pm
SUB TEST	.00	12:30pm
AIR BLK	.00	12:31pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

ONslow COUNTY JACKSONVILLE PD 660

Serial Number: 008930 Test Record Number: 2045  
Test Date: 03/14/2018 Test Time: 12:32pm EDT

System Check: Passed

**Baseline Tests**

Test	Status	Time
IR	Pass	12:32pm
FLO	Pass	12:32pm
FC	Pass	12:32pm

**Temperature Tests**

Test	Status	Time
FCL	Pass	12:32pm
SRC	Pass	12:32pm
DET	Pass	12:32pm
BAR	Pass	12:32pm
BT	Pass	12:32pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:33pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	12:33pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:33pm
CAL	Pass	12:33pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Onslow Instrument Location Onslow County  
Instrument Serial No. 008931 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ON SLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008931  
Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019


Test	g/210L	Time
DIAG	Pass	1:15pm
AIR BLK	.00	1:16pm
ACCY CHK	.08	1:16pm
AIR BLK	.00	1:18pm
SUB TEST	.00	1:18pm
AIR BLK	.00	1:19pm
SUB TEST	.00	1:21pm
AIR BLK	.00	1:22pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008931 Test Record Number: 2809  
Test Date: 03/14/2018 Test Time: 1:25pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:25pm
SRC	Pass	1:25pm
DET	Pass	1:25pm
BAR	Pass	1:25pm
BT	Pass	1:25pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Onslow Instrument Location Onslow County

Instrument Serial No. 008932 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ray E Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008932  
Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	1:18pm
AIR BLK	.00	1:18pm
ACCY CHK	.08	1:19pm
AIR BLK	.00	1:20pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:21pm</b>
AIR BLK	.00	1:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:23pm</b>
AIR BLK	.00	1:24pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008932      Test Record Number: 4395  
Test Date: 03/14/2018      Test Time: 1:26pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:26pm
FLO	Pass	1:26pm
FC	Pass	1:27pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:27pm
SRC	Pass	1:27pm
DET	Pass	1:27pm
BAR	Pass	1:27pm
BT	Pass	1:27pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:27pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	1:27pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:28pm
CAL	Pass	1:28pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County ONslow Instrument Location MCAS New River AFD

Instrument Serial No. 008919

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy Hall  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

ONslow COUNTY MCAS NEW RIVER 660

Serial Number: 008919  
Test Date: 03/14/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702402

Exp Date: 01/24/2019


Test	g/210L	Time
DIAG	Pass	2:30pm
AIR BLK	.00	2:31pm
ACCY CHK	.08	2:31pm
AIR BLK	.00	2:32pm
SUB TEST	.00	2:33pm
AIR BLK	.00	2:34pm
SUB TEST	.00	2:35pm
AIR BLK	.00	2:36pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY MCAS NEW RIVER 660

Serial Number: 008919      Test Record Number: 600  
Test Date: 03/14/2018      Test Time: 2:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:37pm
FLO	Pass	2:37pm
FC	Pass	2:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:37pm
SRC	Pass	2:37pm
DET	Pass	2:37pm
BAR	Pass	2:37pm
BT	Pass	2:37pm

Blank Tests

Test	Status	Time
AIR	Pass	2:38pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:38pm

CRC Tests

Test	Status	Time
COMP	Pass	2:38pm
CAL	Pass	2:38pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE Instrument Location Hillsborough PD  
Instrument Serial No. 008799 127 N. Chanton ST  
Hillsborough, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

642  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799  
Test Date: 03/01/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: BARNES, SIMON S  
Permit Number: 11434E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG716202  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	2:55pm
AIR BLK	.00	2:56pm
ACCY CHK	.08	2:57pm
AIR BLK	.00	2:58pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:59pm
SUB TEST	.00	3:01pm
AIR BLK	.00	3:02pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799      Test Record Number: 2704  
Test Date: 03/01/2018      Test Time: 3:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:02pm
FLO	Pass	3:02pm
FC	Pass	3:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:03pm
SRC	Pass	3:03pm
DET	Pass	3:03pm
BAR	Pass	3:03pm
BT	Pass	3:03pm

Blank Tests

Test	Status	Time
AIR	Pass	3:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:03pm

CRC Tests

Test	Status	Time
COMP	Pass	3:03pm
CAL	Pass	3:03pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Pamlico

Instrument Location BAT Mobile Unit 12


Instrument Serial No. 008788

Pamlico Co So

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

660  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PAMLICO COUNTY BAT MOBILE UNIT 12 680

Serial Number: 008788  
Test Date: 03/10/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JASON R

Permit Number: 19145E

Effective:

03/01/2018-03/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:22pm
AIR BLK	.00	9:23pm
ACCY CHK	.07	9:24pm
AIR BLK	.00	9:24pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:26pm</b>
AIR BLK	.00	9:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:29pm</b>
AIR BLK	.00	9:30pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

**PAMLICO COUNTY BAT MOBILE UNIT 12 680**

Serial Number: 008788      Test Record Number: 1330  
Test Date: 03/10/2018      Test Time: 9:51pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:51pm
FLO	Pass	9:51pm
FC	Pass	9:51pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:52pm
SRC	Pass	9:52pm
DET	Pass	9:52pm
BAR	Pass	9:52pm
BT	Pass	9:52pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:52pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:52pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:52pm
CAL	Pass	9:52pm

Preventive Maintenance  
Status: *Pass*

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Pamlico Instrument Location Pamlico County

Instrument Serial No. 008640 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Holf  
Signature of Certifying Official

354  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: 008640

Test Date: 03/15/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019


Test	g/210L	Time
DIAG	Pass	1:48pm
AIR BLK	.00	1:49pm
ACCY CHK	.08	1:50pm
AIR BLK	.00	1:51pm
SUB TEST	.00	1:52pm
AIR BLK	.00	1:52pm
SUB TEST	.00	1:54pm
AIR BLK	.00	1:55pm

Reported AC: .07 g/210L



Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: 008640 Test Record Number: 1303  
Test Date: 03/15/2018 Test Time: 1:56pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:56pm
FLO	Pass	1:56pm
FC	Pass	1:56pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:56pm
SRC	Pass	1:56pm
DET	Pass	1:56pm
BAR	Pass	1:56pm
BT	Pass	1:56pm

Blank Tests

Test	Status	Time
AIR	Pass	1:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:57pm

CRC Tests

Test	Status	Time
COMP	Pass	1:57pm
CAL	Pass	1:57pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pender Instrument Location Pender County  
Instrument Serial No. 008946 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhodes  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PENDER PENDER COUNTY SD 700

Serial Number: 008946  
Test Date: 03/14/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	2:11pm
AIR BLK	.00	2:12pm
ACCY CHK	.08	2:13pm
AIR BLK	.00	2:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:14pm</b>
AIR BLK	.00	2:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>2:17pm</b>
AIR BLK	.00	2:18pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PENDER PENDER COUNTY SD 700

Serial Number: 008946      Test Record Number: 883  
Test Date: 03/14/2018      Test Time: 2:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:19pm
FLO	Pass	2:19pm
FC	Pass	2:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:19pm
SRC	Pass	2:19pm
DET	Pass	2:19pm
BAR	Pass	2:19pm
BT	Pass	2:19pm

Blank Tests

Test	Status	Time
AIR	Pass	2:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:20pm

CRC Tests

Test	Status	Time
COMP	Pass	2:20pm
CAL	Pass	2:20pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II


County Pender Instrument Location Pender County

Instrument Serial No 008935 Sheriff Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Pender  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

PENDER PENDER CO SD 700

Serial Number: 008935  
Test Date: 03/14/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	1:58pm
AIR BLK	.00	1:59pm
ACCY CHK	.07	1:59pm
AIR BLK	.00	2:01pm
SUB TEST	.00	2:01pm
AIR BLK	.00	2:02pm
SUB TEST	.00	2:04pm
AIR BLK	.00	2:05pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

PENDER PENDER CO SD 700

Serial Number: 008935      Test Record Number: 2252  
Test Date: 03/14/2018      Test Time: 2:06pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:06pm
FLO	Pass	2:06pm
FC	Pass	2:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:07pm
SRC	Pass	2:07pm
DET	Pass	2:07pm
BAR	Pass	2:07pm
BT	Pass	2:07pm

Blank Tests

Test	Status	Time
AIR	Pass	2:07pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:07pm

CRC Tests

Test	Status	Time
COMP	Pass	2:08pm
CAL	Pass	2:08pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pender Instrument Location Pender County  
Instrument Serial No. 008748 Sheriff Department Hanes

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhoden  
Signature of Certifying Official

601  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: 008948

Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVEVTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:00pm
ACCY CHK	.07	12:01pm
AIR BLK	.00	12:02pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:02pm</b>
AIR BLK	.00	12:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:05pm</b>
AIR BLK	.00	12:06pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: 008948      Test Record Number: 802  
Test Date: 03/16/2018      Test Time: 12:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:09pm
FLO	Pass	12:09pm
FC	Pass	12:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:09pm
SRC	Pass	12:09pm
DET	Pass	12:09pm
BAR	Pass	12:09pm
BT	Pass	12:09pm

Blank Tests

Test	Status	Time
AIR	Pass	12:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:10pm

CRC Tests

Test	Status	Time
COMP	Pass	12:10pm
CAL	Pass	12:10pm

Preventive Maintenance

Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County PERQUIMANS Instrument Location PERQUIMANS CO. S. D.  
Instrument Serial No. 008921 110 church ST., HERTFORD, N. C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10<sup>th</sup> day of April, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James Kessel  
Signature of Certifying Official

047  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921

Test Date: 04/10/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA

Permit Number: 11646E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	12:02pm
AIR BLK	.00	12:04pm
ACCY CHK	.08	12:04pm
AIR BLK	.00	12:05pm
SUB TEST	.00	12:06pm
AIR BLK	.00	12:07pm
SUB TEST	.00	12:09pm
AIR BLK	.00	12:09pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921      Test Record Number: 763  
Test Date: 04/10/2018      Test Time: 12:11pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

**Blank Tests**

Test	Status	Time
AIR	Pass	12:12pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	12:12pm

**CRC Tests**

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance  
Status: Pass

  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Person Instrument Location East Mobile Unit 8  
Instrument Serial No. 008601 SHP - Roxsboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Janya B Skinner 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

PERSON COUNTY BATMOBILE UNIT 8 720

Serial Number: 008601  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG702401  
Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	10:02pm
AIR BLK	.00	10:03pm
ACCY CHK	.08	10:04pm
AIR BLK	.00	10:04pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:05pm</b>
AIR BLK	.00	10:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:07pm</b>
AIR BLK	.00	10:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY BATMOBILE UNIT 8 720

Serial Number: 008601      Test Record Number: 1262  
Test Date: 03/16/2018      Test Time: 10:10pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:10pm
FLO	Pass	10:10pm
FC	Pass	10:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:11pm
SRC	Pass	10:11pm
DET	Pass	10:11pm
BAR	Pass	10:11pm
BT	Pass	10:11pm

Blank Tests

Test	Status	Time
AIR	Pass	10:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:11pm

CRC Tests

Test	Status	Time
COMP	Pass	10:11pm
CAL	Pass	10:11pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Person Instrument Location Bat mobile Unit 8  
Instrument Serial No. 008816 SHP - Roxboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner  
Signature of Certifying Official

1044  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008816  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG805801  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	9:59pm
AIR BLK	.00	10:00pm
ACCY CHK	.07	10:00pm
AIR BLK	.00	10:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:02pm</b>
AIR BLK	.00	10:03pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:04pm</b>
AIR BLK	.00	10:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008816 Test Record Number: 7391  
Test Date: 03/16/2018 Test Time: 10:07pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:07pm
FLO	Pass	10:07pm
FC	Pass	10:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:07pm
SRC	Pass	10:07pm
DET	Pass	10:07pm
BAR	Pass	10:07pm
BT	Pass	10:07pm

Blank Tests

Test	Status	Time
AIR	Pass	10:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:08pm

CRC Tests

Test	Status	Time
COMP	Pass	10:08pm
CAL	Pass	10:08pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Person Instrument Location Bat Mobile Unit 8  
Instrument Serial No 008615 SHP. Roxsford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donna B Skinner  
Signature of Certifying Official

444  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008615

Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	9:56pm
AIR BLK	.00	9:57pm
ACCY CHK	.07	9:58pm
AIR BLK	.00	10:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:00pm</b>
AIR BLK	.00	10:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:03pm</b>
AIR BLK	.00	10:04pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008615      Test Record Number: 5528  
Test Date: 03/16/2018      Test Time: 10:12pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:13pm
FLO	Pass	10:13pm
FC	Pass	10:13pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:13pm
SRC	Pass	10:13pm
DET	Pass	10:13pm
BAR	Pass	10:13pm
BT	Pass	10:13pm

**Blank Tests**

Test	Status	Time
AIR	Pass	10:14pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:14pm

**CRC Tests**

Test	Status	Time
COMP	Pass	10:14pm
CAL	Pass	10:14pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Person Instrument Location Bat Mobile Unit 8  
Instrument Serial No. 008775 SHP - Roxboro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donna B Skinnin 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008775  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	10:11pm
AIR BLK	.00	10:12pm
ACCY CHK	.08	10:13pm
AIR BLK	.00	10:14pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:14pm</b>
AIR BLK	.00	10:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:17pm</b>
AIR BLK	.00	10:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

PERSON COUNTY BAT MOBILE UNIT 8 720

Serial Number: 008775      Test Record Number: 1747  
Test Date: 03/16/2018      Test Time: 10:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:20pm
FLO	Pass	10:20pm
FC	Pass	10:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:20pm
SRC	Pass	10:20pm
DET	Pass	10:20pm
BAR	Pass	10:20pm
BT	Pass	10:20pm

Blank Tests

Test	Status	Time
AIR	Pass	10:21pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:21pm

CRC Tests

Test	Status	Time
COMP	Pass	10:21pm
CAL	Pass	10:21pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Polk Instrument Location Polk County SO  
Instrument Serial No. 008832 46 Ward St. Columbus

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hest  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

POLK COUNTY POLK COUNTY SD 740

Serial Number: 008832  
Test Date: 03/22/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	12:56pm
AIR BLK	.00	12:57pm
ACCY CHK	.07	12:57pm
AIR BLK	.00	12:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:59pm</b>
AIR BLK	.00	1:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:02pm</b>
AIR BLK	.00	1:03pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

POLK COUNTY POLK COUNTY SD 740

Serial Number: 008832      Test Record Number: 1450  
Test Date: 03/22/2018      Test Time: 1:04pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:04pm
FLO	Pass	1:04pm
FC	Pass	1:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:04pm
SRC	Pass	1:04pm
DET	Pass	1:04pm
BAR	Pass	1:04pm
BT	Pass	1:04pm

Blank Tests

Test	Status	Time
AIR	Pass	1:05pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:05pm

CRC Tests

Test	Status	Time
COMP	Pass	1:05pm
CAL	Pass	1:05pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Co. Instrument Location Randolph Co. Jail  
Instrument Serial No. 108860 Ashboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2<sup>nd</sup> day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY RANDOLPH COUNTY JAIL  
750

Serial Number: 008860  
Test Date: 03/02/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG702401  
Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	12:53pm
AIR BLK	.00	12:54pm
ACCY CHK	.07	12:55pm
AIR BLK	.00	12:56pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:56pm</b>
AIR BLK	.00	12:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:59pm</b>
AIR BLK	.00	1:00pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY RANDOLPH COUNTY JAIL 750

Serial Number: 008860      Test Record Number: 2586  
Test Date: 03/02/2018      Test Time: 1:03pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:03pm
FLO	Pass	1:03pm
FC	Pass	1:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:04pm
SRC	Pass	1:04pm
DET	Pass	1:04pm
BAR	Pass	1:04pm
BT	Pass	1:04pm

Blank Tests

Test	Status	Time
AIR	Pass	1:04pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:04pm

CRC Tests

Test	Status	Time
COMP	Pass	1:04pm
CAL	Pass	1:04pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County RANDOLPH Co. Instrument Location RANDOLPH Co. Jail

Instrument Serial No. 006899

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

RANDOLPH COUNTY RANDOLPH CO. JAIL 750

Serial Number: 008899  
Test Date: 03/02/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG716201  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:51pm
AIR BLK	.00	12:52pm
ACCY CHK	.08	12:52pm
AIR BLK	.00	12:53pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:54pm</b>
AIR BLK	.00	12:55pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:57pm</b>
AIR BLK	.00	12:58pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

RANDOLPH COUNTY RANDOLPH CO. JAIL 750

Serial Number: 008899      Test Record Number: 2790  
Test Date: 03/02/2018      Test Time: 12:58pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:59pm
FLO	Pass	12:59pm
FC	Pass	12:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:59pm
SRC	Pass	12:59pm
DET	Pass	12:59pm
BAR	Pass	12:59pm
BT	Pass	12:59pm

Blank Tests

Test	Status	Time
AIR	Pass	1:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:00pm

CRC Tests

Test	Status	Time
COMP	Pass	1:00pm
CAL	Pass	1:00pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Richmond Co. Instrument Location Richmond Co. Mag's Office  
Instrument Serial No. 008701 Rockingham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test.**

RICHMOND COUNTY RICHMOND CO. MAG OFF  
760

Serial Number: 008701  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

12/01/2017-12/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	3:07pm
AIR BLK	.00	3:08pm
ACCY CHK	.08	3:08pm
AIR BLK	.00	3:09pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:10pm</b>
AIR BLK	.00	3:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>3:13pm</b>
AIR BLK	.00	3:14pm

Reported AC1 .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008701      Test Record Number: 1152  
Test Date: 03/07/2018      Test Time: 3:15pm EST

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	3:16pm
FLO	Pass	3:16pm
FC	Pass	3:16pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:16pm
SRC	Pass	3:16pm
DET	Pass	3:16pm
BAR	Pass	3:16pm
BT	Pass	3:16pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:17pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:17pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:17pm
CAL	Pass	3:17pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Richmond Co. Instrument Location Richmond Co. Mag's Office

Instrument Serial No. 008840 Lockingtons, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5<sup>th</sup> day of MARCH, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

654  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

RICHMOND COUNTY RICHMOND CO. MAG OFF  
760

Serial Number: 008840  
Test Date: 03/07/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C  
Permit Number: 7682E  
Effective:  
12/01/2017-12/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805802  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	3:04pm
AIR BLK	.00	3:05pm
ACCY CHK	.08	3:05pm
AIR BLK	.00	3:06pm
SUB TEST	.00	3:07pm
AIR BLK	.00	3:08pm
SUB TEST	.00	3:10pm
AIR BLK	.00	3:11pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008840      Test Record Number: 2073  
Test Date: 03/07/2018      Test Time: 3:12pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:12pm
FLO	Pass	3:12pm
FC	Pass	3:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:13pm
SRC	Pass	3:13pm
DET	Pass	3:13pm
BAR	Pass	3:13pm
BT	Pass	3:13pm

Blank Tests

Test	Status	Time
AIR	Pass	3:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:13pm

CRC Tests

Test	Status	Time
COMP	Pass	3:13pm
CAL	Pass	3:13pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County SAMPSON Instrument Location SAMPSON COUNTY

Instrument Serial No. 008825 SHERIFF DEPT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alm R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: 008825  
Test Date: 03/08/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R.  
Permit Number: 15671E  
Effective:  
07/01/2017-07/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG702401  
Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	10:00am
AIR BLK	.00	10:01am
ACCY CHK	.07	10:01am
AIR BLK	.00	10:02am
<b>SUB TEST</b>	<b>.00</b>	<b>10:03am</b>
AIR BLK	.00	10:04am
<b>SUB TEST</b>	<b>.00</b>	<b>10:05am</b>
AIR BLK	.00	10:06am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: 008825      Test Record Number: 2416  
Test Date: 03/08/2018      Test Time: 10:10am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:11am
FLO	Pass	10:11am
FC	Pass	10:11am

Temperature Tests

Test	Status	Time
FC1	Pass	10:11am
SRC	Pass	10:11am
DET	Pass	10:11am
BAR	Pass	10:11am
BT	Pass	10:11am

Blank Tests

Test	Status	Time
AIR	Pass	10:12am


Printer Tests

Test	Status	Time
PRNT	Pass	10:12am

CRC Tests

Test	Status	Time
COMP	Pass	10:12am
CAL	Pass	10:12am

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County SAMPSON Instrument Location SAMPSON COUNTY  
Instrument Serial No. 008877 SHERIFF DEPT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Barnes  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: 008877  
Test Date: 03/08/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG805802


Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	9:45am
AIR BLK	.00	9:45am
ACCY CHK	.08	9:46am
AIR BLK	.00	9:47am
<b>SUB TEST</b>	<b>.00</b>	<b>9:48am</b>
AIR BLK	.00	9:49am
<b>SUB TEST</b>	<b>.00</b>	<b>9:50am</b>
AIR BLK	.00	9:51am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: 008877      Test Record Number: 2786  
Test Date: 03/08/2018      Test Time: 9:53am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:53am
FLO	Pass	9:53am
FC	Pass	9:53am

Temperature Tests

Test	Status	Time
FC1	Pass	9:53am
SRC	Pass	9:53am
DET	Pass	9:53am
BAR	Pass	9:53am
BT	Pass	9:53am

Blank Tests

Test	Status	Time
AIR	Pass	9:54am


Printer Tests

Test	Status	Time
PRNT	Pass	9:54am

CRC Tests

Test	Status	Time
COMP	Pass	9:54am
CAL	Pass	9:54am

Preventive Maintenance  
Status: Pass



Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Stanly County SD  
Instrument Serial No. 008674 1265.3<sup>rd</sup> St, Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1<sup>st</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008824  
Test Date: 03/01/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG621403  
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:38am
AIR BLK	.00	11:39am
ACCY CHK	.08	11:39am
AIR BLK	.00	11:41am
<b>SUB TEST</b>	<b>.00</b>	<b>11:41am</b>
AIR BLK	.00	11:42am
<b>SUB TEST</b>	<b>.00</b>	<b>11:44am</b>
AIR BLK	.00	11:45am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008824      Test Record Number: 1268  
Test Date: 03/01/2018      Test Time: 11:46am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:46am
FLO	Pass	11:46am
FC	Pass	11:46am

Temperature Tests

Test	Status	Time
FC1	Pass	11:47am
SRC	Pass	11:47am
DET	Pass	11:47am
BAR	Pass	11:47am
BT	Pass	11:47am

Blank Tests

Test	Status	Time
AIR	Pass	11:47am


Printer Tests

Test	Status	Time
PRNT	Pass	11:47am

CRC Tests

Test	Status	Time
COMP	Pass	11:47am
CAL	Pass	11:47am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Locust PD  
Instrument Serial No. 008706 186 Roy Kennedy Dr, Locust

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13<sup>th</sup> day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

STANLY LOCUST PD 830

Serial Number: 008706

Test Date: 03/13/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805801

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	9:13am
AIR BLK	.00	9:14am
ACCY CHK	.08	9:15am
AIR BLK	.00	9:16am
<b>SUB TEST</b>	<b>.00</b>	<b>9:17am</b>
AIR BLK	.00	9:18am
<b>SUB TEST</b>	<b>.00</b>	<b>9:19am</b>
AIR BLK	.00	9:20am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

STANLY LOCUST PD 830

Serial Number: 008706      Test Record Number: 3476  
Test Date: 03/13/2018      Test Time: 9:22am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:22am
FLO	Pass	9:22am
FC	Pass	9:22am

Temperature Tests

Test	Status	Time
FC1	Pass	9:22am
SRC	Pass	9:22am
DET	Pass	9:22am
BAR	Pass	9:22am
BT	Pass	9:22am

Blank Tests

Test	Status	Time
AIR	Pass	9:23am


Printer Tests

Test	Status	Time
PRNT	Pass	9:23am

CRC Tests

Test	Status	Time
COMP	Pass	9:24am
CAL	Pass	9:24am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Surry County Jail  
Instrument Serial No. 008934 Dobson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official 657 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SURRY COUNTY SURRY CO JAIL 850

Serial Number: 008934  
Test Date: 03/24/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R.  
Permit Number: 22067E  
Effective:  
07/01/2016-07/01/2018

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG702302  
Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	11:48am
AIR BLK	.00	11:49am
ACCY CHK	.07	11:49am
AIR BLK	.00	11:50am
SUB TEST	.00	11:51am
AIR BLK	.00	11:52am
SUB TEST	.00	11:53am
AIR BLK	.00	11:54am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**



Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY SURRY CO JAIL 850

Serial Number: 008934 Test Record Number: 1865  
Test Date: 03/24/2018 Test Time: 11:55am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:55am
FLO	Pass	11:55am
FC	Pass	11:55am

Temperature Tests

Test	Status	Time
FC1	Pass	11:56am
SRC	Pass	11:56am
DET	Pass	11:56am
BAR	Pass	11:56am
BT	Pass	11:56am

Blank Tests

Test	Status	Time
AIR	Pass	11:56am


Printer Tests

Test	Status	Time
PRNT	Pass	11:56am

CRC Tests

Test	Status	Time
COMP	Pass	11:56am
CAL	Pass	11:56am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Surry Instrument Location Mount Airy  
Instrument Serial No. 008943 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kent R. Bennett  
Signature of Certifying Official

657  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

SURRY COUNTY MOUNT AIRY PD 850

Serial Number: 008943  
Test Date: 03/23/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R  
Permit Number: 22067E  
Effective:  
07/01/2016-07/01/2018

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG702302  
Exp Date: 01/23/2019

Test	g/210L	Time
DIAG	Pass	9:03am
AIR BLK	.00	9:04am
ACCY CHK	.08	9:05am
AIR BLK	.00	9:06am
<b>SUB TEST</b>	<b>.00</b>	<b>9:06am</b>
AIR BLK	.00	9:07am
<b>SUB TEST</b>	<b>.00</b>	<b>9:09am</b>
AIR BLK	.00	9:10am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

*SURRY COUNTY MOUNT AIRY PD 850*

Serial Number: 008943      Test Record Number: 1997  
Test Date: 03/23/2018      Test Time: 9:10am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:10am
FLO	Pass	9:10am
FC	Pass	9:10am

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:11am
SRC	Pass	9:11am
DET	Pass	9:11am
BAR	Pass	9:11am
BT	Pass	9:11am

**Blank Tests**

Test	Status	Time
AIR	Pass	9:11am


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:11am

**CRC Tests**

Test	Status	Time
COMP	Pass	9:11am
CAL	Pass	9:11am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County Swain Instrument Location Swain Co. Jail

Instrument Serial No. 008723 Bryson City, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cutler

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723  
Test Date: 03/22/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R  
Permit Number: 08457  
Effective:  
09/01/2017-09/01/2019

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG734102  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	11:10am
AIR BLK	.00	11:11am
ACCY CHK	.08	11:12am
AIR BLK	.00	11:13am
<b>SUB TEST</b>	<b>.00</b>	<b>11:13am</b>
AIR BLK	.00	11:14am
<b>SUB TEST</b>	<b>.00</b>	<b>11:16am</b>
AIR BLK	.00	11:17am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**Intox EC/IR-II: Preventive Maintenance**

SWAIN COUNTY SWAIN COUNTY JAIL 860

Serial Number: 008723      Test Record Number: 725  
Test Date: 03/22/2018      Test Time: 11:18am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:18am
FLO	Pass	11:18am
FC	Pass	11:18am

Temperature Tests

Test	Status	Time
FC1	Pass	11:18am
SRC	Pass	11:18am
DET	Pass	11:18am
BAR	Pass	11:18am
BT	Pass	11:18am

Blank Tests

Test	Status	Time
AIR	Pass	11:19am


Printer Tests

Test	Status	Time
PRNT	Pass	11:19am

CRC Tests

Test	Status	Time
COMP	Pass	11:19am
CAL	Pass	11:19am

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Union County SD  
Instrument Serial No. 008876 3344 Passon Rd, Monro

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12th day of March, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

656  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876  
Test Date: 03/12/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D  
Permit Number: 15924E  
Effective:  
01/01/2018-01/01/2020

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	10:26am
AIR BLK	.00	10:27am
ACCY CHK	.07	10:28am
AIR BLK	.00	10:29am
<b>SUB TEST</b>	<b>.00</b>	<b>10:30am</b>
AIR BLK	.00	10:31am
<b>SUB TEST</b>	<b>.00</b>	<b>10:32am</b>
AIR BLK	.00	10:33am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876      Test Record Number: 4751  
Test Date: 03/12/2018      Test Time: 10:36am EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	10:36am
FLO	Pass	10:36am
FC	Pass	10:37am

**Temperature Tests**

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

**Blank Tests**

Test	Status	Time
AIR	Pass	10:37am


**Printer Tests**

Test	Status	Time
PRNT	Pass	10:37am

**CRC Tests**

Test	Status	Time
COMP	Pass	10:37am
CAL	Pass	10:37am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Union County SD  
Instrument Serial No. 008866 3344 Presson Rd, Monroe

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]

Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-11: Subject Test

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008866  
Test Date: 03/12/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2018-01/01/2020

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG805802

Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	10:11am
AIR BLK	.00	10:12am
ACCY CHK	.08	10:13am
AIR BLK	.00	10:14am
SUB TEST	.00	10:15am
AIR BLK	.00	10:16am
SUB TEST	.00	10:17am
AIR BLK	.00	10:18am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008866 Test Record Number: 2832

Test Date: 03/12/2018 Test Time: 10:20am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:20am
FLO	Pass	10:20am
FC	Pass	10:20am

Temperature Tests

Test	Status	Time
FC1	Pass	10:20am
SRC	Pass	10:20am
DET	Pass	10:20am
BAR	Pass	10:20am
BT	Pass	10:20am

Blank Tests

Test	Status	Time
AIR	Pass	10:21am

Printer Tests


Test	Status	Time
PRNT	Pass	10:21am

CRC Tests

Test	Status	Time
COMP	Pass	10:21am
CAL	Pass	10:21am

Preventive Maintenance

Status: Pass

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Waxhaw PD  
Instrument Serial No. 008598 3620 Providence Rd, Waxhaw

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22nd day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Holt  
Signature of Certifying Official

650  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

UNION COUNTY WAXHAW PD 890

Serial Number: 008598  
Test Date: 03/22/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH

Permit Number: 19951E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG722408

Exp Date: 08/12/2019

Test	g/210L	Time
DIAG	Pass	03:18pm
AIR BLK	.00	03:19pm
ACCY CHK	.07	03:20pm
AIR BLK	.00	03:21pm
<b>SUB TEST</b>	<b>.00</b>	<b>03:22pm</b>
AIR BLK	.00	03:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>03:26pm</b>
AIR BLK	.00	03:27pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

UNION COUNTY WAXHAW PD 890

Serial Number: 008598      Test Record Number: 723  
Test Date: 03/22/2018      Test Time: 03:28pm

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	03:28pm
FLO	Pass	03:28pm
FC	Pass	03:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	03:28pm
SRC	Pass	03:28pm
DET	Pass	03:28pm
BAR	Pass	03:28pm
BT	Pass	03:28pm

Blank Tests

Test	Status	Time
AIR	Pass	03:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	03:29pm

CRC Tests

Test	Status	Time
COMP	Pass	03:29pm
CAL	Pass	03:29pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location WAKE CO DETENTION CTR

Instrument Serial No. 008960 3301 HAMMOND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 05 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith  
Signature of Certifying Official

637  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760  
Test Date: 03/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA


Agency: DHHS

Test Type: Breath Test


Lot Number: AG734101  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:26pm
AIR BLK	.00	1:26pm
ACCY CHK	.08	1:27pm
AIR BLK	.00	1:28pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:28pm</b>
AIR BLK	.00	1:29pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:31pm</b>
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760      Test Record Number: 2840  
Test Date: 03/05/2018      Test Time: 1:35pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:35pm
FLO	Pass	1:35pm
FC	Pass	1:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:36pm
SRC	Pass	1:36pm
DET	Pass	1:36pm
BAR	Pass	1:36pm
BT	Pass	1:36pm

Blank Tests

Test	Status	Time
AIR	Pass	1:36pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:36pm

CRC Tests

Test	Status	Time
COMP	Pass	1:37pm
CAL	Pass	1:37pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Wake Co. Detention  
Instrument Serial No. 008778 3301 HAMMOND RD  
Raleigh, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

492

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778

Test Date: 03/05/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2017-05/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG734101

Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	1:29pm
AIR BLK	.00	1:30pm
ACCY CHK	.08	1:30pm
AIR BLK	.00	1:31pm
SUB TEST	.00	1:32pm
AIR BLK	.00	1:33pm
SUB TEST	.00	1:34pm
AIR BLK	.00	1:35pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778      Test Record Number: 3016  
Test Date: 03/05/2018      Test Time: 1:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:36pm
FLO	Pass	1:36pm
FC	Pass	1:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:36pm
SRC	Pass	1:36pm
DET	Pass	1:36pm
BAR	Pass	1:36pm
BT	Pass	1:36pm

Blank Tests

Test	Status	Time
AIR	Pass	1:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:37pm

CRC Tests

Test	Status	Time
COMP	Pass	1:37pm
CAL	Pass	1:37pm

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE FOREST PD  
Instrument Serial No. 008700 225 S. TAYLOR ST. WAKE FOREST, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith  
Signature of Certifying Official

637  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700  
Test Date: 03/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702402


Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	1:12pm
AIR BLK	.00	1:13pm
ACCY CHK	.07	1:13pm
AIR BLK	.00	1:15pm
SUB TEST	.00	1:15pm
AIR BLK	.00	1:16pm
SUB TEST	.00	1:18pm
AIR BLK	.00	1:19pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR

  
Analyst



Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700 Test Record Number: 1328  
Test Date: 03/09/2018 Test Time: 1:19pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:20pm
FLO	Pass	1:20pm
FC	Pass	1:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:20pm
SRC	Pass	1:20pm
DET	Pass	1:20pm
BAR	Pass	1:20pm
BT	Pass	1:20pm

Blank Tests

Test	Status	Time
AIR	Pass	1:21pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:21pm

CRC Tests

Test	Status	Time
COMP	Pass	1:21pm
CAL	Pass	1:21pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location STAT MODEL UNIT 10  
Instrument Serial No. 008637

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008637  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	8:53pm
AIR BLK	.00	8:54pm
ACCY CHK	.07	8:55pm
AIR BLK	.00	8:56pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:56pm</b>
AIR BLK	.00	8:57pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:58pm</b>
AIR BLK	.00	8:59pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008637      Test Record Number: 2899  
Test Date: 03/16/2018      Test Time: 9:03pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:03pm
FLO	Pass	9:03pm
FC	Pass	9:03pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:04pm
SRC	Pass	9:04pm
DET	Pass	9:04pm
BAR	Pass	9:04pm
BT	Pass	9:04pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:04pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	9:04pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:04pm
CAL	Pass	9:04pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location SAT MOBILE UNIT 10

Instrument Serial No. 008590

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox, EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008580  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:53pm
AIR BLK	.00	8:54pm
ACCY CHK	.07	8:54pm
AIR BLK	.00	8:55pm
SUB TEST	.00	8:56pm
AIR BLK	.00	8:57pm
SUB TEST	.00	8:58pm
AIR BLK	.00	8:59pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008580      Test Record Number: 2421  
Test Date: 03/16/2018      Test Time: 9:02pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:02pm
FLO	Pass	9:02pm
FC	Pass	9:02pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:02pm
SRC	Pass	9:02pm
DET	Pass	9:02pm
BAR	Pass	9:02pm
BT	Pass	9:02pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:03pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:03pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:03pm
CAL	Pass	9:03pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE

Instrument Location SAT MODEL UNIT 10


Instrument Serial No. 008779

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.





Signature of Certifying Official

663  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008779  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	8:56pm
AIR BLK	.00	8:57pm
ACCY CHK	.07	8:57pm
AIR BLK	.00	8:58pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:58pm</b>
AIR BLK	.00	8:59pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:01pm</b>
AIR BLK	.00	9:02pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008779      Test Record Number: 3471  
Test Date: 03/16/2018      Test Time: 9:06pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:06pm
FLO	Pass	9:06pm
FC	Pass	9:06pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:06pm
SRC	Pass	9:06pm
DET	Pass	9:06pm
BAR	Pass	9:06pm
BT	Pass	9:06pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:07pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:07pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:07pm
CAL	Pass	9:07pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE


Instrument Location SAT MONITOR UNIT 10

Instrument Serial No. 008686

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

663

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686  
Test Date: 03/16/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401


Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	8:57pm
AIR BLK	.00	8:58pm
ACCY CHK	.07	8:59pm
AIR BLK	.00	9:00pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:00pm</b>
AIR BLK	.00	9:01pm
<b>SUB TEST</b>	<b>.00</b>	<b>9:03pm</b>
AIR BLK	.00	9:04pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686      Test Record Number: 6554  
Test Date: 03/16/2018      Test Time: 9:05pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:05pm
FLO	Pass	9:05pm
FC	Pass	9:05pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:05pm
SRC	Pass	9:05pm
DET	Pass	9:05pm
BAR	Pass	9:05pm
BT	Pass	9:05pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:06pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:06pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:06pm
CAL	Pass	9:06pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location SAT MONITOR UNIT 10

Instrument Serial No. 008776

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

663  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008776  
Test Date: 03/16/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: VARNELL, BRYON L  
Permit Number: 16896E  
Effective:  
09/22/2017-09/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	8:51pm
AIR BLK	.00	8:52pm
ACCY CHK	.07	8:53pm
AIR BLK	.00	8:54pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:54pm</b>
AIR BLK	.00	8:55pm
<b>SUB TEST</b>	<b>.00</b>	<b>8:57pm</b>
AIR BLK	.00	8:58pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008776      Test Record Number: 3432  
Test Date: 03/16/2018      Test Time: 9:02pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	9:03pm
FLO	Pass	9:03pm
FC	Pass	9:03pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	9:03pm
SRC	Pass	9:03pm
DET	Pass	9:03pm
BAR	Pass	9:03pm
BT	Pass	9:03pm

**Blank Tests**

Test	Status	Time
AIR	Pass	9:03pm

**Printer Tests**

Test	Status	Time
PRNT	Pass	9:03pm

**CRC Tests**

Test	Status	Time
COMP	Pass	9:04pm
CAL	Pass	9:04pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Bat mobile Unit 8  
Instrument Serial No. 008816 Raleigh PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donya B Skinner 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test.**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008816  
Test Date: 03/17/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG805801  
Exp Date: 02/27/2020

Test	g/210L	Time
DIAG	Pass	11:47pm
AIR BLK	.00	11:48pm
ACCY CHK	.07	11:48pm
AIR BLK	.00	11:49pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:50pm</b>
AIR BLK	.00	11:51pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:52pm</b>
AIR BLK	.00	11:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008816      Test Record Number: 7393  
Test Date: 03/17/2018      Test Time: 11:57pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:57pm
FLO	Pass	11:57pm
FC	Pass	11:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:57pm
SRC	Pass	11:57pm
DET	Pass	11:57pm
BAR	Pass	11:57pm
BT	Pass	11:57pm

Blank Tests

Test	Status	Time
AIR	Pass	11:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:58pm

CRC Tests

Test	Status	Time
COMP	Pass	11:58pm
CAL	Pass	11:58pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Bat Mobile Unit 8  
Instrument Serial No. 008736 Raleigh PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donya B Skinner  
Signature of Certifying Official

6441  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008736  
Test Date: 03/18/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG716201  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:16am
AIR BLK	.00	12:17am
ACCY CHK	.07	12:18am
AIR BLK	.00	12:19am
<b>SUB TEST</b>	<b>.00</b>	<b>12:20am</b>
AIR BLK	.00	12:20am
<b>SUB TEST</b>	<b>.00</b>	<b>12:22am</b>
AIR BLK	.00	12:23am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008736      Test Record Number: 893  
Test Date: 03/18/2018      Test Time: 12:42am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:42am
FLO	Pass	12:42am
FC	Pass	12:43am

Temperature Tests

Test	Status	Time
FC1	Pass	12:43am
SRC	Pass	12:43am
DET	Pass	12:43am
BAR	Pass	12:43am
BT	Pass	12:43am

Blank Tests

Test	Status	Time
AIR	Pass	12:43am

Printer Tests

Test	Status	Time
PRNT	Pass	12:43am

CRC Tests

Test	Status	Time
COMP	Pass	12:43am
CAL	Pass	12:43am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Bat Mobile Unit 8

Instrument Serial No. 008715 Raleigh, PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner  
Signature of Certifying Official

144  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008775  
Test Date: 03/18/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Female  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA  
Permit Number: 13651E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG734101  
Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	12:39am
AIR BLK	.00	12:40am
ACCY CHK	.07	12:41am
AIR BLK	.00	12:42am
SUB TEST	.00	12:42am
AIR BLK	.00	12:43am
SUB TEST	.00	12:46am
AIR BLK	.00	12:47am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008775      Test Record Number: 1751  
Test Date: 03/18/2018      Test Time: 12:50am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:50am
FLO	Pass	12:50am
FC	Pass	12:50am

Temperature Tests

Test	Status	Time
FC1	Pass	12:50am
SRC	Pass	12:50am
DET	Pass	12:50am
BAR	Pass	12:50am
BT	Pass	12:50am

Blank Tests

Test	Status	Time
AIR	Pass	12:51am

Printer Tests

Test	Status	Time
PRNT	Pass	12:51am

CRC Tests

Test	Status	Time
COMP	Pass	12:51am
CAL	Pass	12:51am

Preventive Maintenance  
Status: Pass

  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Unit 8  
Instrument Serial No. 608615 Raleigh PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donya B Skinner 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008615  
Test Date: 03/18/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG716201

Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:56am
AIR BLK	.00	12:57am
ACCY CHK	.07	12:58am
AIR BLK	.00	12:59am
<b>SUB TEST</b>	<b>.00</b>	<b>1:00am</b>
AIR BLK	.00	1:01am
<b>SUB TEST</b>	<b>.00</b>	<b>1:02am</b>
AIR BLK	.00	1:03am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 8 910

Serial Number: 008615      Test Record Number: 5532  
Test Date: 03/18/2018      Test Time: 1:04am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:04am
FLO	Pass	1:04am
FC	Pass	1:04am

Temperature Tests

Test	Status	Time
FC1	Pass	1:04am
SRC	Pass	1:04am
DET	Pass	1:04am
BAR	Pass	1:04am
BT	Pass	1:04am

Blank Tests

Test	Status	Time
AIR	Pass	1:05am

Printer Tests

Test	Status	Time
PRNT	Pass	1:05am

CRC Tests

Test	Status	Time
COMP	Pass	1:05am
CAL	Pass	1:05am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Pat Mobile Unit 8  
Instrument Serial No. 008601 Raleigh PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2018, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donya B Skinner 644  
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAKE COUNTY BATMOBILE UNIT 8 910

Serial Number: 008601  
Test Date: 03/18/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA

Permit Number: 13651E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	12:19am
AIR BLK	.00	12:20am
ACCY CHK	.08	12:21am
AIR BLK	.00	12:22am
<b>SUB TEST</b>	<b>.00</b>	<b>12:23am</b>
AIR BLK	.00	12:23am
<b>SUB TEST</b>	<b>.00</b>	<b>12:25am</b>
AIR BLK	.00	12:26am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WAKE COUNTY BATMOBILE UNIT 8 910

Serial Number: 008601      Test Record Number: 1267  
Test Date: 03/18/2018      Test Time: 12:27am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:28am
FLO	Pass	12:28am
FC	Pass	12:28am

Temperature Tests

Test	Status	Time
FC1	Pass	12:28am
SRC	Pass	12:28am
DET	Pass	12:28am
BAR	Pass	12:28am
BT	Pass	12:28am

Blank Tests

Test	Status	Time
AIR	Pass	12:28am


Printer Tests

Test	Status	Time
PRNT	Pass	12:28am

CRC Tests

Test	Status	Time
COMP	Pass	12:29am
CAL	Pass	12:29am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WARREN Instrument Location NORLINA PD

Instrument Serial No. 008945 101 MAIN ST. NORLINA, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 20 18 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



B. D. Smith  
Signature of Certifying Official

637  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945  
Test Date: 03/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG721401

Exp Date: 08/02/2019

Test	g/210L	Time
DIAG	Pass	5:04pm
AIR BLK	.00	5:05pm
ACCY CHK	.08	5:05pm
AIR BLK	.00	5:06pm
<b>SUB TEST</b>	<b>.00</b>	<b>5:07pm</b>
AIR BLK	.00	5:08pm
<b>SUB TEST</b>	<b>.00</b>	<b>5:09pm</b>
AIR BLK	.00	5:10pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945      Test Record Number: 371  
Test Date: 03/09/2018      Test Time: 5:12pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	5:12pm
FLO	Pass	5:12pm
FC	Pass	5:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:13pm
SRC	Pass	5:13pm
DET	Pass	5:13pm
BAR	Pass	5:13pm
BT	Pass	5:13pm

Blank Tests

Test	Status	Time
AIR	Pass	5:13pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:13pm

CRC Tests

Test	Status	Time
COMP	Pass	5:13pm
CAL	Pass	5:13pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WARREN Instrument Location WARREN CO. LEC

Instrument Serial No. 008793 128 RAFTERS LN WARRENTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith  
Signature of Certifying Official

637  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793  
Test Date: 03/09/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN

Permit Number: 8937E

Effective:

08/01/2017-08/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG734101


Exp Date: 12/07/2019

Test	g/210L	Time
DIAG	Pass	3:35pm
AIR BLK	.00	3:36pm
ACCY CHK	.07	3:36pm
AIR BLK	.00	3:37pm
SUB TEST	.00	3:38pm
AIR BLK	.00	3:39pm
SUB TEST	.00	3:40pm
AIR BLK	.00	3:41pm

Reported AC: .00 g/210L

  
Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007**

**Intox EC/IR-II: Preventive Maintenance**

WARREN COUNTY WARREN COUNTY JAIL 920

Serial Number: 008793      Test Record Number: 1272  
Test Date: 03/09/2018      Test Time: 3:42pm EST

System Check: Passed

**Baseline Tests**

Test	Status	Time
IR	Pass	3:43pm
FLO	Pass	3:43pm
FC	Pass	3:43pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	3:43pm
SRC	Pass	3:43pm
DET	Pass	3:43pm
BAR	Pass	3:43pm
BT	Pass	3:43pm

**Blank Tests**

Test	Status	Time
AIR	Pass	3:44pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	3:44pm

**CRC Tests**

Test	Status	Time
COMP	Pass	3:44pm
CAL	Pass	3:44pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Washington Instrument Location Washington Co. S.O.  
Instrument Serial No. 008829 Adams St., Plymouth, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20<sup>th</sup> day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key [Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WASHINGTON COUNTY SHERIFF'S OFFICE 930

Serial Number: 008829  
Test Date: 03/20/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G  
Permit Number: 12955E  
Effective:  
06/01/2017-06/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	10:14am
AIR BLK	.00	10:15am
ACCY CHK	.07	10:15am
AIR BLK	.00	10:17am
<b>SUB TEST</b>	<b>.00</b>	<b>10:17am</b>
AIR BLK	.00	10:18am
<b>SUB TEST</b>	<b>.00</b>	<b>10:20am</b>
AIR BLK	.00	10:21am

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

Intox EC/IR-II: Preventive Maintenance

WASHINGTON COUNTY SHERIFF'S OFFICE 930

Serial Number: 008829      Test Record Number: 836  
Test Date: 03/20/2018      Test Time: 10:23am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:23am
FLO	Pass	10:23am
FC	Pass	10:23am

Temperature Tests

Test	Status	Time
FC1	Pass	10:23am
SRC	Pass	10:23am
DET	Pass	10:23am
BAR	Pass	10:23am
BT	Pass	10:23am

Blank Tests

Test	Status	Time
AIR	Pass	10:24am

Printer Tests

Test	Status	Time
PRNT	Pass	10:24am

CRC Tests

Test	Status	Time
COMP	Pass	10:24am
CAL	Pass	10:24am

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst



DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAYNE Instrument Location BAT MOBILE UNIT 9  
Instrument Serial No. 008575 GOLDSBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...  
Signature of Certifying Official

648  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY BAT MOBILE UNIT 9 950

Serial Number: 008575

Test Date: 03/17/2018

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

07/01/2017-07/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG710701


Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	11:19pm
AIR BLK	.00	11:20pm
ACCY CHK	.07	11:21pm
AIR BLK	.00	11:22pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:22pm</b>
AIR BLK	.00	11:23pm
<b>SUB TEST</b>	<b>.00</b>	<b>11:25pm</b>
AIR BLK	.00	11:26pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures  
Forensic Tests for Alcohol Branch  
Department of Health and Human Services  
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY BAT MOBILE UNIT 9 950

Serial Number: 008575      Test Record Number: 1077  
Test Date: 03/17/2018      Test Time: 11:29pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:29pm
FLO	Pass	11:29pm
FC	Pass	11:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:29pm
SRC	Pass	11:29pm
DET	Pass	11:29pm
BAR	Pass	11:29pm
BT	Pass	11:29pm

Blank Tests

Test	Status	Time
AIR	Pass	11:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:30pm

CRC Tests

Test	Status	Time
COMP	Pass	11:30pm
CAL	Pass	11:30pm

Preventive Maintenance  
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County Wayne Instrument Location Wayne Co. Detention Ctr.

Instrument Serial No. 008847 207 E. CHESTNUT ST., GOLDSBORO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31<sup>ST</sup> day of MARCH, 2015 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

647  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008847  
Test Date: 03/31/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA  
Permit Number: 11646E  
Effective:  
08/01/2017-08/01/2019

Officer's Name: NONE, NONE  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test


Lot Number: AG710701  
Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	1:22pm
AIR BLK	.00	1:23pm
ACCY CHK	.08	1:24pm
AIR BLK	.00	1:25pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:26pm</b>
AIR BLK	.00	1:27pm
<b>SUB TEST</b>	<b>.00</b>	<b>1:28pm</b>
AIR BLK	.00	1:29pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
**Analyst**

**Intox EC/IR-II: Preventive Maintenance**

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008847      Test Record Number: 613  
Test Date: 03/31/2018      Test Time: 1:19pm EDT

System Check: *Passed*

**Baseline Tests**

Test	Status	Time
IR	Pass	1:19pm
FLO	Pass	1:19pm
FC	Pass	1:19pm

**Temperature Tests**

Test	Status	Time
FC1	Pass	1:19pm
SRC	Pass	1:19pm
DET	Pass	1:19pm
BAR	Pass	1:19pm
BT	Pass	1:19pm

**Blank Tests**

Test	Status	Time
AIR	Pass	1:20pm


**Printer Tests**

Test	Status	Time
PRNT	Pass	1:20pm

**CRC Tests**

Test	Status	Time
COMP	Pass	1:20pm
CAL	Pass	1:20pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II

County WILSON

Instrument Location AT MOBILE UNIT 10

Instrument Serial No. 008776

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

663  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008776  
Test Date: 03/17/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	10:07pm
AIR BLK	.00	10:08pm
ACCY CHK	.07	10:09pm
AIR BLK	.00	10:10pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:10pm</b>
AIR BLK	.00	10:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:13pm</b>
AIR BLK	.00	10:14pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst



**Intox EC/IR-II: Preventive Maintenance**

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008776      Test Record Number: 3436  
Test Date: 03/17/2018      Test Time: 10:15pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:15pm
FLO	Pass	10:15pm
FC	Pass	10:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:15pm
SRC	Pass	10:15pm
DET	Pass	10:15pm
BAR	Pass	10:15pm
BT	Pass	10:15pm

Blank Tests

Test	Status	Time
AIR	Pass	10:16pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:16pm

CRC Tests

Test	Status	Time
COMP	Pass	10:16pm
CAL	Pass	10:16pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County WILSON Instrument Location ISAT MODEL UNIT 10

Instrument Serial No. 008540

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

663  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008580  
Test Date: 03/17/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG710701

Exp Date: 04/17/2019

Test	g/210L	Time
DIAG	Pass	10:08pm
AIR BLK	.00	10:09pm
ACCY CHK	.07	10:10pm
AIR BLK	.00	10:11pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:11pm</b>
AIR BLK	.00	10:12pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:14pm</b>
AIR BLK	.00	10:15pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008580      Test Record Number: 2424  
Test Date: 03/17/2018      Test Time: 10:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:16pm
FLO	Pass	10:16pm
FC	Pass	10:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:16pm
SRC	Pass	10:16pm
DET	Pass	10:16pm
BAR	Pass	10:16pm
BT	Pass	10:16pm

Blank Tests

Test	Status	Time
AIR	Pass	10:17pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:17pm

CRC Tests

Test	Status	Time
COMP	Pass	10:17pm
CAL	Pass	10:17pm

Preventive Maintenance  
Status: Pass



**Analyst**

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**


County WILSON Instrument Location SAT MOBILE UNIT 10

Instrument Serial No. 008646

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2014, the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

663  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

**Intox EC/IR-II: Subject Test**

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008686  
Test Date: 03/17/2018

Citation Number: M0000000-0  
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: VARNELL, BRYON L

Permit Number: 16896E

Effective:

09/22/2017-09/01/2019

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG702401

Exp Date: 01/24/2019

Test	g/210L	Time
DIAG	Pass	10:11pm
AIR BLK	.00	10:12pm
ACCY CHK	.07	10:12pm
AIR BLK	.00	10:13pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:14pm</b>
AIR BLK	.00	10:15pm
<b>SUB TEST</b>	<b>.00</b>	<b>10:16pm</b>
AIR BLK	.00	10:17pm

Reported AC: .00 g/210L

  
\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
\_\_\_\_\_  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

WILSON COUNTY BAT MOBILE UNIT 10 970

Serial Number: 008686      Test Record Number: 6558  
Test Date: 03/17/2018      Test Time: 10:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:19pm
FLO	Pass	10:19pm
FC	Pass	10:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:19pm
SRC	Pass	10:19pm
DET	Pass	10:19pm
BAR	Pass	10:19pm
BT	Pass	10:19pm

Blank Tests

Test	Status	Time
AIR	Pass	10:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:19pm

CRC Tests

Test	Status	Time
COMP	Pass	10:20pm
CAL	Pass	10:20pm

Preventive Maintenance  
Status: Pass

  
\_\_\_\_\_  
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES  
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD  
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yancey Instrument Location Yancey Co. Jail  
Instrument Serial No. 008653 Burnsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 20 day of March, 2018 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]  
Signature of Certifying Official

649  
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.



**Intox EC/IR-II: Subject Test**

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653  
Test Date: 03/20/2018

Citation Number: M0000000-0  
Subject's Name:  
PREVENTIVE, MAINTENANCE  
Subject's Date of Birth: 11/11/1911  
Subject's Sex: Male  
Driver's License State: XX  
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J  
Permit Number: 11304E  
Effective:  
05/01/2017-05/01/2019

Officer's Name: NONE,  
Type of Agency: FTA  
Agency: DHHS  
Test Type: Breath Test

Lot Number: AG716202  
Exp Date: 06/11/2019

Test	g/210L	Time
DIAG	Pass	12:40pm
AIR BLK	.00	12:41pm
ACCY CHK	.07	12:42pm
AIR BLK	.00	12:43pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:43pm</b>
AIR BLK	.00	12:44pm
<b>SUB TEST</b>	<b>.00</b>	<b>12:46pm</b>
AIR BLK	.00	12:47pm

Reported AC: .00 g/210L

\_\_\_\_\_  
Signature of Chemical Analyst

Court CVR

  
Analyst

**Intox EC/IR-II: Preventive Maintenance**

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653      Test Record Number: 1322  
Test Date: 03/20/2018      Test Time: 12:48pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:48pm
FLO	Pass	12:48pm
FC	Pass	12:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:48pm
SRC	Pass	12:48pm
DET	Pass	12:48pm
BAR	Pass	12:48pm
BT	Pass	12:48pm

Blank Tests

Test	Status	Time
AIR	Pass	12:49pm

Printer Tests


Test	Status	Time
PRNT	Pass	12:49pm

CRC Tests

Test	Status	Time
COMP	Pass	12:49pm
CAL	Pass	12:49pm

Preventive Maintenance

Status: Pass

  
\_\_\_\_\_  
Analyst