

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Alleghany Instrument Location Alleghany County Jail
Instrument Serial No. 008890 Sparta, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: 008890
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

09/01/2014-09/01/2016

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	3:02pm
AIR BLK	.00	3:03pm
ACCY CHK	.07	3:03pm
AIR BLK	.00	3:04pm
SUB TEST	.00	3:05pm
AIR BLK	.00	3:05pm
SUB TEST	.00	3:07pm
AIR BLK	.00	3:08pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALLEGHANY COUNTY ALLEGHANY CO JAIL 020

Serial Number: 008890 Test Record Number: 583
Test Date: 03/11/2016 Test Time: 3:09pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:09pm
FLO	Pass	3:09pm
FC	Pass	3:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:09pm
SRC	Pass	3:09pm
DET	Pass	3:09pm
BAR	Pass	3:09pm
BT	Pass	3:09pm

Blank Tests

Test	Status	Time
AIR	Pass	3:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:10pm

CRC Tests

Test	Status	Time
COMP	Pass	3:10pm
CAL	Pass	3:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Ashe Instrument Location Ashe County Jail
Instrument Serial No. 008849 Jefferson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

09/01/2014-09/01/2016

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG525303


Exp Date: 09/10/2017

Test	g/210L	Time
DIAG	Pass	1:40pm
AIR BLK	.00	1:40pm
ACCY CHK	.07	1:41pm
AIR BLK	.00	1:42pm
SUB TEST	.00	1:43pm
AIR BLK	.00	1:44pm
SUB TEST	.00	1:45pm
AIR BLK	.00	1:46pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ASHE COUNTY ASHE COUNTY JAIL 040

Serial Number: 008849 Test Record Number: 967
Test Date: 03/11/2016 Test Time: 1:52pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:53pm
FLO	Pass	1:53pm
FC	Pass	1:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:53pm
SRC	Pass	1:53pm
DET	Pass	1:53pm
BAR	Pass	1:53pm
BT	Pass	1:53pm

Blank Tests

Test	Status	Time
AIR	Pass	1:53pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:53pm

CRC Tests

Test	Status	Time
COMP	Pass	1:54pm
CAL	Pass	1:54pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Avery Instrument Location Avery Co. Jail
Instrument Serial No. 008664 Newland, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664
Test Date: 03/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:20am
AIR BLK	.00	10:21am
ACCY CHK	.07	10:22am
AIR BLK	.00	10:23am
SUB TEST	.00	10:24am
AIR BLK	.00	10:24am
SUB TEST	.00	10:26am
AIR BLK	.00	10:27am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

AVERY COUNTY AVERY COUNTY JAIL 050

Serial Number: 008664 Test Record Number: 789
Test Date: 03/09/2016 Test Time: 10:28am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:28am
FLO	Pass	10:28am
FC	Pass	10:28am

Temperature Tests

Test	Status	Time
FC1	Pass	10:28am
SRC	Pass	10:28am
DET	Pass	10:28am
BAR	Pass	10:28am
BT	Pass	10:28am

Blank Tests

Test	Status	Time
AIR	Pass	10:29am

Printer Tests

Test	Status	Time
PRNT	Pass	10:29am

CRC Tests

Test	Status	Time
COMP	Pass	10:29am
CAL	Pass	10:29am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. 008586 102 E. 2nd St, Washington, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. O.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008586
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:19am
AIR BLK	.00	10:20am
ACCY CHK	.08	10:21am
AIR BLK	.00	10:22am
SUB TEST	.00	10:22am
AIR BLK	.00	10:23am
SUB TEST	.00	10:25am
AIR BLK	.00	10:26am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008586 Test Record Number: 1196
Test Date: 03/18/2016 Test Time: 10:28am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:29am
FLO	Pass	10:29am
FC	Pass	10:29am

Temperature Tests

Test	Status	Time
FC1	Pass	10:29am
SRC	Pass	10:29am
DET	Pass	10:29am
BAR	Pass	10:29am
BT	Pass	10:29am

Blank Tests

Test	Status	Time
AIR	Pass	10:30am

Printer Tests

Test	Status	Time
PRNT	Pass	10:30am

CRC Tests

Test	Status	Time
COMP	Pass	10:30am
CAL	Pass	10:30am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Beaufort Instrument Location Beaufort Co. Courthouse

Instrument Serial No. DD8909 102 E. 2nd St., Washington, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. P.
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008909
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	10:04am
AIR BLK	.00	10:05am
ACCY CHK	.08	10:06am
AIR BLK	.00	10:07am
SUB TEST	.00	10:08am
AIR BLK	.00	10:09am
SUB TEST	.00	10:11am
AIR BLK	.00	10:11am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY COURTHOUSE 060

Serial Number: 008909 Test Record Number: 2416
Test Date: 03/18/2016 Test Time: 10:13am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

Temperature Tests

Test	Status	Time
FC1	Pass	10:13am
SRC	Pass	10:13am
DET	Pass	10:13am
BAR	Pass	10:13am
BT	Pass	10:13am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am

Printer Tests


Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Bertie Instrument Location Bertie Co. S.O.

Instrument Serial No. 008897 104 Dundee St., Windsor, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BERTIE COUNTY BERTIE CO SO 070

Serial Number: 008897

Test Date: 03/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG525303

Exp Date: 09/10/2017

Test	g/210L	Time
DIAG	Pass	9:32am
AIR BLK	.00	9:32am
ACCY CHK	.08	9:33am
AIR BLK	.00	9:34am
SUB TEST	.00	9:35am
AIR BLK	.00	9:36am
SUB TEST	.00	9:38am
AIR BLK	.00	9:39am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BERTIE COUNTY BERTIE CO SO 070

Serial Number: 008897 Test Record Number: 984
Test Date: 03/08/2016 Test Time: 9:41am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:41am
FLO	Pass	9:41am
FC	Pass	9:41am

Temperature Tests

Test	Status	Time
FC1	Pass	9:41am
SRC	Pass	9:41am
DET	Pass	9:41am
BAR	Pass	9:41am
BT	Pass	9:41am

Blank Tests

Test	Status	Time
AIR	Pass	9:42am


Printer Tests

Test	Status	Time
PRNT	Pass	9:42am

CRC Tests

Test	Status	Time
COMP	Pass	9:42am
CAL	Pass	9:42am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Bladen Instrument Location Bladen County

Instrument Serial No. 008894 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



B.C. Rhoden 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008894
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534902
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:35am
AIR BLK	.00	10:36am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
SUB TEST	.00	10:39am
AIR BLK	.00	10:40am
SUB TEST	.00	10:42am
AIR BLK	.00	10:43am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008894 Test Record Number: 766
Test Date: 03/16/2016 Test Time: 10:44am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:45am
FLO	Pass	10:45am
FC	Pass	10:45am

Temperature Tests

Test	Status	Time
FC1	Pass	10:45am
SRC	Pass	10:45am
DET	Pass	10:45am
BAR	Pass	10:45am
BT	Pass	10:45am

Blank Tests

Test	Status	Time
AIR	Pass	10:46am


Printer Tests

Test	Status	Time
PRNT	Pass	10:46am

CRC Tests

Test	Status	Time
COMP	Pass	10:46am
CAL	Pass	10:46am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Bladen Instrument Location Bladen County
Instrument Serial No. 008818 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature] 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008818

Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534902

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:09am
AIR BLK	.00	10:10am
ACCY CHK	.07	10:11am
AIR BLK	.00	10:12am
SUB TEST	.00	10:12am
AIR BLK	.00	10:13am
SUB TEST	.00	10:15am
AIR BLK	.00	10:16am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

BLADEN COUNTY BLADEN COUNTY SD 080

Serial Number: 008818 Test Record Number: 1205
Test Date: 03/16/2016 Test Time: 10:16am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:17am
FLO	Pass	10:17am
FC	Pass	10:17am

Temperature Tests

Test	Status	Time
FC1	Pass	10:17am
SRC	Pass	10:17am
DET	Pass	10:17am
BAR	Pass	10:17am
BT	Pass	10:17am

Blank Tests

Test	Status	Time
AIR	Pass	10:17am


Printer Tests

Test	Status	Time
PRNT	Pass	10:17am

CRC Tests

Test	Status	Time
COMP	Pass	10:18am
CAL	Pass	10:18am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Brunswick Instrument Location Brunswick

Instrument Serial No. 008602 County Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. [Signature] 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008602
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG425303
Exp Date: 09/10/2016

Test	g/210L	Time
DIAG	Pass	3:43pm
AIR BLK	.00	3:44pm
ACCY CHK	.08	3:44pm
AIR BLK	.00	3:46pm
SUB TEST	.00	3:47pm
AIR BLK	.00	3:48pm
SUB TEST	.00	3:49pm
AIR BLK	.00	3:50pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008602 Test Record Number: 3467
Test Date: 03/15/2016 Test Time: 3:53pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:54pm
FLO	Pass	3:54pm
FC	Pass	3:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:54pm
SRC	Pass	3:54pm
DET	Pass	3:54pm
BAR	Pass	3:54pm
BT	Pass	3:54pm

Blank Tests

Test	Status	Time
AIR	Pass	3:54pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:54pm

CRC Tests

Test	Status	Time
COMP	Pass	3:55pm
CAL	Pass	3:55pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location Brunswick County
Instrument Serial No. 008585 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Proctor 681
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008585
Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	3:42pm
AIR BLK	.00	3:43pm
ACCY CHK	.08	3:43pm
AIR BLK	.00	3:44pm
SUB TEST	.00	3:45pm
AIR BLK	.00	3:46pm
SUB TEST	.00	3:48pm
AIR BLK	.00	3:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BRUNSWICK CO SD 090

Serial Number: 008585 Test Record Number: 3398
Test Date: 03/15/2016 Test Time: 3:50pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:51pm
FLO	Pass	3:51pm
FC	Pass	3:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:51pm
SRC	Pass	3:51pm
DET	Pass	3:51pm
BAR	Pass	3:51pm
BT	Pass	3:51pm

Blank Tests

Test	Status	Time
AIR	Pass	3:51pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:51pm

CRC Tests

Test	Status	Time
COMP	Pass	3:52pm
CAL	Pass	3:52pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County BRUNSWICK Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008616 LELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008616
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:28pm
AIR BLK	.00	11:29pm
ACCY CHK	.08	11:29pm
AIR BLK	.00	11:30pm
SUB TEST	.00	11:31pm
AIR BLK	.00	11:32pm
SUB TEST	.00	11:33pm
AIR BLK	.00	11:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008616 Test Record Number: 2185

Test Date: 03/17/2016 Test Time: 11:39pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:39pm
FLO	Pass	11:39pm
FC	Pass	11:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:39pm
SRC	Pass	11:39pm
DET	Pass	11:39pm
BAR	Pass	11:39pm
BT	Pass	11:39pm

Blank Tests

Test	Status	Time
AIR	Pass	11:40pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:40pm

CRC Tests

Test	Status	Time
COMP	Pass	11:40pm
CAL	Pass	11:40pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County BRUNSWICK Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008647 LELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bennis

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008647
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	11:25pm
AIR BLK	.00	11:26pm
ACCY CHK	.07	11:26pm
AIR BLK	.00	11:27pm
SUB TEST	.00	11:28pm
AIR BLK	.00	11:29pm
SUB TEST	.00	11:31pm
AIR BLK	.00	11:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008647 Test Record Number: 2192
Test Date: 03/17/2016 Test Time: 11:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:37pm
FLO	Pass	11:37pm
FC	Pass	11:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:37pm
SRC	Pass	11:37pm
DET	Pass	11:37pm
BAR	Pass	11:37pm
BT	Pass	11:37pm

Blank Tests

Test	Status	Time
AIR	Pass	11:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:38pm

CRC Tests

Test	Status	Time
COMP	Pass	11:38pm
CAL	Pass	11:38pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Brunswick Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008707 LELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan Ray Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008707
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:28pm
AIR BLK	.00	11:29pm
ACCY CHK	.08	11:29pm
AIR BLK	.00	11:30pm
SUB TEST	.00	11:31pm
AIR BLK	.00	11:32pm
SUB TEST	.00	11:33pm
AIR BLK	.00	11:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008707 Test Record Number: 2279
Test Date: 03/17/2016 Test Time: 11:35pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35pm
FLO	Pass	11:35pm
FC	Pass	11:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:36pm
SRC	Pass	11:36pm
DET	Pass	11:36pm
BAR	Pass	11:36pm
BT	Pass	11:36pm

Blank Tests

Test	Status	Time
AIR	Pass	11:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:36pm

CRC Tests

Test	Status	Time
COMP	Pass	11:36pm
CAL	Pass	11:36pm

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location BAT MOBILE UNIT 9

Instrument Serial No. 008595 HELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin R. Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008575

Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:35pm
AIR BLK	.00	11:36pm
ACCY CHK	.08	11:37pm
AIR BLK	.00	11:37pm
SUB TEST	.00	11:38pm
AIR BLK	.00	11:39pm
SUB TEST	.00	11:41pm
AIR BLK	.00	11:42pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008575 Test Record Number: 897
Test Date: 03/17/2016 Test Time: 11:42pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:43pm
FLO	Pass	11:43pm
FC	Pass	11:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:43pm
SRC	Pass	11:43pm
DET	Pass	11:43pm
BAR	Pass	11:43pm
BT	Pass	11:43pm

Blank Tests

Test	Status	Time
AIR	Pass	11:44pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:44pm

CRC Tests

Test	Status	Time
COMP	Pass	11:44pm
CAL	Pass	11:44pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County BRUNSWICK

Instrument Location BAT MOBILE UNIT 9

Instrument Serial No. 008826

LELAND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex R. Berry

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008826
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:33pm
AIR BLK	.00	11:34pm
ACCY CHK	.08	11:34pm
AIR BLK	.00	11:35pm
SUB TEST	.00	11:36pm
AIR BLK	.00	11:36pm
SUB TEST	.00	11:38pm
AIR BLK	.00	11:39pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY BAT MOBILE UNIT 9 090

Serial Number: 008826 Test Record Number: 7899
Test Date: 03/17/2016 Test Time: 11:39pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:40pm
FLO	Pass	11:40pm
FC	Pass	11:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:40pm
SRC	Pass	11:40pm
DET	Pass	11:40pm
BAR	Pass	11:40pm
BT	Pass	11:40pm

Blank Tests

Test	Status	Time
AIR	Pass	11:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:41pm

CRC Tests

Test	Status	Time
COMP	Pass	11:41pm
CAL	Pass	11:41pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Brunswick Instrument Location Sunset Beach

Instrument Serial No. 008874 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature] 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY SUNSET BEACH PD 090

Serial Number: 008874

Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	5:49pm
AIR BLK	.00	5:50pm
ACCY CHK	.07	5:50pm
AIR BLK	.00	5:51pm
SUB TEST	.00	5:52pm
AIR BLK	.00	5:53pm
SUB TEST	.00	5:54pm
AIR BLK	.00	5:55pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY SUNSET BEACH PD 090

Serial Number: 008874 Test Record Number: 503
Test Date: 03/15/2016 Test Time: 5:56pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	5:56pm
FLO	Pass	5:56pm
EC	Pass	5:56pm

Temperature Tests

Test	Status	Time
PC1	Pass	5:56pm
SRC	Pass	5:56pm
DET	Pass	5:56pm
BAR	Pass	5:56pm
BT	Pass	5:56pm

Blank Tests

Test	Status	Time
AIR	Pass	5:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:57pm

CRC Tests

Test	Status	Time
COMP	Pass	5:57pm
CAL	Pass	5:57pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Brunswick Instrument Location OK Island
Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: 008648
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	4:34pm
AIR BLK	.00	4:34pm
ACCY CHK	.07	4:35pm
AIR BLK	.00	4:36pm
SUB TEST	.00	4:37pm
AIR BLK	.00	4:38pm
SUB TEST	.00	4:39pm
AIR BLK	.00	4:40pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

BRUNSWICK COUNTY OAK ISLAND PD 090

Serial Number: 008648 Test Record Number: 1393
Test Date: 03/15/2016 Test Time: 4:41pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:41pm
FLO	Pass	4:41pm
FC	Pass	4:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:41pm
SRC	Pass	4:41pm
DET	Pass	4:41pm
BAR	Pass	4:41pm
BT	Pass	4:41pm

Blank Tests

Test	Status	Time
AIR	Pass	4:42pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:42pm

CRC Tests

Test	Status	Time
COMP	Pass	4:42pm
CAL	Pass	4:42pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Buncombe Instrument Location Buncombe Co Jail
Instrument Serial No. 008631 Asheville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.




Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL
100

Serial Number: 008631
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J
Permit Number: 11304E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:21pm
AIR BLK	.00	12:22pm
ACCY CHK	.07	12:23pm
AIR BLK	.00	12:24pm
SUB TEST	.00	12:25pm
AIR BLK	.00	12:26pm
SUB TEST	.00	12:27pm
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE COUNTY BUNCOMBE COUNTY JAIL 100

Serial Number: 008631 Test Record Number: 4374
Test Date: 03/18/2016 Test Time: 12:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:30pm
FLO	Pass	12:30pm
FC	Pass	12:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:30pm
SRC	Pass	12:30pm
DET	Pass	12:30pm
BAR	Pass	12:30pm
BT	Pass	12:30pm

Blank Tests

Test	Status	Time
AIR	Pass	12:31pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:31pm

CRC Tests

Test	Status	Time
COMP	Pass	12:31pm
CAL	Pass	12:31pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Buncombe Instrument Location Bad mobile Unit 11

Instrument Serial No. 008970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008970
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG425303

Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	8:10pm
AIR BLK	.00	8:11pm
ACCY CHK	.08	8:11pm
AIR BLK	.00	8:12pm
SUB TEST	.00	8:13pm
AIR BLK	.00	8:14pm
SUB TEST	.00	8:15pm
AIR BLK	.00	8:16pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008970 Test Record Number: 127
Test Date: 03/17/2016 Test Time: 8:17pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:17pm
FLO	Pass	8:17pm
FC	Pass	8:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:18pm
SRC	Pass	8:18pm
DET	Pass	8:18pm
BAR	Pass	8:18pm
BT	Pass	8:18pm

Blank Tests

Test	Status	Time
AIR	Pass	8:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	8:18pm

CRC Tests

Test	Status	Time
COMP	Pass	8:18pm
CAL	Pass	8:18pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Buncombe Instrument Location Bot mobile Unit 11
Instrument Serial No. 008090

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chl Vdo
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008090
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V
Permit Number: 26632E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	8:12pm
AIR BLK	.00	8:13pm
ACCY CHK	.08	8:13pm
AIR BLK	.00	8:14pm
SUB TEST	.00	8:15pm
AIR BLK	.00	8:15pm
SUB TEST	.00	8:17pm
AIR BLK	.00	8:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008090 Test Record Number: 53
Test Date: 03/17/2016 Test Time: 8:19pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:19pm
FLO	Pass	8:19pm
FC	Pass	8:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:19pm
SRC	Pass	8:19pm
DET	Pass	8:19pm
BAR	Pass	8:19pm
BT	Pass	8:19pm

Blank Tests

Test	Status	Time
AIR	Pass	8:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	8:20pm

CRC Tests

Test	Status	Time
COMP	Pass	8:20pm
CAL	Pass	8:20pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Burke Instrument Location Bat mobile Unit 11

Instrument Serial No. 008090

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 26 day of March, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE BAT MOBILE UNIT 11 110

Serial Number: 008090
Test Date: 03/26/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: INGLE, LARRY W
Permit Number: 7281E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	6:11pm
AIR BLK	.00	6:12pm
ACCY CHK	.08	6:13pm
AIR BLK	.00	6:14pm
SUB TEST	.00	6:14pm
AIR BLK	.00	6:15pm
SUB TEST	.00	6:17pm
AIR BLK	.00	6:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BURKE BAT MOBILE UNIT 11 110

Serial Number: 008090 Test Record Number: 68
Test Date: 03/26/2016 Test Time: 6:20pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:20pm
FLO	Pass	6:20pm
FC	Pass	6:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:21pm
SRC	Pass	6:21pm
DET	Pass	6:21pm
BAR	Pass	6:21pm
BT	Pass	6:21pm

Blank Tests

Test	Status	Time
AIR	Pass	6:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:21pm

CRC Tests

Test	Status	Time
COMP	Pass	6:21pm
CAL	Pass	6:21pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Burke Instrument Location Bat Mobile Unit 11

Instrument Serial No. 008970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24 day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE BAT MOBILE UNIT 11 110

Serial Number: 008970

Test Date: 03/26/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG425303

Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	4:27pm
AIR BLK	.00	4:29pm
ACCY CHK	.08	4:30pm
AIR BLK	.00	4:30pm
SUB TEST	.00	4:31pm
AIR BLK	.00	4:32pm
SUB TEST	.00	4:33pm
AIR BLK	.00	4:34pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BURKE BAT MOBILE UNIT 11 110

Serial Number: 008970 Test Record Number: 141
Test Date: 03/26/2016 Test Time: 4:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:37pm
FLO	Pass	4:37pm
FC	Pass	4:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:37pm
SRC	Pass	4:37pm
DET	Pass	4:37pm
BAR	Pass	4:37pm
BT	Pass	4:37pm

Blank Tests

Test	Status	Time
AIR	Pass	4:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:37pm

CRC Tests

Test	Status	Time
COMP	Pass	4:38pm
CAL	Pass	4:38pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CABARRUS Instrument Location BAT MOBILE 7

Instrument Serial No. 008968

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CABARRUS BAT MOBILE UNIT 7 120

Serial Number: 008968

Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 7281E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG425303

Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	10:05pm
AIR BLK	.00	10:06pm
ACCY CHK	.07	10:06pm
AIR BLK	.00	10:07pm
SUB TEST	.00	10:07pm
AIR BLK	.00	10:08pm
SUB TEST	.00	10:10pm
AIR BLK	.00	10:11pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CABARRUS BAT MOBILE UNIT 7 120

Serial Number: 008968 Test Record Number: 139
Test Date: 03/11/2016 Test Time: 10:12pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:12pm
FLO	Pass	10:12pm
FC	Pass	10:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:12pm
SRC	Pass	10:12pm
DET	Pass	10:12pm
BAR	Pass	10:12pm
BT	Pass	10:12pm

Blank Tests

Test	Status	Time
AIR	Pass	10:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:13pm

CRC Tests

Test	Status	Time
COMP	Pass	10:13pm
CAL	Pass	10:13pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cabarrus Instrument Location Cabarrus County SD
Instrument Serial No. 008590 30 Corban Ave, Concord

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. J. [Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590 Test Record Number: 2668
Test Date: 03/30/2016 Test Time: 11:43am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:44am
FLO	Pass	11:44am
FC	Pass	11:44am

Temperature Tests

Test	Status	Time
FC1	Pass	11:44am
SRC	Pass	11:44am
DET	Pass	11:44am
BAR	Pass	11:44am
BT	Pass	11:44am

Blank Tests

Test	Status	Time
AIR	Pass	11:45am


Printer Tests

Test	Status	Time
PRNT	Pass	11:45am

CRC Tests

Test	Status	Time
COMP	Pass	11:45am
CAL	Pass	11:45am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CABARRUS COUNTY CABARRUS COUNTY SD 120

Serial Number: 008590
Test Date: 03/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG414801

Exp Date: 05/28/2016


Test	g/210L	Time
DIAG	Pass	11:48am
AIR BLK	.00	11:48am
ACCY CHK	.07	11:49am
AIR BLK	.00	11:50am
SUB TEST	.00	11:51am
AIR BLK	.00	11:52am
SUB TEST	.00	11:53am
AIR BLK	.00	11:54am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Camden Instrument Location Camden Co. S.O.

Instrument Serial No. 008940 113 Hwy 343, Camden, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940
Test Date: 03/03/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017


Test	g/210L	Time
DIAG	Pass	11:53am
AIR BLK	.00	11:54am
ACCY CHK	.07	11:54am
AIR BLK	.00	11:55am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am
SUB TEST	.00	11:58am
AIR BLK	.00	11:59am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CAMDEN COUNTY CAMDEN CO SO 140

Serial Number: 008940 Test Record Number: 766
Test Date: 03/03/2016 Test Time: 12:00pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:01pm
SRC	Pass	12:01pm
DET	Pass	12:01pm
BAR	Pass	12:01pm
BT	Pass	12:01pm

Blank Tests

Test	Status	Time
AIR	Pass	12:01pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:01pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret Instrument Location Carteret County
Instrument Serial No. 008882 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008882
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	9:56am
AIR BLK	.00	9:57am
ACCY CHK	.08	9:57am
AIR BLK	.00	9:59am
SUB TEST	.00	9:59am
AIR BLK	.00	10:00am
SUB TEST	.00	10:01am
AIR BLK	.00	10:02am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008882 Test Record Number: 1392
Test Date: 03/16/2016 Test Time: 10:03am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:03am
FLO	Pass	10:03am
FC	Pass	10:03am

Temperature Tests

Test	Status	Time
FC1	Pass	10:03am
SRC	Pass	10:03am
DET	Pass	10:03am
BAR	Pass	10:03am
BT	Pass	10:03am

Blank Tests

Test	Status	Time
AIR	Pass	10:04am

Printer Tests

Test	Status	Time
PRNT	Pass	10:04am

CRC Tests

Test	Status	Time
COMP	Pass	10:04am
CAL	Pass	10:04am

Preventive Maintenance
Status: Pass

Randy E Hall

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Carteret Instrument Location Carteret County
Instrument Serial No. 008605 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008605
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HALL, RANDY E
Permit Number: 3462E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017


Test	g/210L	Time
DIAG	Pass	9:56am
AIR BLK	.00	9:57am
ACCY CHK	.08	9:58am
AIR BLK	.00	9:59am
SUB TEST	.00	9:59am
AIR BLK	.00	10:00am
SUB TEST	.00	10:02am
AIR BLK	.00	10:03am

Reported AC: 1.00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY CARTERET COUNTY SD 150

Serial Number: 008605 Test Record Number: 3428
Test Date: 03/16/2016 Test Time: 10:06am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:06am
FLO	Pass	10:06am
FC	Pass	10:06am

Temperature Tests

Test	Status	Time
FC1	Pass	10:06am
SRC	Pass	10:06am
DET	Pass	10:06am
BAR	Pass	10:06am
BT	Pass	10:06am

Blank Tests

Test	Status	Time
AIR	Pass	10:07am


Printer Tests

Test	Status	Time
PRNT	Pass	10:07am

CRC Tests

Test	Status	Time
COMP	Pass	10:07am
CAL	Pass	10:07am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret


Instrument Location Morehead City PD

Instrument Serial No. 008731

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: ,
Permit Number:
Effective:
00/00/0000-00/00/0000

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:42am
AIR BLK	.00	10:42am
ACCY CHK	.08	10:43am
AIR BLK	.00	10:44am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:47am
AIR BLK	.00	10:48am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY MOREHEAD CITY PD 150

Serial Number: 008731 Test Record Number: 1759
Test Date: 03/16/2016 Test Time: 10:48am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:48am
FLO	Pass	10:48am
FC	Pass	10:49am

Temperature Tests

Test	Status	Time
FC1	Pass	10:49am
SRC	Pass	10:49am
DET	Pass	10:49am
BAR	Pass	10:49am
BT	Pass	10:49am

Blank Tests

Test	Status	Time
AIR	Pass	10:49am

Printer Tests

Test	Status	Time
PRNT	Pass	10:49am

CRC Tests

Test	Status	Time
COMP	Pass	10:49am
CAL	Pass	10:49am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Carteret


Instrument Location ATLANTIC BEACH AD

Instrument Serial No. 008785

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: 008785
Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:12am
AIR BLK	.00	11:12am
ACCY CHK	.07	11:13am
AIR BLK	.00	11:14am
SUB TEST	.00	11:14am
AIR BLK	.00	11:15am
SUB TEST	.00	11:17am
AIR BLK	.00	11:17am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY ATLANTIC BEACH PD 150

Serial Number: 008785 Test Record Number: 862
Test Date: 03/16/2016 Test Time: 11:18am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:18am
FLO	Pass	11:18am
FC	Pass	11:19am

Temperature Tests

Test	Status	Time
FC1	Pass	11:19am
SRC	Pass	11:19am
DET	Pass	11:19am
BAR	Pass	11:19am
BT	Pass	11:19am

Blank Tests

Test	Status	Time
AIR	Pass	11:19am

Printer Tests

Test	Status	Time
PRNT	Pass	11:19am

CRC Tests

Test	Status	Time
COMP	Pass	11:20am
CAL	Pass	11:20am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Carteret Instrument Location EMERALD ISLE PD

Instrument Serial No. 008620

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CARTERET COUNTY EMERALD ISLE PD 150

Serial Number: 008620
Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	12:19pm
AIR BLK	.00	12:19pm
ACCY CHK	.08	12:20pm
AIR BLK	.00	12:21pm
SUB TEST	.00	12:22pm
AIR BLK	.00	12:23pm
SUB TEST	.00	12:24pm
AIR BLK	.00	12:25pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CARTERET COUNTY EMERALD ISLE PD 150

Serial Number: 008620 Test Record Number: 1774
Test Date: 03/16/2016 Test Time: 12:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:26pm
FLO	Pass	12:26pm
FC	Pass	12:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:26pm
SRC	Pass	12:26pm
DET	Pass	12:26pm
BAR	Pass	12:26pm
BT	Pass	12:26pm

Blank Tests

Test	Status	Time
AIR	Pass	12:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:27pm

CRC Tests

Test	Status	Time
COMP	Pass	12:27pm
CAL	Pass	12:27pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County CASWELL Instrument Location CASWELL CO. DETENTION STR

Instrument Serial No. 008593 231 COUNTY PARK RD YANCEYVILLE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Brian D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CASWELL COUNTY DETENTION CENTER 160

Serial Number: 008593
Test Date: 03/09/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	1:46pm
AIR BLK	.00	1:47pm
ACCY CHK	.08	1:48pm
AIR BLK	.00	1:49pm
SUB TEST	.00	1:49pm
AIR BLK	.00	1:50pm
SUB TEST	.00	1:52pm
AIR BLK	.00	1:53pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CASWELL COUNTY DETENTION CENTER 160

Serial Number: 008593 Test Record Number: 1308
Test Date: 03/09/2016 Test Time: 1:55pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:55pm
FLO	Pass	1:55pm
FC	Pass	1:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:55pm
SRC	Pass	1:55pm
DET	Pass	1:55pm
BAR	Pass	1:55pm
BT	Pass	1:55pm

Blank Tests

Test	Status	Time
AIR	Pass	1:56pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:56pm

CRC Tests

Test	Status	Time
COMP	Pass	1:56pm
CAL	Pass	1:56pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Catawba County SD
Instrument Serial No. 008821 100 B Southwest Blvd, Newton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. J. [Signature]
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821 Test Record Number: 1588
Test Date: 03/21/2016 Test Time: 10:12am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

Temperature Tests

Test	Status	Time
FC1	Pass	10:13am
SRC	Pass	10:13am
DET	Pass	10:13am
BAR	Pass	10:13am
BT	Pass	10:13am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am


Printer Tests

Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821
Test Date: 03/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:17am
AIR BLK	.00	10:17am
ACCY CHK	.07	10:18am
AIR BLK	.00	10:19am
SUB TEST	.00	10:20am
AIR BLK	.00	10:20am
SUB TEST	.00	10:22am
AIR BLK	.00	10:23am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Hickory PD.
Instrument Serial No. 008841 347 2nd Ave. S.W. Hickory

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841 Test Record Number: 1645
Test Date: 03/21/2016 Test Time: 11:40am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:41am
FLO	Pass	11:41am
FC	Pass	11:41am

Temperature Tests

Test	Status	Time
FC1	Pass	11:41am
SRC	Pass	11:41am
DET	Pass	11:41am
BAR	Pass	11:41am
BT	Pass	11:41am

Blank Tests

Test	Status	Time
AIR	Pass	11:42am


Printer Tests

Test	Status	Time
PRNT	Pass	11:42am

CRC Tests

Test	Status	Time
COMP	Pass	11:42am
CAL	Pass	11:42am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841
Test Date: 03/21/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	11:47am
AIR BLK	.00	11:47am
ACCY CHK	.08	11:48am
AIR BLK	.00	11:50am
SUB TEST	.00	11:50am
AIR BLK	.00	11:51am
SUB TEST	.00	11:52am
AIR BLK	.00	11:53am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cleveland Instrument Location Cleveland County S D-Anney
Instrument Serial No. 008893 407 McBrayer St., Shelby

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of April, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mark D. King
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: 008893 Test Record Number: 1425
Test Date: 04/11/2016 Test Time: 10:25am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:26am
FLO	Pass	10:26am
FC	Pass	10:26am

Temperature Tests

Test	Status	Time
FC1	Pass	10:26am
SRC	Pass	10:26am
DET	Pass	10:26am
BAR	Pass	10:26am
BT	Pass	10:26am

Blank Tests

Test	Status	Time
AIR	Pass	10:26am


Printer Tests

Test	Status	Time
PRNT	Pass	10:26am

CRC Tests

Test	Status	Time
COMP	Pass	10:27am
CAL	Pass	10:27am

Preventive Maintenance
Status: *Pass*



Analyst

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SD-ANNEX
220

Serial Number: 008893
Test Date: 04/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:30am
AIR BLK	.00	10:30am
ACCY CHK	.08	10:31am
AIR BLK	.00	10:32am
SUB TEST	.00	10:33am
AIR BLK	.00	10:34am
SUB TEST	.00	10:36am
AIR BLK	.00	10:37am

Reported AC: 00 g/210L



Signature of Chemical Analyst
Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Columbus Instrument Location Columbus County
Instrument Serial No. 008875 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Phelan 681
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008875
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: .5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	8:34am
AIR BLK	.00	8:35am
ACCY CHK	.07	8:35am
AIR BLK	.00	8:36am
SUB TEST	.00	8:37am
AIR BLK	.00	8:38am
SUB TEST	.00	8:39am
AIR BLK	.00	8:40am

Reported AC: 00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008875 Test Record Number: 1712
Test Date: 03/16/2016 Test Time: 8:41am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	8:41am
FLO	Pass	8:41am
FC	Pass	8:41am

Temperature Tests

Test	Status	Time
FC1	Pass	8:42am
SRC	Pass	8:42am
DET	Pass	8:42am
BAR	Pass	8:42am
BT	Pass	8:42am

Blank Tests

Test	Status	Time
AIR	Pass	8:42am


Printer Tests

Test	Status	Time
PRNT	Pass	8:42am

CRC Tests

Test	Status	Time
COMP	Pass	8:42am
CAL	Pass	8:42am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Columbus Instrument Location Columbus County
Instrument Serial No. 008886 Sperritt Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Phelan
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008886
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG534902
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	8:35am
AIR BLK	.00	8:36am
ACCY CHK	.08	8:36am
AIR BLK	.00	8:37am
SUB TEST	.00	8:38am
AIR BLK	.00	8:39am
SUB TEST	.00	8:40am
AIR BLK	.00	8:41am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

COLUMBUS COUNTY COLUMBUS COUNTY SD 230

Serial Number: 008886 Test Record Number: 1167
Test Date: 03/16/2016 Test Time: 8:42am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:42am
FLO	Pass	8:42am
FC	Pass	8:42am

Temperature Tests

Test	Status	Time
FC1	Pass	8:42am
SRC	Pass	8:42am
DET	Pass	8:42am
BAR	Pass	8:42am
BT	Pass	8:42am

Blank Tests

Test	Status	Time
AIR	Pass	8:43am


Printer Tests

Test	Status	Time
PRNT	Pass	8:43am

CRC Tests

Test	Status	Time
COMP	Pass	8:43am
CAL	Pass	8:43am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County CHATHAM Instrument Location CHATHAM Co Jail

Instrument Serial No. 008591 627 RENAISSANCE DR, PITTSBORO NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHATHAM COUNTY JAIL 180

Serial Number: 008591

Test Date: 03/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	5:37pm
AIR BLK	.00	5:38pm
ACCY CHK	.07	5:39pm
AIR BLK	.00	5:40pm
SUB TEST	.00	5:41pm
AIR BLK	.00	5:42pm
SUB TEST	.00	5:43pm
AIR BLK	.00	5:44pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CHATHAM COUNTY JAIL 180

Serial Number: 008591 Test Record Number: 1678
Test Date: 03/18/2016 Test Time: 5:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:46pm
FLO	Pass	5:46pm
FC	Pass	5:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:47pm
SRC	Pass	5:47pm
DET	Pass	5:47pm
BAR	Pass	5:47pm
BT	Pass	5:47pm

Blank Tests

Test	Status	Time
AIR	Pass	5:47pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:47pm

CRC Tests

Test	Status	Time
COMP	Pass	5:47pm
CAL	Pass	5:47pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chowan Instrument Location Chowan Co. Public Safety Ctr.
Instrument Serial No. 008895 305 W. Freemason St., Edenton, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Janet A. Reese
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895

Test Date: 03/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG425303

Exp Date: 09/10/2016


Test	g/210L	Time
DIAG	Pass	11:25am
AIR BLK	.00	11:26am
ACCY CHK	.07	11:27am
AIR BLK	.00	11:28am
SUB TEST	.00	11:28am
AIR BLK	.00	11:29am
SUB TEST	.00	11:31am
AIR BLK	.00	11:32am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895 Test Record Number: 702
Test Date: 03/02/2016 Test Time: 11:34am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:34am

Temperature Tests

Test	Status	Time
FC1	Pass	11:34am
SRC	Pass	11:34am
DET	Pass	11:34am
BAR	Pass	11:34am
BT	Pass	11:34am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am


Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008616 NEW BERN, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 9 240

Serial Number: 008616
Test Date: 03/25/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:45pm
AIR BLK	.00	11:46pm
ACCY CHK	.08	11:46pm
AIR BLK	.00	11:47pm
SUB TEST	.00	11:48pm
AIR BLK	.00	11:49pm
SUB TEST	.00	11:50pm
AIR BLK	.00	11:51pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 9 240

Serial Number: 008616 Test Record Number: 2192
Test Date: 03/25/2016 Test Time: 11:56pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:57pm
FLO	Pass	11:57pm
FC	Pass	11:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:57pm
SRC	Pass	11:57pm
DET	Pass	11:57pm
BAR	Pass	11:57pm
BT	Pass	11:57pm

Blank Tests

Test	Status	Time
AIR	Pass	11:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:57pm

CRC Tests

Test	Status	Time
COMP	Pass	11:58pm
CAL	Pass	11:58pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 0081647 NEW BERN, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Barnes

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY BAT MOBILE UNIT 9 240

Serial Number: 008647

Test Date: 03/25/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434201


Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	10:01pm
AIR BLK	.00	10:03pm
ACCY CHK	.07	10:03pm
AIR BLK	.00	10:04pm
SUB TEST	.00	10:05pm
AIR BLK	.00	10:05pm
SUB TEST	.00	10:07pm
AIR BLK	.00	10:08pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY BAT MOBILE UNIT 9 240

Serial Number: 008647 Test Record Number: 2200
Test Date: 03/25/2016 Test Time: 10:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:09pm
FLO	Pass	10:09pm
FC	Pass	10:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:09pm
SRC	Pass	10:09pm
DET	Pass	10:09pm
BAR	Pass	10:09pm
BT	Pass	10:09pm

Blank Tests

Test	Status	Time
AIR	Pass	10:10pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:10pm

CRC Tests

Test	Status	Time
COMP	Pass	10:10pm
CAL	Pass	10:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County CRAVEN Instrument Location HAVELock PD

Instrument Serial No. 008800

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800

Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	1:26pm
AIR BLK	.00	1:26pm
ACCY CHK	.08	1:27pm
AIR BLK	.00	1:28pm
SUB TEST	.00	1:28pm
AIR BLK	.00	1:29pm
SUB TEST	.00	1:31pm
AIR BLK	.00	1:32pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall
Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY HAVELOCK PD 240

Serial Number: 008800 Test Record Number: 1019
Test Date: 03/16/2016 Test Time: 1:32pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:32pm
FLO	Pass	1:32pm
FC	Pass	1:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:33pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:33pm

CRC Tests

Test	Status	Time
COMP	Pass	1:33pm
CAL	Pass	1:33pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN

Instrument Location METS CHERRY POINT

Instrument Serial No. 010819

AMO

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:02pm
AIR BLK	.00	2:03pm
ACCY CHK	.08	2:03pm
AIR BLK	.00	2:04pm
SUB TEST	.00	2:05pm
AIR BLK	.00	2:06pm
SUB TEST	.00	2:07pm
AIR BLK	.00	2:08pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY MCAS CHERRY POINT 240

Serial Number: 010819 Test Record Number: 450
Test Date: 03/16/2016 Test Time: 2:08pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:09pm
FLO	Pass	2:09pm
FC	Pass	2:09pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:09pm
SRC	Pass	2:09pm
DET	Pass	2:09pm
BAR	Pass	2:09pm
BT	Pass	2:09pm

Blank Tests

Test	Status	Time
AIR	Pass	2:10pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:10pm

CRC Tests

Test	Status	Time
COMP	Pass	2:10pm
CAL	Pass	2:10pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location New Bern AD
Instrument Serial No. 008817

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:25am
AIR BLK	.00	10:26am
ACCY CHK	.08	10:27am
AIR BLK	.00	10:28am
SUB TEST	.00	10:28am
AIR BLK	.00	10:29am
SUB TEST	.00	10:31am
AIR BLK	.00	10:32am

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY NEW BERN PD 240

Serial Number: 008817 Test Record Number: 1164
Test Date: 03/17/2016 Test Time: 10:32am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:33am
FLO	Pass	10:33am
FC	Pass	10:33am

Temperature Tests

Test	Status	Time
FC1	Pass	10:33am
SRC	Pass	10:33am
DET	Pass	10:33am
BAR	Pass	10:33am
BT	Pass	10:33am

Blank Tests

Test	Status	Time
AIR	Pass	10:34am


Printer Tests

Test	Status	Time
PRNT	Pass	10:34am

CRC Tests

Test	Status	Time
COMP	Pass	10:34am
CAL	Pass	10:34am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CRAVEN Instrument Location CRAVEN County
Instrument Serial No. 008732 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hill

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: 008732
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	11:13am
AIR BLK	.00	11:14am
ACCY CHK	.07	11:15am
AIR BLK	.00	11:16am
SUB TEST	.00	11:16am
AIR BLK	.00	11:18am
SUB TEST	.00	11:19am
AIR BLK	.00	11:20am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CRAVEN COUNTY CRAVEN COUNTY SD 240

Serial Number: 008732 Test Record Number: 1612
Test Date: 03/17/2016 Test Time: 11:20am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:21am
FLO	Pass	11:21am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:21am
SRC	Pass	11:21am
DET	Pass	11:21am
BAR	Pass	11:21am
BT	Pass	11:21am

Blank Tests

Test	Status	Time
AIR	Pass	11:21am


Printer Tests

Test	Status	Time
PRNT	Pass	11:21am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location FT. BRAGG
Instrument Serial No. 008787 P.M.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND CO. FORT BRAGG LEC. 250

Serial Number: 008787
Test Date: 03/21/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	4:14pm
AIR BLK	.00	4:16pm
ACCY CHK	.07	4:16pm
AIR BLK	.00	4:17pm
SUB TEST	.00	4:18pm
AIR BLK	.00	4:19pm
SUB TEST	.00	4:21pm
AIR BLK	.00	4:22pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND CO. FORT BRAGG LEC. 250

Serial Number: 008787 Test Record Number: 566
Test Date: 03/21/2016 Test Time: 4:22pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:23pm
FLO	Pass	4:23pm
FC	Pass	4:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:23pm
SRC	Pass	4:23pm
DET	Pass	4:23pm
BAR	Pass	4:23pm
BT	Pass	4:23pm

Blank Tests

Test	Status	Time
AIR	Pass	4:24pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:24pm

CRC Tests

Test	Status	Time
COMP	Pass	4:24pm
CAL	Pass	4:24pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location FT. BRAGG
Instrument Serial No. 008903 P.M.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: 008903
Test Date: 03/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017


Test	g/210L	Time
DIAG	Pass	4:18pm
AIR BLK	.00	4:19pm
ACCY CHK	.08	4:20pm
AIR BLK	.00	4:21pm
SUB TEST	.00	4:22pm
AIR BLK	.00	4:23pm
SUB TEST	.00	4:24pm
AIR BLK	.00	4:25pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: 008903 Test Record Number: 1594
Test Date: 03/21/2016 Test Time: 4:42pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:43pm
FLO	Pass	4:43pm
FC	Pass	4:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:43pm
SRC	Pass	4:43pm
DET	Pass	4:43pm
BAR	Pass	4:43pm
BT	Pass	4:43pm

Blank Tests

Test	Status	Time
AIR	Pass	4:44pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:44pm

CRC Tests

Test	Status	Time
COMP	Pass	4:44pm
CAL	Pass	4:44pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Currituck Instrument Location Currituck Co. S.O.

Instrument Serial No. 008947 407-A Maple Rd., Maple, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CURRITUCK COUNTY CURRITUCK SO-MAPLE
260

Serial Number: 008947
Test Date: 03/03/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:11am
AIR BLK	.00	11:12am
ACCY CHK	.08	11:13am
AIR BLK	.00	11:14am
SUB TEST	.00	11:14am
AIR BLK	.00	11:15am
SUB TEST	.00	11:17am
AIR BLK	.00	11:18am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CURRITUCK COUNTY CURRITUCK SO-MAPLE 260

Serial Number: 008947 Test Record Number: 1834
Test Date: 03/03/2016 Test Time: 11:20am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:20am
FLO	Pass	11:21am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:21am
SRC	Pass	11:21am
DET	Pass	11:21am
BAR	Pass	11:21am
BT	Pass	11:21am

Blank Tests

Test	Status	Time
AIR	Pass	11:21am


Printer Tests

Test	Status	Time
PRNT	Pass	11:21am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Currituck Instrument Location Currituck Co. S.O. - Corolla

Instrument Serial No. 008949 1123 Ocean Trail, Corolla, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: 008949
Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: *GUARD, KELLY G*

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:26am
AIR BLK	.00	11:27am
ACCY CHK	.07	11:27am
AIR BLK	.00	11:28am
SUB TEST	.00	11:29am
AIR BLK	.00	11:30am
SUB TEST	.00	11:31am
AIR BLK	.00	11:32am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CURRITUCK COUNTY SO-COROLLA 260

Serial Number: 008949 Test Record Number: 394
Test Date: 03/28/2016 Test Time: 11:34am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35am
FLO	Pass	11:35am
FC	Pass	11:35am

Temperature Tests

Test	Status	Time
FC1	Pass	11:35am
SRC	Pass	11:35am
DET	Pass	11:35am
BAR	Pass	11:35am
BT	Pass	11:35am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am

Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:36am
CAL	Pass	11:36am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Kill Devil Hills P.D.

Instrument Serial No. 008844 102 Town Hall Dr., Kill Devil Hills, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844
Test Date: 03/28/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534802
Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	1:37pm
AIR BLK	.00	1:38pm
ACCY CHK	.07	1:38pm
AIR BLK	.00	1:39pm
SUB TEST	.00	1:40pm
AIR BLK	.00	1:41pm
SUB TEST	.00	1:42pm
AIR BLK	.00	1:43pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844 Test Record Number: 1749
Test Date: 03/28/2016 Test Time: 1:45pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:45pm
FLO	Pass	1:45pm
FC	Pass	1:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:45pm
SRC	Pass	1:45pm
DET	Pass	1:45pm
BAR	Pass	1:45pm
BT	Pass	1:45pm

Blank Tests

Test	Status	Time
AIR	Pass	1:46pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:46pm

CRC Tests

Test	Status	Time
COMP	Pass	1:46pm
CAL	Pass	1:46pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Wallace Police
Instrument Serial No. 008858 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Proctor
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANC
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534902
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	2:23pm
AIR BLK	.00	2:24pm
ACCY CHK	.07	2:24pm
AIR BLK	.00	2:25pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:28pm
AIR BLK	.00	2:29pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY WALLACE PD 300

Serial Number: 008858 Test Record Number: 747
Test Date: 03/16/2016 Test Time: 2:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:30pm
FLO	Pass	2:30pm
FC	Pass	2:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:31pm
SRC	Pass	2:31pm
DET	Pass	2:31pm
BAR	Pass	2:31pm
BT	Pass	2:31pm

Blank Tests

Test	Status	Time
AIR	Pass	2:31pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:31pm

CRC Tests

Test	Status	Time
COMP	Pass	2:31pm
CAL	Pass	2:31pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Duplin Instrument Location Duplin County
Instrument Serial No. 008844 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: 008864
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: *RHODES, KENNETH C*
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG425303
Exp Date: 09/10/2016

Test	g/210L	Time
DIAG	Pass	1:18pm
AIR BLK	.00	1:19pm
ACCY CHK	.08	1:19pm
AIR BLK	.00	1:20pm
SUB TEST	.00	1:21pm
AIR BLK	.00	1:22pm
SUB TEST	.00	1:24pm
AIR BLK	.00	1:25pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DUPLIN COUNTY DUPLIN CO SD 300

Serial Number: 008864 Test Record Number: 2781
Test Date: 03/16/2016 Test Time: 1:25pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:25pm
FLO	Pass	1:25pm
FC	Pass	1:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:26pm
SRC	Pass	1:26pm
DET	Pass	1:26pm
BAR	Pass	1:26pm
BT	Pass	1:26pm

Blank Tests

Test	Status	Time
AIR	Pass	1:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:26pm

CRC Tests

Test	Status	Time
COMP	Pass	1:26pm
CAL	Pass	1:26pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County FORSYTH Instrument Location BAT MOBILE 7
Instrument Serial No. 008968

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH BAT MOBILE UNIT 7 330

Serial Number: 008968
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W
Permit Number: 7281E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG425303
Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	9:18pm
AIR BLK	.00	9:19pm
ACCY CHK	.07	9:19pm
AIR BLK	.00	9:20pm
SUB TEST	.00	9:21pm
AIR BLK	.00	9:22pm
SUB TEST	.00	9:23pm
AIR BLK	.00	9:24pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

Intox EC/IR-II: Preventive Maintenance

FORSYTH BAT MOBILE UNIT 7 330

Serial Number: 008968 Test Record Number: 143
Test Date: 03/17/2016 Test Time: 9:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:26pm
FLO	Pass	9:26pm
FC	Pass	9:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:26pm
SRC	Pass	9:26pm
DET	Pass	9:26pm
BAR	Pass	9:26pm
BT	Pass	9:26pm

Blank Tests

Test	Status	Time
AIR	Pass	9:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:27pm

CRC Tests

Test	Status	Time
COMP	Pass	9:27pm
CAL	Pass	9:27pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County FORSYTH Instrument Location BAT MOBILE 7
Instrument Serial No. 008972

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

~~CABARRUS~~ ^{FORKSTN} BAT MOBILE UNIT 7 ~~128~~ 330

Serial Number: 008972
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: INGLE, LARRY W
Permit Number: 7281E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG425303
Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	9:25pm
AIR BLK	.00	9:26pm
ACCY CHK	.07	9:26pm
AIR BLK	.00	9:27pm
SUB TEST	.00	9:28pm
AIR BLK	.00	9:29pm
SUB TEST	.00	9:30pm
AIR BLK	.00	9:31pm

Reported AC: .00 g/210L



Signature of Chemical Analyst
Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

~~FOLYSON~~
~~CABARRUS~~ BAT MOBILE UNIT 7 ~~120~~ 330

Serial Number: 008972 Test Record Number: 186
Test Date: 03/17/2016 Test Time: 9:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:37pm
FLO	Pass	9:37pm
FC	Pass	9:37pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:37pm
SRC	Pass	9:37pm
DET	Pass	9:37pm
BAR	Pass	9:37pm
BT	Pass	9:37pm

Blank Tests

Test	Status	Time
AIR	Pass	9:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:37pm

CRC Tests

Test	Status	Time
COMP	Pass	9:38pm
CAL	Pass	9:38pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County FORSYTH Instrument Location BAT MOBILE 7
Instrument Serial No. 008971

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH BAT MOBILE UNIT 7 330

Serial Number: 008971
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: INGLE, LARRY W
Permit Number: 7281E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG425303
Exp Date: 10/10/2016

Test	g/210L	Time
DIAG	Pass	12:49am
AIR BLK	.00	12:49am
ACCY CHK	.08	12:50am
AIR BLK	.00	12:51am
SUB TEST	.00	12:51am
AIR BLK	.00	12:52am
SUB TEST	.00	12:54am
AIR BLK	.00	12:54am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH BAT MOBILE UNIT 7 330

Serial Number: 008971 Test Record Number: 98
Test Date: 03/18/2016 Test Time: 12:56am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:56am
FLO	Pass	12:56am
FC	Pass	12:56am

Temperature Tests

Test	Status	Time
FC1	Pass	12:56am
SRC	Pass	12:56am
DET	Pass	12:56am
BAR	Pass	12:56am
BT	Pass	12:56am

Blank Tests

Test	Status	Time
AIR	Pass	12:57am

Printer Tests

Test	Status	Time
PRNT	Pass	12:57am

CRC Tests

Test	Status	Time
COMP	Pass	12:57am
CAL	Pass	12:57am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLIN CO. JAIL

Instrument Serial No. 008933 285 T KEMP RD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:27pm
AIR BLK	.00	2:28pm
ACCY CHK	.07	2:29pm
AIR BLK	.00	2:30pm
SUB TEST	.00	2:31pm
AIR BLK	.00	2:32pm
SUB TEST	.00	2:33pm
AIR BLK	.00	2:34pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008933 Test Record Number: 764
Test Date: 03/17/2016 Test Time: 2:35pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:36pm
FLO	Pass	2:36pm
FC	Pass	2:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:36pm
SRC	Pass	2:36pm
DET	Pass	2:36pm
BAR	Pass	2:36pm
BT	Pass	2:36pm

Blank Tests

Test	Status	Time
AIR	Pass	2:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:37pm

CRC Tests

Test	Status	Time
COMP	Pass	2:37pm
CAL	Pass	2:37pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLIN CO. JAIL

Instrument Serial No. 008942 285 T. KEMPERD LOUISBURG, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:26pm
AIR BLK	.00	2:27pm
ACCY CHK	.07	2:28pm
AIR BLK	.00	2:29pm
SUB TEST	.00	2:30pm
AIR BLK	.00	2:30pm
SUB TEST	.00	2:32pm
AIR BLK	.00	2:33pm

Reported AC: 100 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLIN CO. JAIL 340

Serial Number: 008942 Test Record Number: 979
Test Date: 03/17/2016 Test Time: 2:34pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:34pm
FLO	Pass	2:35pm
FC	Pass	2:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:35pm
SRC	Pass	2:35pm
DET	Pass	2:35pm
BAR	Pass	2:35pm
BT	Pass	2:35pm

Blank Tests

Test	Status	Time
AIR	Pass	2:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:35pm

CRC Tests

Test	Status	Time
COMP	Pass	2:36pm
CAL	Pass	2:36pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County FRANKLIN Instrument Location FRANKLINTON PD

Instrument Serial No. 008815 #7 W. MASON ST. FRANKLINTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: 008815
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	11:39am
AIR BLK	.00	11:40am
ACCY CHK	.08	11:41am
AIR BLK	.00	11:42am
SUB TEST	.00	11:42am
AIR BLK	.00	11:43am
SUB TEST	.00	11:45am
AIR BLK	.00	11:46am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FRANKLIN COUNTY FRANKLINTON PD 340

Serial Number: 008815 Test Record Number: 1016
Test Date: 03/17/2016 Test Time: 12:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:02pm
FLO	Pass	12:02pm
FC	Pass	12:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:03pm
SRC	Pass	12:03pm
DET	Pass	12:03pm
BAR	Pass	12:03pm
BT	Pass	12:03pm

Blank Tests

Test	Status	Time
AIR	Pass	12:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:03pm

CRC Tests

Test	Status	Time
COMP	Pass	12:03pm
CAL	Pass	12:03pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gaston Instrument Location Red mobile Unit 11

Instrument Serial No. 008970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chl V. Day

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008970 Test Record Number: 139
Test Date: 03/25/2016 Test Time: 9:27pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:27pm
FLO	Pass	9:27pm
FC	Pass	9:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:28pm
SRC	Pass	9:28pm
DET	Pass	9:28pm
BAR	Pass	9:28pm
BT	Pass	9:28pm

Blank Tests

Test	Status	Time
AIR	Pass	9:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:28pm

CRC Tests

Test	Status	Time
COMP	Pass	9:28pm
CAL	Pass	9:28pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gaston Instrument Location Pat Mobile Unit

Instrument Serial No. 008090

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 25 day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008090
Test Date: 03/25/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V
Permit Number: 26632E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	9:21pm
AIR BLK	.00	9:22pm
ACCY CHK	.08	9:23pm
AIR BLK	.00	9:24pm
SUB TEST	.00	9:24pm
AIR BLK	.00	9:25pm
SUB TEST	.00	9:27pm
AIR BLK	.00	9:28pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GASTON BAT MOBILE UNIT 11 350

Serial Number: 008090 Test Record Number: 64
Test Date: 03/25/2016 Test Time: 9:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:29pm
FLO	Pass	9:29pm
FC	Pass	9:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:30pm
SRC	Pass	9:30pm
DET	Pass	9:30pm
BAR	Pass	9:30pm
BT	Pass	9:30pm

Blank Tests

Test	Status	Time
AIR	Pass	9:30pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:30pm

CRC Tests

Test	Status	Time
COMP	Pass	9:31pm
CAL	Pass	9:31pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Gaston Instrument Location Belmont PD

Instrument Serial No. 008733 201 Chronicle St., Belmont

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Marc J. King
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GASTON COUNTY BELMONT PD 350

Serial Number: 008733 Test Record Number: 944
Test Date: 03/03/2016 Test Time: 12:33pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:34pm
FLO	Pass	12:34pm
FC	Pass	12:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:34pm
SRC	Pass	12:34pm
DET	Pass	12:34pm
BAR	Pass	12:34pm
BT	Pass	12:34pm

Blank Tests

Test	Status	Time
AIR	Pass	12:35pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:35pm

CRC Tests

Test	Status	Time
COMP	Pass	12:35pm
CAL	Pass	12:35pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services**

Intox EC/IR-II: Subject Test

GASTON COUNTY BELMONT PD 350

Serial Number: 008733
Test Date: 03/03/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517501
Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	12:44pm
AIR BLK	.00	12:45pm
ACCY CHK	.08	12:46pm
AIR BLK	.00	12:47pm
SUB TEST	.00	12:48pm
AIR BLK	.00	12:48pm
SUB TEST	.00	12:50pm
AIR BLK	.00	12:51pm

Reported AG: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Greene Instrument Location Greene Co. S.O.

Instrument Serial No. 008670 301 W. Greene St., Snow Hill, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key M O
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670
Test Date: 03/29/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513103
Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	10:50am
AIR BLK	.00	10:50am
ACCY CHK	.07	10:51am
AIR BLK	.00	10:52am
SUB TEST	.00	10:53am
AIR BLK	.00	10:54am
SUB TEST	.00	10:55am
AIR BLK	.00	10:56am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670 Test Record Number: 1580
Test Date: 03/29/2016 Test Time: 10:57am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:57am
FLO	Pass	10:57am
FC	Pass	10:57am

Temperature Tests

Test	Status	Time
FC1	Pass	10:57am
SRC	Pass	10:57am
DET	Pass	10:57am
BAR	Pass	10:57am
BT	Pass	10:57am

Blank Tests

Test	Status	Time
AIR	Pass	10:58am


Printer Tests

Test	Status	Time
PRNT	Pass	10:58am

CRC Tests

Test	Status	Time
COMP	Pass	10:58am
CAL	Pass	10:58am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location Bat Mobile Unit 8

Instrument Serial No. 008816 SHP-Guilford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donna B Skinner
Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008816
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:08pm
AIR BLK	.00	11:09pm
ACCY CHK	.07	11:09pm
AIR BLK	.00	11:10pm
SUB TEST	.00	11:11pm
AIR BLK	.00	11:12pm
SUB TEST	.00	11:13pm
AIR BLK	.00	11:14pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008816 Test Record Number: 7198
Test Date: 03/17/2016 Test Time: 11:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:16pm
FLO	Pass	11:16pm
FC	Pass	11:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:16pm
SRC	Pass	11:16pm
DET	Pass	11:16pm
BAR	Pass	11:16pm
BT	Pass	11:16pm

Blank Tests

Test	Status	Time
AIR	Pass	11:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:17pm

CRC Tests

Test	Status	Time
COMP	Pass	11:17pm
CAL	Pass	11:17pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Bat Mobile Unit
Instrument Serial No. D08615 SHP-Guilford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016, the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner 6444
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008615
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:09pm
AIR BLK	.00	11:10pm
ACCY CHK	.07	11:11pm
AIR BLK	.00	11:11pm
SUB TEST	.00	11:12pm
AIR BLK	.00	11:13pm
SUB TEST	.00	11:15pm
AIR BLK	.00	11:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008615 Test Record Number: 5365
Test Date: 03/17/2016 Test Time: 11:17pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:17pm
FLO	Pass	11:17pm
FC	Pass	11:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:17pm
SRC	Pass	11:17pm
DET	Pass	11:17pm
BAR	Pass	11:17pm
BT	Pass	11:17pm

Blank Tests

Test	Status	Time
AIR	Pass	11:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	11:18pm

CRC Tests

Test	Status	Time
COMP	Pass	11:18pm
CAL	Pass	11:18pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Bat Mobile Unit 8
Instrument Serial No. D08736 SHP-Guilford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinnin
Signature of Certifying Official

1044
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008736
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B
Permit Number: 13651E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:27pm
AIR BLK	.00	11:28pm
ACCY CHK	.07	11:28pm
AIR BLK	.00	11:29pm
SUB TEST	.00	11:30pm
AIR BLK	.00	11:31pm
SUB TEST	.00	11:32pm
AIR BLK	.00	11:33pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008736 Test Record Number: 808
Test Date: 03/17/2016 Test Time: 11:35pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35pm
FLO	Pass	11:35pm
FC	Pass	11:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:35pm
SRC	Pass	11:35pm
DET	Pass	11:35pm
BAR	Pass	11:35pm
BT	Pass	11:35pm

Blank Tests

Test	Status	Time
AIR	Pass	11:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:36pm

CRC Tests

Test	Status	Time
COMP	Pass	11:36pm
CAL	Pass	11:36pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Bat mobile unit
Instrument Serial No. DO8601 SHP-Guilford

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016, the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner 644
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:30pm
AIR BLK	.00	11:31pm
ACCY CHK	.08	11:32pm
AIR BLK	.00	11:33pm
SUB TEST	.00	11:33pm
AIR BLK	.00	11:34pm
SUB TEST	.00	11:36pm
AIR BLK	.00	11:36pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601 Test Record Number: 1156
Test Date: 03/17/2016 Test Time: 11:37pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:38pm
FLO	Pass	11:38pm
FC	Pass	11:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:38pm
SRC	Pass	11:38pm
DET	Pass	11:38pm
BAR	Pass	11:38pm
BT	Pass	11:38pm

Blank Tests

Test	Status	Time
AIR	Pass	11:38pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:39pm

CRC Tests

Test	Status	Time
COMP	Pass	11:39pm
CAL	Pass	11:39pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location UNC - Greensboro
Instrument Serial No. 008604 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: 008604
Test Date: 03/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517501
Exp Date: 06/24/2017


Test	g/210L	Time
DIAG	Pass	2:59pm
AIR BLK	.00	2:59pm
ACCY CHK	.08	3:00pm
AIR BLK	.00	3:01pm
SUB TEST	.00	3:02pm
AIR BLK	.00	3:03pm
SUB TEST	.00	3:04pm
AIR BLK	.00	3:05pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY UNC-G POLICE DEPT 400

Serial Number: 008604 Test Record Number: 1425
Test Date: 03/11/2016 Test Time: 3:07pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:07pm
FLO	Pass	3:07pm
FC	Pass	3:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:08pm
SRC	Pass	3:08pm
DET	Pass	3:08pm
BAR	Pass	3:08pm
BT	Pass	3:08pm

Blank Tests

Test	Status	Time
AIR	Pass	3:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:08pm

CRC Tests

Test	Status	Time
COMP	Pass	3:08pm
CAL	Pass	3:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Greensboro PD
Instrument Serial No. 008725 100 Police Plaza, Greensboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725
Test Date: 03/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG517501
Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	2:07pm
AIR BLK	.00	2:08pm
ACCY CHK	.08	2:08pm
AIR BLK	.00	2:09pm
SUB TEST	.00	2:10pm
AIR BLK	.00	2:11pm
SUB TEST	.00	2:12pm
AIR BLK	.00	2:13pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO PD 400

Serial Number: 008725 Test Record Number: 3590
Test Date: 03/11/2016 Test Time: 2:14pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:14pm
FLO	Pass	2:14pm
FC	Pass	2:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:14pm
SRC	Pass	2:14pm
DET	Pass	2:14pm
BAR	Pass	2:14pm
BT	Pass	2:14pm

Blank Tests

Test	Status	Time
AIR	Pass	2:15pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:15pm

CRC Tests

Test	Status	Time
COMP	Pass	2:15pm
CAL	Pass	2:15pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Bat mobile Unit 8
Instrument Serial No 008736 Guilford CO 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 05 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya Skinner 644
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008736 Test Record Number: 802
Test Date: 03/05/2016 Test Time: 12:16am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:16am
FLO	Pass	12:16am
FC	Pass	12:16am

Temperature Tests

Test	Status	Time
FC1	Pass	12:17am
SRC	Pass	12:17am
DET	Pass	12:17am
BAR	Pass	12:17am
BT	Pass	12:17am

Blank Tests

Test	Status	Time
AIR	Pass	12:17am

Printer Tests

Test	Status	Time
PRNT	Pass	12:17am

CRC Tests

Test	Status	Time
COMP	Pass	12:17am
CAL	Pass	12:17am

Preventive Maintenance
Status: Pass


Danya B Skinner
Analyst

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008736
Test Date: 03/05/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENACE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Female
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B
Permit Number: 13651E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:07am
AIR BLK	.00	12:08am
ACCY CHK	.07	12:08am
AIR BLK	.00	12:09am
SUB TEST	.00	12:10am
AIR BLK	.00	12:11am
SUB TEST	.00	12:12am
AIR BLK	.00	12:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Pat Mobile Units
Instrument Serial No. 008601 Guilford CO 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



George B. Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601 Test Record Number: 1150
Test Date: 03/05/2016 Test Time: 12:28am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:28am
FLO	Pass	12:28am
FC	Pass	12:28am

Temperature Tests

Test	Status	Time
FC1	Pass	12:29am
SRC	Pass	12:29am
DET	Pass	12:29am
BAR	Pass	12:29am
BT	Pass	12:29am

Blank Tests

Test	Status	Time
AIR	Pass	12:29am

Printer Tests

Test	Status	Time
PRNT	Pass	12:29am

CRC Tests

Test	Status	Time
COMP	Pass	12:29am
CAL	Pass	12:29am

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601
Test Date: 03/05/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE'

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: FTA

Test Type: Breath Test

Lot Number: AG517402


Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:15am
AIR BLK	.00	12:15am
ACCY CHK	.08	12:16am
AIR BLK	.00	12:17am
SUB TEST	.00	12:17am
AIR BLK	.00	12:18am
SUB TEST	.00	12:20am
AIR BLK	.00	12:21am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Pat mobile Unit 8
Instrument Serial No. 0081001 Guilford CO 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601
Test Date: 03/05/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE'

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: FTA

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:15am
AIR BLK	.00	12:15am
ACCY CHK	.08	12:16am
AIR BLK	.00	12:17am
SUB TEST	.00	12:17am
AIR BLK	.00	12:18am
SUB TEST	.00	12:20am
AIR BLK	.00	12:21am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BATMOBILE UNIT 8 400

Serial Number: 008601 Test Record Number: 1150
Test Date: 03/05/2016 Test Time: 12:28am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:28am
FLO	Pass	12:28am
FC	Pass	12:28am

Temperature Tests

Test	Status	Time
FC1	Pass	12:29am
SRC	Pass	12:29am
DET	Pass	12:29am
BAR	Pass	12:29am
BT	Pass	12:29am

Blank Tests

Test	Status	Time
AIR	Pass	12:29am

Printer Tests

Test	Status	Time
PRNT	Pass	12:29am

CRC Tests

Test	Status	Time
COMP	Pass	12:29am
CAL	Pass	12:29am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Batmobile Unit 8
Instrument Serial No 208615 Guilford CO 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B. Skott
Signature of Certifying Official

6411
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008615
Test Date: 03/05/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Female
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B
Permit Number: 13651E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	1:00am
AIR BLK	.00	1:01am
ACCY CHK	.07	1:02am
AIR BLK	.00	1:02am
SUB TEST	.00	1:04am
AIR BLK	.00	1:05am
SUB TEST	.00	1:06am
AIR BLK	.00	1:07am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GREENSBORO BAT MOBILE UNIT 8 400

Serial Number: 008615 Test Record Number: 5363
Test Date: 03/05/2016 Test Time: 1:09am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:09am
FLO	Pass	1:09am
FC	Pass	1:09am

Temperature Tests

Test	Status	Time
FC1	Pass	1:09am
SRC	Pass	1:09am
DET	Pass	1:09am
BAR	Pass	1:09am
BT	Pass	1:09am

Blank Tests

Test	Status	Time
AIR	Pass	1:10am

Printer Tests

Test	Status	Time
PRNT	Pass	1:10am

CRC Tests

Test	Status	Time
COMP	Pass	1:10am
CAL	Pass	1:10am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HARNETT Instrument Location DUNN POLICE DEPT.
Instrument Serial No. 008644 DUNN, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: 008644
Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	5:25pm
AIR BLK	.00	5:26pm
ACCY CHK	.07	5:27pm
AIR BLK	.00	5:28pm
SUB TEST	.00	5:28pm
AIR BLK	.00	5:29pm
SUB TEST	.00	5:31pm
AIR BLK	.00	5:32pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DUNN POLICE DEPT. 420

Serial Number: 008644 Test Record Number: 1183
Test Date: 03/28/2016 Test Time: 5:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:33pm
FLO	Pass	5:33pm
FC	Pass	5:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:33pm
SRC	Pass	5:33pm
DET	Pass	5:33pm
BAR	Pass	5:33pm
BT	Pass	5:33pm

Blank Tests

Test	Status	Time
AIR	Pass	5:34pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:34pm

CRC Tests

Test	Status	Time
COMP	Pass	5:34pm
CAL	Pass	5:34pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Halifax Co. Instrument Location Halifax Co. S.O.

Instrument Serial No. 008696 355 Fennell Ln Halifax NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HALIFAX CO. HALIFAX CO. SD 410

Serial Number: 008695
Test Date: 03/08/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, GRAYHAM C
Permit Number: 7682E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:07am
AIR BLK	.00	10:08am
ACCY CHK	.08	10:09am
AIR BLK	.00	10:10am
SUB TEST	.00	10:11am
AIR BLK	.00	10:11am
SUB TEST	.00	10:13am
AIR BLK	.00	10:14am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO. HALIFAX CO. SD 410

Serial Number: 008695 Test Record Number: 1981
Test Date: 03/08/2016 Test Time: 10:16am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:16am
FLO	Pass	10:16am
FC	Pass	10:16am

Temperature Tests

Test	Status	Time
FC1	Pass	10:16am
SRC	Pass	10:16am
DET	Pass	10:16am
BAR	Pass	10:16am
BT	Pass	10:16am

Blank Tests

Test	Status	Time
AIR	Pass	10:17am


Printer Tests

Test	Status	Time
PRNT	Pass	10:17am

CRC Tests

Test	Status	Time
COMP	Pass	10:17am
CAL	Pass	10:17am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Halifax Co. Instrument Location Roanoke Rapids P.D.
Instrument Serial No. 008635 1040 Roanoke Ave Roanoke Rapids NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

634
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HALIFAX CO ROANOKE RAPIDS PD 410

Serial Number: 008635
Test Date: 03/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG411202

Exp Date: 04/22/2016


Test	g/210L	Time
DIAG	Pass	12:51pm
AIR BLK	.00	12:52pm
ACCY CHK	.07	12:53pm
AIR BLK	.00	12:54pm
SUB TEST	.00	12:54pm
AIR BLK	.00	12:55pm
SUB TEST	.00	12:57pm
AIR BLK	.00	12:58pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO ROANOKE RAPIDS PD 410

Serial Number: 008635 Test Record Number: 1498
Test Date: 03/08/2016 Test Time: 12:59pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:00pm
FLO	Pass	1:00pm
FC	Pass	1:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:00pm
SRC	Pass	1:00pm
DET	Pass	1:00pm
BAR	Pass	1:00pm
BT	Pass	1:00pm

Blank Tests

Test	Status	Time
AIR	Pass	1:01pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:01pm

CRC Tests

Test	Status	Time
COMP	Pass	1:01pm
CAL	Pass	1:01pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Halifax Co. Instrument Location Roanoke Rapids P.D.

Instrument Serial No. 009656 1040 Roanoke Ave Roanoke Rapids NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656
Test Date: 03/08/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG411202


Exp Date: 04/22/2016

Test	g/210L	Time
DIAG	Pass	12:52pm
AIR BLK	.00	12:53pm
ACCY CHK	.07	12:54pm
AIR BLK	.00	12:55pm
SUB TEST	.00	12:55pm
AIR BLK	.00	12:56pm
SUB TEST	.00	12:58pm
AIR BLK	.00	1:00pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

HALIFAX CO. ROANOKE RAPIDS PD 410

Serial Number: 008656 Test Record Number: 533
Test Date: 03/08/2016 Test Time: 1:00pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:01pm
FLO	Pass	1:01pm
FC	Pass	1:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:01pm
SRC	Pass	1:01pm
DET	Pass	1:01pm
BAR	Pass	1:01pm
BT	Pass	1:01pm

Blank Tests

Test	Status	Time
AIR	Pass	1:02pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:02pm

CRC Tests

Test	Status	Time
COMP	Pass	1:02pm
CAL	Pass	1:02pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Haywood Instrument Location Haywood Co. Jail
Instrument Serial No. 008714 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714
Test Date: 03/14/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101


Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	11:01am
AIR BLK	.00	11:02am
ACCY CHK	.07	11:03am
AIR BLK	.00	11:04am
SUB TEST	.00	11:04am
AIR BLK	.00	11:05am
SUB TEST	.00	11:07am
AIR BLK	.00	11:07am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008714 Test Record Number: 1329
Test Date: 03/14/2016 Test Time: 11:08am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:09am
FLO	Pass	11:09am
FC	Pass	11:09am

Temperature Tests

Test	Status	Time
FC1	Pass	11:09am
SRC	Pass	11:09am
DET	Pass	11:09am
BAR	Pass	11:09am
BT	Pass	11:09am

Blank Tests

Test	Status	Time
AIR	Pass	11:09am

Printer Tests

Test	Status	Time
PRNT	Pass	11:09am

CRC Tests

Test	Status	Time
COMP	Pass	11:10am
CAL	Pass	11:10am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Haywood Instrument Location Haywood Co. Jail
Instrument Serial No. 008712 Waynesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Orvil R. Arthur
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008712
Test Date: 03/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:42am
AIR BLK	.00	10:44am
ACCY CHK	.08	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:46am
AIR BLK	.00	10:47am
SUB TEST	.00	10:48am
AIR BLK	.00	10:49am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HAYWOOD COUNTY HAYWOOD COUNTY JAIL 430

Serial Number: 008712 Test Record Number: 1690
Test Date: 03/14/2016 Test Time: 10:50am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:51am


Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Hertford Instrument Location Murfreesboro P.D.

Instrument Serial No. 008906 115 E. Broad St., Murfreesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linch A. Keele
Signature of Certifying Official

467
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906

Test Date: 03/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	4:16pm
AIR BLK	.00	4:17pm
ACCY CHK	.08	4:18pm
AIR BLK	.00	4:19pm
SUB TEST	.00	4:19pm
AIR BLK	.00	4:20pm
SUB TEST	.00	4:21pm
AIR BLK	.00	4:22pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906 Test Record Number: 543
Test Date: 03/30/2016 Test Time: 4:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:28pm
FLO	Pass	4:28pm
FC	Pass	4:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:29pm
SRC	Pass	4:29pm
DET	Pass	4:29pm
BAR	Pass	4:29pm
BT	Pass	4:29pm

Blank Tests

Test	Status	Time
AIR	Pass	4:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	4:29pm

CRC Tests

Test	Status	Time
COMP	Pass	4:29pm
CAL	Pass	4:29pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Jackson Instrument Location Jackson Co. Jail

Instrument Serial No. 008708 Sylva, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Daniel R. Cutler
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008708
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: *CUTLER, DANIEL R*
Permit Number: 8457E
Effective:
09/01/2015-09/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG409709
Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:52am
ACCY CHK	.07	11:52am
AIR BLK	.00	11:53am
SUB TEST	.00	11:54am
AIR BLK	.00	11:55am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008708 Test Record Number: 1095
Test Date: 03/15/2016 Test Time: 11:59am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:59am
FLO	Pass	11:59am
FC	Pass	11:59am

Temperature Tests

Test	Status	Time
FC1	Pass	11:59am
SRC	Pass	11:59am
DET	Pass	11:59am
BAR	Pass	11:59am
BT	Pass	11:59am

Blank Tests

Test	Status	Time
AIR	Pass	12:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:00pm

CRC Tests

Test	Status	Time
COMP	Pass	12:00pm
CAL	Pass	12:00pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Jackson Instrument Location Jackson Co. Jail

Instrument Serial No. 008722 Sylva, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuthbert
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008722
Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG409709


Exp Date: 04/07/2016

Test	g/210L	Time
DIAG	Pass	11:50am
AIR BLK	.00	11:51am
ACCY CHK	.07	11:51am
AIR BLK	.00	11:53am
SUB TEST	.00	11:53am
AIR BLK	.00	11:54am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008722 Test Record Number: 816
Test Date: 03/15/2016 Test Time: 11:58am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:59am

Printer Tests

Test	Status	Time
PRNT	Pass	11:59am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008722

Test Date: 03/30/2016

Citation Number: M0000000-0

Subject's Name: CANISTER, CHANGE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	12:06pm
AIR BLK	.00	12:07pm
ACCY CHK	.08	12:07pm
AIR BLK	.00	12:09pm
SUB TEST	.00	12:09pm
AIR BLK	.00	12:10pm
SUB TEST	.00	12:12pm
AIR BLK	.00	12:13pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008708
Test Date: 03/30/2016

Citation Number: M0000000-0
Subject's Name: CANISTER, CHANGE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R
Permit Number: 8457E
Effective:
09/01/2015-09/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG534802
Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	12:17pm
AIR BLK	.00	12:18pm
ACCY CHK	.08	12:18pm
AIR BLK	.00	12:19pm
SUB TEST	.00	12:20pm
AIR BLK	.00	12:21pm
SUB TEST	.00	12:22pm
AIR BLK	.00	12:23pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County JOHNSTON Instrument Location BENSON POLICE DEPT.
Instrument Serial No. 008885 BENSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 04 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY BENSON POLICE DEPT.
500

Serial Number: 008885
Test Date: 03/04/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	3:41pm
AIR BLK	.00	3:42pm
ACCY CHK	.07	3:43pm
AIR BLK	.00	3:44pm
SUB TEST	.00	3:45pm
AIR BLK	.00	3:46pm
SUB TEST	.00	3:47pm
AIR BLK	.00	3:48pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY BENSON POLICE DEPT. 500

Serial Number: 008885 Test Record Number: 437
Test Date: 03/04/2016 Test Time: 3:56pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:57pm
FLO	Pass	3:57pm
FC	Pass	3:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:57pm
SRC	Pass	3:57pm
DET	Pass	3:57pm
BAR	Pass	3:57pm
BT	Pass	3:57pm

Blank Tests

Test	Status	Time
AIR	Pass	3:57pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:58pm

CRC Tests

Test	Status	Time
COMP	Pass	3:58pm
CAL	Pass	3:58pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County JOHNSTON Instrument Location JOHNSTON Co. Jail

Instrument Serial No. 008846 SMITHFIELD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY JOHNSTON CO. JAIL 500

Serial Number: 008846

Test Date: 03/29/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	1:17pm
AIR BLK	.00	1:18pm
ACCY CHK	.08	1:18pm
AIR BLK	.00	1:20pm
SUB TEST	.00	1:20pm
AIR BLK	.00	1:21pm
SUB TEST	.00	1:23pm
AIR BLK	.00	1:24pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY JOHNSTON CO. JAIL 500
Serial Number: 008846 Test Record Number: 3850
Test Date: 03/29/2016 Test Time: 1:27pm EDT
System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:27pm
FLO	Pass	1:27pm
FC	Pass	1:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:28pm
SRC	Pass	1:28pm
DET	Pass	1:28pm
BAR	Pass	1:28pm
BT	Pass	1:28pm

Blank Tests

Test	Status	Time
AIR	Pass	1:28pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:28pm

CRC Tests

Test	Status	Time
COMP	Pass	1:28pm
CAL	Pass	1:28pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County JOHNSTON Instrument Location JOHNSTON Co. JAIL
Instrument Serial No. 008810 SMITHFIELD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY JOHNSTON CO. JAIL 500

Serial Number: 008810
Test Date: 03/29/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	1:52pm
AIR BLK	.00	1:53pm
ACCY CHK	.08	1:53pm
AIR BLK	.00	1:55pm
SUB TEST	.00	1:56pm
AIR BLK	.00	1:56pm
SUB TEST	.00	1:58pm
AIR BLK	.00	1:59pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY JOHNSTON CO. JAIL 500

Serial Number: 008810 Test Record Number: 2192
Test Date: 03/29/2016 Test Time: 1:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:00pm
FLO	Pass	2:00pm
FC	Pass	2:00pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:00pm
SRC	Pass	2:00pm
DET	Pass	2:00pm
BAR	Pass	2:00pm
BT	Pass	2:00pm

Blank Tests

Test	Status	Time
AIR	Pass	2:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:01pm

CRC Tests

Test	Status	Time
COMP	Pass	2:01pm
CAL	Pass	2:01pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County JOHNSTON Instrument Location SELMA POLICE DEPT.

Instrument Serial No. 008595 SELMA NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595
Test Date: 03/28/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434901
Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	4:00pm
AIR BLK	.00	4:00pm
ACCY CHK	.07	4:01pm
AIR BLK	.00	4:02pm
SUB TEST	.00	4:03pm
AIR BLK	.00	4:04pm
SUB TEST	.00	4:05pm
AIR BLK	.00	4:06pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY SELMA PD 500

Serial Number: 008595 Test Record Number: 979
Test Date: 03/28/2016 Test Time: 4:07pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:07pm
FLO	Pass	4:07pm
FC	Pass	4:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:07pm
SRC	Pass	4:07pm
DET	Pass	4:07pm
BAR	Pass	4:07pm
BT	Pass	4:07pm

Blank Tests

Test	Status	Time
AIR	Pass	4:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:08pm

CRC Tests

Test	Status	Time
COMP	Pass	4:08pm
CAL	Pass	4:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Jones Instrument Location Jones County
Instrument Serial No. 008705 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JONES COUNTY JONES COUNTY SD 510

Serial Number: 008705
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:01pm
AIR BLK	.00	12:02pm
ACCY CHK	.07	12:02pm
AIR BLK	.00	12:03pm
SUB TEST	.00	12:04pm
AIR BLK	.00	12:04pm
SUB TEST	.00	12:06pm
AIR BLK	.00	12:07pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

JONES COUNTY JONES COUNTY SD 510

Serial Number: 008705 Test Record Number: 1045
Test Date: 03/17/2016 Test Time: 12:07pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:07pm
FLO	Pass	12:07pm
FC	Pass	12:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:08pm
SRC	Pass	12:08pm
DET	Pass	12:08pm
BAR	Pass	12:08pm
BT	Pass	12:08pm

Blank Tests

Test	Status	Time
AIR	Pass	12:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:08pm

CRC Tests

Test	Status	Time
COMP	Pass	12:09pm
CAL	Pass	12:09pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Iredell Instrument Location Mooreville PD

Instrument Serial No. 008685 750 W. Iredell Ave., Mooreville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: 008685 Test Record Number: 2495
Test Date: 03/08/2016 Test Time: 9:38am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:39am
FLO	Pass	9:39am
FC	Pass	9:39am

Temperature Tests

Test	Status	Time
FC1	Pass	9:39am
SRC	Pass	9:39am
DET	Pass	9:39am
BAR	Pass	9:39am
BT	Pass	9:39am

Blank Tests

Test	Status	Time
AIR	Pass	9:40am


Printer Tests

Test	Status	Time
PRNT	Pass	9:40am

CRC Tests

Test	Status	Time
COMP	Pass	9:40am
CAL	Pass	9:40am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

IREDELL COUNTY MOORESVILLE PD 480

Serial Number: 008685

Test Date: 03/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG425303

Exp Date: 09/10/2016

Test	g/210L	Time
DIAG	Pass	9:43am
AIR BLK	.00	9:44am
ACCY CHK	.07	9:44am
AIR BLK	.00	9:45am
SUB TEST	.00	9:47am
AIR BLK	.00	9:48am
SUB TEST	.00	9:49am
AIR BLK	.00	9:50am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County LEE Instrument Location SANFORD POLICE DEPT.

Instrument Serial No. 008867 SANFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867
Test Date: 03/30/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513103
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:46pm
AIR BLK	.00	1:46pm
ACCY CHK	.07	1:47pm
AIR BLK	.00	1:48pm
SUB TEST	.00	1:49pm
AIR BLK	.00	1:50pm
SUB TEST	.00	1:51pm
AIR BLK	.00	1:52pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

LEE COUNTY SANFORD POLICE DEPT 520

Serial Number: 008867 Test Record Number: 930
Test Date: 03/30/2016 Test Time: 1:53pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:53pm
FLO	Pass	1:53pm
FC	Pass	1:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:53pm
SRC	Pass	1:53pm
DET	Pass	1:53pm
BAR	Pass	1:53pm
BT	Pass	1:53pm

Blank Tests

Test	Status	Time
AIR	Pass	1:54pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:54pm

CRC Tests

Test	Status	Time
COMP	Pass	1:54pm
CAL	Pass	1:54pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lenoir Instrument Location Kinston P.D.
Instrument Serial No. 008624 205 E. King St., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keen
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624

Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	3:15pm
AIR BLK	.00	3:16pm
ACCY CHK	.07	3:17pm
AIR BLK	.00	3:18pm
SUB TEST	.00	3:19pm
AIR BLK	.00	3:20pm
SUB TEST	.00	3:21pm
AIR BLK	.00	3:22pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624 Test Record Number: 1528
Test Date: 03/28/2016 Test Time: 3:23pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:23pm
FLO	Pass	3:23pm
FC	Pass	3:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:23pm
SRC	Pass	3:23pm
DET	Pass	3:23pm
BAR	Pass	3:23pm
BT	Pass	3:23pm

Blank Tests

Test	Status	Time
AIR	Pass	3:24pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:24pm

CRC Tests

Test	Status	Time
COMP	Pass	3:24pm
CAL	Pass	3:24pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lenoir Instrument Location Lenoir Co. S.O.
Instrument Serial No. 008639 130 Queen St., Kinston, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Reese
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639
Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016


Test	g/210L	Time
DIAG	Pass	2:56pm
AIR BLK	.00	2:57pm
ACCY CHK	.07	2:58pm
AIR BLK	.00	2:59pm
SUB TEST	.00	3:00pm
AIR BLK	.00	3:01pm
SUB TEST	.00	3:02pm
AIR BLK	.00	3:03pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639 Test Record Number: 2771
Test Date: 03/28/2016 Test Time: 3:04pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:04pm
FLO	Pass	3:04pm
FC	Pass	3:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:05pm
SRC	Pass	3:05pm
DET	Pass	3:05pm
BAR	Pass	3:05pm
BT	Pass	3:05pm

Blank Tests

Test	Status	Time
AIR	Pass	3:05pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:05pm

CRC Tests

Test	Status	Time
COMP	Pass	3:05pm
CAL	Pass	3:05pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lincoln Instrument Location Lincoln County Courthouse
Instrument Serial No. 008910 #1 Courthouse Square, Lincolnton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of April, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. D. [Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008910 Test Record Number: 491
Test Date: 04/04/2016 Test Time: 12:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17pm
FLO	Pass	12:17pm
FC	Pass	12:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:17pm
SRC	Pass	12:17pm
DET	Pass	12:17pm
BAR	Pass	12:17pm
BT	Pass	12:17pm

Blank Tests

Test	Status	Time
AIR	Pass	12:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:18pm

CRC Tests

Test	Status	Time
COMP	Pass	12:18pm
CAL	Pass	12:18pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008910
Test Date: 04/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	12:21pm
AIR BLK	.00	12:22pm
ACCY CHK	.08	12:22pm
AIR BLK	.00	12:23pm
SUB TEST	.00	12:24pm
AIR BLK	.00	12:25pm
SUB TEST	.00	12:26pm
AIR BLK	.00	12:27pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mecklenburg Instrument Location Bedmobile Unit 11

Instrument Serial No. 008090

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chad Zou
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008090
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	7:55pm
AIR BLK	.00	7:56pm
ACCY CHK	.08	7:57pm
AIR BLK	.00	7:57pm
SUB TEST	.00	7:58pm
AIR BLK	.00	7:59pm
SUB TEST	.00	8:00pm
AIR BLK	.00	8:01pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008090 Test Record Number: 61
Test Date: 03/19/2016 Test Time: 8:03pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:03pm
FLO	Pass	8:03pm
FC	Pass	8:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:03pm
SRC	Pass	8:03pm
DET	Pass	8:03pm
BAR	Pass	8:03pm
BT	Pass	8:03pm

Blank Tests

Test	Status	Time
AIR	Pass	8:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:04pm

CRC Tests

Test	Status	Time
COMP	Pass	8:04pm
CAL	Pass	8:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Mecklenburg Instrument Location BAT mobile Unit 11

Instrument Serial No. 068970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of March, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris Wynn
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008970
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V
Permit Number: 26632E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG425303
Exp Date: 10/10/2016


Test	g/210L	Time
DIAG	Pass	7:46pm
AIR BLK	.00	7:47pm
ACCY CHK	.08	7:47pm
AIR BLK	.00	7:48pm
SUB TEST	.00	7:49pm
AIR BLK	.00	7:49pm
SUB TEST	.00	7:51pm
AIR BLK	.00	7:52pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008970 Test Record Number: 136
Test Date: 03/19/2016 Test Time: 7:53pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:54pm
FLO	Pass	7:54pm
FC	Pass	7:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:54pm
SRC	Pass	7:54pm
DET	Pass	7:54pm
BAR	Pass	7:54pm
BT	Pass	7:54pm

Blank Tests

Test	Status	Time
AIR	Pass	7:54pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:54pm

CRC Tests

Test	Status	Time
COMP	Pass	7:55pm
CAL	Pass	7:55pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location CMPD-LEC
Instrument Serial No. 008594 601 E. Trade St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mark D. King
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008594 Test Record Number: 3216
Test Date: 03/18/2016 Test Time: 1:41pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:42pm
FLO	Pass	1:42pm
FC	Pass	1:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:42pm
SRC	Pass	1:42pm
DET	Pass	1:42pm
BAR	Pass	1:42pm
BT	Pass	1:42pm

Blank Tests

Test	Status	Time
AIR	Pass	1:43pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:43pm

CRC Tests

Test	Status	Time
COMP	Pass	1:43pm
CAL	Pass	1:43pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CMPD LEC 590

Serial Number: 008594
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:47pm
AIR BLK	.00	1:47pm
ACCY CHK	.07	1:48pm
AIR BLK	.00	1:49pm
SUB TEST	.00	1:50pm
AIR BLK	.00	1:51pm
SUB TEST	.00	1:52pm
AIR BLK	.00	1:53pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Mecklenburg County SD
Instrument Serial No. 008690 801 E. 4th St., Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mark J. Ryan
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY SHERIFFS DEPARTMENT 590

Serial Number: 008690 Test Record Number: 5176
Test Date: 03/18/2016 Test Time: 2:54pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:55pm
FLO	Pass	2:55pm
FC	Pass	2:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:55pm
SRC	Pass	2:55pm
DET	Pass	2:55pm
BAR	Pass	2:55pm
BT	Pass	2:55pm

Blank Tests

Test	Status	Time
AIR	Pass	2:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:56pm

CRC Tests

Test	Status	Time
COMP	Pass	2:56pm
CAL	Pass	2:56pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY SHERIFFS DEPARTMENT
590

Serial Number: 008690
Test Date: 03/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	3:01pm
AIR BLK	.00	3:01pm
ACCY CHK	.08	3:02pm
AIR BLK	.00	3:03pm
SUB TEST	.00	3:03pm
AIR BLK	.00	3:04pm
SUB TEST	.00	3:06pm
AIR BLK	.00	3:07pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Mecklenburg County SD
Instrument Serial No. 008665 801 E 4th St. Charlotte

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Marc S. Ray
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY SHERIFFS DEPARTMENT 590

Serial Number: 008665 Test Record Number: 3951
Test Date: 03/18/2016 Test Time: 2:50pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:51pm
FLO	Pass	2:51pm
FC	Pass	2:51pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:51pm
SRC	Pass	2:51pm
DET	Pass	2:51pm
BAR	Pass	2:51pm
BT	Pass	2:51pm

Blank Tests

Test	Status	Time
AIR	Pass	2:51pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:52pm

CRC Tests

Test	Status	Time
COMP	Pass	2:52pm
CAL	Pass	2:52pm

Preventive Maintenance
Status: *Pass*



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY SHERIFFS DEPARTMENT
590

Serial Number: 008665
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:56pm
AIR BLK	.00	2:57pm
ACCY CHK	.07	2:58pm
AIR BLK	.00	2:59pm
SUB TEST	.00	2:59pm
AIR BLK	.00	3:00pm
SUB TEST	.00	3:02pm
AIR BLK	.00	3:03pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Huntersville PD
Instrument Serial No. 008747 9630 Julian Clark Ave., Huntersville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mandy [Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747 Test Record Number: 2503
Test Date: 03/28/2016 Test Time: 11:11am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:12am
FLO	Pass	11:12am
FC	Pass	11:12am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am


Printer Tests

Test	Status	Time
PRNT	Pass	11:13am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747
Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017


Test	g/210L	Time
DIAG	Pass	11:15am
AIR BLK	.00	11:16am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
SUB TEST	.00	11:18am
AIR BLK	.00	11:19am
SUB TEST	.00	11:21am
AIR BLK	.00	11:22am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Cornelius P.D.
Instrument Serial No. 008692 21440 Catawba Ave., Cornelius

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Mark D. Ryan
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692 Test Record Number: 2479
Test Date: 03/28/2016 Test Time: 12:28pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:29pm
FLO	Pass	12:29pm
FC	Pass	12:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:29pm
SRC	Pass	12:29pm
DET	Pass	12:29pm
BAR	Pass	12:29pm
BT	Pass	12:29pm

Blank Tests

Test	Status	Time
AIR	Pass	12:30pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:30pm

CRC Tests

Test	Status	Time
COMP	Pass	12:30pm
CAL	Pass	12:30pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692
Test Date: 03/28/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:33pm
AIR BLK	.00	12:33pm
ACCY CHK	.07	12:34pm
AIR BLK	.00	12:35pm
SUB TEST	.00	12:36pm
AIR BLK	.00	12:37pm
SUB TEST	.00	12:38pm
AIR BLK	.00	12:39pm

Reported AC: .00 g/210L



Signature of Chemical Analyst
Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mecklenburg Instrument Location Bat mobile Unit 4

Instrument Serial No. 008090

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ch V Dow
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008090
Test Date: 03/09/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: TOWERY, CHAD V
Permit Number: 26632E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	9:19pm
AIR BLK	.00	9:20pm
ACCY CHK	.08	9:20pm
AIR BLK	.00	9:21pm
SUB TEST	.00	9:21pm
AIR BLK	.00	9:22pm
SUB TEST	.00	9:24pm
AIR BLK	.00	9:25pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 11 590

Serial Number: 008090 Test Record Number: 46
Test Date: 03/09/2016 Test Time: 9:26pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:27pm
FLO	Pass	9:27pm
FC	Pass	9:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:27pm
SRC	Pass	9:27pm
DET	Pass	9:27pm
BAR	Pass	9:27pm
BT	Pass	9:27pm

Blank Tests

Test	Status	Time
AIR	Pass	9:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:28pm

CRC Tests

Test	Status	Time
COMP	Pass	9:28pm
CAL	Pass	9:28pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Mitchell Instrument Location Spruce Pine PD
Instrument Serial No. 008726 Spruce Pine, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726
Test Date: 03/28/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	3:25pm
AIR BLK	.00	3:26pm
ACCY CHK	.08	3:27pm
AIR BLK	.00	3:28pm
SUB TEST	.00	3:28pm
AIR BLK	.00	3:29pm
SUB TEST	.00	3:30pm
AIR BLK	.00	3:31pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726 Test Record Number: 814
Test Date: 03/28/2016 Test Time: 3:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:33pm
FLO	Pass	3:33pm
FC	Pass	3:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:33pm
SRC	Pass	3:33pm
DET	Pass	3:33pm
BAR	Pass	3:33pm
BT	Pass	3:33pm

Blank Tests

Test	Status	Time
AIR	Pass	3:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:34pm

CRC Tests

Test	Status	Time
COMP	Pass	3:34pm
CAL	Pass	3:34pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Moore Co. Instrument Location Moore Co. Jail
Instrument Serial No. 008735 302 McNeill St. Carthage NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

634
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MOORE COUNTY MOORE COUNTY JAIL 620

Serial Number: 008735
Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	11:03am
AIR BLK	.00	11:04am
ACCY CHK	.07	11:04am
AIR BLK	.00	11:05am
SUB TEST	.00	11:06am
AIR BLK	.00	11:07am
SUB TEST	.00	11:08am
AIR BLK	.00	11:09am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY MOORE COUNTY JAIL 620

Serial Number: 008735 Test Record Number: 1765
Test Date: 03/16/2016 Test Time: 11:10am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:10am
FLO	Pass	11:10am
FC	Pass	11:10am

Temperature Tests

Test	Status	Time
FC1	Pass	11:11am
SRC	Pass	11:11am
DET	Pass	11:11am
BAR	Pass	11:11am
BT	Pass	11:11am

Blank Tests

Test	Status	Time
AIR	Pass	11:11am

Printer Tests


Test	Status	Time
PRNT	Pass	11:11am

CRC Tests

Test	Status	Time
COMP	Pass	11:11am
CAL	Pass	11:11am

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County MOORE Instrument Location PINEHURST POLICE DEPT.

Instrument Serial No. 008710 PINEHURST, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 01 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710
Test Date: 03/01/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	3:31pm
AIR BLK	.00	3:32pm
ACCY CHK	.07	3:32pm
AIR BLK	.00	3:33pm
SUB TEST	.00	3:34pm
AIR BLK	.00	3:35pm
SUB TEST	.00	3:36pm
AIR BLK	.00	3:37pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710 Test Record Number: 1263
Test Date: 03/01/2016 Test Time: 3:38pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:38pm
FLO	Pass	3:38pm
FC	Pass	3:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:38pm
SRC	Pass	3:38pm
DET	Pass	3:38pm
BAR	Pass	3:38pm
BT	Pass	3:38pm

Blank Tests

Test	Status	Time
AIR	Pass	3:39pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:39pm

CRC Tests

Test	Status	Time
COMP	Pass	3:39pm
CAL	Pass	3:39pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County MOORE Instrument Location ROBBINS POLICE DEPT.
Instrument Serial No. 008728 Robbins, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 07 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MOORE COUNTY ROBBINS PD 620

Serial Number: 008728
Test Date: 03/07/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	5:31pm
AIR BLK	.00	5:32pm
ACCY CHK	.08	5:32pm
AIR BLK	.00	5:33pm
SUB TEST	.00	5:34pm
AIR BLK	.00	5:35pm
SUB TEST	.00	5:36pm
AIR BLK	.00	5:37pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY ROBBINS PD 620

Serial Number: 008728 Test Record Number: 281
Test Date: 03/07/2016 Test Time: 5:39pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:39pm
FLO	Pass	5:39pm
FC	Pass	5:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:40pm
SRC	Pass	5:40pm
DET	Pass	5:40pm
BAR	Pass	5:40pm
BT	Pass	5:40pm

Blank Tests

Test	Status	Time
AIR	Pass	5:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:40pm

CRC Tests

Test	Status	Time
COMP	Pass	5:40pm
CAL	Pass	5:40pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Nash Co. Instrument Location Nashville P. D.

Instrument Serial No. 008630 501 S. BARNES ST Nashville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NASH COUNTY NASHVILLE PD 630

Serial Number: 008630
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434201

Exp Date: 12/08/2016


Test	g/210L	Time
DIAG	Pass	3:28pm
AIR BLK	.00	3:29pm
ACCY CHK	.08	3:30pm
AIR BLK	.00	3:31pm
SUB TEST	.00	3:31pm
AIR BLK	.00	3:32pm
SUB TEST	.00	3:34pm
AIR BLK	.00	3:35pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY NASHVILLE PD 630

Serial Number: 008630 Test Record Number: 3550
Test Date: 03/17/2016 Test Time: 3:35pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:36pm
FLO	Pass	3:36pm
FC	Pass	3:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:36pm
SRC	Pass	3:36pm
DET	Pass	3:36pm
BAR	Pass	3:36pm
BT	Pass	3:36pm

Blank Tests

Test	Status	Time
AIR	Pass	3:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:37pm

CRC Tests

Test	Status	Time
COMP	Pass	3:37pm
CAL	Pass	3:37pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Nash Co. Instrument Location Rocky Mount P.D.
Instrument Serial No. 008140 #1 Government Plaza Rocky Mount NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008740
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG409709

Exp Date: 04/07/2016


Test	g/210L	Time
DIAG	Pass	2:41pm
AIR BLK	.00	2:42pm
ACCY CHK	.08	2:43pm
AIR BLK	.00	2:44pm
SUB TEST	.00	2:45pm
AIR BLK	.00	2:46pm
SUB TEST	.00	2:47pm
AIR BLK	.00	2:48pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008740 Test Record Number: 571
Test Date: 03/17/2016 Test Time: 2:49pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:50pm
FLO	Pass	2:50pm
FC	Pass	2:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:50pm
SRC	Pass	2:50pm
DET	Pass	2:50pm
BAR	Pass	2:50pm
BT	Pass	2:50pm

Blank Tests

Test	Status	Time
AIR	Pass	2:50pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:51pm

CRC Tests

Test	Status	Time
COMP	Pass	2:51pm
CAL	Pass	2:51pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Nash Co. Instrument Location Rocky Mount P. D.

Instrument Serial No. 008141 # 1 Coker's Plaza Rocky Mount, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008741
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	2:42pm
AIR BLK	.00	2:43pm
ACCY CHK	.07	2:44pm
AIR BLK	.00	2:45pm
SUB TEST	.00	2:46pm
AIR BLK	.00	2:46pm
SUB TEST	.00	2:48pm
AIR BLK	.00	2:49pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008741 Test Record Number: 1840
Test Date: 03/17/2016 Test Time: 2:50pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:50pm
FLO	Pass	2:50pm
FC	Pass	2:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:50pm
SRC	Pass	2:50pm
DET	Pass	2:50pm
BAR	Pass	2:50pm
BT	Pass	2:50pm

Blank Tests

Test	Status	Time
AIR	Pass	2:51pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:51pm

CRC Tests

Test	Status	Time
COMP	Pass	2:51pm
CAL	Pass	2:51pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008575 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Bowers
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 9
640

Serial Number: 008575
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:03am
AIR BLK	.00	12:04am
ACCY CHK	.07	12:05am
AIR BLK	.00	12:06am
SUB TEST	.00	12:07am
AIR BLK	.00	12:08am
SUB TEST	.00	12:09am
AIR BLK	.00	12:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008575 Test Record Number: 901
Test Date: 03/19/2016 Test Time: 12:11am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:11am
FLO	Pass	12:11am
FC	Pass	12:11am

Temperature Tests

Test	Status	Time
FC1	Pass	12:11am
SRC	Pass	12:11am
DET	Pass	12:11am
BAR	Pass	12:11am
BT	Pass	12:11am

Blank Tests

Test	Status	Time
AIR	Pass	12:12am

Printer Tests

Test	Status	Time
PRNT	Pass	12:12am

CRC Tests

Test	Status	Time
COMP	Pass	12:12am
CAL	Pass	12:12am

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008647 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Benson

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 9
640

Serial Number: 008647
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	12:05am
AIR BLK	.00	12:06am
ACCY CHK	.07	12:06am
AIR BLK	.00	12:07am
SUB TEST	.00	12:08am
AIR BLK	.00	12:09am
SUB TEST	.00	12:10am
AIR BLK	.00	12:11am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008647 Test Record Number: 2197
Test Date: 03/19/2016 Test Time: 12:12am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:12am
FLO	Pass	12:12am
FC	Pass	12:12am

Temperature Tests

Test	Status	Time
FC1	Pass	12:12am
SRC	Pass	12:12am
DET	Pass	12:12am
BAR	Pass	12:12am
BT	Pass	12:12am

Blank Tests

Test	Status	Time
AIR	Pass	12:13am

Printer Tests

Test	Status	Time
PRNT	Pass	12:13am

CRC Tests

Test	Status	Time
COMP	Pass	12:13am
CAL	Pass	12:13am

Preventive Maintenance
Status: Pass

Alvin Ray Barnes
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008616 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Deane
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 9
640

Serial Number: 008616
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:07am
AIR BLK	.00	12:07am
ACCY CHK	.08	12:08am
AIR BLK	.00	12:09am
SUB TEST	.00	12:09am
AIR BLK	.00	12:10am
SUB TEST	.00	12:12am
AIR BLK	.00	12:13am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008616 Test Record Number: 2189
Test Date: 03/19/2016 Test Time: 12:17am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:18am
FLO	Pass	12:18am
FC	Pass	12:18am

Temperature Tests

Test	Status	Time
FC1	Pass	12:18am
SRC	Pass	12:18am
DET	Pass	12:18am
BAR	Pass	12:18am
BT	Pass	12:18am

Blank Tests

Test	Status	Time
AIR	Pass	12:19am

Printer Tests

Test	Status	Time
PRNT	Pass	12:19am

CRC Tests

Test	Status	Time
COMP	Pass	12:19am
CAL	Pass	12:19am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County NEW HANOVER Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008826 WILMINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of MARCH, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alan R. Benson
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY BAT MOBILE UNIT 9
640

Serial Number: 008826
Test Date: 03/19/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R
Permit Number: 15671E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:04am
AIR BLK	.00	12:05am
ACCY CHK	.08	12:06am
AIR BLK	.00	12:06am
SUB TEST	.00	12:07am
AIR BLK	.00	12:08am
SUB TEST	.00	12:09am
AIR BLK	.00	12:10am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Alvin Ray Barnes
Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY BAT MOBILE UNIT 9 640

Serial Number: 008826 Test Record Number: 7904
Test Date: 03/19/2016 Test Time: 12:11am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11am
FLO	Pass	12:11am
FC	Pass	12:12am

Temperature Tests

Test	Status	Time
FC1	Pass	12:12am
SRC	Pass	12:12am
DET	Pass	12:12am
BAR	Pass	12:12am
BT	Pass	12:12am

Blank Tests

Test	Status	Time
AIR	Pass	12:12am

Printer Tests

Test	Status	Time
PRNT	Pass	12:12am

CRC Tests

Test	Status	Time
COMP	Pass	12:13am
CAL	Pass	12:13am

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Wrightsville Beach
Instrument Serial No. 008667 Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhoda
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD
640

Serial Number: 008667
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:24pm
AIR BLK	.00	2:24pm
ACCY CHK	.08	2:25pm
AIR BLK	.00	2:26pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:28pm
AIR BLK	.00	2:29pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WRIGHTSVILLE BCH PD 640

Serial Number: 008667 Test Record Number: 1472
Test Date: 03/15/2016 Test Time: 2:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:30pm
FLO	Pass	2:30pm
FC	Pass	2:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:31pm
SRC	Pass	2:31pm
DET	Pass	2:31pm
BAR	Pass	2:31pm
BT	Pass	2:31pm

Blank Tests

Test	Status	Time
AIR	Pass	2:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:31pm

CRC Tests

Test	Status	Time
COMP	Pass	2:31pm
CAL	Pass	2:31pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Carolina Beach
Instrument Serial No. 008661 Police Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature] 601
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY CAROLINA BEACH PD
640

Serial Number: 008661
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG434901
Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	1:25pm
AIR BLK	.00	1:26pm
ACCY CHK	.07	1:26pm
AIR BLK	.00	1:27pm
SUB TEST	.00	1:28pm
AIR BLK	.00	1:29pm
SUB TEST	.00	1:30pm
AIR BLK	.00	1:31pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY CAROLINA BEACH PD 640

Serial Number: 008661 Test Record Number: 2128
Test Date: 03/15/2016 Test Time: 1:32pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:32pm
FLO	Pass	1:32pm
FC	Pass	1:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:33pm
SRC	Pass	1:33pm
DET	Pass	1:33pm
BAR	Pass	1:33pm
BT	Pass	1:33pm

Blank Tests

Test	Status	Time
AIR	Pass	1:33pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:33pm

CRC Tests

Test	Status	Time
COMP	Pass	1:34pm
CAL	Pass	1:34pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location Wilmington
Instrument Serial No. 008628 Police Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature] 681
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: 008628
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 04/11/2017

Test	g/210L	Time
DIAG	Pass	12:22pm
AIR BLK	.00	12:22pm
ACCY CHK	.07	12:23pm
AIR BLK	.00	12:24pm
SUB TEST	.00	12:25pm
AIR BLK	.00	12:26pm
SUB TEST	.00	12:27pm
AIR BLK	.00	12:28pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY WILMINGTON PD 640

Serial Number: 008628 Test Record Number: 3763
Test Date: 03/15/2016 Test Time: 12:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:29pm
FLO	Pass	12:29pm
FC	Pass	12:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:30pm
SRC	Pass	12:30pm
DET	Pass	12:30pm
BAR	Pass	12:30pm
BT	Pass	12:30pm

Blank Tests

Test	Status	Time
AIR	Pass	12:30pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:30pm

CRC Tests

Test	Status	Time
COMP	Pass	12:30pm
CAL	Pass	12:30pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location New Hanover

Instrument Serial No. 008617 County Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhoads
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD
640

Serial Number: 008617
Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901


Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	11:34am
AIR BLK	.00	11:34am
ACCY CHK	.07	11:35am
AIR BLK	.00	11:36am
SUB TEST	.00	11:37am
AIR BLK	.00	11:38am
SUB TEST	.00	11:39am
AIR BLK	.00	11:40am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: 008617 Test Record Number: 2565
Test Date: 03/15/2016 Test Time: 11:41am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:41am
FLO	Pass	11:41am
FC	Pass	11:41am

Temperature Tests

Test	Status	Time
FC1	Pass	11:41am
SRC	Pass	11:41am
DET	Pass	11:41am
BAR	Pass	11:41am
BT	Pass	11:41am

Blank Tests

Test	Status	Time
AIR	Pass	11:42am


Printer Tests

Test	Status	Time
PRNT	Pass	11:42am

CRC Tests

Test	Status	Time
COMP	Pass	11:42am
CAL	Pass	11:42am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County New Hanover Instrument Location New Hanover
Instrument Serial No. 008626 County Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Phelan
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NEW HANOVER COUNTY NEW HANOVER CO SD
640

Serial Number: 008626
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:31am
AIR BLK	.00	11:32am
ACCY CHK	.07	11:33am
AIR BLK	.00	11:34am
SUB TEST	.00	11:34am
AIR BLK	.00	11:36am
SUB TEST	.00	11:37am
AIR BLK	.00	11:38am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

NEW HANOVER COUNTY NEW HANOVER CO SD 640

Serial Number: 008626 Test Record Number: 6340
Test Date: 03/15/2016 Test Time: 11:39am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:39am
FLO	Pass	11:39am
FC	Pass	11:39am

Temperature Tests

Test	Status	Time
FC1	Pass	11:39am
SRC	Pass	11:39am
DET	Pass	11:39am
BAR	Pass	11:39am
BT	Pass	11:39am

Blank Tests

Test	Status	Time
AIR	Pass	11:40am

Printer Tests

Test	Status	Time
PRNT	Pass	11:40am

CRC Tests

Test	Status	Time
COMP	Pass	11:40am
CAL	Pass	11:40am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Northampton Co. Instrument Location Northampton Co Sheriff Dept.

Instrument Serial No. 008605 105 W. Jefferson St. Jackson, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT
650

Serial Number: 008607
Test Date: 03/08/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, GRAYHAM C
Permit Number: 7682E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG411202
Exp Date: 04/22/2016

Test	g/210L	Time
DIAG	Pass	10:56am
AIR BLK	.00	10:57am
ACCY CHK	.07	10:57am
AIR BLK	.00	10:58am
SUB TEST	.00	10:59am
AIR BLK	.00	11:00am
SUB TEST	.00	11:02am
AIR BLK	.00	11:02am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

NORTHAMPTON COUNTY SHERIFFS DEPARTMENT 650

Serial Number: 008607 Test Record Number: 824
Test Date: 03/08/2016 Test Time: 11:04am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	11:04am
FLO	Pass	11:04am
FC	Pass	11:04am

Temperature Tests

Test	Status	Time
FC1	Pass	11:04am
SRC	Pass	11:04am
DET	Pass	11:04am
BAR	Pass	11:04am
BT	Pass	11:04am

Blank Tests

Test	Status	Time
AIR	Pass	11:05am


Printer Tests

Test	Status	Time
PRNT	Pass	11:05am

CRC Tests

Test	Status	Time
COMP	Pass	11:05am
CAL	Pass	11:05am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE Instrument Location HILLSBOROUGH PD

Instrument Serial No. 008799 127 N. CHURTON ST. HILLSBOROUGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	3:38pm
AIR BLK	.00	3:39pm
ACCY CHK	.08	3:40pm
AIR BLK	.00	3:41pm
SUB TEST	.00	3:42pm
AIR BLK	.00	3:43pm
SUB TEST	.00	3:44pm
AIR BLK	.00	3:45pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008799 Test Record Number: 2092
Test Date: 03/16/2016 Test Time: 3:47pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:47pm
FLO	Pass	3:47pm
FC	Pass	3:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:47pm
SRC	Pass	3:47pm
DET	Pass	3:47pm
BAR	Pass	3:47pm
BT	Pass	3:47pm

Blank Tests

Test	Status	Time
AIR	Pass	3:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:48pm

CRC Tests

Test	Status	Time
COMP	Pass	3:48pm
CAL	Pass	3:48pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ORANGE Instrument Location CHAPEL HILL PD

Instrument Serial No. 008839 828 MARTIN LUTHER KING, JR BLVD
CHAPEL HILL, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Russell Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:12pm
AIR BLK	.00	2:13pm
ACCY CHK	.08	2:14pm
AIR BLK	.00	2:15pm
SUB TEST	.00	2:16pm
AIR BLK	.00	2:16pm
SUB TEST	.00	2:18pm
AIR BLK	.00	2:19pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839 Test Record Number: 1440
Test Date: 03/16/2016 Test Time: 2:21pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:21pm
FLO	Pass	2:21pm
FC	Pass	2:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:21pm
SRC	Pass	2:21pm
DET	Pass	2:21pm
BAR	Pass	2:21pm
BT	Pass	2:21pm

Blank Tests

Test	Status	Time
AIR	Pass	2:22pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:22pm

CRC Tests

Test	Status	Time
COMP	Pass	2:22pm
CAL	Pass	2:22pm

Preventive Maintenance
Status: Pass

Bruce D. Smith

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ORANGE Instrument Location CHAPEL HILL PD

Instrument Serial No. 008856 828 MARTIN LUTHER KING JR BLVD
CHAPEL HILL, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bud D. Arnold
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008856

Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:11pm
AIR BLK	.00	2:12pm
ACCY CHK	.08	2:13pm
AIR BLK	.00	2:14pm
SUB TEST	.00	2:15pm
AIR BLK	.00	2:16pm
SUB TEST	.00	2:17pm
AIR BLK	.00	2:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008856 Test Record Number: 1944
Test Date: 03/16/2016 Test Time: 2:21pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:21pm
FLO	Pass	2:21pm
FC	Pass	2:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:21pm
SRC	Pass	2:21pm
DET	Pass	2:21pm
BAR	Pass	2:21pm
BT	Pass	2:21pm

Blank Tests

Test	Status	Time
AIR	Pass	2:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:22pm

CRC Tests

Test	Status	Time
COMP	Pass	2:22pm
CAL	Pass	2:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ORANGE Instrument Location HILLSBOROUGH PD

Instrument Serial No. 208651 127 N. CHURTON ST. HILLSBOROUGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bud D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008651

Test Date: 03/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG525303

Exp Date: 09/10/2017


Test	g/210L	Time
DIAG	Pass	11:00am
AIR BLK	.00	11:00am
ACCY CHK	.08	11:01am
AIR BLK	.00	11:02am
SUB TEST	.00	11:02am
AIR BLK	.00	11:03am
SUB TEST	.00	11:05am
AIR BLK	.00	11:06am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY HILLSBOROUGH PD 670

Serial Number: 008651 Test Record Number: 1209
Test Date: 03/30/2016 Test Time: 11:08am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:09am
FLO	Pass	11:09am
FC	Pass	11:09am

Temperature Tests

Test	Status	Time
FC1	Pass	11:09am
SRC	Pass	11:09am
DET	Pass	11:09am
BAR	Pass	11:09am
BT	Pass	11:09am

Blank Tests

Test	Status	Time
AIR	Pass	11:10am


Printer Tests

Test	Status	Time
PRNT	Pass	11:10am

CRC Tests

Test	Status	Time
COMP	Pass	11:10am
CAL	Pass	11:10am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County Onslow Instrument Location Camp Lejeune APO

Instrument Serial No. 008922

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONSLOW COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008922
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HALL, RANDY E
Permit Number: 3462E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	10:26am
AIR BLK	.00	10:26am
ACCY CHK	.08	10:27am
AIR BLK	.00	10:28am
SUB TEST	.00	10:28am
AIR BLK	.00	10:29am
SUB TEST	.00	10:31am
AIR BLK	.00	10:32am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008922 Test Record Number: 301
Test Date: 03/15/2016 Test Time: 10:32am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:32am
FLO	Pass	10:32am
FC	Pass	10:33am

Temperature Tests

Test	Status	Time
FC1	Pass	10:33am
SRC	Pass	10:33am
DET	Pass	10:33am
BAR	Pass	10:33am
BT	Pass	10:33am

Blank Tests

Test	Status	Time
AIR	Pass	10:33am

Printer Tests

Test	Status	Time
PRNT	Pass	10:33am

CRC Tests

Test	Status	Time
COMP	Pass	10:34am
CAL	Pass	10:34am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Onslow Instrument Location Jacksonville PA

Instrument Serial No. 008930

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY JACKSONVILLE PD 660

Serial Number: 008930
Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:17am
AIR BLK	.00	11:18am
ACCY CHK	.08	11:18am
AIR BLK	.00	11:19am
SUB TEST	.00	11:20am
AIR BLK	.00	11:21am
SUB TEST	.00	11:22am
AIR BLK	.00	11:23am

Reported AC: .00 g/210L

RE Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY JACKSONVILLE PD 660

Serial Number: 008930 Test Record Number: 2020
Test Date: 03/15/2016 Test Time: 11:26am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:27am
FLO	Pass	11:27am
FC	Pass	11:27am

Temperature Tests

Test	Status	Time
FC1	Pass	11:27am
SRC	Pass	11:27am
DET	Pass	11:27am
BAR	Pass	11:27am
BT	Pass	11:27am

Blank Tests

Test	Status	Time
AIR	Pass	11:27am


Printer Tests

Test	Status	Time
PRNT	Pass	11:27am

CRC Tests

Test	Status	Time
COMP	Pass	11:28am
CAL	Pass	11:28am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Owenslow Instrument Location Owenslow County
Instrument Serial No. 008932 Sheriff's Office

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Raney E. Hall

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY ONslow COUNTY SD 660

Serial Number: 008932
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	11:58am
AIR BLK	.00	11:58am
ACCY CHK	.08	11:59am
AIR BLK	.00	12:00pm
SUB TEST	.00	12:01pm
AIR BLK	.00	12:02pm
SUB TEST	.00	12:03pm
AIR BLK	.00	12:04pm

Reported AC: .00 g/210L

RE Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall

Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008932 Test Record Number: 3480
Test Date: 03/15/2016 Test Time: 12:05pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:06pm
FLO	Pass	12:06pm
FC	Pass	12:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:06pm
SRC	Pass	12:06pm
DET	Pass	12:06pm
BAR	Pass	12:06pm
BT	Pass	12:06pm

Blank Tests

Test	Status	Time
AIR	Pass	12:06pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:06pm

CRC Tests

Test	Status	Time
COMP	Pass	12:07pm
CAL	Pass	12:07pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Onslow Instrument Location Onslow County
Instrument Serial No. 008931 SHERIFF'S OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Rouge Hall

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY ONslow COUNTY SD 660

Serial Number: 008931
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	12:11pm
AIR BLK	.00	12:12pm
ACCY CHK	.08	12:13pm
AIR BLK	.00	12:14pm
SUB TEST	.00	12:15pm
AIR BLK	.00	12:15pm
SUB TEST	.00	12:17pm
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L

R E Hall

Signature of Chemical Analyst

Court CVR

Randy E Hall
Analyst

Intox EC/IR-II: Preventive Maintenance

ONLOW COUNTY ONSLOW COUNTY SD 660

Serial Number: 008931 Test Record Number: 2380
Test Date: 03/15/2016 Test Time: 12:18pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:19pm
FLO	Pass	12:19pm
FC	Pass	12:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:19pm
SRC	Pass	12:19pm
DET	Pass	12:19pm
BAR	Pass	12:19pm
BT	Pass	12:19pm

Blank Tests

Test	Status	Time
AIR	Pass	12:19pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:19pm

CRC Tests

Test	Status	Time
COMP	Pass	12:20pm
CAL	Pass	12:20pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ONSLOW


Instrument Location MCA5 New River AAO

Instrument Serial No. 008919

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy Hall

Signature of Certifying Official

354

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY MCAS NEW RIVER 660

Serial Number: 008919
Test Date: 03/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017


Test	g/210L	Time
DIAG	Pass	12:48pm
AIR BLK	.00	12:49pm
ACCY CHK	.08	12:49pm
AIR BLK	.00	12:50pm
SUB TEST	.00	12:51pm
AIR BLK	.00	12:52pm
SUB TEST	.00	12:53pm
AIR BLK	.00	12:54pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY MCAS NEW RIVER 660

Serial Number: 008919 Test Record Number: 512
Test Date: 03/15/2016 Test Time: 12:55pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:55pm
FLO	Pass	12:55pm
FC	Pass	12:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:55pm
SRC	Pass	12:55pm
DET	Pass	12:55pm
BAR	Pass	12:55pm
BT	Pass	12:55pm

Blank Tests

Test	Status	Time
AIR	Pass	12:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:56pm

CRC Tests

Test	Status	Time
COMP	Pass	12:56pm
CAL	Pass	12:56pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ON SLOW Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008616 RICHMOND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Albert B.
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008616
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:31pm
AIR BLK	.00	11:32pm
ACCY CHK	.08	11:32pm
AIR BLK	.00	11:33pm
SUB TEST	.00	11:34pm
AIR BLK	.00	11:35pm
SUB TEST	.00	11:36pm
AIR BLK	.00	11:37pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

ONslow COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008616 Test Record Number: 2179
Test Date: 03/11/2016 Test Time: 11:42pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:42pm
FLO	Pass	11:42pm
FC	Pass	11:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:43pm
SRC	Pass	11:43pm
DET	Pass	11:43pm
BAR	Pass	11:43pm
BT	Pass	11:43pm

Blank Tests

Test	Status	Time
AIR	Pass	11:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:43pm

CRC Tests

Test	Status	Time
COMP	Pass	11:43pm
CAL	Pass	11:43pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ONslow Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008909 RICHMOND, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alvin B.
Signature of Certifying Official

648
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONSLOW COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008707
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	11:27pm
AIR BLK	.00	11:28pm
ACCY CHK	.08	11:29pm
AIR BLK	.00	11:30pm
SUB TEST	.00	11:31pm
AIR BLK	.00	11:32pm
SUB TEST	.00	11:33pm
AIR BLK	.00	11:34pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLow COUNTY BAT MOBILE UNIT 9 660

Serial Number: 008707 Test Record Number: 2273
Test Date: 03/11/2016 Test Time: 11:34pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:35pm
FLO	Pass	11:35pm
FC	Pass	11:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:35pm
SRC	Pass	11:35pm
DET	Pass	11:35pm
BAR	Pass	11:35pm
BT	Pass	11:35pm

Blank Tests

Test	Status	Time
AIR	Pass	11:36pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:36pm

CRC Tests

Test	Status	Time
COMP	Pass	11:36pm
CAL	Pass	11:36pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ONSLOW


Instrument Location CAMP LEJUEVE PMO

Instrument Serial No. 008920

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 13 day of APRIL, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ONslow COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920
Test Date: 04/13/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HALL, RANDY E
Permit Number: 3462E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534902
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:43am
AIR BLK	.00	10:43am
ACCY CHK	.08	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:45am
AIR BLK	.00	10:46am
SUB TEST	.00	10:47am
AIR BLK	.00	10:48am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ONSLOW COUNTY CAMP LEJEUNE PMO 660

Serial Number: 008920 Test Record Number: 1240
Test Date: 04/13/2016 Test Time: 10:49am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:50am
FLO	Pass	10:50am
FC	Pass	10:50am

Temperature Tests

Test	Status	Time
FC1	Pass	10:50am
SRC	Pass	10:50am
DET	Pass	10:50am
BAR	Pass	10:50am
BT	Pass	10:50am

Blank Tests

Test	Status	Time
AIR	Pass	10:50am


Printer Tests

Test	Status	Time
PRNT	Pass	10:50am

CRC Tests

Test	Status	Time
COMP	Pass	10:51am
CAL	Pass	10:51am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pamlico Instrument Location Pamlico County
Instrument Serial No. 008648 SHERIFFS OFFICE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Randy E. Hall
Signature of Certifying Official

354
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: 008640
Test Date: 03/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HALL, RANDY E

Permit Number: 3462E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017


Test	g/210L	Time
DIAG	Pass	1:14pm
AIR BLK	.00	1:15pm
ACCY CHK	.08	1:15pm
AIR BLK	.00	1:16pm
SUB TEST	.00	1:17pm
AIR BLK	.00	1:18pm
SUB TEST	.00	1:19pm
AIR BLK	.00	1:20pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PAMLICO COUNTY PAMLICO COUNTY SD 680

Serial Number: 008640 Test Record Number: 1168
Test Date: 03/17/2016 Test Time: 1:21pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:21pm
FLO	Pass	1:21pm
FC	Pass	1:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:21pm
SRC	Pass	1:21pm
DET	Pass	1:21pm
BAR	Pass	1:21pm
BT	Pass	1:21pm

Blank Tests

Test	Status	Time
AIR	Pass	1:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:22pm

CRC Tests

Test	Status	Time
COMP	Pass	1:22pm
CAL	Pass	1:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County Pender Instrument Location Pender County

Instrument Serial No. 008948 Sheriff Dept. Anson

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K.C. Rhoder
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: 008948
Test Date: 03/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:35am
AIR BLK	.00	10:36am
ACCY CHK	.07	10:37am
AIR BLK	.00	10:38am
SUB TEST	.00	10:38am
AIR BLK	.00	10:39am
SUB TEST	.00	10:40am
AIR BLK	.00	10:41am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PENDER COUNTY SHERIFF DEPT ANNEX 700

Serial Number: 008948 Test Record Number: 721
Test Date: 03/15/2016 Test Time: 10:42am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:42am
FLO	Pass	10:42am
FC	Pass	10:42am

Temperature Tests

Test	Status	Time
FC1	Pass	10:42am
SRC	Pass	10:42am
DET	Pass	10:42am
BAR	Pass	10:42am
BT	Pass	10:42am

Blank Tests

Test	Status	Time
AIR	Pass	10:43am


Printer Tests

Test	Status	Time
PRNT	Pass	10:43am

CRC Tests

Test	Status	Time
COMP	Pass	10:43am
CAL	Pass	10:43am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pender Instrument Location Pender County
Instrument Serial No. 008946 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PENDER PENDER COUNTY, SD 700

Serial Number: 008946

Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NPNE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	3:28pm
AIR BLK	.00	3:29pm
ACCY CHK	.08	3:29pm
AIR BLK	.00	3:30pm
SUB TEST	.00	3:31pm
AIR BLK	.00	3:32pm
SUB TEST	.00	3:33pm
AIR BLK	.00	3:34pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER PENDER COUNTY SD 700

Serial Number: 008946 Test Record Number: 788
Test Date: 03/16/2016 Test Time: 3:42pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:42pm
FLO	Pass	3:42pm
FC	Pass	3:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:42pm
SRC	Pass	3:42pm
DET	Pass	3:42pm
BAR	Pass	3:42pm
BT	Pass	3:42pm

Blank Tests

Test	Status	Time
AIR	Pass	3:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:43pm

CRC Tests

Test	Status	Time
COMP	Pass	3:43pm
CAL	Pass	3:43pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pender Instrument Location Pender County
Instrument Serial No. 008935 Sheriff Dept.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. Rhoads
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PENDER PENDER CO SD 700

Serial Number: 008935
Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	3:26pm
AIR BLK	.00	3:27pm
ACCY CHK	.07	3:28pm
AIR BLK	.00	3:29pm
SUB TEST	.00	3:30pm
AIR BLK	.00	3:31pm
SUB TEST	.00	3:33pm
AIR BLK	.00	3:34pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PENDER PENDER CO SD 700

Serial Number: 008935 Test Record Number: 1836
Test Date: 03/16/2016 Test Time: 3:35pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:35pm
FLO	Pass	3:35pm
FC	Pass	3:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:35pm
SRC	Pass	3:35pm
DET	Pass	3:35pm
BAR	Pass	3:35pm
BT	Pass	3:35pm

Blank Tests

Test	Status	Time
AIR	Pass	3:36pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:36pm

CRC Tests

Test	Status	Time
COMP	Pass	3:36pm
CAL	Pass	3:36pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Perquimans Instrument Location Perquimans Co. S.O.
Instrument Serial No. 008921 110 N. Church St., Hertford, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key M. O.

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921
Test Date: 03/30/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	11:16am
AIR BLK	.00	11:17am
ACCY CHK	.08	11:17am
AIR BLK	.00	11:18am
SUB TEST	.00	11:19am
AIR BLK	.00	11:20am
SUB TEST	.00	11:22am
AIR BLK	.00	11:23am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921 Test Record Number: 611
Test Date: 03/30/2016 Test Time: 11:24am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:25am
FLO	Pass	11:25am
FC	Pass	11:25am

Temperature Tests

Test	Status	Time
FC1	Pass	11:25am
SRC	Pass	11:25am
DET	Pass	11:25am
BAR	Pass	11:25am
BT	Pass	11:25am

Blank Tests

Test	Status	Time
AIR	Pass	11:26am


Printer Tests

Test	Status	Time
PRNT	Pass	11:26am

CRC Tests

Test	Status	Time
COMP	Pass	11:26am
CAL	Pass	11:26am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County PERSON Instrument Location PERSON CO. LEC

Instrument Serial No. 008880 170 COURT ST. ROXBORO NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bud D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880

Test Date: 03/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:13pm
AIR BLK	.00	12:14pm
ACCY CHK	.08	12:15pm
AIR BLK	.00	12:16pm
SUB TEST	.00	12:16pm
AIR BLK	.00	12:17pm
SUB TEST	.00	12:19pm
AIR BLK	.00	12:20pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008880 Test Record Number: 945
Test Date: 03/09/2016 Test Time: 12:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:21pm
FLO	Pass	12:21pm
FC	Pass	12:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:21pm
SRC	Pass	12:21pm
DET	Pass	12:21pm
BAR	Pass	12:21pm
BT	Pass	12:21pm

Blank Tests

Test	Status	Time
AIR	Pass	12:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:22pm

CRC Tests

Test	Status	Time
COMP	Pass	12:22pm
CAL	Pass	12:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County PERSON Instrument Location PERSON CO. LEC

Instrument Serial No. 008693 120 COURT ST. ROXBORO, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 09 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bud D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693

Test Date: 03/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:12pm
AIR BLK	.00	12:13pm
ACCY CHK	.07	12:14pm
AIR BLK	.00	12:14pm
SUB TEST	.00	12:15pm
AIR BLK	.00	12:16pm
SUB TEST	.00	12:17pm
AIR BLK	.00	12:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PERSON COUNTY PERSON CO. LEC 720

Serial Number: 008693 Test Record Number: 1217
Test Date: 03/09/2016 Test Time: 12:20pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:20pm
FLO	Pass	12:20pm
FC	Pass	12:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:20pm
SRC	Pass	12:20pm
DET	Pass	12:20pm
BAR	Pass	12:20pm
BT	Pass	12:20pm

Blank Tests

Test	Status	Time
AIR	Pass	12:21pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:21pm

CRC Tests

Test	Status	Time
COMP	Pass	12:21pm
CAL	Pass	12:21pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Pitt Instrument Location Pitt Co. Detention Center

Instrument Serial No. 008662 124 Detention Dr., Greenville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008662
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:41am
AIR BLK	.00	11:41am
ACCY CHK	.08	11:42am
AIR BLK	.00	11:43am
SUB TEST	.00	11:44am
AIR BLK	.00	11:45am
SUB TEST	.00	11:47am
AIR BLK	.00	11:47am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008662 Test Record Number: 895
Test Date: 03/16/2016 Test Time: 11:48am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:49am
FLO	Pass	11:49am
FC	Pass	11:49am

Temperature Tests

Test	Status	Time
FC1	Pass	11:49am
SRC	Pass	11:49am
DET	Pass	11:49am
BAR	Pass	11:49am
BT	Pass	11:49am

Blank Tests

Test	Status	Time
AIR	Pass	11:50am

Printer Tests

Test	Status	Time
PRNT	Pass	11:50am

CRC Tests

Test	Status	Time
COMP	Pass	11:50am
CAL	Pass	11:50am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. 0081668 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG414801
Exp Date: 05/28/2016

Test	g/210L	Time
DIAG	Pass	11:25am
AIR BLK	.00	11:26am
ACCY CHK	.07	11:26am
AIR BLK	.00	11:27am
SUB TEST	.00	11:28am
AIR BLK	.00	11:29am
SUB TEST	.00	11:30am
AIR BLK	.00	11:31am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668 Test Record Number: 2691
Test Date: 03/16/2016 Test Time: 11:32am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:32am
FLO	Pass	11:32am
FC	Pass	11:32am

Temperature Tests

Test	Status	Time
FC1	Pass	11:32am
SRC	Pass	11:32am
DET	Pass	11:32am
BAR	Pass	11:32am
BT	Pass	11:32am

Blank Tests

Test	Status	Time
AIR	Pass	11:33am


Printer Tests

Test	Status	Time
PRNT	Pass	11:33am

CRC Tests

Test	Status	Time
COMP	Pass	11:33am
CAL	Pass	11:33am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RANDOLPH Instrument Location RANDOLPH CO. JAIL
Instrument Serial No. 008819 Asheboro NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY RANDOLPH CO. JAIL 750

Serial Number: 008899
Test Date: 03/11/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401


Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	5:24pm
AIR BLK	.00	5:25pm
ACCY CHK	.08	5:26pm
AIR BLK	.00	5:27pm
SUB TEST	.00	5:28pm
AIR BLK	.00	5:28pm
SUB TEST	.00	5:30pm
AIR BLK	.00	5:31pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY RANDOLPH CO. JAIL 750

Serial Number: 008899 Test Record Number: 2289
Test Date: 03/11/2016 Test Time: 5:32pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:32pm
FLO	Pass	5:32pm
FC	Pass	5:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:32pm
SRC	Pass	5:32pm
DET	Pass	5:32pm
BAR	Pass	5:32pm
BT	Pass	5:32pm

Blank Tests

Test	Status	Time
AIR	Pass	5:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:33pm

CRC Tests

Test	Status	Time
COMP	Pass	5:33pm
CAL	Pass	5:33pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RANDOLPH Instrument Location RANDOLPH Co. JAIL
Instrument Serial No. 00 8860 Asheboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY RANDOLPH COUNTY JAIL
750

Serial Number: 008860
Test Date: 03/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	5:34pm
AIR BLK	.00	5:35pm
ACCY CHK	.08	5:35pm
AIR BLK	.00	5:36pm
SUB TEST	.00	5:37pm
AIR BLK	.00	5:38pm
SUB TEST	.00	5:39pm
AIR BLK	.00	5:40pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY RANDOLPH COUNTY JAIL 750

Serial Number: 008860 Test Record Number: 2378
Test Date: 03/11/2016 Test Time: 5:42pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:42pm
FLO	Pass	5:42pm
FC	Pass	5:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:42pm
SRC	Pass	5:42pm
DET	Pass	5:42pm
BAR	Pass	5:42pm
BT	Pass	5:42pm

Blank Tests

Test	Status	Time
AIR	Pass	5:43pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:43pm

CRC Tests

Test	Status	Time
COMP	Pass	5:43pm
CAL	Pass	5:43pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Bot Mobile Unit 8
Instrument Serial No 008814 Asheboro PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

644

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH CO BAT MOBILE UNIT 8 750

Serial Number: 008816
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403


Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	9:37pm
AIR BLK	.00	9:38pm
ACCY CHK	.08	9:38pm
AIR BLK	.00	9:39pm
SUB TEST	.00	9:39pm
AIR BLK	.00	9:40pm
SUB TEST	.00	9:42pm
AIR BLK	.00	9:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH CO BAT MOBILE UNIT 8 750

Serial Number: 008816 Test Record Number: 7194
Test Date: 03/11/2016 Test Time: 9:44pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:44pm
FLO	Pass	9:44pm
FC	Pass	9:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:45pm
SRC	Pass	9:45pm
DET	Pass	9:45pm
BAR	Pass	9:45pm
BT	Pass	9:45pm

Blank Tests

Test	Status	Time
AIR	Pass	9:45pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:45pm

CRC Tests

Test	Status	Time
COMP	Pass	9:45pm
CAL	Pass	9:45pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Randolph Instrument Location Arldale
Instrument Serial No. 008791 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:37pm
AIR BLK	.00	2:38pm
ACCY CHK	.08	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	.00	2:41pm
AIR BLK	.00	2:42pm
SUB TEST	.00	2:43pm
AIR BLK	.00	2:45pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY ARCHDALE PD 750

Serial Number: 008791 Test Record Number: 1108
Test Date: 03/18/2016 Test Time: 2:46pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:47pm
FLO	Pass	2:47pm
FC	Pass	2:47pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:47pm
SRC	Pass	2:47pm
DET	Pass	2:47pm
BAR	Pass	2:47pm
BT	Pass	2:47pm

Blank Tests

Test	Status	Time
AIR	Pass	2:48pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:48pm

CRC Tests

Test	Status	Time
COMP	Pass	2:48pm
CAL	Pass	2:48pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rowan Instrument Location Salisbury
Instrument Serial No. 008835 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835
Test Date: 03/23/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	11:24am
AIR BLK	.00	11:25am
ACCY CHK	.07	11:25am
AIR BLK	.00	11:26am
SUB TEST	.00	11:27am
AIR BLK	.00	11:28am
SUB TEST	.00	11:29am
AIR BLK	.00	11:30am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ROWAN COUNTY SALISBURY PD 790

Serial Number: 008835 Test Record Number: 1767
Test Date: 03/23/2016 Test Time: 11:31am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:31am
FLO	Pass	11:31am
FC	Pass	11:31am

Temperature Tests

Test	Status	Time
FC1	Pass	11:31am
SRC	Pass	11:31am
DET	Pass	11:31am
BAR	Pass	11:31am
BT	Pass	11:31am

Blank Tests

Test	Status	Time
AIR	Pass	11:32am


Printer Tests

Test	Status	Time
PRNT	Pass	11:32am

CRC Tests

Test	Status	Time
COMP	Pass	11:32am
CAL	Pass	11:32am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Reidsville
Instrument Serial No. 008784 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: 008784
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG534802
Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	3:40pm
AIR BLK	.00	3:41pm
ACCY CHK	.08	3:41pm
AIR BLK	.00	3:43pm
SUB TEST	.00	3:43pm
AIR BLK	.00	3:44pm
SUB TEST	.00	3:46pm
AIR BLK	.00	3:47pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY REIDSVILLE PD 780

Serial Number: 008784 Test Record Number: 918
Test Date: 03/17/2016 Test Time: 3:47pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:48pm
FLO	Pass	3:48pm
FC	Pass	3:48pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:48pm
SRC	Pass	3:48pm
DET	Pass	3:48pm
BAR	Pass	3:48pm
BT	Pass	3:48pm

Blank Tests

Test	Status	Time
AIR	Pass	3:49pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:49pm

CRC Tests

Test	Status	Time
COMP	Pass	3:49pm
CAL	Pass	3:49pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Rockingham Instrument Location Eden
Instrument Serial No. 008718 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008718
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG517501
Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	2:16pm
AIR BLK	.00	2:17pm
ACCY CHK	.08	2:17pm
AIR BLK	.00	2:19pm
SUB TEST	.00	2:19pm
AIR BLK	.00	2:20pm
SUB TEST	.00	2:21pm
AIR BLK	.00	2:23pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008718 Test Record Number: 1449
Test Date: 03/17/2016 Test Time: 2:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:13pm
FLO	Pass	2:13pm
FC	Pass	2:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:14pm
SRC	Pass	2:14pm
DET	Pass	2:14pm
BAR	Pass	2:14pm
BT	Pass	2:14pm

Blank Tests

Test	Status	Time
AIR	Pass	2:14pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:14pm

CRC Tests

Test	Status	Time
COMP	Pass	2:14pm
CAL	Pass	2:14pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Rockingham Court
Instrument Serial No. 008796 Wentworth, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL
780

Serial Number: 008796
Test Date: 03/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	12:36pm
AIR BLK	.00	12:37pm
ACCY CHK	.07	12:37pm
AIR BLK	.00	12:38pm
SUB TEST	.00	12:39pm
AIR BLK	.00	12:40pm
SUB TEST	.00	12:41pm
AIR BLK	.00	12:42pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY ROCKINGHAM CO JAIL 780

Serial Number: 008796 Test Record Number: 2110
Test Date: 03/17/2016 Test Time: 12:43pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:44pm
FLO	Pass	12:44pm
FC	Pass	12:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:44pm
SRC	Pass	12:44pm
DET	Pass	12:44pm
BAR	Pass	12:44pm
BT	Pass	12:44pm

Blank Tests

Test	Status	Time
AIR	Pass	12:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:45pm

CRC Tests

Test	Status	Time
COMP	Pass	12:45pm
CAL	Pass	12:45pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rutherford Instrument Location Forest City PD
Instrument Serial No. 008889 1875 Church St., Forest City

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889 Test Record Number: 697
Test Date: 03/16/2016 Test Time: 8:29am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:30am
FLO	Pass	8:30am
FC	Pass	8:30am

Temperature Tests

Test	Status	Time
FC1	Pass	8:30am
SRC	Pass	8:30am
DET	Pass	8:30am
BAR	Pass	8:30am
BT	Pass	8:30am

Blank Tests

Test	Status	Time
AIR	Pass	8:30am


Printer Tests

Test	Status	Time
PRNT	Pass	8:30am

CRC Tests

Test	Status	Time
COMP	Pass	8:31am
CAL	Pass	8:31am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	8:36am
AIR BLK	.00	8:37am
ACCY CHK	.07	8:38am
AIR BLK	.00	8:39am
SUB TEST	.00	8:40am
AIR BLK	.00	8:40am
SUB TEST	.00	8:42am
AIR BLK	.00	8:43am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Sampson Instrument Location Sampson County
Instrument Serial No. 008877 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: 008877
Test Date: 03/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RHODES, KENNETH C
Permit Number: 5329E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	12:05pm
AIR BLK	.00	12:06pm
ACCY CHK	.08	12:06pm
AIR BLK	.00	12:08pm
SUB TEST	.00	12:08pm
AIR BLK	.00	12:09pm
SUB TEST	.00	12:10pm
AIR BLK	.00	12:11pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON COUNTY SD 810

Serial Number: 008877 Test Record Number: 2188
Test Date: 03/16/2016 Test Time: 12:13pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:13pm
FLO	Pass	12:13pm
FC	Pass	12:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:13pm
SRC	Pass	12:13pm
DET	Pass	12:13pm
BAR	Pass	12:13pm
BT	Pass	12:13pm

Blank Tests

Test	Status	Time
AIR	Pass	12:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:14pm

CRC Tests

Test	Status	Time
COMP	Pass	12:14pm
CAL	Pass	12:14pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Sampson Instrument Location Sampson County
Instrument Serial No. 008825 Sheriff Dept

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



K. C. [Signature]
Signature of Certifying Official

601
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: 008825
Test Date: 03/16/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RHODES, KENNETH C

Permit Number: 5329E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:03pm
AIR BLK	.00	12:03pm
ACCY CHK	.07	12:04pm
AIR BLK	.00	12:05pm
SUB TEST	.00	12:05pm
AIR BLK	.00	12:06pm
SUB TEST	.00	12:08pm
AIR BLK	.00	12:09pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SAMPSON COUNTY SAMPSON CO SD 810

Serial Number: 008825 Test Record Number: 2100
Test Date: 03/16/2016 Test Time: 12:11pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:11pm
FLO	Pass	12:11pm
FC	Pass	12:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:11pm
SRC	Pass	12:11pm
DET	Pass	12:11pm
BAR	Pass	12:11pm
BT	Pass	12:11pm

Blank Tests

Test	Status	Time
AIR	Pass	12:12pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:12pm

CRC Tests

Test	Status	Time
COMP	Pass	12:12pm
CAL	Pass	12:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Mount Airy
Instrument Serial No. 028943 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY MOUNT AIRY PD 850

Serial Number: 008943
Test Date: 03/10/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R
Permit Number: 22067E
Effective:
09/01/2014-09/01/2016

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:00pm
AIR BLK	.00	12:00pm
ACCY CHK	.08	12:01pm
AIR BLK	.00	12:02pm
SUB TEST	.00	12:03pm
AIR BLK	.00	12:04pm
SUB TEST	.00	12:05pm
AIR BLK	.00	12:06pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY MOUNT AIRY PD 850

Serial Number: 008943 Test Record Number: 1795
Test Date: 03/10/2016 Test Time: 12:07pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:07pm
FLO	Pass	12:07pm
FC	Pass	12:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:07pm
SRC	Pass	12:07pm
DET	Pass	12:07pm
BAR	Pass	12:07pm
BT	Pass	12:07pm

Blank Tests

Test	Status	Time
AIR	Pass	12:08pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:08pm

CRC Tests

Test	Status	Time
COMP	Pass	12:08pm
CAL	Pass	12:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Pilot Mountain
Instrument Serial No. 008938 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: 008938
Test Date: 03/10/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: BENFIELD II, KENNETH R
Permit Number: 22067E
Effective:
09/01/2014-09/01/2016

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016


Test	g/210L	Time
DIAG	Pass	12:45pm
AIR BLK	.00	12:45pm
ACCY CHK	.07	12:46pm
AIR BLK	.00	12:47pm
SUB TEST	.00	12:48pm
AIR BLK	.00	12:49pm
SUB TEST	.00	12:51pm
AIR BLK	.00	12:52pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY PILOT MOUNTAIN PD 850

Serial Number: 008938 Test Record Number: 538
Test Date: 03/10/2016 Test Time: 12:53pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:53pm
FLO	Pass	12:53pm
FC	Pass	12:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:53pm
SRC	Pass	12:53pm
DET	Pass	12:53pm
BAR	Pass	12:53pm
BT	Pass	12:53pm

Blank Tests

Test	Status	Time
AIR	Pass	12:54pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:54pm

CRC Tests

Test	Status	Time
COMP	Pass	12:54pm
CAL	Pass	12:54pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Surry County Jail
Instrument Serial No. 008934 Dobson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY SURRY CO JAIL 850

Serial Number: 008934

Test Date: 03/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

09/01/2014-09/01/2016

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:45pm
AIR BLK	.00	1:46pm
ACCY CHK	.07	1:46pm
AIR BLK	.00	1:48pm
SUB TEST	.00	1:48pm
AIR BLK	.00	1:49pm
SUB TEST	.00	1:51pm
AIR BLK	.00	1:52pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY SURRY CO JAIL 850

Serial Number: 008934 Test Record Number: 1622
Test Date: 03/10/2016 Test Time: 1:53pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:53pm
FLO	Pass	1:53pm
FC	Pass	1:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:53pm
SRC	Pass	1:53pm
DET	Pass	1:53pm
BAR	Pass	1:53pm
BT	Pass	1:53pm

Blank Tests

Test	Status	Time
AIR	Pass	1:54pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:54pm

CRC Tests

Test	Status	Time
COMP	Pass	1:54pm
CAL	Pass	1:54pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Tyrrell Instrument Location Tyrrell Co. S.O.
Instrument Serial No. 008902 402 Main St., Columbia, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keele
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

TYRRELL COUNTY SHERIFF'S OFFICE 880

Serial Number: 008902
Test Date: 03/30/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	1:19pm
AIR BLK	.00	1:20pm
ACCY CHK	.07	1:21pm
AIR BLK	.00	1:22pm
SUB TEST	.00	1:23pm
AIR BLK	.00	1:24pm
SUB TEST	.00	1:26pm
AIR BLK	.00	1:26pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

TYRRELL COUNTY SHERIFF'S OFFICE 880

Serial Number: 008902 Test Record Number: 600
Test Date: 03/30/2016 Test Time: 1:27pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:28pm
FLO	Pass	1:28pm
FC	Pass	1:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:28pm
SRC	Pass	1:28pm
DET	Pass	1:28pm
BAR	Pass	1:28pm
BT	Pass	1:28pm

Blank Tests

Test	Status	Time
AIR	Pass	1:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:29pm

CRC Tests

Test	Status	Time
COMP	Pass	1:29pm
CAL	Pass	1:29pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County VANCE Instrument Location VANCE CO. SHERIFF'S OFFICE

Instrument Serial No. 008937 156 CHURCH ST. HENDERSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937
Test Date: 03/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:54pm
AIR BLK	.00	1:55pm
ACCY CHK	.08	1:56pm
AIR BLK	.00	1:57pm
SUB TEST	.00	1:58pm
AIR BLK	.00	1:59pm
SUB TEST	.00	2:01pm
AIR BLK	.00	2:01pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008937 Test Record Number: 2149
Test Date: 03/21/2016 Test Time: 2:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:03pm
FLO	Pass	2:03pm
FC	Pass	2:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:03pm
SRC	Pass	2:03pm
DET	Pass	2:03pm
BAR	Pass	2:03pm
BT	Pass	2:03pm

Blank Tests

Test	Status	Time
AIR	Pass	2:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:04pm

CRC Tests

Test	Status	Time
COMP	Pass	2:04pm
CAL	Pass	2:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County VANCE Instrument Location VANCE CO. SHERIFF'S OFFICE

Instrument Serial No. 00887D 156 CHURCH ST. HENDERSON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870
Test Date: 03/21/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG434201
Exp Date: 12/08/2016


Test	g/210L	Time
DIAG	Pass	1:54pm
AIR BLK	.00	1:55pm
ACCY CHK	.07	1:55pm
AIR BLK	.00	1:56pm
SUB TEST	.00	1:57pm
AIR BLK	.00	1:58pm
SUB TEST	.00	1:59pm
AIR BLK	.00	2:00pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

VANCE COUNTY SHERIFF'S DEPARTMENT 900

Serial Number: 008870 Test Record Number: 1262
Test Date: 03/21/2016 Test Time: 2:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:02pm
FLO	Pass	2:02pm
FC	Pass	2:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:02pm
SRC	Pass	2:02pm
DET	Pass	2:02pm
BAR	Pass	2:02pm
BT	Pass	2:02pm

Blank Tests

Test	Status	Time
AIR	Pass	2:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:03pm

CRC Tests

Test	Status	Time
COMP	Pass	2:03pm
CAL	Pass	2:03pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location WAKE FOREST PD

Instrument Serial No. 008700 225 S. TAYLOR ST. WAKE FOREST, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 28 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700
Test Date: 03/28/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	3:43pm
AIR BLK	.00	3:44pm
ACCY CHK	.07	3:44pm
AIR BLK	.00	3:45pm
SUB TEST	.00	3:46pm
AIR BLK	.00	3:47pm
SUB TEST	.00	3:48pm
AIR BLK	.00	3:49pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700 Test Record Number: 1063

Test Date: 03/28/2016 Test Time: 3:52pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:53pm
FLO	Pass	3:53pm
FC	Pass	3:53pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:53pm
SRC	Pass	3:53pm
DET	Pass	3:53pm
BAR	Pass	3:53pm
BT	Pass	3:53pm

Blank Tests

Test	Status	Time
AIR	Pass	3:53pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:53pm

CRC Tests

Test	Status	Time
COMP	Pass	3:54pm
CAL	Pass	3:54pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT Mobile Unit #10
Instrument Serial No. 008686 Roleville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686 Test Record Number: 6358
Test Date: 03/11/2016 Test Time: 9:29pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:30pm
FLO	Pass	9:30pm
FC	Pass	9:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:30pm
SRC	Pass	9:30pm
DET	Pass	9:30pm
BAR	Pass	9:30pm
BT	Pass	9:30pm

Blank Tests

Test	Status	Time
AIR	Pass	9:31pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:31pm

CRC Tests

Test	Status	Time
COMP	Pass	9:31pm
CAL	Pass	9:31pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686
Test Date: 03/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 9372E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	9:14pm
AIR BLK	.00	9:15pm
ACCY CHK	.07	9:16pm
AIR BLK	.00	9:17pm
SUB TEST	.00	9:17pm
AIR BLK	.00	9:18pm
SUB TEST	.00	9:20pm
AIR BLK	.00	9:21pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location BAT MOBILE (W.I.T #10)
Instrument Serial No. 008776 Rolesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11th day of January, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] Signature of Certifying Official 030 Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008776 Test Record Number: 3289
Test Date: 03/11/2016 Test Time: 10:43pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:43pm
FLO	Pass	10:43pm
FC	Pass	10:43pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:44pm
SRC	Pass	10:44pm
DET	Pass	10:44pm
BAR	Pass	10:44pm
BT	Pass	10:44pm

Blank Tests

Test	Status	Time
AIR	Pass	10:44pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:44pm

CRC Tests

Test	Status	Time
COMP	Pass	10:44pm
CAL	Pass	10:44pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008776
Test Date: 03/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: MORGART, STEPHEN G
Permit Number: 9372E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017


Test	g/210L	Time
DIAG	Pass	10:32pm
AIR BLK	.00	10:33pm
ACCY CHK	.08	10:34pm
AIR BLK	.00	10:35pm
SUB TEST	.00	10:35pm
AIR BLK	.00	10:36pm
SUB TEST	.00	10:37pm
AIR BLK	.00	10:39pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bat Mobile Intox #10
Instrument Serial No. 008588 Ganner

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008580 Test Record Number: 2262

Test Date: 03/18/2016 Test Time: 10:22pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:22pm
FLO	Pass	10:22pm
FC	Pass	10:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:23pm
SRC	Pass	10:23pm
DET	Pass	10:23pm
BAR	Pass	10:23pm
BT	Pass	10:23pm

Blank Tests

Test	Status	Time
AIR	Pass	10:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:23pm

CRC Tests

Test	Status	Time
COMP	Pass	10:23pm
CAL	Pass	10:23pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008580
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 9372E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017


Test	g/210L	Time
DIAG	Pass	10:07pm
AIR BLK	.00	10:08pm
ACCY CHK	.07	10:09pm
AIR BLK	.00	10:10pm
SUB TEST	.00	10:11pm
AIR BLK	.00	10:11pm
SUB TEST	.00	10:13pm
AIR BLK	.00	10:14pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wake Instrument Location Bot WDB, L&L, IT #10
Instrument Serial No. DD 8779 Ganna

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature] 636
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008779 Test Record Number: 3405
Test Date: 03/18/2016 Test Time: 10:33pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:33pm
FLO	Pass	10:33pm
FC	Pass	10:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:33pm
SRC	Pass	10:33pm
DET	Pass	10:33pm
BAR	Pass	10:33pm
BT	Pass	10:33pm

Blank Tests

Test	Status	Time
AIR	Pass	10:34pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:34pm

CRC Tests

Test	Status	Time
COMP	Pass	10:34pm
CAL	Pass	10:34pm

Preventive Maintenance
Status: Pass

Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008779
Test Date: 03/18/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 9372E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:25pm
AIR BLK	.00	10:25pm
ACCY CHK	.07	10:26pm
AIR BLK	.00	10:27pm
SUB TEST	.00	10:27pm
AIR BLK	.00	10:28pm
SUB TEST	.00	10:30pm
AIR BLK	.00	10:30pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake Instrument Location Pat Mobile Unit #10
Instrument Serial No. 005584 Garner

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910
Serial Number: 008584 Test Record Number: 2090
Test Date: 03/18/2016 Test Time: 10:21pm EDT
System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:22pm
FLO	Pass	10:22pm
FC	Pass	10:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:22pm
SRC	Pass	10:22pm
DET	Pass	10:22pm
BAR	Pass	10:22pm
BT	Pass	10:22pm

Blank Tests

Test	Status	Time
AIR	Pass	10:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:23pm

CRC Tests

Test	Status	Time
COMP	Pass	10:23pm
CAL	Pass	10:23pm

Preventive Maintenance
Status: Pass


Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008584

Test Date: 03/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 9372E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:00pm
AIR BLK	.00	10:01pm
ACCY CHK	.08	10:01pm
AIR BLK	.00	10:02pm
SUB TEST	.00	10:03pm
AIR BLK	.00	10:04pm
SUB TEST	.00	10:06pm
AIR BLK	.00	10:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wake

Instrument Location Bot Mobile Unit #10


Instrument Serial No. 008686

G. Amador

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

636
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686 Test Record Number: 6362
Test Date: 03/18/2016 Test Time: 10:20pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:20pm
FLO	Pass	10:20pm
FC	Pass	10:20pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:20pm
SRC	Pass	10:20pm
DET	Pass	10:20pm
BAR	Pass	10:20pm
BT	Pass	10:20pm

Blank Tests

Test	Status	Time
AIR	Pass	10:21pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:21pm

CRC Tests

Test	Status	Time
COMP	Pass	10:21pm
CAL	Pass	10:21pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 10 910

Serial Number: 008686
Test Date: 03/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: MORGART, STEPHEN G

Permit Number: 9372E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:05pm
AIR BLK	.00	10:06pm
ACCY CHK	.07	10:07pm
AIR BLK	.00	10:08pm
SUB TEST	.00	10:08pm
AIR BLK	.00	10:09pm
SUB TEST	.00	10:11pm
AIR BLK	.00	10:12pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location RALEIGH P.D. NORTHEAST DISTRICT

Instrument Serial No. 008623 5228 GREEN'S DAIRY RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

6317
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY NORTH EAST DISTRICT 910

Serial Number: 008623

Test Date: 03/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	2:23pm
AIR BLK	.00	2:24pm
ACCY CHK	.08	2:25pm
AIR BLK	.00	2:26pm
SUB TEST	.00	2:27pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:29pm
AIR BLK	.00	2:30pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY NORTH EAST DISTRICT 910

Serial Number: 008623 Test Record Number: 3280

Test Date: 03/14/2016 Test Time: 2:32pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:32pm
FLO	Pass	2:32pm
FC	Pass	2:33pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:33pm
SRC	Pass	2:33pm
DET	Pass	2:33pm
BAR	Pass	2:33pm
BT	Pass	2:33pm

Blank Tests

Test	Status	Time
AIR	Pass	2:33pm

Printer Tests


Test	Status	Time
PRNT	Pass	2:33pm

CRC Tests

Test	Status	Time
COMP	Pass	2:33pm
CAL	Pass	2:33pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Watauga Instrument Location Watauga Co. Jail
Instrument Serial No. 008715 Boone, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31 day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.




Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715
Test Date: 03/31/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J
Permit Number: 11304E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG434201
Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	10:47am
AIR BLK	.00	10:48am
ACCY CHK	.08	10:49am
AIR BLK	.00	10:50am
SUB TEST	.00	10:51am
AIR BLK	.00	10:52am
SUB TEST	.00	10:53am
AIR BLK	.00	10:54am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

WATAUGA COUNTY WATAUGA JAIL 940

Serial Number: 008715 Test Record Number: 1790
Test Date: 03/31/2016 Test Time: 10:55am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:55am
FLO	Pass	10:55am
FC	Pass	10:55am

Temperature Tests

Test	Status	Time
FC1	Pass	10:56am
SRC	Pass	10:56am
DET	Pass	10:56am
BAR	Pass	10:56am
BT	Pass	10:56am

Blank Tests

Test	Status	Time
AIR	Pass	10:56am

Printer Tests

Test	Status	Time
PRNT	Pass	10:56am

CRC Tests

Test	Status	Time
COMP	Pass	10:56am
CAL	Pass	10:56am

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention Ctr.
Instrument Serial No. 008671 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of MARCH, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Keese
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671
Test Date: 03/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017


Test	g/210L	Time
DIAG	Pass	11:46am
AIR BLK	.00	11:47am
ACCY CHK	.07	11:47am
AIR BLK	.00	11:49am
SUB TEST	.00	11:49am
AIR BLK	.00	11:50am
SUB TEST	.00	11:51am
AIR BLK	.00	11:52am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671 Test Record Number: 4187
Test Date: 03/03/2016 Test Time: 11:57am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:58am


Printer Tests

Test	Status	Time
PRNT	Pass	11:58am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention Ctr.
Instrument Serial No. 008649 207 E. Chestnut ST., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of MARCH, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649
Test Date: 03/03/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA A
Permit Number: 11646E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG507902
Exp Date: 03/20/2017


Test	g/210L	Time
DIAG	Pass	11:40am
AIR BLK	.00	11:41am
ACCY CHK	.07	11:41am
AIR BLK	.00	11:42am
SUB TEST	.00	11:44am
AIR BLK	.00	11:44am
SUB TEST	.00	11:46am
AIR BLK	.00	11:47am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649 Test Record Number: 2840
Test Date: 03/03/2016 Test Time: 11:49am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:50am
FLO	Pass	11:50am
FC	Pass	11:50am

Temperature Tests

Test	Status	Time
FC1	Pass	11:50am
SRC	Pass	11:50am
DET	Pass	11:50am
BAR	Pass	11:50am
BT	Pass	11:50am

Blank Tests

Test	Status	Time
AIR	Pass	11:50am


Printer Tests

Test	Status	Time
PRNT	Pass	11:51am

CRC Tests

Test	Status	Time
COMP	Pass	11:51am
CAL	Pass	11:51am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilson Instrument Location Wilson Co. Detention Center
Instrument Serial No. 008588 100 E. Green St., Wilson, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: 008588
Test Date: 03/07/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016


Test	g/210L	Time
DIAG	Pass	9:49am
AIR BLK	.00	9:50am
ACCY CHK	.07	9:50am
AIR BLK	.00	9:51am
SUB TEST	.00	9:52am
AIR BLK	.00	9:53am
SUB TEST	.00	9:54am
AIR BLK	.00	9:55am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: 008588 Test Record Number: 880
Test Date: 03/07/2016 Test Time: 9:59am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:59am
FLO	Pass	9:59am
FC	Pass	9:59am

Temperature Tests

Test	Status	Time
FC1	Pass	10:00am
SRC	Pass	10:00am
DET	Pass	10:00am
BAR	Pass	10:00am
BT	Pass	10:00am

Blank Tests

Test	Status	Time
AIR	Pass	10:00am


Printer Tests

Test	Status	Time
PRNT	Pass	10:00am

CRC Tests

Test	Status	Time
COMP	Pass	10:00am
CAL	Pass	10:00am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wilson Instrument Location Wilson Co. Detention Center

Instrument Serial No. 008627 100 E. Green St, Wilson, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 31st day of March, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. [Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: 008627
Test Date: 03/31/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	10:28am
AIR BLK	.00	10:29am
ACCY CHK	.07	10:29am
AIR BLK	.00	10:30am
SUB TEST	.00	10:31am
AIR BLK	.00	10:32am
SUB TEST	.00	10:34am
AIR BLK	.00	10:34am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WILSON COUNTY WILSON CO DETENTION 970

Serial Number: 008627 Test Record Number: 1895
Test Date: 03/31/2016 Test Time: 10:36am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:36am
FLO	Pass	10:36am
FC	Pass	10:36am

Temperature Tests

Test	Status	Time
FC1	Pass	10:36am
SRC	Pass	10:36am
DET	Pass	10:36am
BAR	Pass	10:36am
BT	Pass	10:36am

Blank Tests

Test	Status	Time
AIR	Pass	10:37am

Printer Tests

Test	Status	Time
PRNT	Pass	10:37am

CRC Tests

Test	Status	Time
COMP	Pass	10:37am
CAL	Pass	10:37am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lincoln Instrument Location Lincoln County Courthouse
Instrument Serial No. 008827 #1 Court Square, Lincolnton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 24th day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hatcher
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008827
Test Date: 03/24/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HUTCHINSON, JOSEPH E
Permit Number: 19951E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017


Test	g/210L	Time
DIAG	Pass	12:28pm
AIR BLK	.00	12:28pm
ACCY CHK	.08	12:29pm
AIR BLK	.00	12:30pm
SUB TEST	.00	12:31pm
AIR BLK	.00	12:32pm
SUB TEST	.00	12:34pm
AIR BLK	.00	12:34pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LINCOLN COUNTY COURTHOUSE 540

Serial Number: 008827 Test Record Number: 2324
Test Date: 03/24/2016 Test Time: 12:24pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:24pm
FLO	Pass	12:24pm
FC	Pass	12:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:25pm
SRC	Pass	12:25pm
DET	Pass	12:25pm
BAR	Pass	12:25pm
BT	Pass	12:25pm

Blank Tests

Test	Status	Time
AIR	Pass	12:25pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:25pm

CRC Tests

Test	Status	Time
COMP	Pass	12:25pm
CAL	Pass	12:25pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Union Instrument Location Union County SD
Instrument Serial No. 008876 3344 Presson Road, Monroe
704-283-3770

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of March, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Hutto
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876
Test Date: 03/23/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HUTCHINSON, JOSEPH E
Permit Number: 19951E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	1:21pm
AIR BLK	.00	1:22pm
ACCY CHK	.07	1:23pm
AIR BLK	.00	1:25pm
SUB TEST	.00	1:25pm
AIR BLK	.00	1:26pm
SUB TEST	.00	1:28pm
AIR BLK	.00	1:29pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

UNION COUNTY UNION COUNTY SD 890

Serial Number: 008876 Test Record Number: 3965
Test Date: 03/23/2016 Test Time: 1:30pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:30pm
FLO	Pass	1:30pm
FC	Pass	1:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:30pm
SRC	Pass	1:30pm
DET	Pass	1:30pm
BAR	Pass	1:30pm
BT	Pass	1:30pm

Blank Tests

Test	Status	Time
AIR	Pass	1:31pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:31pm

CRC Tests

Test	Status	Time
COMP	Pass	1:31pm
CAL	Pass	1:31pm

Preventive Maintenance
Status: Pass


Analyst