

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ANSON Instrument Location ANSON Co. Sheriff's Office
Instrument Serial No. 008597 WADESBORO NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597
Test Date: 11/29/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: *RUSSELL, LARRY H*
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534901
Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	11:51am
AIR BLK	.00	11:51am
ACCY CHK	.07	11:52am
AIR BLK	.00	11:53am
SUB TEST	.00	11:53am
AIR BLK	.00	11:54am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON COUNTY SO. 030

Serial Number: 008597 Test Record Number: 1524
Test Date: 11/29/2016 Test Time: 11:58am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:58am
FLO	Pass	11:58am
FC	Pass	11:58am

Temperature Tests

Test	Status	Time
FC1	Pass	11:58am
SRC	Pass	11:58am
DET	Pass	11:58am
BAR	Pass	11:58am
BT	Pass	11:58am

Blank Tests

Test	Status	Time
AIR	Pass	11:59am

Printer Tests

Test	Status	Time
PRNT	Pass	11:59am

CRC Tests

Test	Status	Time
COMP	Pass	11:59am
CAL	Pass	11:59am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ANSON Instrument Location ANSON Co Sheriff's Office
Instrument Serial No. 008739 WADESBORO NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ANSON COUNTY ANSON COUNTY S.O. 030

Serial Number: 008739
Test Date: 11/29/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607601

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	11:53am
AIR BLK	.00	11:54am
ACCY CHK	.07	11:54am
AIR BLK	.00	11:55am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am
SUB TEST	.00	11:58am
AIR BLK	.00	11:59am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ANSON COUNTY ANSON COUNTY S.O. 030

Serial Number: 008739 Test Record Number: 238
Test Date: 11/29/2016 Test Time: 12:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:02pm
FLO	Pass	12:02pm
FC	Pass	12:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:02pm
SRC	Pass	12:02pm
DET	Pass	12:02pm
BAR	Pass	12:02pm
BT	Pass	12:02pm

Blank Tests

Test	Status	Time
AIR	Pass	12:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:03pm

CRC Tests

Test	Status	Time
COMP	Pass	12:03pm
CAL	Pass	12:03pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Avery

Instrument Location Banner ETK PD

Instrument Serial No. 008724

Banner ETK, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

AVERY COUNTY BANNER ELK PD 050

Serial Number: 008724
Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	5:12pm
AIR BLK	.00	5:13pm
ACCY CHK	.07	5:13pm
AIR BLK	.00	5:14pm
SUB TEST	.00	5:15pm
AIR BLK	.00	5:16pm
SUB TEST	.00	5:17pm
AIR BLK	.00	5:18pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

AVERY COUNTY BANNER ELK PD 050

Serial Number: 008724 Test Record Number: 481
Test Date: 11/09/2016 Test Time: 5:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:21pm
FLO	Pass	5:21pm
FC	Pass	5:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:21pm
SRC	Pass	5:21pm
DET	Pass	5:21pm
BAR	Pass	5:21pm
BT	Pass	5:21pm

Blank Tests

Test	Status	Time
AIR	Pass	5:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:22pm

CRC Tests

Test	Status	Time
COMP	Pass	5:22pm
CAL	Pass	5:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO JAIL

Instrument Serial No. 008873 109 S. Maple ST
GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of NOV, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alexis Barnes
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO JAIL 000

Serial Number: 008873

Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	2:37pm
AIR BLK	.00	2:37pm
ACCY CHK	.08	2:38pm
AIR BLK	.00	2:39pm
SUB TEST	.00	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	.00	2:42pm
AIR BLK	.00	2:43pm

Reported AC: 00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

AAMANCE COUNTY ALAMANCE CO JAIL 000

Serial Number: 008873 Test Record Number: 1349
Test Date: 11/10/2016 Test Time: 2:43pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:44pm
FLO	Pass	2:44pm
FC	Pass	2:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:44pm
SRC	Pass	2:44pm
DET	Pass	2:44pm
BAR	Pass	2:44pm
BT	Pass	2:44pm

Blank Tests

Test	Status	Time
AIR	Pass	2:44pm

Printer Tests


Test	Status	Time
PRNT	Pass	2:45pm

CRC Tests

Test	Status	Time
COMP	Pass	2:45pm
CAL	Pass	2:45pm

Preventive Maintenance

Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ALAMANCE Instrument Location ALAMANCE CO. JAIL

Instrument Serial No. 008453 109 S. MAPLE ST
GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Brues

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008853
Test Date: 11/23/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S
Permit Number: 11434E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607501
Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	10:20am
AIR BLK	.00	10:21am
ACCY CHK	.08	10:22am
AIR BLK	.00	10:23am
SUB TEST	.00	10:24am
AIR BLK	.00	10:25am
SUB TEST	.00	10:26am
AIR BLK	.00	10:27am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008853 Test Record Number: 2053
Test Date: 11/23/2016 Test Time: 10:28am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:29am
FLO	Pass	10:29am
FC	Pass	10:29am

Temperature Tests

Test	Status	Time
FC1	Pass	10:29am
SRC	Pass	10:29am
DET	Pass	10:29am
BAR	Pass	10:29am
BT	Pass	10:29am

Blank Tests

Test	Status	Time
AIR	Pass	10:29am

Printer Tests

Test	Status	Time
PRNT	Pass	10:29am

CRC Tests

Test	Status	Time
COMP	Pass	10:30am
CAL	Pass	10:30am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County ALAMANCE Instrument Location ALAMANCE CO. JAIL

Instrument Serial No. 008913 109 S. Maple St
GRAHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James Barnes

Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913
Test Date: 11/23/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: *BARNES, SIMON S*

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:41am
AIR BLK	.00	10:42am
ACCY CHK	.08	10:42am
AIR BLK	.00	10:43am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:46am
AIR BLK	.00	10:47am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures

Forensic Tests for Alcohol Branch

Department of Health and Human Services

Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ALAMANCE COUNTY ALAMANCE CO. JAIL 000

Serial Number: 008913 Test Record Number: 2682
Test Date: 11/23/2016 Test Time: 10:48am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:48am
FLO	Pass	10:48am
FC	Pass	10:48am

Temperature Tests

Test	Status	Time
FC1	Pass	10:48am
SRC	Pass	10:48am
DET	Pass	10:48am
BAR	Pass	10:48am
BT	Pass	10:48am

Blank Tests

Test	Status	Time
AIR	Pass	10:49am

Printer Tests

Test	Status	Time
PRNT	Pass	10:49am

CRC Tests

Test	Status	Time
COMP	Pass	10:49am
CAL	Pass	10:49am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Buncombe Instrument Location BAF mobile Unit 11

Instrument Serial No. 008970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

659

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008970
Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	6:22pm
AIR BLK	.00	6:23pm
ACCY CHK	.08	6:24pm
AIR BLK	.00	6:25pm
SUB TEST	.00	6:25pm
AIR BLK	.00	6:26pm
SUB TEST	.00	6:28pm
AIR BLK	.00	6:29pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008970 Test Record Number: 249
Test Date: 11/04/2016 Test Time: 6:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:30pm
FLO	Pass	6:30pm
FC	Pass	6:30pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:30pm
SRC	Pass	6:30pm
DET	Pass	6:30pm
BAR	Pass	6:30pm
BT	Pass	6:30pm

Blank Tests

Test	Status	Time
AIR	Pass	6:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:31pm

CRC Tests

Test	Status	Time
COMP	Pass	6:31pm
CAL	Pass	6:31pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Buncombe Instrument Location BA7 Mobile Unit 11

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008973
Test Date: 11/04/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	6:29pm
AIR BLK	.00	6:30pm
ACCY CHK	.08	6:30pm
AIR BLK	.00	6:31pm
SUB TEST	.00	6:31pm
AIR BLK	.00	6:32pm
SUB TEST	.00	6:34pm
AIR BLK	.00	6:35pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BUNCOMBE BAT MOBILE UNIT 11 100

Serial Number: 008973 Test Record Number: 216
Test Date: 11/04/2016 Test Time: 6:36pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	6:36pm
FLO	Pass	6:36pm
FC	Pass	6:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	6:36pm
SRC	Pass	6:36pm
DET	Pass	6:36pm
BAR	Pass	6:36pm
BT	Pass	6:36pm

Blank Tests

Test	Status	Time
AIR	Pass	6:37pm

Printer Tests

Test	Status	Time
PRNT	Pass	6:37pm

CRC Tests

Test	Status	Time
COMP	Pass	6:37pm
CAL	Pass	6:37pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Burke Instrument Location Burke-Catawba Jail

Instrument Serial No. 008904 Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008904
Test Date: 11/01/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: *BURNETTE, ANTHONY J*

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: *Breath Test*

Lot Number: AG434201

Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	3:08pm
AIR BLK	.00	3:09pm
ACCY CHK	.08	3:09pm
AIR BLK	.00	3:10pm
SUB TEST	.00	3:11pm
AIR BLK	.00	3:12pm
SUB TEST	.00	3:14pm
AIR BLK	.00	3:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008904 Test Record Number: 1926
Test Date: 11/01/2016 Test Time: 3:16pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:16pm
FLO	Pass	3:16pm
FC	Pass	3:16pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:16pm
SRC	Pass	3:16pm
DET	Pass	3:16pm
BAR	Pass	3:16pm
BT	Pass	3:16pm

Blank Tests

Test	Status	Time
AIR	Pass	3:17pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:17pm

CRC Tests

Test	Status	Time
COMP	Pass	3:17pm
CAL	Pass	3:17pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Burke

Instrument Location Burke - Catawba Jail


Instrument Serial No. 008931

Morganton, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008831
Test Date: 11/01/2016

Citation Number: M0000000-0
Subject's Name: PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J
Permit Number: 11304E
Effective: 05/01/2015-05/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	3:08pm
AIR BLK	.00	3:09pm
ACCY CHK	.08	3:10pm
AIR BLK	.00	3:11pm
SUB TEST	.00	3:12pm
AIR BLK	.00	3:13pm
SUB TEST	.00	3:14pm
AIR BLK	.00	3:15pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BURKE COUNTY BURKE-CATAWBA JAIL 110

Serial Number: 008831 Test Record Number: 1716
Test Date: 11/01/2016 Test Time: 3:17pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:17pm
FLO	Pass	3:17pm
FC	Pass	3:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:17pm
SRC	Pass	3:17pm
DET	Pass	3:17pm
BAR	Pass	3:17pm
BT	Pass	3:17pm

Blank Tests

Test	Status	Time
AIR	Pass	3:18pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:18pm

CRC Tests

Test	Status	Time
COMP	Pass	3:18pm
CAL	Pass	3:18pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Beaufort Instrument Location Belhaven PD

Instrument Serial No. 008928 Belhaven, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key M
Signature of Certifying Official

6043
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928

Test Date: 11/17/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	11:03am
AIR BLK	.00	11:04am
ACCY CHK	.08	11:04am
AIR BLK	.00	11:06am
SUB TEST	.00	11:06am
AIR BLK	.00	11:07am
SUB TEST	.00	11:09am
AIR BLK	.00	11:09am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

BEAUFORT COUNTY BELHAVEN PD 060

Serial Number: 008928 Test Record Number: 308
Test Date: 11/17/2016 Test Time: 11:11am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:11am
FLO	Pass	11:11am
FC	Pass	11:11am

Temperature Tests

Test	Status	Time
FC1	Pass	11:12am
SRC	Pass	11:12am
DET	Pass	11:12am
BAR	Pass	11:12am
BT	Pass	11:12am

Blank Tests

Test	Status	Time
AIR	Pass	11:12am

Printer Tests

Test	Status	Time
PRNT	Pass	11:12am

CRC Tests

Test	Status	Time
COMP	Pass	11:13am
CAL	Pass	11:13am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County CUMBERLAND Instrument Location CUMBERLAND Co. DET. CENTRE

Instrument Serial No. 008633 Fayetteville NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633
Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	10:04am
AIR BLK	.00	10:06am
ACCY CHK	.07	10:06am
AIR BLK	.00	10:08am
SUB TEST	.00	10:08am
AIR BLK	.00	10:09am
SUB TEST	.00	10:11am
AIR BLK	.00	10:12am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008633 Test Record Number: 4038
Test Date: 11/30/2016 Test Time: 10:13am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

Temperature Tests

Test	Status	Time
FC1	Pass	10:13am
SRC	Pass	10:13am
DET	Pass	10:13am
BAR	Pass	10:13am
BT	Pass	10:13am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am

Printer Tests

Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location CUMBERLAND Co. Det. Center
Instrument Serial No. 008672 Fayetteville NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672

Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	10:25am
AIR BLK	.00	10:26am
ACCY CHK	.08	10:26am
AIR BLK	.00	10:27am
SUB TEST	.00	10:28am
AIR BLK	.00	10:29am
SUB TEST	.00	10:31am
AIR BLK	.00	10:31am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008672 Test Record Number: 5366
Test Date: 11/30/2016 Test Time: 10:32am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:32am
FLO	Pass	10:32am
FC	Pass	10:33am

Temperature Tests

Test	Status	Time
FC1	Pass	10:33am
SRC	Pass	10:33am
DET	Pass	10:33am
BAR	Pass	10:33am
BT	Pass	10:33am

Blank Tests

Test	Status	Time
AIR	Pass	10:33am

Printer Tests

Test	Status	Time
PRNT	Pass	10:33am

CRC Tests

Test	Status	Time
COMP	Pass	10:33am
CAL	Pass	10:33am

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location CUMBERLAND CO. DET. CENTER
Instrument Serial No. 008614 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614

Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501

Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	10:49am
AIR BLK	.00	10:50am
ACCY CHK	.08	10:51am
AIR BLK	.00	10:52am
SUB TEST	.00	10:52am
AIR BLK	.00	10:53am
SUB TEST	.00	10:55am
AIR BLK	.00	10:56am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008614 Test Record Number: 3587
Test Date: 11/30/2016 Test Time: 10:59am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:59am
FLO	Pass	10:59am
FC	Pass	11:00am

Temperature Tests

Test	Status	Time
FC1	Pass	11:00am
SRC	Pass	11:00am
DET	Pass	11:00am
BAR	Pass	11:00am
BT	Pass	11:00am

Blank Tests

Test	Status	Time
AIR	Pass	11:00am


Printer Tests

Test	Status	Time
PRNT	Pass	11:00am

CRC Tests

Test	Status	Time
COMP	Pass	11:00am
CAL	Pass	11:00am

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location CUMBERLAND Co. Det. Center
Instrument Serial No. 008632 Fayetteville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008632
Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:55am
AIR BLK	.00	10:56am
ACCY CHK	.07	10:56am
AIR BLK	.00	10:57am
SUB TEST	.00	10:58am
AIR BLK	.00	10:59am
SUB TEST	.00	11:01am
AIR BLK	.00	11:02am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY DETENTION CENTER 250

Serial Number: 008632 Test Record Number: 3737
Test Date: 11/30/2016 Test Time: 11:08am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:09am
FLO	Pass	11:09am
FC	Pass	11:09am

Temperature Tests

Test	Status	Time
FC1	Pass	11:09am
SRC	Pass	11:09am
DET	Pass	11:09am
BAR	Pass	11:09am
BT	Pass	11:09am

Blank Tests

Test	Status	Time
AIR	Pass	11:10am


Printer Tests

Test	Status	Time
PRNT	Pass	11:10am

CRC Tests

Test	Status	Time
COMP	Pass	11:10am
CAL	Pass	11:10am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location F.T. BRAGG
Instrument Serial No. 008787 P.M.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 03 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND CO. FORT BRAGG LEC. 250

Serial Number: 008787
Test Date: 11/03/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517402
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	1:13pm
AIR BLK	.00	1:14pm
ACCY CHK	.07	1:15pm
AIR BLK	.00	1:16pm
SUB TEST	.00	1:17pm
AIR BLK	.00	1:18pm
SUB TEST	.00	1:19pm
AIR BLK	.00	1:20pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND CO. FORT BRAGG LEC. 250

Serial Number: 008787 Test Record Number: 665
Test Date: 11/03/2016 Test Time: 1:31pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:32pm
FLO	Pass	1:32pm
FC	Pass	1:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:32pm
SRC	Pass	1:32pm
DET	Pass	1:32pm
BAR	Pass	1:32pm
BT	Pass	1:32pm

Blank Tests

Test	Status	Time
AIR	Pass	1:33pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:33pm

CRC Tests

Test	Status	Time
COMP	Pass	1:33pm
CAL	Pass	1:33pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County CUMBERLAND Instrument Location FT. BRAGG
Instrument Serial No. 008903 P.M.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 03 day of NOVEMBER, 2016 the foregoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: 008903
Test Date: 11/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	1:17pm
AIR BLK	.00	1:18pm
ACCY CHK	.08	1:19pm
AIR BLK	.00	1:20pm
SUB TEST	.00	1:22pm
AIR BLK	.00	1:23pm
SUB TEST	.00	1:25pm
AIR BLK	.00	1:26pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CUMBERLAND COUNTY FORT BRAGG LEC. 250

Serial Number: 008903 Test Record Number: 1743
Test Date: 11/03/2016 Test Time: 1:29pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:29pm
FLO	Pass	1:29pm
FC	Pass	1:29pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:29pm
SRC	Pass	1:29pm
DET	Pass	1:29pm
BAR	Pass	1:29pm
BT	Pass	1:29pm

Blank Tests

Test	Status	Time
AIR	Pass	1:30pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:30pm

CRC Tests

Test	Status	Time
COMP	Pass	1:30pm
CAL	Pass	1:30pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Caldwell Instrument Location Caldwell Co Jail

Instrument Serial No. 008719 Lenoir, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CALDWELL COUNTY CALDWELL COUNTY JAIL

130

Serial Number: 008719

Test Date: 11/07/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434201

Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	2:23pm
AIR BLK	.00	2:24pm
ACCY CHK	.07	2:24pm
AIR BLK	.00	2:25pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:28pm
AIR BLK	.00	2:29pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CALDWELL COUNTY CALDWELL COUNTY JAIL 130

Serial Number: 008719 Test Record Number: 2075
Test Date: 11/07/2016 Test Time: 2:30pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:31pm
FLO	Pass	2:31pm
FC	Pass	2:31pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:31pm
SRC	Pass	2:31pm
DET	Pass	2:31pm
BAR	Pass	2:31pm
BT	Pass	2:31pm

Blank Tests

Test	Status	Time
AIR	Pass	2:31pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:31pm

CRC Tests

Test	Status	Time
COMP	Pass	2:32pm
CAL	Pass	2:32pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Caldwell Instrument Location Caldwell Co. Jail

Instrument Serial No. 008803 Lenoir, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CALDWELL COUNTY CALDWELL COUNTY JAIL
130

Serial Number: 008803
Test Date: 11/07/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	2:22pm
AIR BLK	.00	2:23pm
ACCY CHK	.07	2:23pm
AIR BLK	.00	2:24pm
SUB TEST	.00	2:25pm
AIR BLK	.00	2:26pm
SUB TEST	.00	2:28pm
AIR BLK	.00	2:28pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Subject Test

CALDWELL COUNTY CALDWELL COUNTY JAIL

130

Serial Number: 008803

Test Date: 11/07/2016

Citation Number: M0000000-0

Subject's Name: CANISTER, CHANGE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FDA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test g/210L Time

DIAG Pass 2:39pm

AIR BLK .00 2:40pm

ACCY CHK .08 2:41pm

AIR BLK .00 2:42pm

SUB TEST .** 2:42pm

BREATH AT IMPROPER TIME

Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Cleveland Instrument Location Cleveland County S.D. Amex
Instrument Serial No. 008694 407 McBrayer St, Shelby

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

CLEVELAND COUNTY CLEVELAND SD-ANNEX 220

Serial Number: 008694 Test Record Number: 1248
Test Date: 11/02/2016 Test Time: 10:35am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:35am
FLO	Pass	10:35am
FC	Pass	10:35am

Temperature Tests

Test	Status	Time
FC1	Pass	10:36am
SRC	Pass	10:36am
DET	Pass	10:36am
BAR	Pass	10:36am
BT	Pass	10:36am

Blank Tests:

Test	Status	Time
AIR	Pass	10:36am


Printer Tests

Test	Status	Time
PRNT	Pass	10:36am

CRC Tests

Test	Status	Time
COMP	Pass	10:36am
CAL	Pass	10:36am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

CLEVELAND COUNTY CLEVELAND SD-ANNEX
220

Serial Number: 008694
Test Date: 11/02/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:39am
AIR BLK	.00	10:39am
ACCY CHK	.08	10:40am
AIR BLK	.00	10:41am
SUB TEST	.00	10:41am
AIR BLK	.00	10:42am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location BAT Mobile Unit 7

Instrument Serial No. 008972

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008972

Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	6:54pm
AIR BLK	.00	6:55pm
ACCY CHK	.07	6:55pm
AIR BLK	.00	6:56pm
SUB TEST	.00	6:57pm
AIR BLK	.00	6:58pm
SUB TEST	.00	7:00pm
AIR BLK	.00	7:00pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CATAWBA BAT MOBILE UNIT 7 170

Serial Number: 008972 Test Record Number: 240
Test Date: 11/04/2016 Test Time: 7:02pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:02pm
FLO	Pass	7:02pm
FC	Pass	7:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:02pm
SRC	Pass	7:02pm
DET	Pass	7:02pm
BAR	Pass	7:02pm
BT	Pass	7:02pm

Blank Tests

Test	Status	Time
AIR	Pass	7:03pm


Printer Tests

Test	Status	Time
PRNT	Pass	7:03pm

CRC Tests

Test	Status	Time
COMP	Pass	7:03pm
CAL	Pass	7:03pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County Cleveland Instrument Location BAT mobile Unit 11

Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 20 14 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris Jay
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CLEVELAND BAT MOBILE UNIT 11 220

Serial Number: 008973

Test Date: 11/12/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	8:19pm
AIR BLK	.00	8:20pm
ACCY CHK	.08	8:20pm
AIR BLK	.00	8:21pm
SUB TEST	.00	8:21pm
AIR BLK	.00	8:22pm
SUB TEST	.00	8:24pm
AIR BLK	.00	8:25pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CLEVELAND BAT MOBILE UNIT 11 220

Serial Number: 008973 Test Record Number: 227
Test Date: 11/12/2016 Test Time: 8:25pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:26pm
FLO	Pass	8:26pm
FC	Pass	8:26pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:26pm
SRC	Pass	8:26pm
DET	Pass	8:26pm
BAR	Pass	8:26pm
BT	Pass	8:26pm

Blank Tests

Test	Status	Time
AIR	Pass	8:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:27pm

CRC Tests

Test	Status	Time
COMP	Pass	8:27pm
CAL	Pass	8:27pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Chowan Instrument Location Chowan Co. Public Safety
Instrument Serial No. 008895 Chv., 305 Freemason St., Edenton, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895

Test Date: 11/07/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	11:36am
AIR BLK	.00	11:37am
ACCY CHK	.07	11:37am
AIR BLK	.00	11:39am
SUB TEST	.00	11:39am
AIR BLK	.00	11:40am
SUB TEST	.00	11:42am
AIR BLK	.00	11:43am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

CHOWAN COUNTY PUBLIC SAFETY CENTER 200

Serial Number: 008895 Test Record Number: 741
Test Date: 11/07/2016 Test Time: 11:44am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:45am
FLO	Pass	11:45am
FC	Pass	11:45am

Temperature Tests

Test	Status	Time
FC1	Pass	11:45am
SRC	Pass	11:45am
DET	Pass	11:45am
BAR	Pass	11:45am
BT	Pass	11:45am

Blank Tests

Test	Status	Time
AIR	Pass	11:45am

Printer Tests

Test	Status	Time
PRNT	Pass	11:45am

CRC Tests

Test	Status	Time
COMP	Pass	11:46am
CAL	Pass	11:46am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Catawba County SD
Instrument Serial No. 008821 100 B Southwest Blvd, Newton

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Heath
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821
Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102


Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:06pm
AIR BLK	.00	2:07pm
ACCY CHK	.07	2:08pm
AIR BLK	.00	2:09pm
SUB TEST	.00	2:09pm
AIR BLK	.00	2:10pm
SUB TEST	.00	2:11pm
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY CATAWBA COUNTY SD 170

Serial Number: 008821 Test Record Number: 1695
Test Date: 11/10/2016 Test Time: 2:13pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	2:14pm
FLO	Pass	2:14pm
FC	Pass	2:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:14pm
SRC	Pass	2:14pm
DET	Pass	2:14pm
BAR	Pass	2:14pm
BT	Pass	2:14pm

Blank Tests

Test	Status	Time
AIR	Pass	2:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:15pm

CRC Tests

Test	Status	Time
COMP	Pass	2:15pm
CAL	Pass	2:15pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Catawba Instrument Location Hickory PD
Instrument Serial No. 008841 347 2nd Ave SW, Hickory

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. H. H.
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841

Test Date: 11/07/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	2:08pm
AIR BLK	.00	2:09pm
ACCY CHK	.08	2:09pm
AIR BLK	.00	2:11pm
SUB TEST	.00	2:11pm
AIR BLK	.00	2:12pm
SUB TEST	.00	2:13pm
AIR BLK	.00	2:14pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CATAWBA COUNTY HICKORY PD 170

Serial Number: 008841 Test Record Number: 1703
Test Date: 11/07/2016 Test Time: 2:15pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:15pm
FLO	Pass	2:15pm
FC	Pass	2:15pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:16pm
SRC	Pass	2:16pm
DET	Pass	2:16pm
BAR	Pass	2:16pm
BT	Pass	2:16pm

Blank Tests

Test	Status	Time
AIR	Pass	2:16pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:16pm

CRC Tests

Test	Status	Time
COMP	Pass	2:16pm
CAL	Pass	2:16pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Clay Instrument Location Clay Co. Jail
Instrument Serial No. 008608 Hayesville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Orin R. Cuthbert
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

CLAY COUNTY CLAY COUNTY JAIL 210

Serial Number: 008608

Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	11:29am
AIR BLK	.00	11:30am
ACCY CHK	.07	11:30am
AIR BLK	.00	11:32am
SUB TEST	.00	11:32am
AIR BLK	.00	11:33am
SUB TEST	.00	11:35am
AIR BLK	.00	11:36am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

CLAY COUNTY CLAY COUNTY JAIL 210

Serial Number: 008608 Test Record Number: 1128
Test Date: 11/08/2016 Test Time: 11:37am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:37am
FLO	Pass	11:37am
FC	Pass	11:37am

Temperature Tests

Test	Status	Time
FC1	Pass	11:37am
SRC	Pass	11:37am
DET	Pass	11:37am
BAR	Pass	11:37am
BT	Pass	11:37am

Blank Tests

Test	Status	Time
AIR	Pass	11:38am

Printer Tests

Test	Status	Time
PRNT	Pass	11:38am

CRC Tests

Test	Status	Time
COMP	Pass	11:38am
CAL	Pass	11:38am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Dare Co. S.O. - KATIERGS
Instrument Serial No. 008807 50346 N.C. HWY 12, FRISCO, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Trick A. Reese
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008807
Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	1:06pm
AIR BLK	.00	1:07pm
ACCY CHK	.08	1:08pm
AIR BLK	.00	1:09pm
SUB TEST	.00	1:09pm
AIR BLK	.00	1:10pm
SUB TEST	.00	1:12pm
AIR BLK	.00	1:13pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008807 Test Record Number: 804
Test Date: 11/21/2016 Test Time: 1:13pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:14pm
FLO	Pass	1:14pm
FC	Pass	1:14pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:14pm
SRC	Pass	1:14pm
DET	Pass	1:14pm
BAR	Pass	1:14pm
BT	Pass	1:14pm

Blank Tests

Test	Status	Time
AIR	Pass	1:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:14pm

CRC Tests

Test	Status	Time
COMP	Pass	1:15pm
CAL	Pass	1:15pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Dare Co. S.O. - Hatteras
Instrument Serial No. DD 8847 50346 NC Hwy 12, Frisco, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

847
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008847
Test Date: 11/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	1:12pm
AIR BLK	.00	1:13pm
ACCY CHK	.08	1:13pm
AIR BLK	.00	1:14pm
SUB TEST	.00	1:15pm
AIR BLK	.00	1:16pm
SUB TEST	.00	1:17pm
AIR BLK	.00	1:18pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY DARE CO SO HATTERAS 270

Serial Number: 008847 Test Record Number: 586
Test Date: 11/14/2016 Test Time: 1:21pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:21pm
FLO	Pass	1:21pm
FC	Pass	1:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:21pm
SRC	Pass	1:21pm
DET	Pass	1:21pm
BAR	Pass	1:21pm
BT	Pass	1:21pm

Blank Tests

Test	Status	Time
AIR	Pass	1:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:22pm

CRC Tests

Test	Status	Time
COMP	Pass	1:22pm
CAL	Pass	1:22pm

Preventive Maintenance
Status: *Pass*


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DURHAM Instrument Location Durham County Jail

Instrument Serial No. 008891 219 S. Mangum St.
Durham, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barnes
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008891

Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	2:47pm
AIR BLK	.00	2:49pm
ACCY CHK	.08	2:49pm
AIR BLK	.00	2:50pm
SUB TEST	.00	2:50pm
AIR BLK	.00	2:51pm
SUB TEST	.00	2:53pm
AIR BLK	.00	2:54pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008891 Test Record Number: 3409
Test Date: 11/04/2016 Test Time: 2:55pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:55pm
FLO	Pass	2:55pm
FC	Pass	2:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:55pm
SRC	Pass	2:55pm
DET	Pass	2:55pm
BAR	Pass	2:55pm
BT	Pass	2:55pm

Blank Tests

Test	Status	Time
AIR	Pass	2:56pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:56pm

CRC Tests

Test	Status	Time
COMP	Pass	2:56pm
CAL	Pass	2:56pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Davie Instrument Location Davie County Jail
Instrument Serial No. 008905 Mocksville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905

Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517501


Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	11:26am
AIR BLK	.00	11:27am
ACCY CHK	.08	11:28am
AIR BLK	.00	11:29am
SUB TEST	.00	11:29am
AIR BLK	.00	11:30am
SUB TEST	.00	11:32am
AIR BLK	.00	11:33am

Reported AC: 00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DAVIE COUNTY DAVIE COUNTY JAIL 290

Serial Number: 008905 Test Record Number: 1934
Test Date: 11/09/2016 Test Time: 11:34am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:34am
FLO	Pass	11:34am
FC	Pass	11:34am

Temperature Tests

Test	Status	Time
FC1	Pass	11:34am
SRC	Pass	11:34am
DET	Pass	11:34am
BAR	Pass	11:34am
BT	Pass	11:34am

Blank Tests

Test	Status	Time
AIR	Pass	11:35am


Printer Tests

Test	Status	Time
PRNT	Pass	11:35am

CRC Tests

Test	Status	Time
COMP	Pass	11:35am
CAL	Pass	11:35am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Durham Instrument Location Bat Mobile Unit 8
Instrument Serial No. DD8736 Durham Co 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 12 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY BAT MOBILE UNIT 8 310

Serial Number: 008736
Test Date: 11/12/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	9:47pm
AIR BLK	.00	9:49pm
ACCY CHK	.07	9:49pm
AIR BLK	.00	9:50pm
SUB TEST	.00	9:51pm
AIR BLK	.00	9:52pm
SUB TEST	.00	9:53pm
AIR BLK	.00	9:54pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DURHAM COUNTY BAT MOBILE UNIT 8 310

Serial Number: 008736 Test Record Number: 858
Test Date: 11/12/2016 Test Time: 9:56pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:57pm
FLO	Pass	9:57pm
FC	Pass	9:57pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:57pm
SRC	Pass	9:57pm
DET	Pass	9:57pm
BAR	Pass	9:57pm
BT	Pass	9:57pm

Blank Tests

Test	Status	Time
AIR	Pass	9:58pm

Printer Tests

Test	Status	Time
PRNT	Pass	9:58pm

CRC Tests

Test	Status	Time
COMP	Pass	9:58pm
CAL	Pass	9:58pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County DURHAM Instrument Location Durham Co Jail

Instrument Serial No. 008924 219 S. Mangum ST

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of NOV, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DURHAM COUNTY DURHAM COUNTY JAIL 310

Serial Number: 008924
Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:34am
AIR BLK	.00	10:35am
ACCY CHK	.08	10:36am
AIR BLK	.00	10:37am
SUB TEST	.00	10:37am
AIR BLK	.00	10:38am
SUB TEST	.00	10:40am
AIR BLK	.00	10:41am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County DAVIDSON Instrument Location DAVIDSON CO JAIL
Instrument Serial No. 008845 Lexington, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845
Test Date: 11/22/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607601

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	3:22pm
AIR BLK	.00	3:23pm
ACCY CHK	.08	3:24pm
AIR BLK	.00	3:25pm
SUB TEST	.00	3:26pm
AIR BLK	.00	3:27pm
SUB TEST	.00	3:29pm
AIR BLK	.00	3:30pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY DAVIDSON CO JAIL 280

Serial Number: 008845 Test Record Number: 2400
Test Date: 11/22/2016 Test Time: 3:31pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:31pm
FLO	Pass	3:31pm
FC	Pass	3:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:32pm
SRC	Pass	3:32pm
DET	Pass	3:32pm
BAR	Pass	3:32pm
BT	Pass	3:32pm

Blank Tests

Test	Status	Time
AIR	Pass	3:32pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:32pm

CRC Tests

Test	Status	Time
COMP	Pass	3:33pm
CAL	Pass	3:33pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County DAVIDSON Instrument Location LEXINGTON
Instrument Serial No. 008718 POLICE DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008718
Test Date: 11/22/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test


Lot Number: AG513102
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:37pm
AIR BLK	.00	2:38pm
ACCY CHK	.08	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	.00	2:40pm
AIR BLK	.00	2:41pm
SUB TEST	.00	2:43pm
AIR BLK	.00	2:44pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY LEXINGTON PD 280

Serial Number: 008718 Test Record Number: 1494
Test Date: 11/22/2016 Test Time: 2:45pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:46pm
FLO	Pass	2:46pm
FC	Pass	2:46pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:46pm
SRC	Pass	2:46pm
DET	Pass	2:46pm
BAR	Pass	2:46pm
BT	Pass	2:46pm

Blank Tests

Test	Status	Time
AIR	Pass	2:47pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:47pm

CRC Tests

Test	Status	Time
COMP	Pass	2:47pm
CAL	Pass	2:47pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County DAVIDSON Instrument Location THOMASVILLE
Instrument Serial No. 008872 POLICE DEPARTMENT

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872

Test Date: 11/22/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	11:58am
AIR BLK	.00	11:59am
ACCY CHK	.07	12:00pm
AIR BLK	.00	12:00pm
SUB TEST	.00	12:01pm
AIR BLK	.00	12:02pm
SUB TEST	.00	12:04pm
AIR BLK	.00	12:05pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

DAVIDSON COUNTY THOMASVILLE PD 280

Serial Number: 008872 Test Record Number: 1289
Test Date: 11/22/2016 Test Time: 12:06pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:06pm
FLO	Pass	12:06pm
FC	Pass	12:06pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:06pm
SRC	Pass	12:06pm
DET	Pass	12:06pm
BAR	Pass	12:06pm
BT	Pass	12:06pm

Blank Tests

Test	Status	Time
AIR	Pass	12:07pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:07pm

CRC Tests

Test	Status	Time
COMP	Pass	12:07pm
GAL	Pass	12:07pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Dare Instrument Location Kill Devil Hills P.D.

Instrument Serial No. 008844 102 Town Hall Dr., Kill Devil Hills, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844
Test Date: 11/30/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607501
Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	11:20am
AIR BLK	.00	11:21am
ACCY CHK	.08	11:21am
AIR BLK	.00	11:22am
SUB TEST	.00	11:23am
AIR BLK	.00	11:24am
SUB TEST	.00	11:25am
AIR BLK	.00	11:26am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

DARE COUNTY KILL DEVIL HILLS PD 270

Serial Number: 008844 Test Record Number: 1867
Test Date: 11/30/2016 Test Time: 11:28am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:28am
FLO	Pass	11:28am
FC	Pass	11:28am

Temperature Tests

Test	Status	Time
FC1	Pass	11:28am
SRC	Pass	11:28am
DET	Pass	11:28am
BAR	Pass	11:28am
BT	Pass	11:28am

Blank Tests

Test	Status	Time
AIR	Pass	11:29am

Printer Tests

Test	Status	Time
PRNT	Pass	11:29am

CRC Tests

Test	Status	Time
COMP	Pass	11:29am
CAL	Pass	11:29am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Edgerombe Instrument Location Edgerombe Co. Magistrate's

Instrument Serial No. 008603 Office, 300 S. Anacounda Rd., Tarboro, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kerry H. P.

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR
320

Serial Number: 008603
Test Date: 11/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:04am
AIR BLK	.00	10:05am
ACCY CHK	.08	10:05am
AIR BLK	.00	10:06am
SUB TEST	.00	10:07am
AIR BLK	.00	10:08am
SUB TEST	.00	10:09am
AIR BLK	.00	10:10am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008603 Test Record Number: 1552
Test Date: 11/14/2016 Test Time: 10:12am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:13am
FLO	Pass	10:13am
FC	Pass	10:13am

Temperature Tests

Test	Status	Time
FC1	Pass	10:13am
SRC	Pass	10:13am
DET	Pass	10:13am
BAR	Pass	10:13am
BT	Pass	10:13am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am


Printer Tests

Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:14am
CAL	Pass	10:14am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Edgemont Instrument Location Edgemont Co. Magistrate's

Instrument Serial No. DD81663 Office, 300 S. Anaconda Rd., Tarboro, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Kelly M. D.

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR
320

Serial Number: 008663

Test Date: 11/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:37am
AIR BLK	.00	9:38am
ACCY CHK	.08	9:39am
AIR BLK	.00	9:40am
SUB TEST	.00	9:40am
AIR BLK	.00	9:41am
SUB TEST	.00	9:43am
AIR BLK	.00	9:44am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

EDGECOMBE COUNTY EDGECOMBE CO MAGISTR 320

Serial Number: 008663 Test Record Number: 2611
Test Date: 11/14/2016 Test Time: 9:45am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:45am
FLO	Pass	9:45am
FC	Pass	9:45am

Temperature Tests

Test	Status	Time
FC1	Pass	9:45am
SRC	Pass	9:45am
DET	Pass	9:45am
BAR	Pass	9:45am
BT	Pass	9:45am

Blank Tests

Test	Status	Time
AIR	Pass	9:46am


Printer Tests

Test	Status	Time
PRNT	Pass	9:46am

CRC Tests

Test	Status	Time
COMP	Pass	9:46am
CAL	Pass	9:46am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Bat mobile Unit 11
Instrument Serial No. 008973

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chloe Jones
Signature of Certifying Official

658
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH BAT MOBILE UNIT 11 330

Serial Number: 008973

Test Date: 11/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501


Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	7:11pm
AIR BLK	.00	7:12pm
ACCY CHK	.08	7:13pm
AIR BLK	.00	7:14pm
SUB TEST	.00	7:14pm
AIR BLK	.00	7:15pm
SUB TEST	.00	7:16pm
AIR BLK	.00	7:17pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

FORSYTH BAT MOBILE UNIT 11 330

Serial Number: 008973 Test Record Number: 222
Test Date: 11/11/2016 Test Time: 7:19pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:19pm
FLO	Pass	7:19pm
FC	Pass	7:19pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:19pm
SRC	Pass	7:19pm
DET	Pass	7:19pm
BAR	Pass	7:19pm
BT	Pass	7:19pm

Blank Tests

Test	Status	Time
AIR	Pass	7:20pm


Printer Tests

Test	Status	Time
PRNT	Pass	7:20pm

CRC Tests

Test	Status	Time
COMP	Pass	7:20pm
CAL	Pass	7:20pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 008925 Winston-Salem, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: 008925
Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607602

Exp Date: 03/16/2018


Test	g/210L	Time
DIAG	Pass	3:29pm
AIR BLK	.00	3:29pm
ACCY CHK	.08	3:30pm
AIR BLK	.00	3:31pm
SUB TEST	.00	3:32pm
AIR BLK	.00	3:33pm
SUB TEST	.00	3:34pm
AIR BLK	.00	3:35pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008925 Test Record Number: 1512
Test Date: 11/08/2016 Test Time: 3:36pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:36pm
FLO	Pass	3:36pm
FC	Pass	3:36pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:36pm
SRC	Pass	3:36pm
DET	Pass	3:36pm
BAR	Pass	3:36pm
BT	Pass	3:36pm

Blank Tests

Test	Status	Time
AIR	Pass	3:37pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:37pm

CRC Tests

Test	Status	Time
COMP	Pass	3:37pm
CAL	Pass	3:37pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 008583 Winston-Salem, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: 008583
Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607602

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	3:26pm
AIR BLK	.00	3:27pm
ACCY CHK	.08	3:28pm
AIR BLK	.00	3:29pm
SUB TEST	.00	3:30pm
AIR BLK	.00	3:31pm
SUB TEST	.00	3:33pm
AIR BLK	.00	3:34pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008583 Test Record Number: 6444
Test Date: 11/08/2016 Test Time: 3:35pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:35pm
FLO	Pass	3:35pm
FC	Pass	3:35pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:35pm
SRC	Pass	3:35pm
DET	Pass	3:35pm
BAR	Pass	3:35pm
BT	Pass	3:35pm

Blank Tests

Test	Status	Time
AIR	Pass	3:36pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:36pm

CRC Tests

Test	Status	Time
COMP	Pass	3:36pm
CAL	Pass	3:36pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Forsyth County Detention
Instrument Serial No. 808659 Winston-Salem, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY FORSYTH CO DETENTION
330

Serial Number: 008659
Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607602

Exp Date: 03/16/2018


Test	g/210L	Time
DIAG	Pass	3:15pm
AIR BLK	.00	3:15pm
ACCY CHK	.08	3:16pm
AIR BLK	.00	3:17pm
SUB TEST	.00	3:17pm
AIR BLK	.00	3:18pm
SUB TEST	.00	3:20pm
AIR BLK	.00	3:21pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY FORSYTH CO DETENTION 330

Serial Number: 008659 Test Record Number: 4178
Test Date: 11/08/2016 Test Time: 3:22pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:22pm
FLO	Pass	3:22pm
FC	Pass	3:22pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:22pm
SRC	Pass	3:22pm
DET	Pass	3:22pm
BAR	Pass	3:22pm
BT	Pass	3:22pm

Blank Tests

Test	Status	Time
AIR	Pass	3:23pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:23pm

CRC Tests

Test	Status	Time
COMP	Pass	3:23pm
CAL	Pass	3:23pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Forsyth Instrument Location Kernersville Police
Instrument Serial No. 008650 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650
Test Date: 11/08/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	2:23pm
AIR BLK	.00	2:24pm
ACCY CHK	.08	2:25pm
AIR BLK	.00	2:26pm
SUB TEST	.00	2:26pm
AIR BLK	.00	2:27pm
SUB TEST	.00	2:29pm
AIR BLK	.00	2:30pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

FORSYTH COUNTY KERNERSVILLE PD 330

Serial Number: 008650 Test Record Number: 1277
Test Date: 11/08/2016 Test Time: 2:32pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:32pm
FLO	Pass	2:32pm
FC	Pass	2:32pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:32pm
SRC	Pass	2:32pm
DET	Pass	2:32pm
BAR	Pass	2:32pm
BT	Pass	2:32pm

Blank Tests

Test	Status	Time
AIR	Pass	2:33pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:33pm

CRC Tests

Test	Status	Time
COMP	Pass	2:33pm
CAL	Pass	2:33pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County GRANVILLE Instrument Location CREEDMOOR PD

Instrument Serial No. 008641 111 MASONIC ST CREEDMOOR, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7 day of NOV, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Brachos

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641
Test Date: 11/07/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:03am
AIR BLK	.00	10:04am
ACCY CHK	.07	10:05am
AIR BLK	.00	10:06am
SUB TEST	.00	10:07am
AIR BLK	.00	10:08am
SUB TEST	.00	10:09am
AIR BLK	.00	10:10am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY CREEDMOOR PD 380

Serial Number: 008641 Test Record Number: 914
Test Date: 11/07/2016 Test Time: 10:11am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:11am
FLO	Pass	10:11am
FC	Pass	10:12am

Temperature Tests

Test	Status	Time
FC1	Pass	10:12am
SRC	Pass	10:12am
DET	Pass	10:12am
BAR	Pass	10:12am
BT	Pass	10:12am

Blank Tests

Test	Status	Time
AIR	Pass	10:12am


Printer Tests

Test	Status	Time
PRNT	Pass	10:12am

CRC Tests

Test	Status	Time
COMP	Pass	10:12am
CAL	Pass	10:12am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County GRANVILLE Instrument Location OXFORD PD

Instrument Serial No. 008923 204 E. McCLAWAHAN ST. OXFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 02 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GRANVILLE COUNTY OXFORD PD 380

Serial Number: 008923
Test Date: 11/02/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA


Agency: DHHS

Test Type: Breath Test

Lot Number: AG525303
Exp Date: 09/10/2017


Test	g/210L	Time
DIAG	Pass	3:31pm
AIR BLK	.00	3:32pm
ACCY CHK	.07	3:33pm
AIR BLK	.00	3:33pm
SUB TEST	.00	3:34pm
AIR BLK	.00	3:35pm
SUB TEST	.00	3:37pm
AIR BLK	.00	3:38pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GRANVILLE COUNTY OXFORD PD 380

Serial Number: 008923 Test Record Number: 1440
Test Date: 11/02/2016 Test Time: 3:38pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:39pm
FLO	Pass	3:39pm
FC	Pass	3:39pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:39pm
SRC	Pass	3:39pm
DET	Pass	3:39pm
BAR	Pass	3:39pm
BT	Pass	3:39pm

Blank Tests

Test	Status	Time
AIR	Pass	3:40pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:40pm

CRC Tests

Test	Status	Time
COMP	Pass	3:40pm
CAL	Pass	3:40pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location Greensboro Jail
Instrument Serial No. 008638

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638

Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	2:52pm
AIR BLK	.00	2:53pm
ACCY CHK	.08	2:53pm
AIR BLK	.00	2:54pm
SUB TEST	.00	2:55pm
AIR BLK	.00	2:56pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:58pm

Reported AC: .00 g/210L

L. K. Dean

Signature of Chemical Analyst

Court CVR

L. K. Dean
Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008638 Test Record Number: 2581
Test Date: 11/02/2016 Test Time: 2:59pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:59pm
FLO	Pass	2:59pm
FC	Pass	2:59pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:00pm
SRC	Pass	3:00pm
DET	Pass	3:00pm
BAR	Pass	3:00pm
BT	Pass	3:00pm

Blank Tests

Test	Status	Time
AIR	Pass	3:00pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:00pm

CRC Tests

Test	Status	Time
COMP	Pass	3:00pm
CAL	Pass	3:00pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford


Instrument Location Greensboro Jail

Instrument Serial No. 008794

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794
Test Date: 11/02/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501


Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	3:36pm
AIR BLK	.00	3:36pm
ACCY CHK	.08	3:37pm
AIR BLK	.00	3:38pm
SUB TEST	.00	3:38pm
AIR BLK	.00	3:39pm
SUB TEST	.00	3:41pm
AIR BLK	.00	3:42pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008794 Test Record Number: 5266
Test Date: 11/02/2016 Test Time: 3:43pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:44pm
FLO	Pass	3:44pm
FC	Pass	3:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:44pm
SRC	Pass	3:44pm
DET	Pass	3:44pm
BAR	Pass	3:44pm
BT	Pass	3:44pm

Blank Tests

Test	Status	Time
AIR	Pass	3:44pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:44pm

CRC Tests

Test	Status	Time
COMP	Pass	3:45pm
CAL	Pass	3:45pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Greensboro Jail

Instrument Serial No. 008790

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Deon
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790
Test Date: 11/02/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG621501
Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	4:05pm
AIR BLK	.00	4:05pm
ACCY CHK	.07	4:06pm
AIR BLK	.00	4:07pm
SUB TEST	.00	4:08pm
AIR BLK	.00	4:09pm
SUB TEST	.00	4:10pm
AIR BLK	.00	4:11pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY GREENSBORO JAIL 400

Serial Number: 008790 Test Record Number: 5716
Test Date: 11/02/2016 Test Time: 4:12pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	4:12pm
FLO	Pass	4:12pm
FC	Pass	4:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	4:12pm
SRC	Pass	4:12pm
DET	Pass	4:12pm
BAR	Pass	4:12pm
BT	Pass	4:12pm

Blank Tests

Test	Status	Time
AIR	Pass	4:13pm

Printer Tests

Test	Status	Time
PRNT	Pass	4:13pm

CRC Tests

Test	Status	Time
COMP	Pass	4:13pm
CAL	Pass	4:13pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location Bat Mobile Unit 8

Instrument Serial No 008736 High Point PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2014 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Jonya B Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD CO BAT MOBILE UNIT 8 410

Serial Number: 008736
Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:19pm
AIR BLK	.00	10:20pm
ACCY CHK	.07	10:21pm
AIR BLK	.00	10:21pm
SUB TEST	.00	10:22pm
AIR BLK	.00	10:23pm
SUB TEST	.00	10:24pm
AIR BLK	.00	10:25pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD CO BAT MOBILE UNIT 8 410

Serial Number: 008736 Test Record Number: 852
Test Date: 11/04/2016 Test Time: 10:27pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27pm
FLO	Pass	10:27pm
FC	Pass	10:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:27pm
SRC	Pass	10:27pm
DET	Pass	10:27pm
BAR	Pass	10:27pm
BT	Pass	10:27pm

Blank Tests

Test	Status	Time
AIR	Pass	10:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:28pm

CRC Tests

Test	Status	Time
COMP	Pass	10:28pm
CAL	Pass	10:28pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location Bat Mobile Unit 8
Instrument Serial No. 008775 High Point PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

6414
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HIGH POINT BAT MOBILE UNIT 8 401

Serial Number: 008775
Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:59pm
AIR BLK	.00	10:00pm
ACCY CHK	.08	10:01pm
AIR BLK	.00	10:02pm
SUB TEST	.00	10:02pm
AIR BLK	.00	10:03pm
SUB TEST	.00	10:05pm
AIR BLK	.00	10:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HIGH POINT BAT MOBILE UNIT 8 401

Serial Number: 008775 Test Record Number: 1682
Test Date: 11/04/2016 Test Time: 10:09pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:10pm
FLO	Pass	10:10pm
FC	Pass	10:10pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:10pm
SRC	Pass	10:10pm
DET	Pass	10:10pm
BAR	Pass	10:10pm
BT	Pass	10:10pm

Blank Tests

Test	Status	Time
AIR	Pass	10:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:11pm

CRC Tests

Test	Status	Time
COMP	Pass	10:11pm
CAL	Pass	10:11pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Guilford Instrument Location Bot mobile unit 8

Instrument Serial No. 668816 High Point PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Danya B Skinner
Signature of Certifying Official

1644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD CO BAT MOBILE UNIT 8 410

Serial Number: 008816

Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:37pm
AIR BLK	.00	10:38pm
ACCY CHK	.07	10:38pm
AIR BLK	.00	10:39pm
SUB TEST	.00	10:40pm
AIR BLK	.00	10:41pm
SUB TEST	.00	10:42pm
AIR BLK	.00	10:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD CO BAT MOBILE UNIT 8 410

Serial Number: 008816 Test Record Number: 7270
Test Date: 11/04/2016 Test Time: 10:44pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:45pm
FLO	Pass	10:45pm
FC	Pass	10:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:45pm
SRC	Pass	10:45pm
DET	Pass	10:45pm
BAR	Pass	10:45pm
BT	Pass	10:45pm

Blank Tests

Test	Status	Time
AIR	Pass	10:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	10:46pm

CRC Tests

Test	Status	Time
COMP	Pass	10:46pm
CAL	Pass	10:46pm

Preventive Maintenance
Status: Pass

Donya B Skunny
Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Gaston Instrument Location Gaston County SD
Instrument Serial No. 008910 425 N. Marietta St., Gastonia

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: 008910 Test Record Number: 541
Test Date: 11/04/2016 Test Time: 9:26am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:27am
FLO	Pass	9:27am
FC	Pass	9:27am

Temperature Tests

Test	Status	Time
FC1	Pass	9:27am
SRC	Pass	9:27am
DET	Pass	9:27am
BAR	Pass	9:27am
BT	Pass	9:27am

Blank Tests

Test	Status	Time
AIR	Pass	9:28am


Printer Tests

Test	Status	Time
PRNT	Pass	9:28am

CRC Tests

Test	Status	Time
COMP	Pass	9:28am
CAL	Pass	9:28am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

GASTON COUNTY GASTON COUNTY SD 350

Serial Number: 008910
Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607502

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:31am
AIR BLK	.00	9:32am
ACCY CHK	.08	9:32am
AIR BLK	.00	9:33am
SUB TEST	.00	9:34am
AIR BLK	.00	9:35am
SUB TEST	.00	9:36am
AIR BLK	.00	9:37am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Greene Instrument Location Greene Co. S.O.

Instrument Serial No. 008670 301 W. Greene St., Snow Hill
N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

697
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670
Test Date: 11/16/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: KEESLER, LINDA A
Permit Number: 11646E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513103
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:04pm
AIR BLK	.00	2:06pm
ACCY CHK	.07	2:06pm
AIR BLK	.00	2:07pm
SUB TEST	.00	2:09pm
AIR BLK	.00	2:09pm
SUB TEST	.00	2:11pm
AIR BLK	.00	2:12pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

GREENE COUNTY GREENE CO SO 390

Serial Number: 008670 Test Record Number: 1617
Test Date: 11/16/2016 Test Time: 2:13pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:13pm
FLO	Pass	2:13pm
FC	Pass	2:13pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:13pm
SRC	Pass	2:13pm
DET	Pass	2:13pm
BAR	Pass	2:13pm
BT	Pass	2:13pm

Blank Tests

Test	Status	Time
AIR	Pass	2:14pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:14pm

CRC Tests

Test	Status	Time
COMP	Pass	2:14pm
CAL	Pass	2:14pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Guilford Instrument Location High Point Jail
Instrument Serial No. 008655

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008655
Test Date: 11/22/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: DEAN, L K
Permit Number: 11598E
Effective:
05/01/2015-05/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG534802
Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	11:08am
AIR BLK	.00	11:09am
ACCY CHK	.07	11:09am
AIR BLK	.00	11:10am
SUB TEST	.00	11:11am
AIR BLK	.00	11:12am
SUB TEST	.00	11:14am
AIR BLK	.00	11:15am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT JAIL 401

Serial Number: 008655 Test Record Number: 3067
Test Date: 11/22/2016 Test Time: 11:16am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:16am
FLO	Pass	11:16am
FC	Pass	11:16am

Temperature Tests

Test	Status	Time
FC1	Pass	11:16am
SRC	Pass	11:16am
DET	Pass	11:16am
BAR	Pass	11:16am
BT	Pass	11:16am

Blank Tests

Test	Status	Time
AIR	Pass	11:17am

Printer Tests

Test	Status	Time
PRNT	Pass	11:17am

CRC Tests

Test	Status	Time
COMP	Pass	11:17am
CAL	Pass	11:17am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Guilford Instrument Location High Point
Instrument Serial No. 008828 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GUILFORD COUNTY HIGH POINT PD 401

Serial Number: 008828
Test Date: 11/22/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621501


Exp Date: 08/02/2018

Test	g/210L	Time
DIAG	Pass	10:19am
AIR BLK	.00	10:19am
ACCY CHK	.08	10:20am
AIR BLK	.00	10:21am
SUB TEST	.00	10:21am
AIR BLK	.00	10:22am
SUB TEST	.00	10:24am
AIR BLK	.00	10:25am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

GUILFORD COUNTY HIGH POINT PD 401

Serial Number: 008828 Test Record Number: 2260
Test Date: 11/22/2016 Test Time: 10:25am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:25am
FLO	Pass	10:25am
FC	Pass	10:26am

Temperature Tests

Test	Status	Time
FC1	Pass	10:26am
SRC	Pass	10:26am
DET	Pass	10:26am
BAR	Pass	10:26am
BT	Pass	10:26am

Blank Tests

Test	Status	Time
AIR	Pass	10:26am

Printer Tests

Test	Status	Time
PRNT	Pass	10:26am

CRC Tests

Test	Status	Time
COMP	Pass	10:27am
CAL	Pass	10:27am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Gates Instrument Location Gates Co. S.O.

Instrument Serial No. DD 8884 202 Court St., Gatesville, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

GATES COUNTY GATES CO SO 360

Serial Number: 008884

Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:20am
AIR BLK	.00	10:21am
ACCY CHK	.07	10:22am
AIR BLK	.00	10:23am
SUB TEST	.00	10:23am
AIR BLK	.00	10:24am
SUB TEST	.00	10:26am
AIR BLK	.00	10:27am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

GATES COUNTY GATES CO SO 360

Serial Number: 008884 Test Record Number: 761
Test Date: 11/21/2016 Test Time: 10:29am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:30am
FLO	Pass	10:30am
FC	Pass	10:30am

Temperature Tests

Test	Status	Time
FC1	Pass	10:30am
SRC	Pass	10:30am
DET	Pass	10:30am
BAR	Pass	10:30am
BT	Pass	10:30am

Blank Tests

Test	Status	Time
AIR	Pass	10:31am

Printer Tests

Test	Status	Time
PRNT	Pass	10:31am

CRC Tests

Test	Status	Time
COMP	Pass	10:31am
CAL	Pass	10:31am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County HARNETT Instrument Location HARNETT CO DET. CENTER

Instrument Serial No. 008730 LILLINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008730
Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	7:47am
AIR BLK	.00	7:47am
ACCY CHK	.07	7:48am
AIR BLK	.00	7:49am
SUB TEST	.00	7:50am
AIR BLK	.00	7:50am
SUB TEST	.00	7:52am
AIR BLK	.00	7:53am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008730 Test Record Number: 2642
Test Date: 11/30/2016 Test Time: 7:53am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:54am
FLO	Pass	7:54am
FC	Pass	7:54am

Temperature Tests

Test	Status	Time
FC1	Pass	7:54am
SRC	Pass	7:54am
DET	Pass	7:54am
BAR	Pass	7:54am
BT	Pass	7:54am

Blank Tests

Test	Status	Time
AIR	Pass	7:55am


Printer Tests

Test	Status	Time
PRNT	Pass	7:55am

CRC Tests

Test	Status	Time
COMP	Pass	7:55am
CAL	Pass	7:55am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County HARNETT Instrument Location HARNETT Co. DET. CTR.

Instrument Serial No. 008729 LILLINGTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008729

Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	7:51am
AIR BLK	.00	7:51am
ACCY CHK	.07	7:52am
AIR BLK	.00	7:53am
SUB TEST	.00	7:54am
AIR BLK	.00	7:55am
SUB TEST	.00	7:56am
AIR BLK	.00	7:57am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT COUNTY DETENTION CENTER 420

Serial Number: 008729 Test Record Number: 2088

Test Date: 11/30/2016 Test Time: 7:59am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:59am
FLO	Pass	7:59am
FC	Pass	7:59am

Temperature Tests

Test	Status	Time
FC1	Pass	7:59am
SRC	Pass	7:59am
DET	Pass	7:59am
BAR	Pass	7:59am
BT	Pass	7:59am

Blank Tests

Test	Status	Time
AIR	Pass	8:00am

Printer Tests

Test	Status	Time
PRNT	Pass	8:00am

CRC Tests

Test	Status	Time
COMP	Pass	8:00am
CAL	Pass	8:00am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Hoke Instrument Location Hoke Co. DETENTION CTR.
Instrument Serial No. 008852 RAEFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008852
Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	2:55pm
AIR BLK	.00	2:56pm
ACCY CHK	.08	2:57pm
AIR BLK	.00	2:58pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:59pm
SUB TEST	.00	3:01pm
AIR BLK	.00	3:02pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008852 Test Record Number: 819
Test Date: 11/10/2016 Test Time: 3:02pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:03pm
FLO	Pass	3:03pm
FC	Pass	3:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:03pm
SRC	Pass	3:03pm
DET	Pass	3:03pm
BAR	Pass	3:03pm
BT	Pass	3:03pm

Blank Tests

Test	Status	Time
AIR	Pass	3:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:04pm

CRC Tests

Test	Status	Time
COMP	Pass	3:04pm
CAL	Pass	3:04pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County HOKE Instrument Location HOKE CO. DETENTION CTR.
Instrument Serial No. 008855 RAEFORD, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008855

Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	3:00pm
AIR BLK	.00	3:01pm
ACCY CHK	.08	3:01pm
AIR BLK	.00	3:02pm
SUB TEST	.00	3:03pm
AIR BLK	.00	3:03pm
SUB TEST	.00	3:05pm
AIR BLK	.00	3:06pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

HOKE COUNTY DETENTION CENTER 460

Serial Number: 008855 Test Record Number: 1190
Test Date: 11/10/2016 Test Time: 3:17pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:18pm
FLO	Pass	3:18pm
FC	Pass	3:18pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:18pm
SRC	Pass	3:18pm
DET	Pass	3:18pm
BAR	Pass	3:18pm
BT	Pass	3:18pm

Blank Tests

Test	Status	Time
AIR	Pass	3:19pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:19pm

CRC Tests

Test	Status	Time
COMP	Pass	3:19pm
CAL	Pass	3:19pm

Preventive Maintenance
Status: *Pass*


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Hertford Instrument Location Murfreesboro, P.P.
Instrument Serial No. 00 8906 115 E. Broad St., Murfreesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Liba N. Kewer
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906

Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:09pm
AIR BLK	.00	12:10pm
ACCY CHK	.08	12:11pm
AIR BLK	.00	12:12pm
SUB TEST	.00	12:12pm
AIR BLK	.00	12:13pm
SUB TEST	.00	12:15pm
AIR BLK	.00	12:15pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY MURFREESBORO PD 450

Serial Number: 008906 Test Record Number: 571
Test Date: 11/02/2016 Test Time: 12:17pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:17pm
FLO	Pass	12:17pm
FC	Pass	12:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:17pm
SRC	Pass	12:17pm
DET	Pass	12:17pm
BAR	Pass	12:17pm
BT	Pass	12:17pm

Blank Tests

Test	Status	Time
AIR	Pass	12:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:18pm

CRC Tests

Test	Status	Time
COMP	Pass	12:18pm
CAL	Pass	12:18pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKEFORD Instrument Location Ahoskie P.O.
Instrument Serial No. 008648 705 W. main st., Ahoskie, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Linda A. Kuehl
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: 008848
Test Date: 11/02/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	11:19am
AIR BLK	.00	11:20am
ACCY CHK	.07	11:21am
AIR BLK	.00	11:22am
SUB TEST	.00	11:22am
AIR BLK	.00	11:23am
SUB TEST	.00	11:25am
AIR BLK	.00	11:26am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HERTFORD COUNTY AHOSKIE PD 450

Serial Number: 008848 Test Record Number: 1224
Test Date: 11/02/2016 Test Time: 11:26am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:27am
FLO	Pass	11:27am
FC	Pass	11:27am

Temperature Tests

Test	Status	Time
FC1	Pass	11:27am
SRC	Pass	11:27am
DET	Pass	11:27am
BAR	Pass	11:27am
BT	Pass	11:27am

Blank Tests

Test	Status	Time
AIR	Pass	11:28am


Printer Tests

Test	Status	Time
PRNT	Pass	11:28am

CRC Tests

Test	Status	Time
COMP	Pass	11:28am
CAL	Pass	11:28am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Harnett Instrument Location Bat mobile Unit 8

Instrument Serial No. 008775 Harnett Co SD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donna B. Kinney
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT CO BAT MOBILE UNIT 8 420

Serial Number: 008775
Test Date: 11/11/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Female
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B
Permit Number: 13651E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607501
Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	10:54pm
AIR BLK	.00	10:55pm
ACCY CHK	.08	10:55pm
AIR BLK	.00	10:56pm
SUB TEST	.00	10:57pm
AIR BLK	.00	10:58pm
SUB TEST	.00	10:59pm
AIR BLK	.00	11:00pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HARNETT CO BAT MOBILE UNIT 8 420

Serial Number: 008775 Test Record Number: 1686
Test Date: 11/11/2016 Test Time: 11:01pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:02pm
FLO	Pass	11:02pm
FC	Pass	11:02pm

Temperature Tests

Test	Status	Time
FC1	Pass	11:02pm
SRC	Pass	11:02pm
DET	Pass	11:02pm
BAR	Pass	11:02pm
BT	Pass	11:02pm

Blank Tests

Test	Status	Time
AIR	Pass	11:03pm

Printer Tests

Test	Status	Time
PRNT	Pass	11:03pm

CRC Tests

Test	Status	Time
COMP	Pass	11:03pm
CAL	Pass	11:03pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Harnett Instrument Location Pat mobile Unit 8
Instrument Serial No 008736 Harnett Co 50

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Donna B. Skinner
Signature of Certifying Official

644
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HARNETT CO BAT MOBILE UNIT 8 420

Serial Number: 008736
Test Date: 11/11/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Female

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SKINNER, TONYA B

Permit Number: 13651E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	8:26pm
AIR BLK	.00	8:27pm
ACCY CHK	.07	8:27pm
AIR BLK	.00	8:28pm
SUB TEST	.00	8:29pm
AIR BLK	.00	8:30pm
SUB TEST	.00	8:31pm
AIR BLK	.00	8:32pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

HARNETT CO BAT MOBILE UNIT 8 420

Serial Number: 008736 Test Record Number: 855
Test Date: 11/11/2016 Test Time: 8:34pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:34pm
FLO	Pass	8:34pm
FC	Pass	8:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:34pm
SRC	Pass	8:34pm
DET	Pass	8:34pm
BAR	Pass	8:34pm
BT	Pass	8:34pm

Blank Tests

Test	Status	Time
AIR	Pass	8:35pm

Printer Tests

Test	Status	Time
PRNT	Pass	8:35pm

CRC Tests

Test	Status	Time
COMP	Pass	8:35pm
CAL	Pass	8:35pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Hyde Instrument Location Hyde Co. S.O. - Ocracoke
Instrument Serial No. 008797 N.C. 17, Ocracoke, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



James Keel
Signature of Certifying Official

697
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HYDE COUNTY HYDE CO SO OCRACOKE 470

Serial Number: 008797

Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	3:41pm
AIR BLK	.00	3:42pm
ACCY CHK	.07	3:43pm
AIR BLK	.00	3:44pm
SUB TEST	.00	3:45pm
AIR BLK	.00	3:46pm
SUB TEST	.00	3:47pm
AIR BLK	.00	3:48pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HYDE COUNTY HYDE CO SO OCRACOKE 470

Serial Number: 008797 Test Record Number: 467
Test Date: 11/21/2016 Test Time: 3:49pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:49pm
FLO	Pass	3:49pm
FC	Pass	3:50pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:50pm
SRC	Pass	3:50pm
DET	Pass	3:50pm
BAR	Pass	3:50pm
BT	Pass	3:50pm

Blank Tests

Test	Status	Time
AIR	Pass	3:50pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:50pm

CRC Tests

Test	Status	Time
COMP	Pass	3:51pm
CAL	Pass	3:51pm

Preventive Maintenance
Status: *Pass*


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Hyde Instrument Location Hyde Co. S.D.

Instrument Serial No. 008801 1233 Main St., Swan Quarter, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 17th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key [Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801
Test Date: 11/17/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG513101
Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:00am
AIR BLK	.00	10:00am
ACCY CHK	.07	10:01am
AIR BLK	.00	10:02am
SUB TEST	.00	10:03am
AIR BLK	.00	10:04am
SUB TEST	.00	10:05am
AIR BLK	.00	10:06am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

HYDE COUNTY HYDE CO SO SWAN QUAR 470

Serial Number: 008801 Test Record Number: 408
Test Date: 11/17/2016 Test Time: 10:07am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:07am
FLO	Pass	10:07am
FC	Pass	10:08am

Temperature Tests

Test	Status	Time
FC1	Pass	10:08am
SRC	Pass	10:08am
DET	Pass	10:08am
BAR	Pass	10:08am
BT	Pass	10:08am

Blank Tests

Test	Status	Time
AIR	Pass	10:08am


Printer Tests

Test	Status	Time
PRNT	Pass	10:08am

CRC Tests

Test	Status	Time
COMP	Pass	10:09am
CAL	Pass	10:09am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson Instrument Location Henderson Co Detention
Instrument Serial No. 008822 Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DETENTION 440

Serial Number: 008822
Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:59pm
AIR BLK	.00	2:00pm
ACCY CHK	.08	2:01pm
AIR BLK	.00	2:02pm
SUB TEST	.00	2:02pm
AIR BLK	.00	2:03pm
SUB TEST	.00	2:05pm
AIR BLK	.00	2:06pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DETENTION 440

Serial Number: 008822 Test Record Number: 2026

Test Date: 11/30/2016 Test Time: 2:10pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:11pm
FLO	Pass	2:11pm
FC	Pass	2:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:11pm
SRC	Pass	2:11pm
DET	Pass	2:11pm
BAR	Pass	2:11pm
BT	Pass	2:11pm

Blank Tests

Test	Status	Time
AIR	Pass	2:11pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:12pm

CRC Tests

Test	Status	Time
COMP	Pass	2:12pm
CAL	Pass	2:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Henderson

Instrument Location Henderson Detention


Instrument Serial No. 008806

Hendersonville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806
Test Date: 11/30/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607601

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	1:58pm
AIR BLK	.00	1:59pm
ACCY CHK	.07	2:00pm
AIR BLK	.00	2:01pm
SUB TEST	.00	2:01pm
AIR BLK	.00	2:02pm
SUB TEST	.00	2:04pm
AIR BLK	.00	2:05pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

HENDERSON COUNTY DENTENTION 440

Serial Number: 008806 Test Record Number: 2071
Test Date: 11/30/2016 Test Time: 2:11pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:12pm
FLO	Pass	2:12pm
FC	Pass	2:12pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:12pm
SRC	Pass	2:12pm
DET	Pass	2:12pm
BAR	Pass	2:12pm
BT	Pass	2:12pm

Blank Tests

Test	Status	Time
AIR	Pass	2:12pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:12pm

CRC Tests

Test	Status	Time
COMP	Pass	2:13pm
CAL	Pass	2:13pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Statesville PD
Instrument Serial No. 008619 300 S. Tradd St., Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY STATESVILLE PD 480

Serial Number: 008619 Test Record Number: 1274
Test Date: 11/30/2016 Test Time: 10:01am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:02am
FLO	Pass	10:02am
FC	Pass	10:02am

Temperature Tests

Test	Status	Time
FC1	Pass	10:02am
SRC	Pass	10:02am
DET	Pass	10:02am
BAR	Pass	10:02am
BT	Pass	10:02am

Blank Tests

Test	Status	Time
AIR	Pass	10:03am


Printer Tests

Test	Status	Time
PRNT	Pass	10:03am

CRC Tests

Test	Status	Time
COMP	Pass	10:03am
CAL	Pass	10:03am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

IREDELL COUNTY STATESVILLE PD 480

Serial Number: 008619
Test Date: 11/30/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: HAYS, MARK D
Permit Number: 15924E
Effective:
01/01/2016-01/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG621403
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:06am
AIR BLK	.00	10:06am
ACCY CHK	.08	10:07am
AIR BLK	.00	10:08am
SUB TEST	.00	10:09am
AIR BLK	.00	10:09am
SUB TEST	.00	10:11am
AIR BLK	.00	10:12am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Iredell Instrument Location Iredell County SD
Instrument Serial No. 008809 201 E. Water St. Statesville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: 008809 Test Record Number: 3511
Test Date: 11/29/2016 Test Time: 10:45am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:45am
FLO	Pass	10:45am
FC	Pass	10:45am

Temperature Tests

Test	Status	Time
FC1	Pass	10:45am
SRC	Pass	10:45am
DET	Pass	10:45am
BAR	Pass	10:45am
BT	Pass	10:45am

Blank Tests

Test	Status	Time
AIR	Pass	10:46am


Printer Tests

Test	Status	Time
PRNT	Pass	10:46am

CRC Tests

Test	Status	Time
COMP	Pass	10:46am
CAL	Pass	10:46am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

IREDELL COUNTY IREDELL COUNTY SD 480

Serial Number: 008809
Test Date: 11/29/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607502

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	10:49am
AIR BLK	.00	10:49am
ACCY CHK	.07	10:50am
AIR BLK	.00	10:51am
SUB TEST	.00	10:52am
AIR BLK	.00	10:53am
SUB TEST	.00	10:54am
AIR BLK	.00	10:55am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Jackson Instrument Location Jackson Co. Jail
Instrument Serial No. 008722 Sylva, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cuthbert

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008722

Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	9:53am
AIR BLK	.00	9:54am
ACCY CHK	.07	9:54am
AIR BLK	.00	9:56am
SUB TEST	.00	9:56am
AIR BLK	.00	9:57am
SUB TEST	.00	9:59am
AIR BLK	.00	9:59am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008722 Test Record Number: 891
Test Date: 11/02/2016 Test Time: 10:01am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:01am
FLO	Pass	10:01am
FC	Pass	10:01am

Temperature Tests

Test	Status	Time
FC1	Pass	10:01am
SRC	Pass	10:01am
DET	Pass	10:01am
BAR	Pass	10:01am
BT	Pass	10:01am

Blank Tests

Test	Status	Time
AIR	Pass	10:02am


Printer Tests

Test	Status	Time
PRNT	Pass	10:02am

CRC Tests

Test	Status	Time
COMP	Pass	10:02am
CAL	Pass	10:02am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Jackson Instrument Location Jackson Co. Jail

Instrument Serial No. 008708 Sylva, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Erin R. Arthur
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008708
Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534802


Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	9:52am
AIR BLK	.00	9:53am
ACCY CHK	.07	9:54am
AIR BLK	.00	9:55am
SUB TEST	.00	9:55am
AIR BLK	.00	9:56am
SUB TEST	.00	9:57am
AIR BLK	.00	9:58am

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JACKSON COUNTY JACKSON COUNTY JAIL 490

Serial Number: 008708 Test Record Number: 1128
Test Date: 11/02/2016 Test Time: 10:00am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:00am
FLO	Pass	10:00am
FC	Pass	10:00am

Temperature Tests

Test	Status	Time
FC1	Pass	10:00am
SRC	Pass	10:00am
DET	Pass	10:00am
BAR	Pass	10:00am
BT	Pass	10:00am

Blank Tests

Test	Status	Time
AIR	Pass	10:01am

Printer Tests

Test	Status	Time
PRNT	Pass	10:01am

CRC Tests

Test	Status	Time
COMP	Pass	10:01am
CAL	Pass	10:01am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County JONES Instrument Location BAT MOBILE UNIT 9
Instrument Serial No. 008907 KINSTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 19 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Celen R. B...

Signature of Certifying Official

648

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JONES COUNTY BAT MOBILE UNIT 9 510

Serial Number: 008707

Test Date: 11/19/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, ALVIN R

Permit Number: 15671E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901


Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	9:46pm
AIR BLK	.00	9:47pm
ACCY CHK	.07	9:48pm
AIR BLK	.00	9:49pm
SUB TEST	.00	9:49pm
AIR BLK	.00	9:50pm
SUB TEST	.00	9:52pm
AIR BLK	.00	9:52pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

JONES COUNTY BAT MOBILE UNIT 9 510

Serial Number: 008707 Test Record Number: 2362
Test Date: 11/19/2016 Test Time: 9:53pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:54pm
FLO	Pass	9:54pm
FC	Pass	9:54pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:54pm
SRC	Pass	9:54pm
DET	Pass	9:54pm
BAR	Pass	9:54pm
BT	Pass	9:54pm

Blank Tests

Test	Status	Time
AIR	Pass	9:55pm


Printer Tests

Test	Status	Time
PRNT	Pass	9:55pm

CRC Tests

Test	Status	Time
COMP	Pass	9:55pm
CAL	Pass	9:55pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Jackson


Instrument Location BAT Mobile Unit 11

Instrument Serial No. 008970

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Chris J. Jolly

Signature of Certifying Official

658

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JACKSON BAT MOBILE UNIT 11 490

Serial Number: 008970

Test Date: 11/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: TOWERY, CHAD V

Permit Number: 26632E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:56pm
AIR BLK	.00	9:57pm
ACCY CHK	.08	9:58pm
AIR BLK	.00	9:59pm
SUB TEST	.00	9:59pm
AIR BLK	.00	10:00pm
SUB TEST	.00	10:01pm
AIR BLK	.00	10:02pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

JACKSON BAT MOBILE UNIT 11 490

Serial Number: 008970 Test Record Number: 246
Test Date: 11/03/2016 Test Time: 10:04pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:04pm
FLO	Pass	10:04pm
FC	Pass	10:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:04pm
SRC	Pass	10:04pm
DET	Pass	10:04pm
BAR	Pass	10:04pm
BT	Pass	10:04pm

Blank Tests

Test	Status	Time
AIR	Pass	10:05pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:05pm

CRC Tests

Test	Status	Time
COMP	Pass	10:05pm
CAL	Pass	10:05pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Johnston Co. Instrument Location Clayton Police Dept.

Instrument Serial No. 108658

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

654
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: 008658
Test Date: 11/23/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, GRAYHAM C

Permit Number: 7682E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	3:30pm
AIR BLK	.00	3:31pm
ACCY CHK	.08	3:32pm
AIR BLK	.00	3:33pm
SUB TEST	.00	3:34pm
AIR BLK	.00	3:34pm
SUB TEST	.00	3:36pm
AIR BLK	.00	3:37pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

JOHNSTON COUNTY CLAYTON PD. 500

Serial Number: 008658 Test Record Number: 1267
Test Date: 11/23/2016 Test Time: 3:41pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:42pm
FLO	Pass	3:42pm
FC	Pass	3:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:42pm
SRC	Pass	3:42pm
DET	Pass	3:42pm
BAR	Pass	3:42pm
BT	Pass	3:42pm

Blank Tests

Test	Status	Time
AIR	Pass	3:43pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:43pm

CRC Tests

Test	Status	Time
COMP	Pass	3:43pm
CAL	Pass	3:43pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lenoir Instrument Location Lenoir Co. S.O.

Instrument Serial No. 008851 130 Queen St., Kinston, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008851
Test Date: 11/01/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	10:13am
AIR BLK	.00	10:14am
ACCY CHK	.08	10:15am
AIR BLK	.00	10:16am
SUB TEST	.00	10:16am
AIR BLK	.00	10:17am
SUB TEST	.00	10:19am
AIR BLK	.00	10:20am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008851 Test Record Number: 570
Test Date: 11/01/2016 Test Time: 10:21am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:21am
FLO	Pass	10:21am
FC	Pass	10:21am

Temperature Tests

Test	Status	Time
FC1	Pass	10:21am
SRC	Pass	10:21am
DET	Pass	10:21am
BAR	Pass	10:21am
BT	Pass	10:21am

Blank Tests

Test	Status	Time
AIR	Pass	10:22am


Printer Tests

Test	Status	Time
PRNT	Pass	10:22am

CRC Tests

Test	Status	Time
COMP	Pass	10:22am
CAL	Pass	10:22am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Lenoir Instrument Location Kinston P.D.

Instrument Serial No. 008624 205 E. King St., Kinston, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 1st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624

Test Date: 11/01/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513103

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	9:22am
AIR BLK	.00	9:23am
ACCY CHK	.07	9:24am
AIR BLK	.00	9:25am
SUB TEST	.00	9:25am
AIR BLK	.00	9:26am
SUB TEST	.00	9:28am
AIR BLK	.00	9:29am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY KINSTON PD 530

Serial Number: 008624 Test Record Number: 1568
Test Date: 11/01/2016 Test Time: 9:30am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	9:31am
FLO	Pass	9:31am
FC	Pass	9:31am

Temperature Tests

Test	Status	Time
FC1	Pass	9:31am
SRC	Pass	9:31am
DET	Pass	9:31am
BAR	Pass	9:31am
BT	Pass	9:31am

Blank Tests

Test	Status	Time
AIR	Pass	9:32am


Printer Tests

Test	Status	Time
PRNT	Pass	9:32am

CRC Tests

Test	Status	Time
COMP	Pass	9:32am
CAL	Pass	9:32am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Lenoir Instrument Location Lenoir Co. S.O.
Instrument Serial No. 008639 130 Queen St., Kinston, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639

Test Date: 11/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	9:19am
AIR BLK	.00	9:19am
ACCY CHK	.07	9:20am
AIR BLK	.00	9:21am
SUB TEST	.00	9:21am
AIR BLK	.00	9:22am
SUB TEST	.00	9:24am
AIR BLK	.00	9:25am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

LENOIR COUNTY LENOIR CO SO 530

Serial Number: 008639 Test Record Number: 2905
Test Date: 11/18/2016 Test Time: 9:26am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:26am
FLO	Pass	9:26am
FC	Pass	9:26am

Temperature Tests

Test	Status	Time
FC1	Pass	9:27am
SRC	Pass	9:27am
DET	Pass	9:27am
BAR	Pass	9:27am
BT	Pass	9:27am

Blank Tests

Test	Status	Time
AIR	Pass	9:27am

Printer Tests

Test	Status	Time
PRNT	Pass	9:27am

CRC Tests

Test	Status	Time
COMP	Pass	9:27am
CAL	Pass	9:27am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Macon

Instrument Location Macon Co. Magistrate


Instrument Serial No. 008795

Highlands, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Paul R. Cotton

Signature of Certifying Official

635

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: 008795
Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG526401


Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	3:20pm
AIR BLK	.00	3:21pm
ACCY CHK	.07	3:22pm
AIR BLK	.00	3:23pm
SUB TEST	.00	3:23pm
AIR BLK	.00	3:24pm
SUB TEST	.00	3:26pm
AIR BLK	.00	3:27pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MACON COUNTY MACON CO MAGISTRATE 550

Serial Number: 008795 Test Record Number: 412
Test Date: 11/21/2016 Test Time: 3:27pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	3:28pm
FLO	Pass	3:28pm
FC	Pass	3:28pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:28pm
SRC	Pass	3:28pm
DET	Pass	3:28pm
BAR	Pass	3:28pm
BT	Pass	3:28pm

Blank Tests

Test	Status	Time
AIR	Pass	3:29pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:29pm

CRC Tests

Test	Status	Time
COMP	Pass	3:29pm
CAL	Pass	3:29pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County MOORE Instrument Location PINEHURST POLICE DEPT.
Instrument Serial No. 008710 PINEHURST NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710

Test Date: 11/18/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:35pm
AIR BLK	.00	1:35pm
ACCY CHK	.07	1:36pm
AIR BLK	.00	1:37pm
SUB TEST	.00	1:37pm
AIR BLK	.00	1:38pm
SUB TEST	.00	1:40pm
AIR BLK	.00	1:41pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MOORE COUNTY PINEHURST PD. 620

Serial Number: 008710 Test Record Number: 1350
Test Date: 11/18/2016 Test Time: 1:41pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:42pm
FLO	Pass	1:42pm
FC	Pass	1:42pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:42pm
SRC	Pass	1:42pm
DET	Pass	1:42pm
BAR	Pass	1:42pm
BT	Pass	1:42pm

Blank Tests

Test	Status	Time
AIR	Pass	1:43pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:43pm

CRC Tests

Test	Status	Time
COMP	Pass	1:43pm
CAL	Pass	1:43pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mitchell Instrument Location Spivee Dine PD
Instrument Serial No. 008726 Spivee Dine, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726
Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434901

Exp Date: 12/15/2016

Test	g/210L	Time
DIAG	Pass	2:44pm
AIR BLK	.00	2:45pm
ACCY CHK	.08	2:45pm
AIR BLK	.00	2:46pm
SUB TEST	.00	2:47pm
AIR BLK	.00	2:48pm
SUB TEST	.00	2:49pm
AIR BLK	.00	2:50pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MITCHELL COUNTY SPRUCE PINE PD 600

Serial Number: 008726 Test Record Number: 841
Test Date: 11/02/2016 Test Time: 2:52pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:52pm
FLO	Pass	2:52pm
FC	Pass	2:52pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:52pm
SRC	Pass	2:52pm
DET	Pass	2:52pm
BAR	Pass	2:52pm
BT	Pass	2:52pm

Blank Tests

Test	Status	Time
AIR	Pass	2:53pm


Printer Tests

Test	Status	Time
PRNT	Pass	2:53pm

CRC Tests

Test	Status	Time
COMP	Pass	2:53pm
CAL	Pass	2:53pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County MECKLENBURG Instrument Location BAT MOBILE 7
Instrument Serial No. 008972

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

659
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008972
Test Date: 11/10/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: INGLE, LARRY W
Permit Number: 7281E
Effective:
02/01/2016-02/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG621404
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	7:31pm
AIR BLK	.00	7:32pm
ACCY CHK	.07	7:33pm
AIR BLK	.00	7:34pm
SUB TEST	.00	7:35pm
AIR BLK	.00	7:36pm
SUB TEST	.00	7:38pm
AIR BLK	.00	7:39pm

Reported AC: 00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008972 Test Record Number: 243
Test Date: 11/10/2016 Test Time: 7:40pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:40pm
FLO	Pass	7:40pm
FC	Pass	7:40pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:41pm
SRC	Pass	7:41pm
DET	Pass	7:41pm
BAR	Pass	7:41pm
BT	Pass	7:41pm

Blank Tests

Test	Status	Time
AIR	Pass	7:41pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:41pm

CRC Tests

Test	Status	Time
COMP	Pass	7:41pm
CAL	Pass	7:41pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County MECKLENBURG Instrument Location RAF MOBILE 7

Instrument Serial No. 008971

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]

Signature of Certifying Official

659

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008971
Test Date: 11/10/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: INGLE, LARRY W

Permit Number: 7281E

Effective:

02/01/2016-02/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	7:16pm
AIR BLK	.00	7:17pm
ACCY CHK	.08	7:17pm
AIR BLK	.00	7:18pm
SUB TEST	.00	7:19pm
AIR BLK	.00	7:20pm
SUB TEST	.00	7:21pm
AIR BLK	.00	7:22pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG BAT MOBILE UNIT 7 590

Serial Number: 008971 Test Record Number: 145
Test Date: 11/10/2016 Test Time: 7:24pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	7:25pm
FLO	Pass	7:25pm
FC	Pass	7:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	7:25pm
SRC	Pass	7:25pm
DET	Pass	7:25pm
BAR	Pass	7:25pm
BT	Pass	7:25pm

Blank Tests

Test	Status	Time
AIR	Pass	7:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	7:26pm

CRC Tests

Test	Status	Time
COMP	Pass	7:26pm
CAL	Pass	7:26pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Martin Instrument Location Martin Co. S.O.
Instrument Serial No. 008912 305 E. Main St., Williamston, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 16th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Keen M O

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008912
Test Date: 11/16/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:07am
AIR BLK	.00	11:07am
ACCY CHK	.08	11:08am
AIR BLK	.00	11:09am
SUB TEST	.00	11:10am
AIR BLK	.00	11:11am
SUB TEST	.00	11:13am
AIR BLK	.00	11:13am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

MARTIN COUNTY SHERIFF'S OFFICE 570

Serial Number: 008912 Test Record Number: 1190
Test Date: 11/16/2016 Test Time: 11:15am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:15am
FLO	Pass	11:15am
FC	Pass	11:15am

Temperature Tests

Test	Status	Time
FC1	Pass	11:16am
SRC	Pass	11:16am
DET	Pass	11:16am
BAR	Pass	11:16am
BT	Pass	11:16am

Blank Tests

Test	Status	Time
AIR	Pass	11:16am

Printer Tests

Test	Status	Time
PRNT	Pass	11:16am

CRC Tests

Test	Status	Time
COMP	Pass	11:16am
CAL	Pass	11:16am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Cornelius PD
Instrument Serial No. 008692 21440 Catawba Ave., Cornelius

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692 Test Record Number: 2559
Test Date: 11/21/2016 Test Time: 9:54am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:55am
FLO	Pass	9:55am
FC	Pass	9:55am

Temperature Tests

Test	Status	Time
FC1	Pass	9:55am
SRC	Pass	9:55am
DET	Pass	9:55am
BAR	Pass	9:55am
BT	Pass	9:55am

Blank Tests

Test	Status	Time
AIR	Pass	9:55am


Printer Tests

Test	Status	Time
PRNT	Pass	9:56am

CRC Tests

Test	Status	Time
COMP	Pass	9:56am
CAL	Pass	9:56am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG COUNTY CORNELIUS PD 590

Serial Number: 008692
Test Date: 11/21/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	9:58am
AIR BLK	.00	9:59am
ACCY CHK	.08	9:59am
AIR BLK	.00	10:00am
SUB TEST	.00	10:01am
AIR BLK	.00	10:02am
SUB TEST	.00	10:03am
AIR BLK	.00	10:04am

Reported AC: 00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Mecklenburg Instrument Location Huntersville PD
Instrument Serial No. 008747 9630 Julian Clarke Ave., Huntersville

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Handwritten Signature]
Signature of Certifying Official

656

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747 Test Record Number: 2552
Test Date: 11/21/2016 Test Time: 10:36am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:37am
FLO	Pass	10:37am
FC	Pass	10:37am

Temperature Tests

Test	Status	Time
FC1	Pass	10:37am
SRC	Pass	10:37am
DET	Pass	10:37am
BAR	Pass	10:37am
BT	Pass	10:37am

Blank Tests

Test	Status	Time
AIR	Pass	10:38am


Printer Tests

Test	Status	Time
PRNT	Pass	10:38am

CRC Tests

Test	Status	Time
COMP	Pass	10:38am
CAL	Pass	10:38am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

MECKLENBURG HUNTERSVILLE PD 590

Serial Number: 008747
Test Date: 11/21/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	10:40am
AIR BLK	.00	10:41am
ACCY CHK	.08	10:41am
AIR BLK	.00	10:42am
SUB TEST	.00	10:43am
AIR BLK	.00	10:44am
SUB TEST	.00	10:45am
AIR BLK	.00	10:46am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County McDowell Instrument Location McDowell Co Jail

Instrument Serial No. 008892 MARION, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 5 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MCDOWELL COUNTY JAIL 580

Serial Number: 008892

Test Date: 11/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	2:56pm
AIR BLK	.00	2:57pm
ACCY CHK	.07	2:58pm
AIR BLK	.00	2:58pm
SUB TEST	.00	2:59pm
AIR BLK	.00	3:00pm
SUB TEST	.00	3:01pm
AIR BLK	.00	3:02pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MCDOWELL COUNTY JAIL 580

Serial Number: 008892 Test Record Number: 445
Test Date: 11/03/2016 Test Time: 3:03pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:03pm
FLO	Pass	3:03pm
FC	Pass	3:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:04pm
SRC	Pass	3:04pm
DET	Pass	3:04pm
BAR	Pass	3:04pm
BT	Pass	3:04pm

Blank Tests

Test	Status	Time
AIR	Pass	3:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:04pm

CRC Tests

Test	Status	Time
COMP	Pass	3:04pm
CAL	Pass	3:04pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County McDowell Instrument Location McDowell Co. Jail

Instrument Serial No. 008888 Marion, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

MCDOWELL COUNTY JAIL 580

Serial Number: 008888
Test Date: 11/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	2:55pm
AIR BLK	.00	2:56pm
ACCY CHK	.07	2:57pm
AIR BLK	.00	2:58pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:59pm
SUB TEST	.00	3:01pm
AIR BLK	.00	3:02pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

MCDOWELL COUNTY JAIL 580

Serial Number: 008888 Test Record Number: 1408
Test Date: 11/03/2016 Test Time: 3:03pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:03pm
FLO	Pass	3:03pm
FC	Pass	3:03pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:03pm
SRC	Pass	3:03pm
DET	Pass	3:03pm
BAR	Pass	3:03pm
BT	Pass	3:03pm

Blank Tests

Test	Status	Time
AIR	Pass	3:04pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:04pm

CRC Tests

Test	Status	Time
COMP	Pass	3:04pm
CAL	Pass	3:04pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County NASH Instrument Location ROCKY MOUNT PD
Instrument Serial No. 008740 # 7 GOVERNMENT PLAZA
ROCKY MOUNT, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008740
Test Date: 11/03/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	8:29am
AIR BLK	.00	8:30am
ACCY CHK	.07	8:30am
AIR BLK	.00	8:31am
SUB TEST	.00	8:31am
AIR BLK	.00	8:32am
SUB TEST	.00	8:34am
AIR BLK	.00	8:35am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY ROCKY MOUNT PD 630

Serial Number: 008740 Test Record Number: 598
Test Date: 11/03/2016 Test Time: 8:42am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:43am
FLO	Pass	8:43am
FC	Pass	8:43am

Temperature Tests

Test	Status	Time
FC1	Pass	8:43am
SRC	Pass	8:43am
DET	Pass	8:43am
BAR	Pass	8:43am
BT	Pass	8:43am

Blank Tests

Test	Status	Time
AIR	Pass	8:44am

Printer Tests

Test	Status	Time
PRNT	Pass	8:44am

CRC Tests

Test	Status	Time
COMP	Pass	8:44am
CAL	Pass	8:44am

Preventive Maintenance
Status: Pass

Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County NASH Instrument Location BAT MOBILE UNIT #6
Instrument Serial No. 005869 NASHVILLE

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of Nov, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

661
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

NASH COUNTY BAT MOBILE UNIT 6 630

Serial Number: 008869 Test Record Number: 929
Test Date: 11/11/2016 Test Time: 8:55pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:55pm
FLO	Pass	8:55pm
FC	Pass	8:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:56pm
SRC	Pass	8:56pm
DET	Pass	8:56pm
BAR	Pass	8:56pm
BT	Pass	8:56pm

Blank Tests

Test	Status	Time
AIR	Pass	8:56pm


Printer Tests

Test	Status	Time
PRNT	Pass	8:56pm

CRC Tests

Test	Status	Time
COMP	Pass	8:56pm
CAL	Pass	8:56pm

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Subject Test

NASH COUNTY BAT MOBILE UNIT 6 630

Serial Number: 008869
Test Date: 11/11/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, JOCK B
Permit Number: 20630E
Effective:
05/01/2016-05/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG621403
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	8:47pm
AIR BLK	.00	8:48pm
ACCY CHK	.07	8:49pm
AIR BLK	.00	8:50pm
SUB TEST	.00	8:50pm
AIR BLK	.00	8:51pm
SUB TEST	.00	8:53pm
AIR BLK	.00	8:54pm

Reported AC:  .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ORANGE Instrument Location CHAPEL HILL PD

Instrument Serial No. 828 MARTIN Luther King, Jr. Blvd
Chapel Hill, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of NOV, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008873

Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	9:00am
AIR BLK	.00	9:01am
ACCY CHK	.08	9:01am
AIR BLK	.00	9:03am
SUB TEST	.00	9:04am
AIR BLK	.00	9:05am
SUB TEST	.00	9:06am
AIR BLK	.00	9:07am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008873 Test Record Number: 1345
Test Date: 11/04/2016 Test Time: 9:10am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:10am
FLO	Pass	9:10am
FC	Pass	9:10am

Temperature Tests

Test	Status	Time
FC1	Pass	9:10am
SRC	Pass	9:10am
DET	Pass	9:10am
BAR	Pass	9:10am
BT	Pass	9:10am

Blank Tests

Test	Status	Time
AIR	Pass	9:11am


Printer Tests

Test	Status	Time
PRNT	Pass	9:11am

CRC Tests

Test	Status	Time
COMP	Pass	9:11am
CAL	Pass	9:11am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County ORANGE Instrument Location CHAPEL HILL PD

Instrument Serial No. 008839 828 MARTIN LUTHER KING, JR. BLVD
CHAPEL HILL, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of NOV, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839

Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017


Test	g/210L	Time
DIAG	Pass	12:51pm
AIR BLK	.00	12:52pm
ACCY CHK	.08	12:52pm
AIR BLK	.00	12:53pm
SUB TEST	.00	12:54pm
AIR BLK	.00	12:55pm
SUB TEST	.00	12:56pm
AIR BLK	.00	12:57pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ORANGE COUNTY CHAPEL HILL PD 670

Serial Number: 008839 Test Record Number: 1498
Test Date: 11/10/2016 Test Time: 1:03pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:04pm
FLO	Pass	1:04pm
FC	Pass	1:04pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:04pm
SRC	Pass	1:04pm
DET	Pass	1:04pm
BAR	Pass	1:04pm
BT	Pass	1:04pm

Blank Tests

Test	Status	Time
AIR	Pass	1:04pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:04pm

CRC Tests

Test	Status	Time
COMP	Pass	1:05pm
CAL	Pass	1:05pm

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Polk Instrument Location Polk County SD
Instrument Serial No. 008893 46 Ward St, Columbus

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



M. A. [Signature]
Signature of Certifying Official

656
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

POLK COUNTY POLK COUNTY SD 740

Serial Number: 008893 Test Record Number: 1438
Test Date: 11/14/2016 Test Time: 10:47am EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:47am
FLO	Pass	10:47am
FC	Pass	10:48am

Temperature Tests

Test	Status	Time
FC1	Pass	10:48am
SRC	Pass	10:48am
DET	Pass	10:48am
BAR	Pass	10:48am
BT	Pass	10:48am

Blank Tests

Test	Status	Time
AIR	Pass	10:48am


Printer Tests

Test	Status	Time
PRNT	Pass	10:48am

CRC Tests

Test	Status	Time
COMP	Pass	10:49am
CAL	Pass	10:49am

Preventive Maintenance
Status: Pass



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Subject Test

POLK COUNTY POLK COUNTY SD 740

Serial Number: 008893
Test Date: 11/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HAYS, MARK D

Permit Number: 15924E

Effective:

01/01/2016-01/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG434201

Exp Date: 12/18/2016

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:51am
ACCY CHK	.08	10:52am
AIR BLK	.00	10:53am
SUB TEST	.00	10:53am
AIR BLK	.00	10:54am
SUB TEST	.00	10:56am
AIR BLK	.00	10:57am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location BAT Mobile Unit 10
Instrument Serial No. 008686 Elizabeth City P.O.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

660
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY BAT MOBILE UNIT 10
690

Serial Number: 008686
Test Date: 11/23/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, JASON R
Permit Number: 19145E
Effective:
03/01/2016-03/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:07pm
AIR BLK	.00	10:08pm
ACCY CHK	.07	10:08pm
AIR BLK	.00	10:09pm
SUB TEST	.00	10:10pm
AIR BLK	.00	10:11pm
SUB TEST	.00	10:12pm
AIR BLK	.00	10:13pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY BAT MOBILE UNIT 10 690

Serial Number: 008686 Test Record Number: 6455
Test Date: 11/23/2016 Test Time: 10:17pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:17pm
FLO	Pass	10:17pm
FC	Pass	10:17pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:17pm
SRC	Pass	10:17pm
DET	Pass	10:17pm
BAR	Pass	10:17pm
BT	Pass	10:17pm

Blank Tests

Test	Status	Time
AIR	Pass	10:18pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:18pm

CRC Tests

Test	Status	Time
COMP	Pass	10:18pm
CAL	Pass	10:18pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. 008668 124 Detention Dr., Greenville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668
Test Date: 11/23/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607501
Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	9:50am
AIR BLK	.00	9:51am
ACCY CHK	.07	9:52am
AIR BLK	.00	9:53am
SUB TEST	.00	9:53am
AIR BLK	.00	9:54am
SUB TEST	.00	9:56am
AIR BLK	.00	9:57am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008668 Test Record Number: 2788
Test Date: 11/23/2016 Test Time: 9:58am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:58am
FLO	Pass	9:58am
FC	Pass	9:58am

Temperature Tests

Test	Status	Time
FC1	Pass	9:58am
SRC	Pass	9:58am
DET	Pass	9:58am
BAR	Pass	9:58am
BT	Pass	9:58am

Blank Tests

Test	Status	Time
AIR	Pass	9:59am

Printer Tests

Test	Status	Time
PRNT	Pass	9:59am

CRC Tests

Test	Status	Time
COMP	Pass	9:59am
CAL	Pass	9:59am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Pitt Co. Detention Center
Instrument Serial No. 008646 124 Detention Dr., Greenville, NC.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23rd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646
Test Date: 11/23/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607601

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	9:33am
AIR BLK	.00	9:33am
ACCY CHK	.08	9:34am
AIR BLK	.00	9:35am
SUB TEST	.00	9:36am
AIR BLK	.00	9:36am
SUB TEST	.00	9:38am
AIR BLK	.00	9:39am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PITT COUNTY PITT CO DETENTION 730

Serial Number: 008646 Test Record Number: 3334
Test Date: 11/23/2016 Test Time: 9:40am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:40am
FLO	Pass	9:40am
FC	Pass	9:40am

Temperature Tests

Test	Status	Time
FC1	Pass	9:41am
SRC	Pass	9:41am
DET	Pass	9:41am
BAR	Pass	9:41am
BT	Pass	9:41am

Blank Tests

Test	Status	Time
AIR	Pass	9:41am


Printer Tests

Test	Status	Time
PRNT	Pass	9:41am

CRC Tests

Test	Status	Time
COMP	Pass	9:41am
CAL	Pass	9:41am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pitt Instrument Location Ayden P.D.

Instrument Serial No. 0086666 4144 West Ave, Ayden, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2nd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key S. D.

Signature of Certifying Official

643

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PITT AYDEN PD 730

Serial Number: 008666

Test Date: 11/22/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517403

Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	10:27am
AIR BLK	.00	10:27am
ACCY CHK	.08	10:28am
AIR BLK	.00	10:29am
SUB TEST	.00	10:30am
AIR BLK	.00	10:30am
SUB TEST	.00	10:32am
AIR BLK	.00	10:33am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PITT AYDEN PD 730

Serial Number: 008666 Test Record Number: 909
Test Date: 11/22/2016 Test Time: 10:34am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:35am
FLO	Pass	10:35am
FC	Pass	10:35am

Temperature Tests

Test	Status	Time
FC1	Pass	10:35am
SRC	Pass	10:35am
DET	Pass	10:35am
BAR	Pass	10:35am
BT	Pass	10:35am

Blank Tests

Test	Status	Time
AIR	Pass	10:35am

Printer Tests

Test	Status	Time
PRNT	Pass	10:35am

CRC Tests

Test	Status	Time
COMP	Pass	10:36am
CAL	Pass	10:36am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rosquotank Instrument Location Rosquotank Co. Public Safety
Instrument Serial No. 008941 Bldg., 200 E. Colonial St., Elizabeth City,
N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

**PASQUOTANK COUNTY PUBLIC SAFETY BLDG
690**

Serial Number: 008941
Test Date: 11/15/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: *GUARD, KELLY G*

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: *Breath Test*

Lot Number: AG621404

Exp Date: 08/01/2018


Test	g/210L	Time
DIAG	Pass	11:52am
AIR BLK	.00	11:52am
ACCY CHK	.08	11:53am
AIR BLK	.00	11:54am
SUB TEST	.00	11:54am
AIR BLK	.00	11:55am
SUB TEST	.00	11:57am
AIR BLK	.00	11:58am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008941 Test Record Number: 1176
Test Date: 11/15/2016 Test Time: 11:59am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:59am
FLO	Pass	11:59am
FC	Pass	11:59am

Temperature Tests

Test	Status	Time
FC1	Pass	12:00pm
SRC	Pass	12:00pm
DET	Pass	12:00pm
BAR	Pass	12:00pm
BT	Pass	12:00pm

Blank Tests

Test	Status	Time
AIR	Pass	12:00pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:00pm

CRC Tests

Test	Status	Time
COMP	Pass	12:01pm
CAL	Pass	12:01pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Pasquotank Instrument Location Pasquotank Co. Public Safety
Instrument Serial No. 008950 Bldg. 200 E. Colonial St., Elizabeth City,
N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PASQUOTANK COUNTY PUBLIC SAFETY BLDG
690

Serial Number: 008950
Test Date: 11/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G
Permit Number: 12955E
Effective:
07/01/2015-07/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG621404
Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:14am
AIR BLK	.00	11:15am
ACCY CHK	.08	11:15am
AIR BLK	.00	11:16am
SUB TEST	.00	11:17am
AIR BLK	.00	11:18am
SUB TEST	.00	11:20am
AIR BLK	.00	11:21am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PASQUOTANK COUNTY PUBLIC SAFETY BLDG 690

Serial Number: 008950 Test Record Number: 1472
Test Date: 11/15/2016 Test Time: 11:22am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:23am
FLO	Pass	11:23am
FC	Pass	11:23am

Temperature Tests

Test	Status	Time
FC1	Pass	11:23am
SRC	Pass	11:23am
DET	Pass	11:23am
BAR	Pass	11:23am
BT	Pass	11:23am

Blank Tests

Test	Status	Time
AIR	Pass	11:24am

Printer Tests

Test	Status	Time
PRNT	Pass	11:24am

CRC Tests

Test	Status	Time
COMP	Pass	11:24am
CAL	Pass	11:24am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Perquimans Instrument Location Perquimans Co. S.O.
Instrument Serial No. 008921 110 N. Church St., Hertford, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Key SA
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921
Test Date: 11/07/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	10:51am
AIR BLK	.00	10:52am
ACCY CHK	.08	10:52am
AIR BLK	.00	10:53am
SUB TEST	.00	10:54am
AIR BLK	.00	10:55am
SUB TEST	.00	10:56am
AIR BLK	.00	10:57am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

PERQUIMANS COUNTY PERQUIMANS CO SO 710

Serial Number: 008921 Test Record Number: 656
Test Date: 11/07/2016 Test Time: 10:59am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:00am
FLO	Pass	11:00am
FC	Pass	11:00am

Temperature Tests

Test	Status	Time
FC1	Pass	11:00am
SRC	Pass	11:00am
DET	Pass	11:00am
BAR	Pass	11:00am
BT	Pass	11:00am

Blank Tests

Test	Status	Time
AIR	Pass	11:01am

Printer Tests

Test	Status	Time
PRNT	Pass	11:01am

CRC Tests

Test	Status	Time
COMP	Pass	11:01am
CAL	Pass	11:01am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ROBESON Instrument Location ST. PAULS POLICE DEPT.
Instrument Serial No. 008814 ST. PAULS NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: 008814
Test Date: 11/30/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513101

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	3:50pm
AIR BLK	.00	3:50pm
ACCY CHK	.07	3:51pm
AIR BLK	.00	3:52pm
SUB TEST	.00	3:52pm
AIR BLK	.00	3:53pm
SUB TEST	.00	3:55pm
AIR BLK	.00	3:56pm

Reported AC₂ .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY ST. PAULS PD. 770

Serial Number: 008814 Test Record Number: 588
Test Date: 11/30/2016 Test Time: 3:57pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:58pm
FLO	Pass	3:58pm
FC	Pass	3:58pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:58pm
SRC	Pass	3:58pm
DET	Pass	3:58pm
BAR	Pass	3:58pm
BT	Pass	3:58pm

Blank Tests

Test	Status	Time
AIR	Pass	3:58pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:58pm

CRC Tests

Test	Status	Time
COMP	Pass	3:59pm
CAL	Pass	3:59pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ROBESON Instrument Location ROBESON Co. JAIL

Instrument Serial No. 008805 LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROBESON COUNTY LUMBERTON, LEC 770

Serial Number: 008805

Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	5:00pm
AIR BLK	.00	5:00pm
ACCY CHK	.07	5:01pm
AIR BLK	.00	5:02pm
SUB TEST	.00	5:03pm
AIR BLK	.00	5:04pm
SUB TEST	.00	5:05pm
AIR BLK	.00	5:06pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY LUMBERTON, LEC 770

Serial Number: 008805 Test Record Number: 3702
Test Date: 11/30/2016 Test Time: 5:07pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:07pm
FLO	Pass	5:07pm
FC	Pass	5:07pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:08pm
SRC	Pass	5:08pm
DET	Pass	5:08pm
BAR	Pass	5:08pm
BT	Pass	5:08pm

Blank Tests

Test	Status	Time
AIR	Pass	5:08pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:08pm

CRC Tests

Test	Status	Time
COMP	Pass	5:08pm
CAL	Pass	5:08pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County ROBESON Instrument Location ROBESON CO. JAIL
Instrument Serial No. 008836 LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 30 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROBESON COUNTY LUMBERTON, LEC. 770

Serial Number: 008836

Test Date: 11/30/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	5:19pm
AIR BLK	.00	5:20pm
ACCY CHK	.07	5:21pm
AIR BLK	.00	5:22pm
SUB TEST	.00	5:22pm
AIR BLK	.00	5:23pm
SUB TEST	.00	5:25pm
AIR BLK	.00	5:26pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY LUMBERTON, LEC. 770

Serial Number: 008836 Test Record Number: 4257
Test Date: 11/30/2016 Test Time: 5:26pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	5:27pm
FLO	Pass	5:27pm
FC	Pass	5:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:27pm
SRC	Pass	5:27pm
DET	Pass	5:27pm
BAR	Pass	5:27pm
BT	Pass	5:27pm

Blank Tests

Test	Status	Time
AIR	Pass	5:27pm

Printer Tests

Test	Status	Time
PRNT	Pass	5:27pm

CRC Tests

Test	Status	Time
COMP	Pass	5:28pm
CAL	Pass	5:28pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County RANDOLPH Instrument Location RANDLEMAN POLICE DEPT.
Instrument Serial No. 008737 RANDLEMAN NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



R. H. Russell
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RANDOLPH COUNTY RANDLEMAN PD 750

Serial Number: 008737
Test Date: 11/29/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607602
Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	3:15pm
AIR BLK	.00	3:15pm
ACCY CHK	.07	3:16pm
AIR BLK	.00	3:17pm
SUB TEST	.00	3:17pm
AIR BLK	.00	3:18pm
SUB TEST	.00	3:20pm
AIR BLK	.00	3:21pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

RANDOLPH COUNTY RANDLEMAN PD 750

Serial Number: 008737 Test Record Number: 943
Test Date: 11/29/2016 Test Time: 3:24pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:24pm
FLO	Pass	3:24pm
FC	Pass	3:24pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:25pm
SRC	Pass	3:25pm
DET	Pass	3:25pm
BAR	Pass	3:25pm
BT	Pass	3:25pm

Blank Tests

Test	Status	Time
AIR	Pass	3:25pm

Printer Tests

Test	Status	Time
PRNT	Pass	3:25pm

CRC Tests

Test	Status	Time
COMP	Pass	3:25pm
CAL	Pass	3:25pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II


County RICHMOND Instrument Location RICHMOND Co. MAG. OFFICE

Instrument Serial No. 008701 ROCKINGHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RICHMOND COUNTY RICHMOND CO. MAG OFF
760

Serial Number: 008701
Test Date: 11/29/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607601

Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	10:39am
AIR BLK	.00	10:39am
ACCY CHK	.08	10:40am
AIR BLK	.00	10:41am
SUB TEST	.00	10:42am
AIR BLK	.00	10:43am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008701 Test Record Number: 1114
Test Date: 11/29/2016 Test Time: 10:46am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:46am
FLO	Pass	10:46am
FC	Pass	10:46am

Temperature Tests

Test	Status	Time
FC1	Pass	10:47am
SRC	Pass	10:47am
DET	Pass	10:47am
BAR	Pass	10:47am
BT	Pass	10:47am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am


Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:47am
CAL	Pass	10:47am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County RICHMOND Instrument Location RICHMOND Co. MAG. OFFICE
Instrument Serial No. 008840 ROCKINGHAM, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 29 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RICHMOND COUNTY RICHMOND CO. MAG OFF.
760

Serial Number: 008840
Test Date: 11/29/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: RUSSELL, LARRY H

Permit Number: 6108E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA


Agency: DHHS

Test Type: Breath Test

Lot Number: AG607602
Exp Date: 03/16/2018

Test	g/210L	Time
DIAG	Pass	10:13am
AIR BLK	.00	10:14am
ACCY CHK	.07	10:14am
AIR BLK	.00	10:16am
SUB TEST	.00	10:16am
AIR BLK	.00	10:17am
SUB TEST	.00	10:19am
AIR BLK	.00	10:20am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

RICHMOND COUNTY RICHMOND CO. MAG OFF 760

Serial Number: 008840 Test Record Number: 1860
Test Date: 11/29/2016 Test Time: 10:26am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27am
FLO	Pass	10:27am
FC	Pass	10:27am

Temperature Tests

Test	Status	Time
FC1	Pass	10:27am
SRC	Pass	10:27am
DET	Pass	10:27am
BAR	Pass	10:27am
BT	Pass	10:27am

Blank Tests

Test	Status	Time
AIR	Pass	10:27am

Printer Tests

Test	Status	Time
PRNT	Pass	10:28am

CRC Tests

Test	Status	Time
COMP	Pass	10:28am
CAL	Pass	10:28am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County ROBESON Instrument Location LUMBERTON POLICE DEPT.

Instrument Serial No. 008629 LUMBERTON, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

371
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROBESON COUNTY LUMBERTON PD 770

Serial Number: 008629
Test Date: 11/10/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: RUSSELL, LARRY H
Permit Number: 6108E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607502
Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	5:41pm
AIR BLK	.00	5:42pm
ACCY CHK	.07	5:42pm
AIR BLK	.00	5:43pm
SUB TEST	.00	5:44pm
AIR BLK	.00	5:45pm
SUB TEST	.00	5:47pm
AIR BLK	.00	5:48pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

ROBESON COUNTY LUMBERTON PD 770

Serial Number: 008629 Test Record Number: 395
Test Date: 11/10/2016 Test Time: 5:48pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	5:49pm
FLO	Pass	5:49pm
FC	Pass	5:49pm

Temperature Tests

Test	Status	Time
FC1	Pass	5:49pm
SRC	Pass	5:49pm
DET	Pass	5:49pm
BAR	Pass	5:49pm
BT	Pass	5:49pm

Blank Tests

Test	Status	Time
AIR	Pass	5:50pm


Printer Tests

Test	Status	Time
PRNT	Pass	5:50pm

CRC Tests

Test	Status	Time
COMP	Pass	5:50pm
CAL	Pass	5:50pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Madison Police
Instrument Serial No. 008802 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802
Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	1:15pm
AIR BLK	.00	1:16pm
ACCY CHK	.08	1:16pm
AIR BLK	.00	1:18pm
SUB TEST	.00	1:18pm
AIR BLK	.00	1:19pm
SUB TEST	.00	1:21pm
AIR BLK	.00	1:21pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY MADISON PD 780

Serial Number: 008802 Test Record Number: 720
Test Date: 11/08/2016 Test Time: 1:22pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:23pm
FLO	Pass	1:23pm
FC	Pass	1:23pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:23pm
SRC	Pass	1:23pm
DET	Pass	1:23pm
BAR	Pass	1:23pm
BT	Pass	1:23pm

Blank Tests

Test	Status	Time
AIR	Pass	1:23pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:23pm

CRC Tests

Test	Status	Time
COMP	Pass	1:24pm
CAL	Pass	1:24pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Richmond Instrument Location BAT mobile unit 10
Instrument Serial No. 008686 Rockingham PD

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 11 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

660
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RICHMOND COUNTY BAT MOBILE UNIT 10 760

Serial Number: 008686
Test Date: 11/11/2016

Citation Number: M0000000-0
Subject's Name: NONE, NONE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: SMITH, JASON R
Permit Number: 19145E
Effective:
03/01/2016-03/01/2018

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG517403
Exp Date: 06/23/2017

Test	g/210L	Time
DIAG	Pass	8:13pm
AIR BLK	.00	8:14pm
ACCY CHK	.07	8:15pm
AIR BLK	.00	8:16pm
SUB TEST	.00	8:16pm
AIR BLK	.00	8:17pm
SUB TEST	.00	8:19pm
AIR BLK	.00	8:20pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RICHMOND COUNTY BAT MOBILE UNIT 10 760

Serial Number: 008686 Test Record Number: 6452
Test Date: 11/11/2016 Test Time: 8:20pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	8:21pm
FLO	Pass	8:21pm
FC	Pass	8:21pm

Temperature Tests

Test	Status	Time
FC1	Pass	8:21pm
SRC	Pass	8:21pm
DET	Pass	8:21pm
BAR	Pass	8:21pm
BT	Pass	8:21pm

Blank Tests

Test	Status	Time
AIR	Pass	8:22pm


Printer Tests

Test	Status	Time
PRNT	Pass	8:22pm

CRC Tests

Test	Status	Time
COMP	Pass	8:22pm
CAL	Pass	8:22pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Rockingham Instrument Location Eden
Instrument Serial No. 008636 Police Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 23 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



L. Kevin Dean
Signature of Certifying Official

642
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008636
Test Date: 11/23/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: DEAN, L K

Permit Number: 11598E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517501

Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	12:27pm
AIR BLK	.00	12:27pm
ACCY CHK	.07	12:28pm
AIR BLK	.00	12:29pm
SUB TEST	.00	12:30pm
AIR BLK	.00	12:31pm
SUB TEST	.00	12:32pm
AIR BLK	.00	12:33pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

ROCKINGHAM COUNTY EDEN PD 780

Serial Number: 008636 Test Record Number: 1675
Test Date: 11/23/2016 Test Time: 12:34pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:34pm
FLO	Pass	12:34pm
FC	Pass	12:34pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:35pm
SRC	Pass	12:35pm
DET	Pass	12:35pm
BAR	Pass	12:35pm
BT	Pass	12:35pm

Blank Tests

Test	Status	Time
AIR	Pass	12:35pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:35pm

CRC Tests

Test	Status	Time
COMP	Pass	12:35pm
CAL	Pass	12:35pm

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Rutherford Instrument Location Forest City PD

Instrument Serial No. 008889 187 S. Church Street, Forest City

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 7th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Josiah E. Helt
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889
Test Date: 11/07/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017


Test	g/210L	Time
DIAG	Pass	11:25am
AIR BLK	.00	11:25am
ACCY CHK	.08	11:26am
AIR BLK	.00	11:27am
SUB TEST	.00	11:27am
AIR BLK	.00	11:28am
SUB TEST	.00	11:30am
AIR BLK	.00	11:31am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

RUTHERFORD COUNTY FOREST CITY PD 800

Serial Number: 008889 Test Record Number: 729
Test Date: 11/07/2016 Test Time: 11:20am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:21am
FLO	Pass	11:21am
FC	Pass	11:21am

Temperature Tests

Test	Status	Time
FC1	Pass	11:21am
SRC	Pass	11:21am
DET	Pass	11:21am
BAR	Pass	11:21am
BT	Pass	11:21am

Blank Tests

Test	Status	Time
AIR	Pass	11:22am


Printer Tests

Test	Status	Time
PRNT	Pass	11:22am

CRC Tests

Test	Status	Time
COMP	Pass	11:22am
CAL	Pass	11:22am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stokes Instrument Location King Police
Instrument Serial No. 008610 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 14 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STOKES COUNTY KING PD 840

Serial Number: 008610

Test Date: 11/14/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG517501

Exp Date: 06/24/2017

Test	g/210L	Time
DIAG	Pass	10:04am
AIR BLK	.00	10:04am
ACCY CHK	.07	10:05am
AIR BLK	.00	10:06am
SUB TEST	.00	10:07am
AIR BLK	.00	10:08am
SUB TEST	.00	10:11am
AIR BLK	.00	10:12am

Reported AC: 00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY KING PD 840

Serial Number: 008610 Test Record Number: 1799
Test Date: 11/14/2016 Test Time: 10:13am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:14am
FLO	Pass	10:14am
FC	Pass	10:14am

Temperature Tests

Test	Status	Time
FC1	Pass	10:14am
SRC	Pass	10:14am
DET	Pass	10:14am
BAR	Pass	10:14am
BT	Pass	10:14am

Blank Tests

Test	Status	Time
AIR	Pass	10:14am


Printer Tests

Test	Status	Time
PRNT	Pass	10:14am

CRC Tests

Test	Status	Time
COMP	Pass	10:15am
CAL	Pass	10:15am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stokes Instrument Location Stokes County Jail
Instrument Serial No. DD8596 Danbury, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: 008596
Test Date: 11/08/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513101


Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	12:17pm
AIR BLK	.00	12:18pm
ACCY CHK	.07	12:19pm
AIR BLK	.00	12:20pm
SUB TEST	.00	12:20pm
AIR BLK	.00	12:21pm
SUB TEST	.00	12:23pm
AIR BLK	.00	12:24pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

STOKES COUNTY STOKES COUNTY JAIL 840

Serial Number: 008596 Test Record Number: 882
Test Date: 11/08/2016 Test Time: 12:24pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	12:25pm
FLO	Pass	12:25pm
FC	Pass	12:25pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:25pm
SRC	Pass	12:25pm
DET	Pass	12:25pm
BAR	Pass	12:25pm
BT	Pass	12:25pm

Blank Tests

Test	Status	Time
AIR	Pass	12:26pm


Printer Tests

Test	Status	Time
PRNT	Pass	12:26pm

CRC Tests

Test	Status	Time
COMP	Pass	12:26pm
CAL	Pass	12:26pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Surry Instrument Location Elkin Police
Instrument Serial No. 008926 Department

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

SURRY COUNTY ELKIN PD 850

Serial Number: 008926
Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607502

Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	2:53pm
AIR BLK	.00	2:53pm
ACCY CHK	.08	2:54pm
AIR BLK	.00	2:55pm
SUB TEST	.00	2:56pm
AIR BLK	.00	2:57pm
SUB TEST	.00	2:58pm
AIR BLK	.00	2:59pm

Reported AC: 00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

SURRY COUNTY ELKIN PD 850

Serial Number: 008926 Test Record Number: 726
Test Date: 11/09/2016 Test Time: 3:01pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	3:01pm
FLO	Pass	3:01pm
FC	Pass	3:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	3:01pm
SRC	Pass	3:01pm
DET	Pass	3:01pm
BAR	Pass	3:01pm
BT	Pass	3:01pm

Blank Tests

Test	Status	Time
AIR	Pass	3:02pm


Printer Tests

Test	Status	Time
PRNT	Pass	3:02pm

CRC Tests

Test	Status	Time
COMP	Pass	3:02pm
CAL	Pass	3:02pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Stanly County SD
Instrument Serial No. 008824 126 S. 3rd Street, Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joyce [Signature]
Signature of Certifying Official

650
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008824
Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG434201


Exp Date: 12/08/2016

Test	g/210L	Time
DIAG	Pass	10:42am
AIR BLK	.00	10:43am
ACCY CHK	.08	10:44am
AIR BLK	.00	10:45am
SUB TEST	.00	10:46am
AIR BLK	.00	10:47am
SUB TEST	.00	10:48am
AIR BLK	.00	10:49am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008824 Test Record Number: 1143
Test Date: 11/21/2016 Test Time: 10:50am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:51am
FLO	Pass	10:51am
FC	Pass	10:51am

Temperature Tests

Test	Status	Time
FC1	Pass	10:51am
SRC	Pass	10:51am
DET	Pass	10:51am
BAR	Pass	10:51am
BT	Pass	10:51am

Blank Tests

Test	Status	Time
AIR	Pass	10:52am


Printer Tests

Test	Status	Time
PRNT	Pass	10:52am

CRC Tests

Test	Status	Time
COMP	Pass	10:52am
CAL	Pass	10:52am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Stanly Instrument Location Stanly County SD
Instrument Serial No. 008842 126 S. 3rd Street, Albemarle

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21st day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Joseph E. Huth 650
Signature of Certifying Official Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008842
Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: HUTCHINSON, JOSEPH E

Permit Number: 19951E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902


Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	10:37am
AIR BLK	.00	10:38am
ACCY CHK	.07	10:39am
AIR BLK	.00	10:40am
SUB TEST	.00	10:41am
AIR BLK	.00	10:42am
SUB TEST	.00	10:44am
AIR BLK	.00	10:45am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

STANLY COUNTY STANLY COUNTY SD 830

Serial Number: 008842 Test Record Number: 1887
Test Date: 11/21/2016 Test Time: 10:46am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:46am
FLO	Pass	10:46am
FC	Pass	10:47am

Temperature Tests

Test	Status	Time
FC1	Pass	10:47am
SRC	Pass	10:47am
DET	Pass	10:47am
BAR	Pass	10:47am
BT	Pass	10:47am

Blank Tests

Test	Status	Time
AIR	Pass	10:47am


Printer Tests

Test	Status	Time
PRNT	Pass	10:47am

CRC Tests

Test	Status	Time
COMP	Pass	10:48am
CAL	Pass	10:48am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Tyrrell Instrument Location Tyrrell Co. SO.
Instrument Serial No. 008902 402 Main St., Columbia, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 3rd day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

643
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

TYRRELL COUNTY SHERIFF'S OFFICE 880

Serial Number: 008902

Test Date: 11/03/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: GUARD, KELLY G

Permit Number: 12955E

Effective:

07/01/2015-07/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	10:13am
AIR BLK	.00	10:13am
ACCY CHK	.07	10:14am
AIR BLK	.00	10:15am
SUB TEST	.00	10:15am
AIR BLK	.00	10:17am
SUB TEST	.00	10:18am
AIR BLK	.00	10:19am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

TYRRELL COUNTY SHERIFF'S OFFICE 880

Serial Number: 008902 Test Record Number: 629
Test Date: 11/03/2016 Test Time: 10:21am EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:21am
FLO	Pass	10:21am
FC	Pass	10:21am

Temperature Tests

Test	Status	Time
FC1	Pass	10:21am
SRC	Pass	10:21am
DET	Pass	10:21am
BAR	Pass	10:21am
BT	Pass	10:21am

Blank Tests

Test	Status	Time
AIR	Pass	10:22am


Printer Tests

Test	Status	Time
PRNT	Pass	10:22am

CRC Tests

Test	Status	Time
COMP	Pass	10:22am
CAL	Pass	10:22am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Transylvania Instrument Location Transylvania Co. Jail

Instrument Serial No. 008609 Brevard, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

TRANSYLVANIA COUNTY TRANSYLVANIA CO
JAIL 870

Serial Number: 008609
Test Date: 11/15/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R
Permit Number: 8457E
Effective:
09/01/2015-09/01/2017

Officer's Name: NONE,
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607502
Exp Date: 03/15/2018

Test	g/210L	Time
DIAG	Pass	2:37pm
AIR BLK	.00	2:38pm
ACCY CHK	.08	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	.00	2:40pm
AIR BLK	.00	2:41pm
SUB TEST	.00	2:42pm
AIR BLK	.00	2:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

TRANSYLVANIA COUNTY TRANSYLVANIA CO JAIL 870

Serial Number: 008609 Test Record Number: 710
Test Date: 11/15/2016 Test Time: 2:44pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:45pm
FLO	Pass	2:45pm
FC	Pass	2:45pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:45pm
SRC	Pass	2:45pm
DET	Pass	2:45pm
BAR	Pass	2:45pm
BT	Pass	2:45pm

Blank Tests

Test	Status	Time
AIR	Pass	2:46pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:46pm

CRC Tests

Test	Status	Time
COMP	Pass	2:46pm
CAL	Pass	2:46pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Transylvania Instrument Location Transylvania Co. Jail
Instrument Serial No. 008820 Brevard, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 15 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



David R. Cuth
Signature of Certifying Official

635
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

TRANSYLVANIA COUNTY TRANSYLVANIA CO
JAIL 870

Serial Number: 008820
Test Date: 11/15/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: CUTLER, DANIEL R

Permit Number: 8457E

Effective:

09/01/2015-09/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2016

Test	g/210L	Time
DIAG	Pass	2:36pm
AIR BLK	.00	2:37pm
ACCY CHK	.07	2:38pm
AIR BLK	.00	2:39pm
SUB TEST	.00	2:39pm
AIR BLK	.00	2:40pm
SUB TEST	.00	2:42pm
AIR BLK	.00	2:43pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

TRANSYLVANIA COUNTY TRANSYLVANIA CO JAIL 870

Serial Number: 008820 Test Record Number: 1011
Test Date: 11/15/2016 Test Time: 2:44pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:44pm
FLO	Pass	2:44pm
FC	Pass	2:44pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:45pm
SRC	Pass	2:45pm
DET	Pass	2:45pm
BAR	Pass	2:45pm
BT	Pass	2:45pm

Blank Tests

Test	Status	Time
AIR	Pass	2:45pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:45pm

CRC Tests

Test	Status	Time
COMP	Pass	2:45pm
CAL	Pass	2:45pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wayne Instrument Location Wayne Co. Detention CTR.
Instrument Serial No. 008879 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Jacob A. Reed
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879
Test Date: 11/10/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG517402

Exp Date: 06/23/2017


Test	g/210L	Time
DIAG	Pass	9:12am
AIR BLK	.00	9:13am
ACCY CHK	.08	9:13am
AIR BLK	.00	9:14am
SUB TEST	.00	9:15am
AIR BLK	.00	9:16am
SUB TEST	.00	9:17am
AIR BLK	.00	9:18am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008879 Test Record Number: 526
Test Date: 11/10/2016 Test Time: 9:19am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:20am
FLO	Pass	9:20am
FC	Pass	9:20am

Temperature Tests

Test	Status	Time
FC1	Pass	9:20am
SRC	Pass	9:20am
DET	Pass	9:20am
BAR	Pass	9:20am
BT	Pass	9:20am

Blank Tests

Test	Status	Time
AIR	Pass	9:21am


Printer Tests

Test	Status	Time
PRNT	Pass	9:21am

CRC Tests

Test	Status	Time
COMP	Pass	9:21am
CAL	Pass	9:21am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II

County Wayne Instrument Location Wayne Co. Detention Ctr.
Instrument Serial No. 008649 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649
Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	9:02am
AIR BLK	.00	9:03am
ACCY CHK	.07	9:04am
AIR BLK	.00	9:04am
SUB TEST	.00	9:05am
AIR BLK	.00	9:06am
SUB TEST	.00	9:09am
AIR BLK	.00	9:10am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008649 Test Record Number: 3067
Test Date: 11/10/2016 Test Time: 9:11am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:11am
FLO	Pass	9:11am
FC	Pass	9:11am

Temperature Tests

Test	Status	Time
FC1	Pass	9:11am
SRC	Pass	9:11am
DET	Pass	9:11am
BAR	Pass	9:11am
BT	Pass	9:11am

Blank Tests

Test	Status	Time
AIR	Pass	9:12am


Printer Tests

Test	Status	Time
PRNT	Pass	9:12am

CRC Tests

Test	Status	Time
COMP	Pass	9:12am
CAL	Pass	9:12am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wayne Instrument Location Wayne Co. Detention Ctr.

Instrument Serial No. 008671 207 E. Chestnut St., Goldsboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 10th day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

647
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671
Test Date: 11/10/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: KEESLER, LINDA A

Permit Number: 11646E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534901

Exp Date: 12/15/2017


Test	g/210L	Time
DIAG	Pass	9:01am
AIR BLK	.00	9:02am
ACCY CHK	.07	9:03am
AIR BLK	.00	9:04am
SUB TEST	.00	9:04am
AIR BLK	.00	9:05am
SUB TEST	.00	9:07am
AIR BLK	.00	9:08am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY WAYNE CO DETENTION 950

Serial Number: 008671 Test Record Number: 4450
Test Date: 11/10/2016 Test Time: 9:10am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:10am
FLO	Pass	9:10am
FC	Pass	9:10am

Temperature Tests

Test	Status	Time
FC1	Pass	9:11am
SRC	Pass	9:11am
DET	Pass	9:11am
BAR	Pass	9:11am
BT	Pass	9:11am

Blank Tests

Test	Status	Time
AIR	Pass	9:11am


Printer Tests

Test	Status	Time
PRNT	Pass	9:11am

CRC Tests

Test	Status	Time
COMP	Pass	9:11am
CAL	Pass	9:11am

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location BAT MOBILE UNIT #16

Instrument Serial No. 008869 RACE11-H

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 18 day of NOV, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

661

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAYNE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008869
Test Date: 11/18/2016

Citation Number: M0000000-0

Subject's Name:
PREVNTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NNE

Analyst's Name: SMITH, JOCK B

Permit Number: 20630E

Effective:

05/01/2016-05/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403


Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	8:59pm
AIR BLK	.00	9:00pm
ACCY CHK	.07	9:01pm
AIR BLK	.00	9:02pm
SUB TEST	.00	9:02pm
AIR BLK	.00	9:03pm
SUB TEST	.00	9:05pm
AIR BLK	.00	9:06pm

Reported AS:  .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAYNE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008869 Test Record Number: 932
Test Date: 11/18/2016 Test Time: 9:07pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:08pm
FLO	Pass	9:08pm
FC	Pass	9:08pm

Temperature Tests

Test	Status	Time
FC1	Pass	9:08pm
SRC	Pass	9:08pm
DET	Pass	9:08pm
BAR	Pass	9:08pm
BT	Pass	9:08pm

Blank Tests

Test	Status	Time
AIR	Pass	9:08pm

Printer Tests


Test	Status	Time
PRNT	Pass	9:09pm

CRC Tests

Test	Status	Time
COMP	Pass	9:09pm
CAL	Pass	9:09pm

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location PAT MOBILE UNIT #6

Instrument Serial No. 008869 HOLLY SPRINGS

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 4 day of NOV, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]

Signature of Certifying Official

6661
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008869
Test Date: 11/04/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, JOCK B

Permit Number: 20630E

Effective:

05/01/2016-05/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:04pm
AIR BLK	.00	10:05pm
ACCY CHK	.08	10:05pm
AIR BLK	.00	10:06pm
SUB TEST	.00	10:07pm
AIR BLK	.00	10:08pm
SUB TEST	.00	10:09pm
AIR BLK	.00	10:10pm

Reported AC:  .00 g/210L

Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY BAT MOBILE UNIT 6 910

Serial Number: 008869 Test Record Number: 925
Test Date: 11/04/2016 Test Time: 10:26pm EDT

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:27pm
FLO	Pass	10:27pm
FC	Pass	10:27pm

Temperature Tests

Test	Status	Time
FC1	Pass	10:27pm
SRC	Pass	10:27pm
DET	Pass	10:27pm
BAR	Pass	10:27pm
BT	Pass	10:27pm

Blank Tests

Test	Status	Time
AIR	Pass	10:28pm


Printer Tests

Test	Status	Time
PRNT	Pass	10:28pm

CRC Tests

Test	Status	Time
COMP	Pass	10:28pm
CAL	Pass	10:28pm

Preventive Maintenance
Status: Pass



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE FOREST PD
Instrument Serial No. 008700 225 S. TAYLOR ST WAKE FOREST NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 2 day of Nov, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Alex Barnes
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700

Test Date: 11/02/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG513102

Exp Date: 05/11/2017

Test	g/210L	Time
DIAG	Pass	1:33pm
AIR BLK	.00	1:34pm
ACCY CHK	.07	1:35pm
AIR BLK	.00	1:36pm
SUB TEST	.00	1:36pm
AIR BLK	.00	1:37pm
SUB TEST	.00	1:39pm
AIR BLK	.00	1:40pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY WAKE FOREST PD 910

Serial Number: 008700 Test Record Number: 1165
Test Date: 11/02/2016 Test Time: 1:41pm EDT

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:41pm
FLO	Pass	1:41pm
FC	Pass	1:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:41pm
SRC	Pass	1:41pm
DET	Pass	1:41pm
BAR	Pass	1:41pm
BT	Pass	1:41pm

Blank Tests

Test	Status	Time
AIR	Pass	1:42pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:42pm

CRC Tests

Test	Status	Time
COMP	Pass	1:42pm
CAL	Pass	1:42pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilkes Instrument Location Wilkes County Detention
Instrument Serial No. 008865 Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865

Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017

Test	g/210L	Time
DIAG	Pass	1:33pm
AIR BLK	.00	1:34pm
ACCY CHK	.07	1:35pm
AIR BLK	.00	1:35pm
SUB TEST	.00	1:36pm
AIR BLK	.00	1:37pm
SUB TEST	.00	1:38pm
AIR BLK	.00	1:39pm

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008865 Test Record Number: 464
Test Date: 11/09/2016 Test Time: 1:40pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:41pm
FLO	Pass	1:41pm
FC	Pass	1:41pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:41pm
SRC	Pass	1:41pm
DET	Pass	1:41pm
BAR	Pass	1:41pm
BT	Pass	1:41pm

Blank Tests

Test	Status	Time
AIR	Pass	1:42pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:42pm

CRC Tests

Test	Status	Time
COMP	Pass	1:42pm
CAL	Pass	1:42pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Wilkes Instrument Location Wilkes County Detention
Instrument Serial No. DD8843 Wilkesboro, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843
Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG534802

Exp Date: 12/14/2017


Test	g/210L	Time
DIAG	Pass	1:31pm
AIR BLK	.00	1:31pm
ACCY CHK	.07	1:32pm
AIR BLK	.00	1:33pm
SUB TEST	.00	1:34pm
AIR BLK	.00	1:35pm
SUB TEST	.00	1:36pm
AIR BLK	.00	1:37pm

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WILKES COUNTY WILKES CO DETENTION 960

Serial Number: 008843 Test Record Number: 1987
Test Date: 11/09/2016 Test Time: 1:38pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	1:38pm
FLO	Pass	1:38pm
FC	Pass	1:38pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:39pm
SRC	Pass	1:39pm
DET	Pass	1:39pm
BAR	Pass	1:39pm
BT	Pass	1:39pm

Blank Tests

Test	Status	Time
AIR	Pass	1:39pm


Printer Tests

Test	Status	Time
PRNT	Pass	1:39pm

CRC Tests

Test	Status	Time
COMP	Pass	1:39pm
CAL	Pass	1:39pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County Wake Instrument Location Apex PD STATION 4

Instrument Serial No. 008621 1615 E. Williams ST
Apex, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Barnes
Signature of Certifying Official

662
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY APEX PD

Serial Number: 008621

Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG607502

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	10:12am
AIR BLK	.00	10:13am
ACCY CHK	.08	10:14am
AIR BLK	.00	10:15am
SUB TEST	.00	10:15am
AIR BLK	.00	10:16am
SUB TEST	.00	10:18am
AIR BLK	.00	10:19am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY APEX PD

Serial Number: 008621 Test Record Number: 2164

Test Date: 11/21/2016 Test Time: 10:20am EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	10:20am
FLO	Pass	10:20am
FC	Pass	10:20am

Temperature Tests

Test	Status	Time
FC1	Pass	10:20am
SRC	Pass	10:20am
DET	Pass	10:20am
BAR	Pass	10:20am
BT	Pass	10:20am

Blank Tests

Test	Status	Time
AIR	Pass	10:21am

Printer Tests


Test	Status	Time
PRNT	Pass	10:21am

CRC Tests

Test	Status	Time
COMP	Pass	10:21am
CAL	Pass	10:21am

Preventive Maintenance

Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location CARY PD

Instrument Serial No. 008587 120 Wilkinson Ave
CARY, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Stokes Barnes

Signature of Certifying Official

662

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY CARY PD 910

Serial Number: 008587

Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BARNES, SIMON S

Permit Number: 11434E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607502

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	11:04am
AIR BLK	.00	11:06am
ACCY CHK	.08	11:06am
AIR BLK	.00	11:07am
SUB TEST	.00	11:08am
AIR BLK	.00	11:09am
SUB TEST	.00	11:10am
AIR BLK	.00	11:11am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY CARY PD 910

Serial Number: 008587 Test Record Number: 3367
Test Date: 11/21/2016 Test Time: 11:12am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:13am
FLO	Pass	11:13am
FC	Pass	11:13am

Temperature Tests

Test	Status	Time
FC1	Pass	11:13am
SRC	Pass	11:13am
DET	Pass	11:13am
BAR	Pass	11:13am
BT	Pass	11:13am

Blank Tests

Test	Status	Time
AIR	Pass	11:13am


Printer Tests

Test	Status	Time
PRNT	Pass	11:14am

CRC Tests

Test	Status	Time
COMP	Pass	11:14am
CAL	Pass	11:14am

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE CO. DETENTION CTR
Instrument Serial No. 008778 3301 HAMMOND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778
Test Date: 11/21/2016

Citation Number: M0000000-0
Subject's Name:
PREVENTIVE, MAINTENANCE
Subject's Date of Birth: 11/11/1911
Subject's Sex: Male
Driver's License State: XX
Driver's License Number: NONE


Analyst's Name: SMITH, BRIAN D
Permit Number: 8937E
Effective:
08/01/2015-08/01/2017

Officer's Name: NONE, NONE
Type of Agency: FTA
Agency: DHHS
Test Type: Breath Test

Lot Number: AG607601
Exp Date: 03/16/2018


Test	g/210L	Time
DIAG	Pass	11:53am
AIR BLK	.00	11:54am
ACCY CHK	.08	11:54am
AIR BLK	.00	11:55am
SUB TEST	.00	11:56am
AIR BLK	.00	11:57am
SUB TEST	.00	11:58am
AIR BLK	.00	11:59am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008778 Test Record Number: 2356
Test Date: 11/21/2016 Test Time: 12:00pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	12:01pm
FLO	Pass	12:01pm
FC	Pass	12:01pm

Temperature Tests

Test	Status	Time
FC1	Pass	12:01pm
SRC	Pass	12:01pm
DET	Pass	12:01pm
BAR	Pass	12:01pm
BT	Pass	12:01pm

Blank Tests

Test	Status	Time
AIR	Pass	12:01pm

Printer Tests

Test	Status	Time
PRNT	Pass	12:01pm

CRC Tests

Test	Status	Time
COMP	Pass	12:02pm
CAL	Pass	12:02pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location WAKE CO. DETENTION CTR

Instrument Serial No. 008760 3301 HAMMOND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760
Test Date: 11/21/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG621403

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	11:34am
AIR BLK	.00	11:35am
ACCY CHK	.08	11:36am
AIR BLK	.00	11:37am
SUB TEST	.00	11:39am
AIR BLK	.00	11:39am
SUB TEST	.00	11:41am
AIR BLK	.00	11:42am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008760 Test Record Number: 1868
Test Date: 11/21/2016 Test Time: 11:43am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:44am
FLO	Pass	11:44am
FC	Pass	11:44am

Temperature Tests

Test	Status	Time
FC1	Pass	11:44am
SRC	Pass	11:44am
DET	Pass	11:44am
BAR	Pass	11:44am
BT	Pass	11:44am

Blank Tests

Test	Status	Time
AIR	Pass	11:44am


Printer Tests

Test	Status	Time
PRNT	Pass	11:45am

CRC Tests

Test	Status	Time
COMP	Pass	11:45am
CAL	Pass	11:45am

Preventive Maintenance
Status: *Pass*



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location WAKE CO. DETENTION CTR

Instrument Serial No. 008577 3301 HAMMOND RD. RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Brian D. Arnold

Signature of Certifying Official

637

Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577
Test Date: 11/21/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG607501

Exp Date: 03/15/2018


Test	g/210L	Time
DIAG	Pass	10:59am
AIR BLK	.00	11:00am
ACCY CHK	.07	11:00am
AIR BLK	.00	11:02am
SUB TEST	.00	11:02am
AIR BLK	.00	11:03am
SUB TEST	.00	11:05am
AIR BLK	.00	11:06am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008577 Test Record Number: 2630
Test Date: 11/21/2016 Test Time: 11:24am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	11:24am
FLO	Pass	11:24am
FC	Pass	11:24am

Temperature Tests

Test	Status	Time
FC1	Pass	11:24am
SRC	Pass	11:24am
DET	Pass	11:24am
BAR	Pass	11:24am
BT	Pass	11:24am

Blank Tests

Test	Status	Time
AIR	Pass	11:25am


Printer Tests

Test	Status	Time
PRNT	Pass	11:25am

CRC Tests

Test	Status	Time
COMP	Pass	11:25am
CAL	Pass	11:25am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**


County WAKE Instrument Location KNIGHTDALE P.S.

Instrument Serial No. 008838 979 STEEPLE SQUARE, CT. KNIGHTDALE, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY KNIGHTDALE PS 910

Serial Number: 008838 Test Record Number: 1421
Test Date: 11/21/2016 Test Time: 9:42am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:42am
FLO	Pass	9:42am
FC	Pass	9:42am

Temperature Tests

Test	Status	Time
FC1	Pass	9:42am
SRC	Pass	9:42am
DET	Pass	9:42am
BAR	Pass	9:42am
BT	Pass	9:42am

Blank Tests

Test	Status	Time
AIR	Pass	9:43am

Printer Tests

Test	Status	Time
PRNT	Pass	9:43am

CRC Tests

Test	Status	Time
COMP	Pass	9:43am
CAL	Pass	9:43am

Preventive Maintenance
Status: Pass



Analyst

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY KNIGHTDALE PS 910

Serial Number: 008838 Test Record Number: 1421
Test Date: 11/21/2016 Test Time: 9:42am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	9:42am
FLO	Pass	9:42am
FC	Pass	9:42am

Temperature Tests

Test	Status	Time
FC1	Pass	9:42am
SRC	Pass	9:42am
DET	Pass	9:42am
BAR	Pass	9:42am
BT	Pass	9:42am

Blank Tests

Test	Status	Time
AIR	Pass	9:43am

Printer Tests

Test	Status	Time
PRNT	Pass	9:43am

CRC Tests

Test	Status	Time
COMP	Pass	9:43am
CAL	Pass	9:43am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WAKE Instrument Location WAKE CO. DETENTION CTR

Instrument Serial No. DO8612 3301 HAMMOND RD RALEIGH, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 21 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Bruce D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612
Test Date: 11/21/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS


Test Type: Breath Test

Lot Number: AG621404

Exp Date: 08/01/2018

Test	g/210L	Time
DIAG	Pass	10:35am
AIR BLK	.00	10:36am
ACCY CHK	.08	10:37am
AIR BLK	.00	10:38am
SUB TEST	.00	10:38am
AIR BLK	.00	10:39am
SUB TEST	.00	10:41am
AIR BLK	.00	10:42am

Reported AC: .00 g/210L



Signature of Chemical Analyst

Court CVR



Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

WAKE COUNTY DETENTION CENTER 910

Serial Number: 008612 Test Record Number: 3288
Test Date: 11/21/2016 Test Time: 10:49am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:50am
FLO	Pass	10:50am
FC	Pass	10:50am

Temperature Tests

Test	Status	Time
FC1	Pass	10:50am
SRC	Pass	10:50am
DET	Pass	10:50am
BAR	Pass	10:50am
BT	Pass	10:50am

Blank Tests

Test	Status	Time
AIR	Pass	10:51am

Printer Tests

Test	Status	Time
PRNT	Pass	10:51am

CRC Tests

Test	Status	Time
COMP	Pass	10:51am
CAL	Pass	10:51am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH

**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County WARREN Instrument Location NORLINA PD

Instrument Serial No. 008945 101 MAIN ST. NORLINA, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 22 day of NOVEMBER, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



Ray D. Smith
Signature of Certifying Official

637
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945
Test Date: 11/22/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: SMITH, BRIAN D

Permit Number: 8937E

Effective:

08/01/2015-08/01/2017

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG534902

Exp Date: 12/15/2017

Test	g/210L	Time
DIAG	Pass	1:02pm
AIR BLK	.00	1:03pm
ACCY CHK	.08	1:04pm
AIR BLK	.00	1:05pm
SUB TEST	.00	1:06pm
AIR BLK	.00	1:07pm
SUB TEST	.00	1:08pm
AIR BLK	.00	1:09pm

Reported AC: .00 g/210L

Brian D Smith

Signature of Chemical Analyst

Court CVR .

Brian D Smith

Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

WARREN COUNTY NORLINA POLICE DEPT 920

Serial Number: 008945 Test Record Number: 325
Test Date: 11/22/2016 Test Time: 1:10pm EST

System Check: Passed

Baseline Tests

Test	Status	Time
IR	Pass	1:11pm
FLO	Pass	1:11pm
FC	Pass	1:11pm

Temperature Tests

Test	Status	Time
FC1	Pass	1:11pm
SRC	Pass	1:11pm
DET	Pass	1:11pm
BAR	Pass	1:11pm
BT	Pass	1:11pm

Blank Tests

Test	Status	Time
AIR	Pass	1:11pm

Printer Tests

Test	Status	Time
PRNT	Pass	1:11pm

CRC Tests

Test	Status	Time
COMP	Pass	1:12pm
CAL	Pass	1:12pm

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yancey Instrument Location Yancey Co. Jail
Instrument Serial No. 008653 Burnsville, NC

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 8 day of November, 2016 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

649
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653
Test Date: 11/08/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BURNETTE, ANTHONY J

Permit Number: 11304E

Effective:

05/01/2015-05/01/2017

Officer's Name: NONE,

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG526401

Exp Date: 09/21/2017

Test	g/210L	Time
DIAG	Pass	2:46pm
AIR BLK	.00	2:47pm
ACCY CHK	.08	2:48pm
AIR BLK	.00	2:49pm
SUB TEST	.00	2:50pm
AIR BLK	.00	2:51pm
SUB TEST	.00	2:52pm
AIR BLK	.00	2:53pm

Reported AC: .00 g/210L

Signature of Chemical Analyst

Court CVR


Analyst

**This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007**

Intox EC/IR-II: Preventive Maintenance

YANCEY COUNTY YANCEY COUNTY JAIL 990

Serial Number: 008653 Test Record Number: 1215
Test Date: 11/08/2016 Test Time: 2:54pm EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	2:54pm
FLO	Pass	2:54pm
FC	Pass	2:55pm

Temperature Tests

Test	Status	Time
FC1	Pass	2:55pm
SRC	Pass	2:55pm
DET	Pass	2:55pm
BAR	Pass	2:55pm
BT	Pass	2:55pm

Blank Tests

Test	Status	Time
AIR	Pass	2:55pm

Printer Tests

Test	Status	Time
PRNT	Pass	2:55pm

CRC Tests

Test	Status	Time
COMP	Pass	2:55pm
CAL	Pass	2:55pm

Preventive Maintenance
Status: Pass


Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yadkin Instrument Location Yadkin County Jail
Instrument Serial No. 008854 Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008854
Test Date: 11/09/2016

Citation Number: M0000000-0
Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902

Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	10:25am
AIR BLK	.00	10:26am
ACCY CHK	.08	10:27am
AIR BLK	.00	10:28am
SUB TEST	.00	10:28am
AIR BLK	.00	10:29am
SUB TEST	.00	10:31am
AIR BLK	.00	10:32am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

This form is used when performing Preventive Maintenance procedures
Forensic Tests for Alcohol Branch
Department of Health and Human Services
Rev. 12/2007

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008854 Test Record Number: 421
Test Date: 11/09/2016 Test Time: 10:33am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:34am
FLO	Pass	10:34am
FC	Pass	10:34am

Temperature Tests

Test	Status	Time
FC1	Pass	10:34am
SRC	Pass	10:34am
DET	Pass	10:34am
BAR	Pass	10:34am
BT	Pass	10:34am

Blank Tests

Test	Status	Time
AIR	Pass	10:34am


Printer Tests

Test	Status	Time
PRNT	Pass	10:35am

CRC Tests

Test	Status	Time
COMP	Pass	10:35am
CAL	Pass	10:35am

Preventive Maintenance
Status: Pass



Analyst

DEPARTMENT OF HEALTH AND HUMAN SERVICES
FORENSIC TESTS FOR ALCOHOL BRANCH


**PREVENTIVE MAINTENANCE RECORD
INTOXIMETERS, MODEL INTOX EC/IR II**

County Yadkin Instrument Location Yadkin County Jail
Instrument Serial No. 008944 Yadkinville, N.C.

The preventive maintenance procedures for the Intoximeters, Model Intox EC/IR II to be followed at least once every four months are:

1. Verify the ethanol gas canister displays pressure, or the alcoholic breath simulator thermometer shows 34 degrees, plus or minus .2 degree centigrade;
2. Verify instrument displays time and date;
3. Initiate breath test sequence;
4. Enter information as prompted;
5. Verify instrument accuracy;
6. When "PLEASE BLOW" appears, collect breath sample;
7. When "PLEASE BLOW" appears, collect breath sample;
8. Print test record;
9. Verify Diagnostic Program; and
10. Verify that the ethanol gas canister is being changed before expiration date, or the alcoholic breath simulator solution is being changed every four months or after 125 Alcoholic Breath Simulator tests, whichever occurs first.

I certify that on the 9 day of November, 20 16 the forgoing preventive maintenance procedures were performed on the instrument indicated above, in accordance with current regulations of the N.C. Department of Health and Human Services, and the instrument is functioning properly.



[Signature]
Signature of Certifying Official

657
Certificate Number

A signed original of the preventive maintenance record shall be kept on file for at least three years.

Intox EC/IR-II: Subject Test

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944
Test Date: 11/09/2016

Citation Number: M0000000-0

Subject's Name:

PREVENTIVE, MAINTENANCE

Subject's Date of Birth: 11/11/1911

Subject's Sex: Male

Driver's License State: XX

Driver's License Number: NONE

Analyst's Name: BENFIELD II, KENNETH R

Permit Number: 22067E

Effective:

07/01/2016-07/01/2018

Officer's Name: NONE, NONE

Type of Agency: FTA

Agency: DHHS

Test Type: Breath Test

Lot Number: AG507902


Exp Date: 03/20/2017

Test	g/210L	Time
DIAG	Pass	10:22am
AIR BLK	.00	10:23am
ACCY CHK	.07	10:24am
AIR BLK	.00	10:25am
SUB TEST	.00	10:26am
AIR BLK	.00	10:26am
SUB TEST	.00	10:28am
AIR BLK	.00	10:29am

Reported AC: .00 g/210L


Signature of Chemical Analyst

Court CVR


Analyst

Intox EC/IR-II: Preventive Maintenance

YADKIN COUNTY YADKIN CO JAIL 980

Serial Number: 008944 Test Record Number: 1332
Test Date: 11/09/2016 Test Time: 10:30am EST

System Check: *Passed*

Baseline Tests

Test	Status	Time
IR	Pass	10:30am
FLO	Pass	10:30am
FC	Pass	10:30am

Temperature Tests

Test	Status	Time
FC1	Pass	10:30am
SRC	Pass	10:30am
DET	Pass	10:30am
BAR	Pass	10:30am
BT	Pass	10:30am

Blank Tests

Test	Status	Time
AIR	Pass	10:31am


Printer Tests

Test	Status	Time
PRNT	Pass	10:31am

CRC Tests

Test	Status	Time
COMP	Pass	10:31am
CAL	Pass	10:31am

Preventive Maintenance
Status: Pass



Analyst